

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006128.D
 Acq On : 07 Jul 2021 11:54
 Operator : CG/JU
 Sample : SSTDIC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDIC010

Manual Integrations
 APPROVED

mohammad
 7/8/2021 12:38:53 PM

Quant Time: Jul 07 13:50:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	167698	20.000	ng	0.00
21) Naphthalene-d8	10.681	136	679606	20.000	ng	# 0.00
39) Acenaphthene-d10	14.510	164	397479	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	751129	20.000	ng	# 0.00
76) Chrysene-d12	21.345	240	719031	20.000	ng	# 0.00
86) Perylene-d12	23.739	264	815835	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.458	112	208217	19.923	ng	0.00
7) Phenol-d6	7.069	99	266025	19.639	ng	0.00
23) Nitrobenzene-d5	9.046	82	217439	15.686	ng	0.00
42) 2,4,6-Tribromophenol	16.010	330	85532	12.539	ng	0.00
45) 2-Fluorobiphenyl	13.140	172	554037	18.303	ng	0.00
79) Terphenyl-d14	19.816	244	729267	18.991	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.323	88	43765	9.269	ng	90
3) Pyridine	3.758	79	96766	8.720	ng	# 85
4) n-Nitrosodimethylamine	3.658	42	40563	7.864	ng	# 90
6) Aniline	7.211	93	140752	9.307	ng	99
8) 2-Chlorophenol	7.446	128	121427	10.134	ng	99
9) Benzaldehyde	7.028	77	80439	13.930	ng	94
10) Phenol	7.093	94	128768	9.516	ng	94
11) bis(2-Chloroethyl)ether	7.305	93	102346	9.981	ng	98
12) 1,3-Dichlorobenzene	7.763	146	129947	9.698	ng	97
13) 1,4-Dichlorobenzene	7.916	146	132219	9.676	ng	98
14) 1,2-Dichlorobenzene	8.228	146	127702	9.777	ng	99
15) Benzyl Alcohol	8.140	79	73768m	7.230	ng	
16) 2,2'-oxybis(1-Chloropr...	8.411	45	120020	12.662	ng	91
17) 2-Methylphenol	8.346	107	87767	9.519	ng	95
18) Hexachloroethane	8.952	117	43695	8.840	ng	91
19) n-Nitroso-di-n-propyla...	8.693	70	74746	8.970	ng	97
20) 3+4-Methylphenols	8.681	107	116875	9.384	ng	94
22) Acetophenone	8.711	105	158844	8.880	ng	98
24) Nitrobenzene	9.093	77	112473	7.814	ng	95
25) Isophorone	9.622	82	207120	8.091	ng	97
26) 2-Nitrophenol	9.805	139	55787	8.123	ng	90
27) 2,4-Dimethylphenol	9.875	122	87118	8.514	ng	98
28) bis(2-Chloroethoxy)met...	10.099	93	115003	8.608	ng	98
29) 2,4-Dichlorophenol	10.369	162	94248	7.665	ng	99
30) 1,2,4-Trichlorobenzene	10.546	180	105150	7.611	ng	94
31) Naphthalene	10.734	128	366357	9.317	ng	99
32) Benzoic acid	10.034	122	35675	5.435	ng	91
33) 4-Chloroaniline	10.863	127	133676	8.415	ng	98
34) Hexachlorobutadiene	11.010	225	56390	6.034	ng	97
35) Caprolactam	11.675	113	29082	8.211	ng	# 74
36) 4-Chloro-3-methylphenol	11.999	107	106960	8.567	ng	97
37) 2-Methylnaphthalene	12.346	142	249174	8.900	ng	97
38) 1-Methylnaphthalene	12.563	142	240519	9.121	ng	99
40) 1,2,4,5-Tetrachloroben...	12.716	216	97519	6.970	ng	97
43) 2,4,6-Trichlorophenol	12.969	196	66625	6.943	ng	93
44) 2,4,5-Trichlorophenol	13.057	196	72788	7.043	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006128.D
 Acq On : 07 Jul 2021 11:54
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Manual Integrations
 APPROVED

mohammad
 7/8/2021 12:38:53 PM

Quant Time: Jul 07 13:50:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.351	154	307792	9.604	ng	99
47) 2-Chloronaphthalene	13.393	162	234476	8.960	ng	95
48) 2-Nitroaniline	13.616	65	56214	8.170	ng	94
49) Acenaphthylene	14.240	152	378100	9.417	ng	99
50) Dimethylphthalate	13.975	163	291003	8.904	ng	99
51) 2,6-Dinitrotoluene	14.098	165	63010	8.483	ng	99
52) Acenaphthene	14.575	154	232808	9.548	ng	98
53) 3-Nitroaniline	14.445	138	61476	8.539	ng	95
54) 2,4-Dinitrophenol	14.693	184	12343	3.258	ng	94
55) Dibenzofuran	14.916	168	356432	8.889	ng	98
56) 4-Nitrophenol	14.822	139	23745	4.069	ng	# 81
57) 2,4-Dinitrotoluene	14.904	165	83365	8.391	ng	# 89
58) Fluorene	15.563	166	288295	9.228	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.151	232	55630	6.107	ng	# 72
60) Diethylphthalate	15.334	149	301193	9.186	ng	98
61) 4-Chlorophenyl-phenyle...	15.551	204	125708	7.760	ng	# 89
62) 4-Nitroaniline	15.610	138	49735	6.712	ng	# 81
63) Azobenzene	15.851	77	252305	8.505	ng	96
65) 4,6-Dinitro-2-methylph...	15.675	198	31610	5.727	ng	84
66) n-Nitrosodiphenylamine	15.775	169	250367	10.051	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	72648	7.535	ng	# 90
68) Hexachlorobenzene	16.569	284	84552	7.317	ng	# 90
69) Atrazine	16.734	200	76064	9.815	ng	90
70) Pentachlorophenol	16.934	266	29592	4.605	ng	99
71) Phenanthrene	17.304	178	437074	9.689	ng	99
72) Anthracene	17.398	178	423472	9.539	ng	99
73) Carbazole	17.681	167	417739	9.801	ng	99
74) Di-n-butylphthalate	18.216	149	497627	10.077	ng	# 99
75) Fluoranthene	19.275	202	467148	8.528	ng	95
77) Benzidine	19.463	184	196282	13.119	ng	98
78) Pyrene	19.628	202	489138	9.744	ng	100
80) Butylbenzylphthalate	20.486	149	244536	11.841	ng	98
81) Benzo(a)anthracene	21.328	228	467552	9.008	ng	99
82) 3,3'-Dichlorobenzidine	21.269	252	130434	7.269	ng	# 97
83) Chrysene	21.380	228	463746	9.225	ng	98
84) Bis(2-ethylhexyl)phtha...	21.245	149	366653	12.106	ng	# 98
85) Di-n-octyl phthalate	22.157	149	660045	12.182	ng	99
87) Indeno(1,2,3-cd)pyrene	26.257	276	597224	8.560	ng	# 87
88) Benzo(b)fluoranthene	23.004	252	487772	8.729	ng	# 94
89) Benzo(k)fluoranthene	23.057	252	511272	9.436	ng	# 95
90) Benzo(a)pyrene	23.633	252	469771	8.942	ng	# 94
91) Dibenzo(a,h)anthracene	26.274	278	510822	8.507	ng	# 92
92) Benzo(g,h,i)perylene	27.039	276	514253	8.669	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006128.D
 Acq On : 07 Jul 2021 11:54
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Manual Integrations
 APPROVED

mohammad
 7/8/2021 12:38:53 PM

Quant Time: Jul 07 13:50:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.875 min Scan# 814
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:152 Resp: 167698
Ion Ratio Lower Upper
152 100
150 153.5 125.6 188.4
115 59.6 48.5 72.7

Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#2
1,4-Dioxane
Concen: 9.269 ng
RT: 3.323 min Scan# 40
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion: 88 Resp: 43765
Ion Ratio Lower Upper
88 100
58 67.5 46.7 70.1
43 23.4 20.3 30.5





#3
Pyridine
Concen: 8.720 ng
RT: 3.758 min Scan# 114
Delta R.T. 0.018 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion: 79 Resp: 96766
Ion Ratio Lower Upper
79 100
52 65.0 41.7 62.5#
51 35.7 24.1 36.1




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#4
n-Nitrosodimethylamine
Concen: 7.864 ng
RT: 3.658 min Scan# 97
Delta R.T. 0.012 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion: 42 Resp: 40563
Ion Ratio Lower Upper
42 100
74 143.2 105.4 158.0
44 12.5 4.2 6.4#









#5
2-Fluorophenol
Concen: 19.923 ng
RT: 5.458 min Scan# 403
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:112 Resp: 208217
Ion Ratio Lower Upper
112 100
64 52.1 41.4 62.2
63 26.8 23.8 35.8




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#6
Aniline
Concen: 9.307 ng
RT: 7.211 min Scan# 701
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion: 93 Resp: 140752
Ion Ratio Lower Upper
93 100
66 36.7 29.7 44.5
65 19.3 15.3 22.9





#7
Phenol-d6
Concen: 19.639 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion: 99 Resp: 266025
Ion Ratio Lower Upper
99 100
42 16.2 13.4 20.0
71 32.8 30.6 46.0

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#8
2-Chlorophenol
Concen: 10.134 ng
RT: 7.446 min Scan# 741
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion: 128 Resp: 121427
Ion Ratio Lower Upper
128 100
130 31.1 12.4 52.4
64 40.3 20.3 60.3





#9
Benzaldehyde
Concen: 13.930 ng
RT: 7.028 min Scan# 670
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion: 77 Resp: 80439
Ion Ratio Lower Upper
77 100
105 105.6 76.7 116.7
106 99.3 76.2 116.2

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#10
Phenol
Concen: 9.516 ng
RT: 7.093 min Scan# 681
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54



Tgt Ion: 94 Resp: 128768
Ion Ratio Lower Upper
94 100
65 25.6 8.8 48.8
66 36.7 20.7 60.7





#11
bis(2-Chloroethyl)ether
Concen: 9.981 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54





Tgt Ion: 93 Resp: 102346

Ion	Ratio	Lower	Upper
93	100		
63	67.8	46.2	86.2
95	31.3	12.2	52.2

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED




mohammad
7/8/2021 12:38:53 PM




#12
1,3-Dichlorobenzene
Concen: 9.698 ng
RT: 7.763 min Scan# 795
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion: 146 Resp: 129947

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.5	77.3
75	27.9	25.7	38.5









#13
1,4-Dichlorobenzene
Concen: 9.676 ng
RT: 7.916 min Scan# 821
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:146 Resp: 132219
Ion Ratio Lower Upper
146 100
148 66.0 51.3 76.9
111 41.5 33.0 49.6

Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#14
1,2-Dichlorobenzene
Concen: 9.777 ng
RT: 8.228 min Scan# 874
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:146 Resp: 127702
Ion Ratio Lower Upper
146 100
148 62.6 50.3 75.5
111 40.9 33.7 50.5






#15
Benzyl Alcohol
Concen: 7.230 ng m
RT: 8.140 min Scan# 859
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED

mohammad
7/8/2021 12:38:53 PM







Tgt Ion: 79 Resp: 73768
Ion Ratio Lower Upper
79 100
108 85.9 62.5 93.7
77 68.3 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 12.662 ng
RT: 8.411 min Scan# 905
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion: 45 Resp: 120020
Ion Ratio Lower Upper
45 100
77 18.0 2.0 42.0
79 12.8 0.0 36.7





#17
2-Methylphenol
Concen: 9.519 ng
RT: 8.346 min Scan# 894
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:	107	Resp:	87767
Ion	Ratio	Lower	Upper
107	100		
108	109.7	87.7	131.5
77	42.8	40.2	60.2
79	42.8	39.0	58.4




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#18
Hexachloroethane
Concen: 8.840 ng
RT: 8.952 min Scan# 997
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:	117	Resp:	43695
Ion	Ratio	Lower	Upper
117	100		
119	97.5	76.2	114.2
201	87.9	83.4	125.0





#19
n-Nitroso-di-n-propylamine
Concen: 8.970 ng
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:	70	Resp:	74746
Ion	Ratio	Lower	Upper
70	100		
42	53.7	40.8	61.2
101	11.2	8.8	13.2
130	24.9	18.9	28.3

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#20
3+4-Methylphenols
Concen: 9.384 ng
RT: 8.681 min Scan# 951
Delta R.T. 0.012 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:	107	Resp:	116875
Ion	Ratio	Lower	Upper
107	100		
108	88.8	71.0	111.0
77	31.9	16.6	56.6
79	25.6	12.7	52.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.681 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:	136	Resp:	679606
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	6.4	4.6	6.8
68	5.5	3.6	5.4




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#22
Acetophenone
Concen: 8.880 ng
RT: 8.711 min Scan# 956
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:	105	Resp:	158844
Ion	Ratio	Lower	Upper
105	100		
71	2.9	1.9	2.9
51	24.4	18.4	27.6
120	22.1	17.9	26.9





#23
Nitrobenzene-d5
Concen: 15.686 ng
RT: 9.046 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion: 82 Resp: 217439
Ion Ratio Lower Upper
82 100
128 47.1 35.7 53.5
54 45.9 35.0 52.4

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#24
Nitrobenzene
Concen: 7.814 ng
RT: 9.093 min Scan# 1021
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion: 77 Resp: 112473
Ion Ratio Lower Upper
77 100
123 51.0 38.2 57.2
65 12.4 11.4 17.2





#25
Isophorone
Concen: 8.091 ng
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion: 82 Resp: 207120
Ion Ratio Lower Upper
82 100
95 7.7 6.5 9.7
138 20.5 15.0 22.4

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#26
2-Nitrophenol
Concen: 8.123 ng
RT: 9.805 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion: 139 Resp: 55787
Ion Ratio Lower Upper
139 100
109 31.6 28.8 43.2
65 42.9 41.0 61.6





#27
2,4-Dimethylphenol
Concen: 8.514 ng
RT: 9.875 min Scan# 1154
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:122 Resp: 87118
Ion Ratio Lower Upper
122 100
107 115.6 94.5 141.7
121 61.3 47.4 71.2

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#28
bis(2-Chloroethoxy)methane
Concen: 8.608 ng
RT: 10.099 min Scan# 1192
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54



Tgt Ion: 93 Resp: 115003
Ion Ratio Lower Upper
93 100
95 31.3 25.9 38.9
123 11.7 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 7.665 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.024 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:	162	Resp:	94248
Ion Ratio	Lower	Upper	
162	100		
164	63.7	44.9	84.9
98	34.4	14.3	54.3

Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#30
1,2,4-Trichlorobenzene
Concen: 7.611 ng
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:	180	Resp:	105150
Ion Ratio	Lower	Upper	
180	100		
182	91.8	77.9	116.9
145	34.6	24.7	37.1





#31
Naphthalene
Concen: 9.317 ng
RT: 10.734 min Scan# 1300
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:128 Resp: 366357
Ion Ratio Lower Upper
128 100
129 10.4 9.0 13.6
127 13.1 10.7 16.1




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#32
Benzoic acid
Concen: 5.435 ng
RT: 10.034 min Scan# 1181
Delta R.T. -0.029 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:122 Resp: 35675
Ion Ratio Lower Upper
122 100
105 113.4 103.2 143.2
77 86.6 76.2 116.2





#33
4-Chloroaniline
Concen: 8.415 ng
RT: 10.863 min Scan# 1322
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:	127	Resp:	133676
Ion	Ratio	Lower	Upper
127	100		
129	32.3	26.1	39.1
65	25.9	22.5	33.7
92	17.8	15.0	22.6




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#34
Hexachlorobutadiene
Concen: 6.034 ng
RT: 11.010 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:	225	Resp:	56390
Ion	Ratio	Lower	Upper
225	100		
223	65.9	51.0	76.6
227	62.1	51.4	77.2





#35
Caprolactam
Concen: 8.211 ng
RT: 11.675 min Scan# 1460
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:113 Resp: 29082
Ion Ratio Lower Upper
113 100
55 149.0 98.9 138.9#
56 114.5 71.7 111.7#




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#36
4-Chloro-3-methylphenol
Concen: 8.567 ng
RT: 11.999 min Scan# 1515
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:107 Resp: 106960
Ion Ratio Lower Upper
107 100
144 24.8 21.1 31.7
142 80.1 66.0 99.0





#37
2-Methylnaphthalene
Concen: 8.900 ng
RT: 12.346 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDICC010



Tgt Ion:142 Resp: 249174
Ion Ratio Lower Upper
142 100
141 90.3 71.8 107.6
115 34.3 31.6 47.4




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#38
1-Methylnaphthalene
Concen: 9.121 ng
RT: 12.563 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:142 Resp: 240519
Ion Ratio Lower Upper
142 100
141 93.0 73.3 109.9
116 3.8 3.5 5.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:164 Resp: 397479
Ion Ratio Lower Upper
164 100
162 102.4 80.2 120.2
160 44.8 34.5 51.7

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 6.970 ng
RT: 12.716 min Scan# 1637
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:216 Resp: 97519
Ion Ratio Lower Upper
216 100
214 78.9 61.9 92.9
179 24.2 16.8 25.2
108 20.9 14.9 22.3





#42
2,4,6-Tribromophenol
Concen: 12.539 ng
RT: 16.010 min Scan# 2197
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:330 Resp: 85532
Ion Ratio Lower Upper
330 100
332 98.1 77.8 116.8
141 40.4 25.4 38.2#

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#43
2,4,6-Trichlorophenol
Concen: 6.943 ng
RT: 12.969 min Scan# 1680
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54



Tgt Ion:196 Resp: 66625
Ion Ratio Lower Upper
196 100
198 90.6 79.5 119.3
200 30.6 25.0 37.6









#44
2,4,5-Trichlorophenol
Concen: 7.043 ng
RT: 13.057 min Scan# 1695
Delta R.T. 0.012 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:196	Resp:	72788
Ion Ratio	Lower	Upper
196	100	
198	95.6	74.7 112.1
97	49.0	37.4 56.0
132	28.3	20.0 30.0




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#45
2-Fluorobiphenyl
Concen: 18.303 ng
RT: 13.140 min Scan# 1709
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:172	Resp:	554037
Ion Ratio	Lower	Upper
172	100	
171	35.3	27.9 41.9
170	22.8	18.6 27.8









#46
1,1'-Biphenyl
Concen: 9.604 ng
RT: 13.351 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:154 Resp: 307792
Ion Ratio Lower Upper
154 100
153 40.5 20.4 60.4
76 11.9 0.0 30.8

Instrument :
BNA_P
ClientSampleId :
SSTDICC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#47
2-Chloronaphthalene
Concen: 8.960 ng
RT: 13.393 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:162 Resp: 234476
Ion Ratio Lower Upper
162 100
127 42.6 31.4 47.0
164 31.3 26.6 39.8





#48
2-Nitroaniline
Concen: 8.170 ng
RT: 13.616 min Scan# 1790
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion: 65 Resp: 56214
Ion Ratio Lower Upper
65 100
92 75.7 57.7 86.5
138 123.0 92.1 138.1




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#49
Acenaphthylene
Concen: 9.417 ng
RT: 14.240 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion: 152 Resp: 378100
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 12.8 10.6 16.0









#50
Dimethylphthalate
Concen: 8.904 ng
RT: 13.975 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:163 Resp: 291003
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.5 8.2 12.2

Instrument :
BNA_P
ClientSampleId :
SSTDICC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#51
2,6-Dinitrotoluene
Concen: 8.483 ng
RT: 14.098 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:165 Resp: 63010
Ion Ratio Lower Upper
165 100
63 52.3 41.3 61.9
89 55.0 44.8 67.2





#52
Acenaphthene
Concen: 9.548 ng
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:154 Resp: 232808
Ion Ratio Lower Upper
154 100
153 114.1 89.4 134.2
152 55.0 43.0 64.4

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#53
3-Nitroaniline
Concen: 8.539 ng
RT: 14.445 min Scan# 1931
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:138 Resp: 61476
Ion Ratio Lower Upper
138 100
108 8.6 8.4 12.6
92 104.6 87.8 131.8





#54
2,4-Dinitrophenol
Concen: 3.258 ng
RT: 14.693 min Scan# 1973
Delta R.T. 0.030 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:184 Resp: 12343
Ion Ratio Lower Upper
184 100
63 55.0 47.0 70.6
154 70.5 52.3 78.5

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#55
Dibenzofuran
Concen: 8.889 ng
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54



Tgt Ion:168 Resp: 356432
Ion Ratio Lower Upper
168 100
139 40.3 31.0 46.6
169 13.7 10.8 16.2









#56
4-Nitrophenol
Concen: 4.069 ng
RT: 14.822 min Scan# 1995
Delta R.T. 0.041 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:139 Resp: 23745
Ion Ratio Lower Upper
139 100
109 68.1 72.3 112.3#
65 81.2 72.5 112.5

Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#57
2,4-Dinitrotoluene
Concen: 8.391 ng
RT: 14.904 min Scan# 2009
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:165 Resp: 83365
Ion Ratio Lower Upper
165 100
63 52.2 32.3 48.5#
89 75.3 55.9 83.9
182 3.1 2.2 3.4





#58
Fluorene
Concen: 9.228 ng
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:166 Resp: 288295
Ion Ratio Lower Upper
166 100
165 98.0 78.0 117.0
167 13.5 10.5 15.7

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#59
2,3,4,6-Tetrachlorophenol
Concen: 6.107 ng
RT: 15.151 min Scan# 2051
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:232 Resp: 55630
Ion Ratio Lower Upper
232 100
131 58.6 32.2 48.4#
130 3.5 2.0 3.0#
166 38.8 19.4 29.2#





#60
Diethylphthalate
Concen: 9.186 ng
RT: 15.334 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:149 Resp: 301193
Ion Ratio Lower Upper
149 100
177 20.5 17.5 26.3
150 12.9 9.9 14.9




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#61
4-Chlorophenyl-phenylether
Concen: 7.760 ng
RT: 15.551 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:204 Resp: 125708
Ion Ratio Lower Upper
204 100
206 32.5 26.6 40.0
141 73.2 48.6 73.0#






#62
4-Nitroaniline
Concen: 6.712 ng
RT: 15.610 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Manual Integrations
APPROVED

mohammad
7/8/2021 12:38:53 PM







Tgt Ion:138 Resp: 49735
Ion Ratio Lower Upper
138 100
92 46.8 31.2 71.2
108 72.5 78.6 118.6#



#63
Azobenzene
Concen: 8.505 ng
RT: 15.851 min Scan# 2170
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion: 77 Resp: 252305
Ion Ratio Lower Upper
77 100
182 29.0 6.0 46.0
105 22.3 0.0 38.8
51 28.4 8.8 48.8





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDICC010



Tgt Ion:188 Resp: 751129
Ion Ratio Lower Upper
188 100
94 10.0 5.6 8.4#
80 10.6 6.3 9.5#




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#65
4,6-Dinitro-2-methylphenol
Concen: 5.727 ng
RT: 15.675 min Scan# 2140
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:198 Resp: 31610
Ion Ratio Lower Upper
198 100
51 44.6 14.0 54.0
105 48.3 19.5 59.5









#66
n-Nitrosodiphenylamine
Concen: 10.051 ng
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Tgt Ion:169 Resp: 250367
Ion Ratio Lower Upper
169 100
168 67.6 54.4 81.6
167 35.1 29.0 43.6

Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#67
4-Bromophenyl-phenylether
Concen: 7.535 ng
RT: 16.451 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:248 Resp: 72648
Ion Ratio Lower Upper
248 100
250 97.1 79.8 119.8
141 89.6 57.8 86.6#





#68
Hexachlorobenzene
Concen: 7.317 ng
RT: 16.569 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:284 Resp: 84552
Ion Ratio Lower Upper
284 100
142 43.4 26.9 40.3#
249 31.2 24.5 36.7

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#69
Atrazine
Concen: 9.815 ng
RT: 16.734 min Scan# 2320
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:200 Resp: 76064
Ion Ratio Lower Upper
200 100
173 30.1 4.1 44.1
215 41.7 27.7 67.7





#70
Pentachlorophenol
Concen: 4.605 ng
RT: 16.934 min Scan# 2354
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:266 Resp: 29592
Ion Ratio Lower Upper
266 100
268 65.0 51.6 77.4
264 64.3 50.3 75.5

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#71
Phenanthrene
Concen: 9.689 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:178 Resp: 437074
Ion Ratio Lower Upper
178 100
176 19.7 15.1 22.7
179 15.9 12.4 18.6





#72
 Anthracene
 Concen: 9.539 ng
 RT: 17.398 min Scan# 2433
 Delta R.T. 0.000 min
 Lab File: BP006128.D
 Acq: 07 Jul 2021 11:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010




Tgt Ion:178 Resp: 423472
 Ion Ratio Lower Upper
 178 100
 176 19.0 14.9 22.3
 179 15.1 12.2 18.2

Manual Integrations
 APPROVED



mohammad
 7/8/2021 12:38:53 PM




#73
 Carbazole
 Concen: 9.801 ng
 RT: 17.681 min Scan# 2481
 Delta R.T. 0.006 min
 Lab File: BP006128.D
 Acq: 07 Jul 2021 11:54




Tgt Ion:167 Resp: 417739
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 14.6 10.7 16.1





#74
Di-n-butylphthalate
Concen: 10.077 ng
RT: 18.216 min Scan# 2572
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:149 Resp: 497627
Ion Ratio Lower Upper
149 100
150 9.0 7.3 10.9
104 4.4 4.5 6.7#

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#75
Fluoranthene
Concen: 8.528 ng
RT: 19.275 min Scan# 2752
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:202 Resp: 467148
Ion Ratio Lower Upper
202 100
101 13.0 0.0 27.6
203 18.1 0.0 37.8





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:240 Resp: 719031
Ion Ratio Lower Upper
240 100
120 13.9 6.5 9.7#
236 25.6 20.6 31.0

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#77
Benzidine
Concen: 13.119 ng
RT: 19.463 min Scan# 2784
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:184 Resp: 196282
Ion Ratio Lower Upper
184 100
185 12.4 11.0 16.4
183 12.4 9.5 14.3





#78
Pyrene
Concen: 9.744 ng
RT: 19.628 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:202 Resp: 489138
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.7 14.2 21.2

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#79
Terphenyl-d14
Concen: 18.991 ng
RT: 19.816 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:244 Resp: 729267
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 13.1 7.1 10.7#





#80
Butylbenzylphthalate
Concen: 11.841 ng
RT: 20.486 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:149 Resp: 244536
Ion Ratio Lower Upper
149 100
91 67.6 54.6 81.8
206 17.3 16.7 25.1

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#81
Benzo(a)anthracene
Concen: 9.008 ng
RT: 21.328 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:228 Resp: 467552
Ion Ratio Lower Upper
228 100
226 26.4 21.6 32.4
229 20.0 15.8 23.6





#82
3,3'-Dichlorobenzidine
Concen: 7.269 ng
RT: 21.269 min Scan# 3091
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:252 Resp: 130434
Ion Ratio Lower Upper
252 100
254 64.2 50.6 76.0
126 14.0 7.0 10.6#

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#83
Chrysene
Concen: 9.225 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:228 Resp: 463746
Ion Ratio Lower Upper
228 100
226 28.8 23.8 35.6
229 20.0 15.4 23.2





#84
Bis(2-ethylhexyl)phthalate
Concen: 12.106 ng
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:149 Resp: 366653
Ion Ratio Lower Upper
149 100
167 26.7 21.9 32.9
279 3.7 4.3 6.5#




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#85
Di-n-octyl phthalate
Concen: 12.182 ng
RT: 22.157 min Scan# 3242
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:149 Resp: 660045
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 7.6 5.6 8.4





#86
Perylene-d12
Concen: 20.000 ng
RT: 23.739 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010




Tgt Ion:264 Resp: 815835
Ion Ratio Lower Upper
264 100
260 23.6 18.6 28.0
265 22.1 17.6 26.4

Manual Integrations
APPROVED



mohammad
7/8/2021 12:38:53 PM




#87
Indeno(1,2,3-cd)pyrene
Concen: 8.560 ng
RT: 26.257 min Scan# 3939
Delta R.T. -0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Tgt Ion:276 Resp: 597224
Ion Ratio Lower Upper
276 100
138 32.7 15.3 22.9#
277 25.7 20.4 30.6





#88
Benzo(b)fluoranthene
Concen: 8.729 ng
RT: 23.004 min Scan# 3386
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:252 Resp: 487772
Ion Ratio Lower Upper
252 100
253 22.6 17.5 26.3
125 14.2 6.2 9.4#




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#89
Benzo(k)fluoranthene
Concen: 9.436 ng
RT: 23.057 min Scan# 3395
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:252 Resp: 511272
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 14.2 6.0 9.0#





#90
Benzo(a)pyrene
Concen: 8.942 ng
RT: 23.633 min Scan# 3493
Delta R.T. 0.000 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54




Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion:252 Resp: 469771
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 15.2 6.5 9.7#




Manual Integrations
APPROVED


mohammad
7/8/2021 12:38:53 PM



#91
Dibenzo(a,h)anthracene
Concen: 8.507 ng
RT: 26.274 min Scan# 3942
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54


Tgt Ion:278 Resp: 510822
Ion Ratio Lower Upper
278 100
139 19.9 9.3 13.9#
279 23.1 18.8 28.2





#92
Benzo(g,h,i)perylene
Concen: 8.669 ng
RT: 27.039 min Scan# 4072
Delta R.T. 0.006 min
Lab File: BP006128.D
Acq: 07 Jul 2021 11:54

Instrument :
BNA_P
ClientSampleId :
SSTDIC010



Tgt Ion: 276 Resp: 514253
Ion Ratio Lower Upper
276 100
277 23.0 18.9 28.3
138 28.2 12.6 18.8#

Manual Integrations
APPROVED

mohammad
7/8/2021 12:38:53 PM

