



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006126.D
 Acq On : 07 Jul 2021 10:46
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 07 16:08:13 2021






AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	72958	PASS
68	69	0.00	2	1.9	1464	PASS
69	198	0.00	100	39.3	76199	PASS
70	69	0.00	2	0.4	303	PASS
127	198	10	80	55.6	107840	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	193973	PASS
199	198	5	9	7.0	13564	PASS
275	198	10	60	26.5	51377	PASS
365	198	1	100	4.1	7891	PASS
441	198	0.01	100	12.2	23609	PASS
442	442	100	100	100.0	152496	PASS
443	442	15	24	19.7	30095	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006126.D
 Acq On : 07 Jul 2021 10:46
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Jul 07 11:28:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 11:23:35 2021
 Response via : Initial Calibration



TIC: BP006126.D\data.ms

(77) Benzidine

19.463min (+ 0.000) 496360.91 ng




response 1445959

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.04
183.00	11.90	11.96
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
Data File : BP006126.D
Acq On : 07 Jul 2021 10:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jul 07 11:28:29 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 07 11:23:35 2021
Response via : Initial Calibration



TIC: BP006126.D\data.ms

(70) Pentachlorophenol (C)

16.933min (+ 0.000) 1157649.88 ng

response 168351

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	65.75
264.00	62.90	65.22
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/7/2021	BNA_P	BP006126.D
Compound Name	Response	Retention Time
DDT	684550	20.65
DDD	2253	20.23
DDE	2479	19.73
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4732	689282	0.69