

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
 Data File : BP006137.D
 Acq On : 08 Jul 2021 13:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 08 14:20:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	183240	20.000	ng	0.00
21) Naphthalene-d8	10.675	136	718031	20.000	ng	# 0.00
39) Acenaphthene-d10	14.516	164	407849	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	759799	20.000	ng	# 0.00
76) Chrysene-d12	21.345	240	717523	20.000	ng	# 0.00
86) Perylene-d12	23.745	264	792326	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.452	112	884164	83.024	ng	0.00
7) Phenol-d6	7.058	99	1168330	82.940	ng	0.00
23) Nitrobenzene-d5	9.046	82	942901	82.947	ng	0.00
42) 2,4,6-Tribromophenol	16.010	330	377157	77.673	ng	0.00
45) 2-Fluorobiphenyl	13.140	172	2217932	80.394	ng	0.00
79) Terphenyl-d14	19.816	244	2944557	80.911	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.317	88	173772	41.587	ng	90
3) Pyridine	3.734	79	469188	44.192	ng	# 88
4) n-Nitrosodimethylamine	3.640	42	178014	42.628	ng	# 89
6) Aniline	7.205	93	612532	41.213	ng	98
8) 2-Chlorophenol	7.446	128	507265	40.670	ng	98
9) Benzaldehyde	7.022	77	308478	40.125	ng	94
10) Phenol	7.087	94	558026	41.247	ng	93
11) bis(2-Chloroethyl)ether	7.305	93	415080	40.603	ng	99
12) 1,3-Dichlorobenzene	7.764	146	531897	39.800	ng	97
13) 1,4-Dichlorobenzene	7.911	146	540146	39.574	ng	98
14) 1,2-Dichlorobenzene	8.228	146	517463	39.824	ng	98
15) Benzyl Alcohol	8.128	79	358179	42.059	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.411	45	499792	41.165	ng	89
17) 2-Methylphenol	8.334	107	379816	41.269	ng	95
18) Hexachloroethane	8.946	117	183745	40.438	ng	91
19) n-Nitroso-di-n-propyla...	8.693	70	317509	41.345	ng	96
20) 3+4-Methylphenols	8.669	107	515206	41.436	ng	94
22) Acetophenone	8.711	105	672508	41.150	ng	# 98
24) Nitrobenzene	9.087	77	480589	41.803	ng	94
25) Isophorone	9.616	82	891296	41.303	ng	97
26) 2-Nitrophenol	9.799	139	264132	41.327	ng	89
27) 2,4-Dimethylphenol	9.869	122	391047	41.449	ng	96
28) bis(2-Chloroethoxy)met...	10.093	93	496427	41.510	ng	98
29) 2,4-Dichlorophenol	10.346	162	426014	41.097	ng	98
30) 1,2,4-Trichlorobenzene	10.546	180	440827	39.942	ng	98
31) Naphthalene	10.728	128	1505531	40.221	ng	100
32) Benzoic acid	10.063	122	264701	38.535	ng	94
33) 4-Chloroaniline	10.852	127	597480	41.350	ng	99
34) Hexachlorobutadiene	11.005	225	237483	39.080	ng	98
35) Caprolactam	11.669	113	137661	42.127	ng	88
36) 4-Chloro-3-methylphenol	11.993	107	465254	41.563	ng	99
37) 2-Methylnaphthalene	12.340	142	1044847	40.231	ng	97
38) 1-Methylnaphthalene	12.557	142	984614	40.060	ng	97
40) 1,2,4,5-Tetrachloroben...	12.710	216	409122	40.439	ng	97
41) Hexachlorocyclopentadiene	12.681	237	61461	39.311	ng	98
43) 2,4,6-Trichlorophenol	12.963	196	300211	41.016	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
 Data File : BP006137.D
 Acq On : 08 Jul 2021 13:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 08 14:20:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.046	196	316164	40.261	ng	# 92
46) 1,1'-Biphenyl	13.351	154	1241169	40.535	ng	99
47) 2-Chloronaphthalene	13.393	162	947552	40.226	ng	96
48) 2-Nitroaniline	13.610	65	250104	42.728	ng	90
49) Acenaphthylene	14.240	152	1540493	40.614	ng	100
50) Dimethylphthalate	13.981	163	1174079	40.033	ng	99
51) 2,6-Dinitrotoluene	14.104	165	272469	41.208	ng	95
52) Acenaphthene	14.575	154	921868	39.955	ng	98
53) 3-Nitroaniline	14.440	138	280359	41.851	ng	96
54) 2,4-Dinitrophenol	14.663	184	115554	38.455	ng	96
55) Dibenzofuran	14.916	168	1443235	40.182	ng	98
56) 4-Nitrophenol	14.781	139	180869	39.041	ng	84
57) 2,4-Dinitrotoluene	14.898	165	367517	41.270	ng	98
58) Fluorene	15.563	166	1165726	40.399	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.151	232	245227	39.909	ng	# 72
60) Diethylphthalate	15.334	149	1215039	40.398	ng	98
61) 4-Chlorophenyl-phenyle...	15.551	204	499618	39.496	ng	# 89
62) 4-Nitroaniline	15.604	138	259954	38.603	ng	# 85
63) Azobenzene	15.851	77	1058660	42.162	ng	95
65) 4,6-Dinitro-2-methylph...	15.663	198	191324	39.436	ng	83
66) n-Nitrosodiphenylamine	15.775	169	1005025	40.624	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	293448	39.178	ng	# 89
68) Hexachlorobenzene	16.569	284	344683	38.769	ng	# 93
69) Atrazine	16.734	200	314829	40.534	ng	94
70) Pentachlorophenol	16.928	266	169799	41.380	ng	99
71) Phenanthrene	17.304	178	1732695	40.380	ng	100
72) Anthracene	17.398	178	1695375	40.330	ng	99
73) Carbazole	17.675	167	1705130	40.934	ng	99
74) Di-n-butylphthalate	18.216	149	2061725	41.118	ng	99
75) Fluoranthene	19.275	202	1908822	40.588	ng	96
77) Benzidine	19.457	184	945508	44.439	ng	99
78) Pyrene	19.622	202	1987487	40.981	ng	99
80) Butylbenzylphthalate	20.486	149	986582	41.012	ng	97
81) Benzo(a)anthracene	21.328	228	1829972	40.324	ng	100
82) 3,3'-Dichlorobenzidine	21.263	252	537414	40.368	ng	# 97
83) Chrysene	21.380	228	1792251	39.963	ng	99
84) Bis(2-ethylhexyl)phtha...	21.245	149	1456918	41.032	ng	# 98
85) Di-n-octyl phthalate	22.157	149	2547825	40.741	ng	99
87) Indeno(1,2,3-cd)pyrene	26.268	276	2380305	40.748	ng	# 88
88) Benzo(b)fluoranthene	23.010	252	2009085	41.802	ng	# 96
89) Benzo(k)fluoranthene	23.057	252	1898435	40.412	ng	# 96
90) Benzo(a)pyrene	23.639	252	1853064	41.086	ng	# 95
91) Dibenzo(a,h)anthracene	26.274	278	2046756	40.691	ng	# 92
92) Benzo(g,h,i)perylene	27.039	276	2004583	40.494	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
 Data File : BP006137.D
 Acq On : 08 Jul 2021 13:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jul 08 14:20:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.875 min Scan# 814
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 183240
Ion Ratio Lower Upper
152 100
150 156.9 125.6 188.4
115 61.0 48.5 72.7



#2
1,4-Dioxane
Concen: 41.587 ng
RT: 3.317 min Scan# 39
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 88 Resp: 173772
Ion Ratio Lower Upper
88 100
58 68.2 46.7 70.1
43 24.6 20.3 30.5









#3
Pyridine
Concen: 44.192 ng
RT: 3.734 min Scan# 110
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 469188
Ion Ratio Lower Upper
79 100
52 63.8 41.7 62.5#
51 32.5 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 42.628 ng
RT: 3.640 min Scan# 94
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 42 Resp: 178014
Ion Ratio Lower Upper
42 100
74 144.6 105.4 158.0
44 7.3 4.2 6.4#









#5
2-Fluorophenol
Concen: 83.024 ng
RT: 5.452 min Scan# 402
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 884164
Ion Ratio Lower Upper
112 100
64 50.3 41.4 62.2
63 25.2 23.8 35.8



#6
Aniline
Concen: 41.213 ng
RT: 7.205 min Scan# 700
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 93 Resp: 612532
Ion Ratio Lower Upper
93 100
66 36.1 29.7 44.5
65 18.4 15.3 22.9









#7
Phenol-d6
Concen: 82.940 ng
RT: 7.058 min Scan# 675
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 1168330
Ion Ratio Lower Upper
99 100
42 15.8 13.4 20.0
71 32.2 30.6 46.0



#8
2-Chlorophenol
Concen: 40.670 ng
RT: 7.446 min Scan# 741
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 128 Resp: 507265
Ion Ratio Lower Upper
128 100
130 32.8 12.4 52.4
64 38.0 20.3 60.3









#9
Benzaldehyde
Concen: 40.125 ng
RT: 7.022 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 308478
Ion Ratio Lower Upper
77 100
105 105.7 76.7 116.7
106 99.4 76.2 116.2



#10
Phenol
Concen: 41.247 ng
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 94 Resp: 558026
Ion Ratio Lower Upper
94 100
65 25.6 8.8 48.8
66 35.5 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 40.603 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 415080
Ion Ratio Lower Upper
93 100
63 67.8 46.2 86.2
95 32.2 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 39.800 ng
RT: 7.764 min Scan# 795
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:146 Resp: 531897
Ion Ratio Lower Upper
146 100
148 65.0 51.5 77.3
75 27.6 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 39.574 ng
RT: 7.911 min Scan# 820
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 540146
Ion Ratio Lower Upper
146 100
148 65.1 51.3 76.9
111 40.0 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 39.824 ng
RT: 8.228 min Scan# 874
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:146 Resp: 517463
Ion Ratio Lower Upper
146 100
148 64.7 50.3 75.5
111 41.6 33.7 50.5









#15
Benzyl Alcohol
Concen: 42.059 ng
RT: 8.128 min Scan# 857
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 358179
Ion Ratio Lower Upper
79 100
108 83.2 62.5 93.7
77 62.5 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.165 ng
RT: 8.411 min Scan# 905
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 45 Resp: 499792
Ion Ratio Lower Upper
45 100
77 16.3 2.0 42.0
79 12.7 0.0 36.7









#17
2-Methylphenol
Concen: 41.269 ng
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	379816
Ion	Ratio	Lower	Upper
107	100		
108	110.5	87.7	131.5
77	41.9	40.2	60.2
79	42.6	39.0	58.4



#18
Hexachloroethane
Concen: 40.438 ng
RT: 8.946 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	117	Resp:	183745
Ion	Ratio	Lower	Upper
117	100		
119	96.4	76.2	114.2
201	88.3	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 41.345 ng
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	317509
Ion	Ratio	Lower	Upper
70	100		
42	53.2	40.8	61.2
101	11.0	8.8	13.2
130	26.6	18.9	28.3



#20
3+4-Methylphenols
Concen: 41.436 ng
RT: 8.669 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	107	Resp:	515206
Ion	Ratio	Lower	Upper
107	100		
108	89.3	71.0	111.0
77	30.6	16.6	56.6
79	26.6	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.675 min Scan# 1290
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	718031
Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.6	12.8
54	7.0	4.6	6.8#
68	5.6	3.6	5.4#



#22
Acetophenone
Concen: 41.150 ng
RT: 8.711 min Scan# 956
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	105	Resp:	672508
Ion	Ratio	Lower	Upper
105	100		
71	1.4	1.9	2.9#
51	23.9	18.4	27.6
120	23.8	17.9	26.9









#23
Nitrobenzene-d5
Concen: 82.947 ng
RT: 9.046 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 942901
Ion Ratio Lower Upper
82 100
128 51.1 35.7 53.5
54 45.9 35.0 52.4



#24
Nitrobenzene
Concen: 41.803 ng
RT: 9.087 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 77 Resp: 480589
Ion Ratio Lower Upper
77 100
123 51.8 38.2 57.2
65 12.8 11.4 17.2









#25
Isophorone
Concen: 41.303 ng
RT: 9.616 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 891296




Ion	Ratio	Lower	Upper
82	100		
95	7.4	6.5	9.7
138	19.9	15.0	22.4




#26
2-Nitrophenol
Concen: 41.327 ng
RT: 9.799 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 139 Resp: 264132

Ion	Ratio	Lower	Upper
139	100		
109	30.0	28.8	43.2
65	43.4	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 41.449 ng
RT: 9.869 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 391047
Ion Ratio Lower Upper
122 100
107 113.1 94.5 141.7
121 61.1 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 41.510 ng
RT: 10.093 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 93 Resp: 496427
Ion Ratio Lower Upper
93 100
95 31.1 25.9 38.9
123 12.0 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 41.097 ng
RT: 10.346 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 426014
Ion Ratio Lower Upper
162 100
164 65.0 44.9 84.9
98 36.7 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 39.942 ng
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:180 Resp: 440827
Ion Ratio Lower Upper
180 100
182 96.5 77.9 116.9
145 33.2 24.7 37.1









#31
Naphthalene
Concen: 40.221 ng
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1505531
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.4 10.7 16.1



#32
Benzoic acid
Concen: 38.535 ng
RT: 10.063 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:122 Resp: 264701
Ion Ratio Lower Upper
122 100
105 119.8 103.2 143.2
77 86.1 76.2 116.2









#33
4-Chloroaniline
Concen: 41.350 ng
RT: 10.852 min Scan# 1320
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	597480
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.1	39.1
65	26.9	22.5	33.7
92	18.1	15.0	22.6



#34
Hexachlorobutadiene
Concen: 39.080 ng
RT: 11.005 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	225	Resp:	237483
Ion	Ratio	Lower	Upper
225	100		
223	61.2	51.0	76.6
227	63.6	51.4	77.2









#35
Caprolactam
Concen: 42.127 ng
RT: 11.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	113	Resp:	137661
Ion	Ratio	Lower	Upper
113	100		
55	132.3	98.9	138.9
56	103.1	71.7	111.7



#36
4-Chloro-3-methylphenol
Concen: 41.563 ng
RT: 11.993 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	107	Resp:	465254
Ion	Ratio	Lower	Upper
107	100		
144	26.2	21.1	31.7
142	83.8	66.0	99.0









#37
2-Methylnaphthalene
Concen: 40.231 ng
RT: 12.340 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 1044847
Ion Ratio Lower Upper
142 100
141 89.0 71.8 107.6
115 33.9 31.6 47.4



#38
1-Methylnaphthalene
Concen: 40.060 ng
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:142 Resp: 984614
Ion Ratio Lower Upper
142 100
141 94.1 73.3 109.9
116 3.7 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 407849
Ion Ratio Lower Upper
164 100
162 102.4 80.2 120.2
160 44.4 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.439 ng
RT: 12.710 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:216 Resp: 409122
Ion Ratio Lower Upper
216 100
214 78.8 61.9 92.9
179 24.8 16.8 25.2
108 20.4 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 39.311 ng
RT: 12.681 min Scan# 1631
Delta R.T. -0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	237	Resp:	61461
Ion Ratio	Lower	Upper	
237	100		
235	66.2	44.5	84.5
272	12.4	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 77.673 ng
RT: 16.010 min Scan# 2197
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	330	Resp:	377157
Ion Ratio	Lower	Upper	
330	100		
332	96.5	77.8	116.8
141	39.2	25.4	38.2#









#43
2,4,6-Trichlorophenol
Concen: 41.016 ng
RT: 12.963 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 300211
Ion Ratio Lower Upper
196 100
198 93.6 79.5 119.3
200 29.9 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 40.261 ng
RT: 13.046 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:196 Resp: 316164
Ion Ratio Lower Upper
196 100
198 98.4 74.7 112.1
97 51.6 37.4 56.0
132 33.9 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 80.394 ng
RT: 13.140 min Scan# 1709
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2217932
Ion Ratio Lower Upper
172 100
171 36.3 27.9 41.9
170 23.3 18.6 27.8



#46
1,1'-Biphenyl
Concen: 40.535 ng
RT: 13.351 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:154 Resp: 1241169
Ion Ratio Lower Upper
154 100
153 40.7 20.4 60.4
76 11.1 0.0 30.8









#47
2-Chloronaphthalene
Concen: 40.226 ng
RT: 13.393 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	162	Resp:	947552
Ion	Ratio	Lower	Upper
162	100		
127	43.0	31.4	47.0
164	33.0	26.6	39.8



#48
2-Nitroaniline
Concen: 42.728 ng
RT: 13.610 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	65	Resp:	250104
Ion	Ratio	Lower	Upper
65	100		
92	74.0	57.7	86.5
138	131.0	92.1	138.1









#49
Acenaphthylene
Concen: 40.614 ng
RT: 14.240 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1540493
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2
153 13.0 10.6 16.0



#50
Dimethylphthalate
Concen: 40.033 ng
RT: 13.981 min Scan# 1852
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:163 Resp: 1174079
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 9.7 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 41.208 ng
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 272469
Ion Ratio Lower Upper
165 100
63 47.2 41.3 61.9
89 52.7 44.8 67.2



#52
Acenaphthene
Concen: 39.955 ng
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:154 Resp: 921868
Ion Ratio Lower Upper
154 100
153 114.2 89.4 134.2
152 53.7 43.0 64.4









#53
3-Nitroaniline
Concen: 41.851 ng
RT: 14.440 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 280359
Ion Ratio Lower Upper
138 100
108 10.2 8.4 12.6
92 104.8 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 38.455 ng
RT: 14.663 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:184 Resp: 115554
Ion Ratio Lower Upper
184 100
63 63.2 47.0 70.6
154 67.1 52.3 78.5









#55
Dibenzenofuran
Concen: 40.182 ng
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1443235
Ion Ratio Lower Upper
168 100
139 39.9 31.0 46.6
169 13.0 10.8 16.2



#56
4-Nitrophenol
Concen: 39.041 ng
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:139 Resp: 180869
Ion Ratio Lower Upper
139 100
109 72.6 72.3 112.3
65 81.3 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 41.270 ng
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	165	Resp:	367517
Ion	Ratio	Lower	Upper
165	100		
63	42.7	32.3	48.5
89	69.2	55.9	83.9
182	2.9	2.2	3.4



#58
Fluorene
Concen: 40.399 ng
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	166	Resp:	1165726
Ion	Ratio	Lower	Upper
166	100		
165	98.5	78.0	117.0
167	13.3	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 39.909 ng
RT: 15.151 min Scan# 2051
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	232	Resp:	245227
Ion Ratio	Lower	Upper	
232	100		
131	58.9	32.2	48.4#
130	3.3	2.0	3.0#
166	37.4	19.4	29.2#



#60
Diethylphthalate
Concen: 40.398 ng
RT: 15.334 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	149	Resp:	1215039
Ion Ratio	Lower	Upper	
149	100		
177	20.5	17.5	26.3
150	12.8	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 39.496 ng
RT: 15.551 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 499618
Ion Ratio Lower Upper
204 100
206 33.0 26.6 40.0
141 73.9 48.6 73.0#



#62
4-Nitroaniline
Concen: 38.603 ng
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:138 Resp: 259954
Ion Ratio Lower Upper
138 100
92 47.6 31.2 71.2
108 78.1 78.6 118.6#









#63
Azobenzene
Concen: 42.162 ng
RT: 15.851 min Scan# 2170
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1058660




Ion	Ratio	Lower	Upper
77	100		
182	28.2	6.0	46.0
105	22.8	0.0	38.8
51	27.7	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion: 188 Resp: 759799

Ion	Ratio	Lower	Upper
188	100		
94	9.7	5.6	8.4#
80	10.4	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 39.436 ng
RT: 15.663 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 191324
Ion Ratio Lower Upper
198 100
51 45.6 14.0 54.0
105 47.9 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 40.624 ng
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:169 Resp: 1005025
Ion Ratio Lower Upper
169 100
168 67.0 54.4 81.6
167 35.9 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 39.178 ng
RT: 16.451 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 293448
Ion Ratio Lower Upper
248 100
250 97.8 79.8 119.8
141 91.2 57.8 86.6#



#68
Hexachlorobenzene
Concen: 38.769 ng
RT: 16.569 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:284 Resp: 344683
Ion Ratio Lower Upper
284 100
142 41.2 26.9 40.3#
249 31.1 24.5 36.7









#69
Atrazine
Concen: 40.534 ng
RT: 16.734 min Scan# 2320
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 314829
Ion Ratio Lower Upper
200 100
173 28.2 4.1 44.1
215 44.9 27.7 67.7



#70
Pentachlorophenol
Concen: 41.380 ng
RT: 16.928 min Scan# 2353
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:266 Resp: 169799
Ion Ratio Lower Upper
266 100
268 64.8 51.6 77.4
264 64.1 50.3 75.5









#71
Phenanthrene
Concen: 40.380 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1732695
Ion Ratio Lower Upper
178 100
176 19.2 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 40.330 ng
RT: 17.398 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:178 Resp: 1695375
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.6 12.2 18.2









#73
 Carbazole
 Concen: 40.934 ng
 RT: 17.675 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BP006137.D
 Acq: 08 Jul 2021 13:30

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:167 Resp: 1705130




Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	14.1	10.7	16.1




#74
 Di-n-butylphthalate
 Concen: 41.118 ng
 RT: 18.216 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BP006137.D
 Acq: 08 Jul 2021 13:30

Tgt Ion:149 Resp: 2061725


Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.7	4.5	6.7








#75
Fluoranthene
Concen: 40.588 ng
RT: 19.275 min Scan# 2752
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 1908822




Ion	Ratio	Lower	Upper
202	100		
101	12.5	0.0	27.6
203	17.6	0.0	37.8




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:240 Resp: 717523

Ion	Ratio	Lower	Upper
240	100		
120	13.8	6.5	9.7#
236	25.3	20.6	31.0









#77
Benzidine
Concen: 44.439 ng
RT: 19.457 min Scan# 2783
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	184	Resp:	945508
Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.0	16.4
183	12.1	9.5	14.3



#78
Pyrene
Concen: 40.981 ng
RT: 19.622 min Scan# 2811
Delta R.T. -0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:	202	Resp:	1987487
Ion	Ratio	Lower	Upper
202	100		
200	21.4	16.3	24.5
203	17.4	14.2	21.2









#79
Terphenyl-d14
Concen: 80.911 ng
RT: 19.816 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 2944557
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 13.4 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 41.012 ng
RT: 20.486 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:149 Resp: 986582
Ion Ratio Lower Upper
149 100
91 66.5 54.6 81.8
206 17.7 16.7 25.1









#81
Benzo(a)anthracene
Concen: 40.324 ng
RT: 21.328 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 1829972
Ion Ratio Lower Upper
228 100
226 27.2 21.6 32.4
229 19.5 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 40.368 ng
RT: 21.263 min Scan# 3090
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:252 Resp: 537414
Ion Ratio Lower Upper
252 100
254 64.4 50.6 76.0
126 14.3 7.0 10.6#









#83
Chrysene
Concen: 39.963 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 1792251
Ion Ratio Lower Upper
228 100
226 29.7 23.8 35.6
229 20.0 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 41.032 ng
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:149 Resp: 1456918
Ion Ratio Lower Upper
149 100
167 26.8 21.9 32.9
279 3.9 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 40.741 ng
RT: 22.157 min Scan# 3242
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 2547825
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 7.5 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.745 min Scan# 3512
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:264 Resp: 792326
Ion Ratio Lower Upper
264 100
260 23.5 18.6 28.0
265 21.8 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 40.748 ng
RT: 26.268 min Scan# 3941
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 2380305
Ion Ratio Lower Upper
276 100
138 31.9 15.3 22.9#
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 41.802 ng
RT: 23.010 min Scan# 3387
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:252 Resp: 2009085
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 12.7 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 40.412 ng
RT: 23.057 min Scan# 3395
Delta R.T. 0.000 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 1898435
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 12.8 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 41.086 ng
RT: 23.639 min Scan# 3494
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:252 Resp: 1853064
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 13.8 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 40.691 ng
RT: 26.274 min Scan# 3942
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2046756
Ion Ratio Lower Upper
278 100
139 20.5 9.3 13.9#
279 23.8 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 40.494 ng
RT: 27.039 min Scan# 4072
Delta R.T. 0.006 min
Lab File: BP006137.D
Acq: 08 Jul 2021 13:30

Tgt Ion:276 Resp: 2004583
Ion Ratio Lower Upper
276 100
277 23.4 18.9 28.3
138 27.4 12.6 18.8#

