



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
 Data File : BP006136.D
 Acq On : 08 Jul 2021 12:51
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 07 16:08:13 2021






AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	70914	PASS
68	69	0.00	2	1.8	1321	PASS
69	198	0.00	100	38.2	74059	PASS
70	69	0.00	2	0.5	387	PASS
127	198	10	80	55.0	106475	PASS
197	198	0.00	2	0.2	443	PASS
198	198	100	100	100.0	193621	PASS
199	198	5	9	6.8	13232	PASS
275	198	10	60	27.2	52637	PASS
365	198	1	100	4.0	7780	PASS
441	198	0.01	100	13.0	25147	PASS
442	442	100	100	100.0	159509	PASS
443	442	15	24	19.7	31477	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
 Data File : BP006136.D
 Acq On : 08 Jul 2021 12:51
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Jul 08 13:47:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration



TIC: BP006136.D\data.ms

(77) Benzidine

19.457min (+ 0.000) 365054.23 ng




response 1374747

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	13.92
183.00	11.90	11.95
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070821\
Data File : BP006136.D
Acq On : 08 Jul 2021 12:51
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jul 08 13:47:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 07 16:08:13 2021
Response via : Initial Calibration



TIC: BP006136.D\data.ms

(70) Pentachlorophenol (C)

16.928min (+ 0.000) 4933800.33 ng

response 293100

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.08
264.00	62.90	62.01
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/7/2021	BNA_P	BP006136.D
Compound Name	Response	Retention Time
DDT	689336	20.64
DDD	2823	20.22
DDE	2333	19.72
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5156	694492	0.74