

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070824\
 Data File : BP020870.D
 Acq On : 09 Jul 2024 19:52
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679

Quant Time: Jul 09 22:58:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 15:01:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	244360	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.487	136	1011802	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	674586	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.174	188	1449397	20.000	ng/u1	0.01
79) Chrysene-d12	21.633	240	1367876	20.000	ng/u1	0.01
88) Perylene-d12	24.980	264	1572650	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	40442	7.107	ng/uL	0.00
4) Pyridine-d5	3.675	84	296136	17.856	ng/u1	0.00
7) Phenol-d5	6.916	99	365088	18.216	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.069	67	214318	16.868	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.269	132	307516	18.774	ng/u1	0.00
15) 4-Methylphenol-d8	8.434	113	295402	18.148	ng/u1	0.00
21) Nitrobenzene-d5	8.875	128	145994	19.544	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.587	143	166425	22.263	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	309030	20.069	ng/u1	0.00
31) 4-Chloroaniline-d4	10.640	131	413963	19.183	ng/u1	0.00
46) Dimethylphthalate-d6	13.769	166	886170	18.894	ng/u1	0.01
49) Acenaphthylene-d8	14.051	160	1013955	18.117	ng/u1	0.01
54) 4-Nitrophenol-d4	14.610	143	150575	21.011	ng/u1	0.03
60) Fluorene-d10	15.375	176	758193	19.210	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.516	200	134900	18.469	ng/u1	0.02
73) Anthracene-d10	17.274	188	1190088	18.295	ng/u1	0.01
81) Pyrene-d10	19.657	212	1428724	17.386	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.762	264	1417913	18.535	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	44464	6.982	ng/uL	95
5) Pyridine	3.693	79	301768	17.566	ng/u1	99
6) Benzaldehyde	6.887	77	226495	22.347	ng/u1	97
8) Phenol	6.946	94	368098	18.010	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.157	93	296964	17.550	ng/u1	97
12) 2-Chlorophenol	7.304	128	306116	18.590	ng/u1	98
13) 2-Methylphenol	8.169	108	282134	17.992	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.234	45	424566	16.511	ng/u1	99
16) Acetophenone	8.546	105	458961	17.656	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.522	70	229236	16.590	ng/u1	98
18) 4-Methylphenol	8.493	108	303439	18.281	ng/u1	97
19) Hexachloroethane	8.787	117	129936	18.178	ng/u1	88
22) Nitrobenzene	8.916	77	348297	17.986	ng/u1	97
23) Isophorone	9.440	82	658286	17.156	ng/u1	99
25) 2-Nitrophenol	9.616	139	177335	21.966	ng/u1	94
26) 2,4-Dimethylphenol	9.681	107	330360	17.570	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.910	93	397671	17.422	ng/u1	99
29) 2,4-Dichlorophenol	10.157	162	300849	19.739	ng/u1	98
30) Naphthalene	10.545	128	973058	17.968	ng/u1	99
32) 4-Chloroaniline	10.657	127	389173	18.731	ng/u1	99
33) Hexachlorobutadiene	10.804	225	217536	19.091	ng/u1	97
34) Caprolactam	11.451	113	78301	16.882	ng/u1	86
35) 4-Chloro-3-methylphenol	11.792	107	328145	19.500	ng/u1	97
36) 2-Methylnaphthalene	12.145	142	665915	18.356	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070824\
 Data File : BP020870.D
 Acq On : 09 Jul 2024 19:52
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679

Quant Time: Jul 09 22:58:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 15:01:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.381	142	682948	18.416	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.522	216	398939	18.497	ng/ul	98
40) Hexachlorocyclopentadiene	12.486	237	160233	11.936	ng/ul	100
41) 2,4,6-Trichlorophenol	12.781	196	253765	19.723	ng/ul	98
42) 2,4,5-Trichlorophenol	12.863	196	273612	20.504	ng/ul	97
43) 1,1'-Biphenyl	13.175	154	894357	17.728	ng/ul	99
44) 2-Chloronaphthalene	13.216	162	712404	17.929	ng/ul	98
45) 2-Nitroaniline	13.439	65	207146	20.697	ng/ul	94
47) Dimethylphthalate	13.816	163	899030	18.817	ng/ul	100
48) 2,6-Dinitrotoluene	13.939	165	188474	22.123	ng/ul	95
50) Acenaphthylene	14.081	152	1139208	18.020	ng/ul	100
51) 3-Nitroaniline	14.281	138	192301	24.286	ng/ul	94
52) Acenaphthene	14.428	153	751594	17.922	ng/ul	99
53) 2,4-Dinitrophenol	14.510	184	76637	17.342	ng/ul	97
55) 4-Nitrophenol	14.622	109	122088	19.037	ng/ul	97
56) Dibenzofuran	14.769	168	1044518	18.484	ng/ul	98
57) 2,4-Dinitrotoluene	14.751	165	273247	23.235	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.004	232	239846	21.762	ng/ul#	92
59) Diethylphthalate	15.198	149	902439	19.149	ng/ul	98
61) Fluorene	15.428	166	852489	18.776	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.422	204	447327	18.880	ng/ul	97
63) 4-Nitroaniline	15.469	138	184624	24.568	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.528	198	144429	18.459	ng/ul	93
67) N-Nitrosodiphenylamine	15.645	169	739655	17.304	ng/ul	99
68) 4-Bromophenyl-phenylether	16.339	248	295356	18.269	ng/ul	96
69) Hexachlorobenzene	16.457	284	343080	18.957	ng/ul	93
70) Atrazine	16.627	200	290101	19.252	ng/ul	99
71) Pentachlorophenol	16.822	266	195613	18.774	ng/ul	94
72) Phenanthrene	17.222	178	1366846	18.009	ng/ul	100
74) Anthracene	17.310	178	1405617	18.004	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.139	216	394985	16.788	ng/uL	98
76) Pentachlorobenzene	14.680	250	394315	17.585	ng/uL	99
77) Carbazole	17.598	167	1260815	19.568	ng/ul	100
78) Di-n-butylphthalate	18.186	149	1590414	19.005	ng/ul	100
80) Fluoranthene	19.310	202	1679159	16.728	ng/ul	99
82) Pyrene	19.692	202	1725649	16.763	ng/ul	100
83) Butylbenzylphthalate	20.633	149	748595	19.583	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.533	252	557624	18.438	ng/ul	97
85) Benzo(a)anthracene	21.610	228	1735243	18.099	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.533	149	1068761	18.007	ng/ul	98
87) Chrysene	21.680	228	1641555	18.114	ng/ul	99
89) Di-n-octyl phthalate	22.815	149	1859563	18.419	ng/ul	100
90) Benzo(b)fluoranthene	23.921	252	1729479	18.153	ng/ul	99
91) Benzo(k)fluoranthene	23.992	252	1693778	17.575	ng/ul	99
93) Benzo(a)pyrene	24.833	252	1643356	18.272	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.791	276	1028760	8.918	ng/ul	96
95) Dibenzo(a,h)anthracene	28.868	278	1688459	17.809	ng/ul	98
96) Benzo(g,h,i)perylene	29.974	276	1649843	17.517	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.982 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 88	Resp: 44464
Ion Ratio	Lower Upper
88	100
43	29.1 28.1 42.1
58	74.6 61.1 91.7



#5
Pyridine
Concen: 17.566 ng/uL
RT: 3.693 min Scan# 120
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion: 79	Resp: 301768
Ion Ratio	Lower Upper
79	100
52	80.5 64.6 97.0
51	38.3 32.0 48.0









#6
Benzaldehyde
Concen: 22.347 ng/ul
RT: 6.887 min Scan# 663
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 77 Resp: 226495
Ion Ratio Lower Upper
77 100
105 100.1 78.2 117.2
106 95.3 73.2 109.8



#8
Phenol
Concen: 18.010 ng/ul
RT: 6.946 min Scan# 673
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion: 94 Resp: 368098
Ion Ratio Lower Upper
94 100
65 28.4 24.2 36.2
66 39.4 33.9 50.9









#10
Bis(2-Chloroethyl)ether
Concen: 17.550 ng/ul
RT: 7.157 min Scan# 709
Delta R.T. -0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 93 Resp: 296964
Ion Ratio Lower Upper
93 100
63 76.2 63.4 95.2
95 32.5 27.0 40.4



#12
2-Chlorophenol
Concen: 18.590 ng/ul
RT: 7.304 min Scan# 734
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:128 Resp: 306116
Ion Ratio Lower Upper
128 100
64 46.9 39.4 59.2
130 32.1 25.9 38.9





#13

2-Methylphenol

Concen: 17.992 ng/ul

RT: 8.169 min Scan# 81

Delta R.T. -0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :





SSTD020679

Tgt Ion:108 Resp: 282134

Ion Ratio Lower Upper

108 100

107 89.3 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.511 ng/ul

RT: 8.234 min Scan# 892

Delta R.T. -0.018 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion: 45 Resp: 424566


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 10.9 9.1 13.7









#16
Acetophenone
Concen: 17.656 ng/ul
RT: 8.546 min Scan# 945
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:	105	Resp:	458961
Ion	Ratio	Lower	Upper
105	100		
77	75.1	62.2	93.4
51	27.4	23.6	35.4



#17
N-Nitroso-di-n-propylamine
Concen: 16.590 ng/ul
RT: 8.522 min Scan# 941
Delta R.T. -0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:	70	Resp:	229236
Ion	Ratio	Lower	Upper
70	100		
42	59.4	46.8	70.2
101	9.8	8.0	12.0
130	22.8	16.4	24.6









#18
4-Methylphenol
Concen: 18.281 ng/ul
RT: 8.493 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:108 Resp: 303439
Ion Ratio Lower Upper
108 100
107 116.0 90.2 135.2



#19
Hexachloroethane
Concen: 18.178 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:117 Resp: 129936
Ion Ratio Lower Upper
117 100
201 118.6 83.0 124.6
199 71.5 51.9 77.9









#22
Nitrobenzene
Concen: 17.986 ng/ul
RT: 8.916 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 77 Resp: 348297
Ion Ratio Lower Upper
77 100
123 46.1 35.0 52.6
65 13.6 11.0 16.4



#23
Isophorone
Concen: 17.156 ng/ul
RT: 9.440 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion: 82 Resp: 658286
Ion Ratio Lower Upper
82 100
95 7.4 6.2 9.2
138 16.8 13.2 19.8









#25
2-Nitrophenol
Concen: 21.966 ng/ul
RT: 9.616 min Scan# 1127
Delta R.T. -0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:139 Resp: 177335
Ion Ratio Lower Upper
139 100
65 51.5 46.2 69.2
109 33.4 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 17.570 ng/ul
RT: 9.681 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:107 Resp: 330360
Ion Ratio Lower Upper
107 100
121 48.0 38.5 57.7
122 80.6 66.7 100.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.422 ng/ul
 RT: 9.910 min Scan# 1177
 Delta R.T. -0.006 min
 Lab File: BP020870.D
 Acq: 09 Jul 2024 19:52




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679


Tgt Ion: 93	Resp: 397671
Ion Ratio	Lower Upper
93 100	
95 32.3	25.4 38.0
123 11.2	8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 19.739 ng/ul
 RT: 10.157 min Scan# 1219
 Delta R.T. 0.006 min
 Lab File: BP020870.D
 Acq: 09 Jul 2024 19:52

Tgt Ion:162	Resp: 300849		
Ion Ratio	Lower	Upper	
162	100		
164	65.6	51.2	76.8
98	36.8	30.3	45.5









#30
Naphthalene
Concen: 17.968 ng/ul
RT: 10.545 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:128 Resp: 973058
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 18.731 ng/ul
RT: 10.657 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:127 Resp: 389173
Ion Ratio Lower Upper
127 100
129 32.7 25.5 38.3









#33
Hexachlorobutadiene
Concen: 19.091 ng/ul
RT: 10.804 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:225 Resp: 217536
Ion Ratio Lower Upper
225 100
223 64.2 49.8 74.8
227 64.9 49.9 74.9



#34
Caprolactam
Concen: 16.882 ng/ul
RT: 11.451 min Scan# 1439
Delta R.T. -0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:113 Resp: 78301
Ion Ratio Lower Upper
113 100
55 165.8 150.0 225.0
56 124.8 112.4 168.6









#35
4-Chloro-3-methylphenol
Concen: 19.500 ng/ul
RT: 11.792 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:107 Resp: 328145
Ion Ratio Lower Upper
107 100
144 26.1 18.3 27.5
142 76.8 60.1 90.1



#36
2-Methylnaphthalene
Concen: 18.356 ng/ul
RT: 12.145 min Scan# 1557
Delta R.T. -0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:142 Resp: 665915
Ion Ratio Lower Upper
142 100
141 89.8 71.1 106.7





#37

1-Methylnaphthalene

Concen: 18.416 ng/ul

RT: 12.381 min Scan# 1597

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:142 Resp: 682948

Ion Ratio Lower Upper

142 100

141 94.0 74.9 112.3

116 3.9 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.497 ng/ul

RT: 12.522 min Scan# 1621

Delta R.T. -0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Tgt Ion:216 Resp: 398939




Ion Ratio Lower Upper


216 100

214 77.2 62.9 94.3

179 20.4 17.5 26.3

108 15.1 13.1 19.7





#40

Hexachlorocyclopentadiene

Concen: 11.936 ng/ul

RT: 12.486 min Scan# 1616

Delta R.T. -0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:237 Resp: 160233

Ion Ratio Lower Upper

237 100

235 63.9 51.1 76.7

272 13.8 10.4 15.6



#41

2,4,6-Trichlorophenol

Concen: 19.723 ng/ul

RT: 12.781 min Scan# 1665

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:196 Resp: 253765


Ion Ratio Lower Upper

196 100

198 98.0 77.0 115.4

200 30.3 25.8 38.8









#42
2,4,5-Trichlorophenol
Concen: 20.504 ng/ul
RT: 12.863 min Scan# 1679
Delta R.T. 0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:196 Resp: 273612
Ion Ratio Lower Upper
196 100
198 97.1 79.5 119.3
200 30.4 25.7 38.5



#43
1,1'-Biphenyl
Concen: 17.728 ng/ul
RT: 13.175 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:154 Resp: 894357
Ion Ratio Lower Upper
154 100
153 40.8 32.6 48.8
76 12.0 10.8 16.2





#44

2-Chloronaphthalene

Concen: 17.929 ng/ul

RT: 13.216 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:162 Resp: 712404

Ion Ratio Lower Upper

162 100

164 32.7 26.5 39.7

127 38.4 32.5 48.7



#45

2-Nitroaniline

Concen: 20.697 ng/ul

RT: 13.439 min Scan# 1777

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion: 65 Resp: 207146


Ion Ratio Lower Upper

65 100

92 68.7 53.7 80.5

138 103.4 76.4 114.6









#47
Dimethylphthalate
Concen: 18.817 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. 0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:163 Resp: 899030
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.4 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 22.123 ng/ul
RT: 13.939 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:165 Resp: 188474
Ion Ratio Lower Upper
165 100
89 55.7 48.7 73.1
121 21.5 17.3 25.9





#50

Acenaphthylene

Concen: 18.020 ng/ul

RT: 14.081 min Scan# 11

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:152 Resp: 1139208

Ion Ratio Lower Upper

152 100

151 20.0 16.2 24.4

153 13.4 10.8 16.2



#51

3-Nitroaniline

Concen: 24.286 ng/ul

RT: 14.281 min Scan# 1920

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:138 Resp: 192301


Ion Ratio Lower Upper

138 100

108 11.3 10.8 16.2

92 115.5 98.3 147.5









#52
Acenaphthene
Concen: 17.922 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. 0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:153 Resp: 751594
Ion Ratio Lower Upper
153 100
152 46.9 38.2 57.2
154 87.7 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 17.342 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. 0.029 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:184 Resp: 76637
Ion Ratio Lower Upper
184 100
63 63.8 49.5 74.3
154 66.1 51.4 77.0









#55
4-Nitrophenol
Concen: 19.037 ng/ul
RT: 14.622 min Scan# 1977
Delta R.T. 0.023 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:109 Resp: 122088
Ion Ratio Lower Upper
109 100
139 122.5 97.4 146.2
65 128.2 98.2 147.4



#56
Dibenzofuran
Concen: 18.484 ng/ul
RT: 14.769 min Scan# 2003
Delta R.T. 0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:168 Resp: 1044518
Ion Ratio Lower Upper
168 100
139 37.4 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 23.235 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.017 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:165 Resp: 273247
Ion Ratio Lower Upper
165 100
63 50.6 38.3 57.5
182 2.9 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.762 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:232 Resp: 239846
Ion Ratio Lower Upper
232 100
131 39.8 37.3 55.9
130 2.0 2.2 3.2#
166 28.9 25.4 38.2





#59

Diethylphthalate

Concen: 19.149 ng/ul

RT: 15.198 min Scan# 2076

Delta R.T. 0.011 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:149 Resp: 902439

Ion Ratio Lower Upper

149 100

177 22.5 17.2 25.8

150 12.3 10.2 15.4



#61

Fluorene

Concen: 18.776 ng/ul

RT: 15.428 min Scan# 2115

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:166 Resp: 852489


Ion Ratio Lower Upper

166 100

165 98.2 78.8 118.2

167 13.5 11.2 16.8









#62
4-Chlorophenyl-phenylether
Concen: 18.880 ng/ul
RT: 15.422 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:204 Resp: 447327
Ion Ratio Lower Upper
204 100
206 33.2 26.0 39.0
141 58.7 49.0 73.4



#63
4-Nitroaniline
Concen: 24.568 ng/ul
RT: 15.469 min Scan# 2122
Delta R.T. 0.023 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:138 Resp: 184624
Ion Ratio Lower Upper
138 100
92 52.2 43.7 65.5
108 84.8 73.1 109.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.459 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. 0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:198 Resp: 144429
Ion Ratio Lower Upper
198 100
182 4.7 3.9 5.9
77 24.4 23.0 34.6



#67
N-Nitrosodiphenylamine
Concen: 17.304 ng/ul
RT: 15.645 min Scan# 2152
Delta R.T. 0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:169 Resp: 739655
Ion Ratio Lower Upper
169 100
168 66.7 52.7 79.1
167 36.6 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 18.269 ng/ul
RT: 16.339 min Scan# 21
Delta R.T. 0.011 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:248 Resp: 295356
Ion Ratio Lower Upper
248 100
250 97.4 78.4 117.6
141 72.0 63.0 94.4



#69
Hexachlorobenzene
Concen: 18.957 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. 0.012 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:284 Resp: 343080
Ion Ratio Lower Upper
284 100
142 31.1 29.6 44.4
249 30.9 26.5 39.7





#70

Atrazine

Concen: 19.252 ng/ul

RT: 16.627 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:200 Resp: 290101

Ion Ratio Lower Upper

200 100

173 26.8 21.2 31.8

215 49.1 39.0 58.4



#71

Pentachlorophenol

Concen: 18.774 ng/ul

RT: 16.822 min Scan# 2352

Delta R.T. 0.017 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:266 Resp: 195613


Ion Ratio Lower Upper

266 100

264 63.1 48.1 72.1

268 67.8 49.8 74.8





#72

Phenanthrene

Concen: 18.009 ng/ul

RT: 17.222 min Scan# 2420

Delta R.T. 0.017 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:178 Resp: 1366846

Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.1 15.4 23.2



#74

Anthracene

Concen: 18.004 ng/ul

RT: 17.310 min Scan# 2435

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:178 Resp: 1405617


Ion Ratio Lower Upper

178 100

179 15.6 12.1 18.1

176 18.2 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.788 ng/uL
RT: 13.139 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:216 Resp: 394985
Ion Ratio Lower Upper
216 100
214 78.7 61.1 91.7
218 48.5 38.0 57.0



#76
Pentachlorobenzene
Concen: 17.585 ng/uL
RT: 14.680 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP020870.D
Acq: 09 Jul 2024 19:52

Tgt Ion:250 Resp: 394315
Ion Ratio Lower Upper
250 100
248 63.1 50.6 75.8
252 63.5 51.8 77.8





#77

Carbazole

Concen: 19.568 ng/ul

RT: 17.598 min Scan# 2484

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:167 Resp: 1260815

Ion Ratio Lower Upper

167 100

166 21.6 17.2 25.8

139 13.0 10.6 16.0



#78

Di-n-butylphthalate

Concen: 19.005 ng/ul

RT: 18.186 min Scan# 2584

Delta R.T. 0.017 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:149 Resp: 1590414


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.0

104 5.0 4.1 6.1





#80

Fluoranthene

Concen: 16.728 ng/ul

RT: 19.310 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:202 Resp: 1679159

Ion Ratio Lower Upper

202 100

101 10.8 8.2 12.4

100 8.2 6.6 10.0



#82

Pyrene

Concen: 16.763 ng/ul

RT: 19.692 min Scan# 2840

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:202 Resp: 1725649


Ion Ratio Lower Upper

202 100

101 12.2 9.8 14.8

100 9.7 7.9 11.9





#83

Butylbenzylphthalate

Concen: 19.583 ng/ul

RT: 20.633 min Scan# 3000

Delta R.T. -0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:149 Resp: 748595

Ion Ratio Lower Upper

149 100

91 74.3 58.0 87.0

206 21.8 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 18.438 ng/ul

RT: 21.533 min Scan# 3153

Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:252 Resp: 557624


Ion Ratio Lower Upper

252 100

254 63.6 53.0 79.6

126 10.3 8.5 12.7





#85

Benzo(a)anthracene

Concen: 18.099 ng/ul

RT: 21.610 min Scan# 3166

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:228 Resp: 1735243

Ion Ratio Lower Upper

228 100

229 19.7 15.2 22.8

226 26.9 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.007 ng/ul

RT: 21.533 min Scan# 3153

Delta R.T. -0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:149 Resp: 1068761


Ion Ratio Lower Upper

149 100

167 28.3 22.0 33.0

279 4.8 3.3 4.9





#87

Chrysene

Concen: 18.114 ng/ul

RT: 21.680 min Scan# 3178

Delta R.T. 0.012 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:228 Resp: 1641555

Ion Ratio Lower Upper

228 100

226 29.8 23.8 35.6

229 19.4 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 18.419 ng/ul




RT: 22.815 min Scan# 3371


Delta R.T. 0.006 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Tgt Ion:149 Resp: 1859563





#90

Benzo(b)fluoranthene

Concen: 18.153 ng/ul

RT: 23.921 min Scan# 3559

Delta R.T. 0.017 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:252 Resp: 1729479

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 8.8 7.9 11.9



#91

Benzo(k)fluoranthene

Concen: 17.575 ng/ul

RT: 23.992 min Scan# 3571

Delta R.T. 0.017 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:252 Resp: 1693778


Ion Ratio Lower Upper

252 100

253 21.5 17.6 26.4

125 9.0 7.9 11.9





#93

Benzo(a)pyrene

Concen: 18.272 ng/ul

RT: 24.833 min Scan# 31

Delta R.T. 0.023 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:252 Resp: 1643356

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 9.2 8.7 13.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.918 ng/ul

RT: 28.791 min Scan# 4387

Delta R.T. 0.029 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52




Tgt Ion:276 Resp: 1028760


Ion Ratio Lower Upper

276 100

138 16.0 15.2 22.8

277 24.4 19.0 28.4





#95

Dibenzo(a,h)anthracene

Concen: 17.809 ng/ul

RT: 28.868 min Scan# 4400

Delta R.T. 0.029 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:278 Resp: 1688459

Ion Ratio Lower Upper

278 100

139 14.9 11.1 16.7

279 24.0 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 17.517 ng/ul

RT: 29.974 min Scan# 4588

Delta R.T. 0.035 min

Lab File: BP020870.D

Acq: 09 Jul 2024 19:52

Tgt Ion:276 Resp: 1649843

Ion Ratio Lower Upper

276 100

138 19.6 16.4 24.6

277 23.4 18.7 28.1

