

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070921\
 Data File : BP006160.D
 Acq On : 09 Jul 2021 14:36
 Operator : CG/JU
 Sample : PB137592BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB137592BS

Quant Time: Jul 09 15:29:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	183048	20.000	ng	0.00
21) Naphthalene-d8	10.675	136	733166	20.000	ng	0.00
39) Acenaphthene-d10	14.510	164	405611	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	747524	20.000	ng	# 0.00
76) Chrysene-d12	21.345	240	683139	20.000	ng	# 0.00
86) Perylene-d12	23.745	264	732141	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.446	112	1537854	144.558	ng	-0.01
7) Phenol-d6	7.057	99	1811364	128.724	ng	0.00
23) Nitrobenzene-d5	9.046	82	1035373	89.201	ng	0.00
42) 2,4,6-Tribromophenol	16.010	330	611112	126.549	ng	0.00
45) 2-Fluorobiphenyl	13.134	172	2398072	87.403	ng	0.00
79) Terphenyl-d14	19.816	244	2991064	86.326	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.305	88	163451	39.158	ng	92
3) Pyridine	3.722	79	404665	38.155	ng	# 88
4) n-Nitrosodimethylamine	3.628	42	199526	47.830	ng	# 88
6) Aniline	7.205	93	587193	39.550	ng	98
8) 2-Chlorophenol	7.440	128	614881	49.350	ng	98
9) Benzaldehyde	7.022	77	27331	3.559	ng	95
10) Phenol	7.087	94	682618	50.510	ng	92
11) bis(2-Chloroethyl)ether	7.299	93	504017	49.355	ng	98
12) 1,3-Dichlorobenzene	7.757	146	625730	46.870	ng	96
13) 1,4-Dichlorobenzene	7.904	146	638203	46.808	ng	99
14) 1,2-Dichlorobenzene	8.222	146	611511	47.111	ng	99
15) Benzyl Alcohol	8.128	79	455821	53.581	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.410	45	572633	47.214	ng	91
17) 2-Methylphenol	8.334	107	465250	50.605	ng	94
18) Hexachloroethane	8.946	117	219994	48.467	ng	95
19) n-Nitroso-di-n-propyla...	8.687	70	356717	46.499	ng	97
20) 3+4-Methylphenols	8.669	107	624889	50.310	ng	95
22) Acetophenone	8.704	105	787784	47.209	ng	99
24) Nitrobenzene	9.087	77	531473	45.274	ng	94
25) Isophorone	9.616	82	963593	43.731	ng	96
26) 2-Nitrophenol	9.799	139	312592	47.900	ng	87
27) 2,4-Dimethylphenol	9.863	122	525772	54.579	ng	96
28) bis(2-Chloroethoxy)met...	10.093	93	619036	50.694	ng	99
29) 2,4-Dichlorophenol	10.346	162	508895	48.079	ng	99
30) 1,2,4-Trichlorobenzene	10.540	180	515725	45.763	ng	97
31) Naphthalene	10.728	128	1724152	45.111	ng	99
32) Benzoic acid	10.069	122	348246	47.659	ng	91
33) 4-Chloroaniline	10.857	127	413149	28.003	ng	98
34) Hexachlorobutadiene	11.004	225	274473	44.235	ng	99
35) Caprolactam	11.669	113	166540	49.913	ng	87
36) 4-Chloro-3-methylphenol	11.993	107	552071	48.301	ng	99
37) 2-Methylnaphthalene	12.340	142	1201466	45.307	ng	97
38) 1-Methylnaphthalene	12.557	142	1093215	43.560	ng	98
40) 1,2,4,5-Tetrachloroben...	12.710	216	474332	47.143	ng	96
41) Hexachlorocyclopentadiene	12.681	237	258138	90.675	ng	97
43) 2,4,6-Trichlorophenol	12.957	196	348105	47.822	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070921\
 Data File : BP006160.D
 Acq On : 09 Jul 2021 14:36
 Operator : CG/JU
 Sample : PB137592BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 PB137592BS


Quant Time: Jul 09 15:29:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.045	196	374626	47.969	ng	# 94
46) 1,1'-Biphenyl	13.345	154	1435099	47.128	ng	99
47) 2-Chloronaphthalene	13.392	162	1079748	46.091	ng	98
48) 2-Nitroaniline	13.610	65	293584	50.433	ng	87
49) Acenaphthylene	14.234	152	1768676	46.888	ng	99
50) Dimethylphthalate	13.975	163	1348607	46.237	ng	99
51) 2,6-Dinitrotoluene	14.104	165	300943	45.766	ng	93
52) Acenaphthene	14.575	154	1030928	44.928	ng	98
53) 3-Nitroaniline	14.439	138	231105	34.689	ng	93
54) 2,4-Dinitrophenol	14.657	184	309405	88.507	ng	98
55) Dibenzofuran	14.910	168	1589951	44.511	ng	98
56) 4-Nitrophenol	14.769	139	477625	91.872	ng	# 84
57) 2,4-Dinitrotoluene	14.898	165	421180	47.557	ng	97
58) Fluorene	15.563	166	1301200	45.343	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.151	232	284136	46.496	ng	# 74
60) Diethylphthalate	15.334	149	1386203	46.343	ng	98
61) 4-Chlorophenyl-phenyle...	15.551	204	571798	45.451	ng	89
62) 4-Nitroaniline	15.604	138	297607	43.991	ng	# 82
63) Azobenzene	15.845	77	1126693	45.119	ng	96
65) 4,6-Dinitro-2-methylph...	15.669	198	196641	41.000	ng	91
66) n-Nitrosodiphenylamine	15.775	169	1121198	46.064	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	334493	45.391	ng	# 89
68) Hexachlorobenzene	16.569	284	396999	45.386	ng	# 92
69) Atrazine	16.733	200	361533	47.311	ng	95
70) Pentachlorophenol	16.922	266	431630	106.917	ng	98
71) Phenanthrene	17.310	178	1944163	46.052	ng	99
72) Anthracene	17.398	178	1958969	47.366	ng	100
73) Carbazole	17.675	167	1856159	45.291	ng	99
74) Di-n-butylphthalate	18.216	149	2372025	48.084	ng	99
75) Fluoranthene	19.274	202	2108450	45.568	ng	96
77) Benzidine	19.457	184	897881	44.325	ng	99
78) Pyrene	19.627	202	2168710	46.968	ng	99
80) Butylbenzylphthalate	20.492	149	1114033	48.641	ng	93
81) Benzo(a)anthracene	21.333	228	1954172	45.228	ng	99
82) 3,3'-Dichlorobenzidine	21.263	252	623148	49.164	ng	# 96
83) Chrysene	21.386	228	1930596	45.214	ng	99
84) Bis(2-ethylhexyl)phtha...	21.245	149	1651915	48.866	ng	# 98
85) Di-n-octyl phthalate	22.157	149	2883138	48.424	ng	100
87) Indeno(1,2,3-cd)pyrene	26.280	276	2907819	53.871	ng	# 88
88) Benzo(b)fluoranthene	23.015	252	2263205	50.960	ng	# 96
89) Benzo(k)fluoranthene	23.063	252	2060100	47.458	ng	# 96
90) Benzo(a)pyrene	23.645	252	2165226	51.954	ng	# 95
91) Dibenzo(a,h)anthracene	26.286	278	2495820	53.698	ng	# 92
92) Benzo(g,h,i)perylene	27.051	276	2507657	54.821	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 09 15:29:48 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 07 16:08:13 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.869 min Scan# 813
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:152 Resp: 183048
Ion Ratio Lower Upper
152 100
150 157.5 125.6 188.4
115 62.2 48.5 72.7



#2
1,4-Dioxane
Concen: 39.158 ng
RT: 3.305 min Scan# 37
Delta R.T. -0.012 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 88 Resp: 163451
Ion Ratio Lower Upper
88 100
58 66.1 46.7 70.1
43 26.8 20.3 30.5









#3
Pyridine
Concen: 38.155 ng
RT: 3.722 min Scan# 108
Delta R.T. -0.018 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 79 Resp: 404665
Ion Ratio Lower Upper
79 100
52 64.4 41.7 62.5#
51 32.0 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 47.830 ng
RT: 3.628 min Scan# 92
Delta R.T. -0.018 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 42 Resp: 199526
Ion Ratio Lower Upper
42 100
74 145.6 105.4 158.0
44 7.5 4.2 6.4#









#5
2-Fluorophenol
Concen: 144.558 ng
RT: 5.446 min Scan# 401
Delta R.T. -0.012 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:112 Resp: 1537854
Ion Ratio Lower Upper
112 100
64 49.4 41.4 62.2
63 25.5 23.8 35.8



#6
Aniline
Concen: 39.550 ng
RT: 7.205 min Scan# 700
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 93 Resp: 587193
Ion Ratio Lower Upper
93 100
66 36.4 29.7 44.5
65 18.0 15.3 22.9









#7
Phenol-d6
Concen: 128.724 ng
RT: 7.057 min Scan# 675
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 99 Resp: 1811364
Ion Ratio Lower Upper
99 100
42 15.1 13.4 20.0
71 32.6 30.6 46.0



#8
2-Chlorophenol
Concen: 49.350 ng
RT: 7.440 min Scan# 740
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 128 Resp: 614881
Ion Ratio Lower Upper
128 100
130 32.4 12.4 52.4
64 38.1 20.3 60.3









#9
Benzaldehyde
Concen: 3.559 ng
RT: 7.022 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 77 Resp: 27331
Ion Ratio Lower Upper
77 100
105 103.1 76.7 116.7
106 98.7 76.2 116.2



#10
Phenol
Concen: 50.510 ng
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 94 Resp: 682618
Ion Ratio Lower Upper
94 100
65 24.9 8.8 48.8
66 35.5 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 49.355 ng
RT: 7.299 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 93 Resp: 504017
Ion Ratio Lower Upper
93 100
63 68.0 46.2 86.2
95 31.8 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 46.870 ng
RT: 7.757 min Scan# 794
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 146 Resp: 625730
Ion Ratio Lower Upper
146 100
148 63.9 51.5 77.3
75 25.9 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 46.808 ng
RT: 7.904 min Scan# 819
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:146 Resp: 638203
Ion Ratio Lower Upper
146 100
148 63.6 51.3 76.9
111 40.1 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 47.111 ng
RT: 8.222 min Scan# 873
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:146 Resp: 611511
Ion Ratio Lower Upper
146 100
148 63.0 50.3 75.5
111 41.4 33.7 50.5









#15
Benzyl Alcohol
Concen: 53.581 ng
RT: 8.128 min Scan# 857
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 79 Resp: 455821
Ion Ratio Lower Upper
79 100
108 83.9 62.5 93.7
77 61.2 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.214 ng
RT: 8.410 min Scan# 905
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 45 Resp: 572633
Ion Ratio Lower Upper
45 100
77 16.9 2.0 42.0
79 13.4 0.0 36.7









#17
2-Methylphenol
Concen: 50.605 ng
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:	107	Resp:	465250
Ion	Ratio	Lower	Upper
107	100		
108	111.3	87.7	131.5
77	42.6	40.2	60.2
79	41.3	39.0	58.4



#18
Hexachloroethane
Concen: 48.467 ng
RT: 8.946 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	117	Resp:	219994
Ion	Ratio	Lower	Upper
117	100		
119	95.1	76.2	114.2
201	95.2	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 46.499 ng
RT: 8.687 min Scan# 952
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:	70	Resp:	356717
Ion	Ratio	Lower	Upper
70	100		
42	52.4	40.8	61.2
101	11.0	8.8	13.2
130	26.1	18.9	28.3



#20
3+4-Methylphenols
Concen: 50.310 ng
RT: 8.669 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	107	Resp:	624889
Ion	Ratio	Lower	Upper
107	100		
108	90.2	71.0	111.0
77	30.6	16.6	56.6
79	27.9	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.675 min Scan# 1290
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:	136	Resp:	733166
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.6	12.8
54	6.6	4.6	6.8
68	5.2	3.6	5.4



#22
Acetophenone
Concen: 47.209 ng
RT: 8.704 min Scan# 955
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	105	Resp:	787784
Ion Ratio	Lower	Upper	
105	100		
71	2.4	1.9	2.9
51	23.0	18.4	27.6
120	23.5	17.9	26.9









#23
Nitrobenzene-d5
Concen: 89.201 ng
RT: 9.046 min Scan# 1013
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 82 Resp: 1035373
Ion Ratio Lower Upper
82 100
128 52.3 35.7 53.5
54 44.6 35.0 52.4



#24
Nitrobenzene
Concen: 45.274 ng
RT: 9.087 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 77 Resp: 531473
Ion Ratio Lower Upper
77 100
123 52.6 38.2 57.2
65 13.0 11.4 17.2









#25
Isophorone
Concen: 43.731 ng
RT: 9.616 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion: 82 Resp: 963593
Ion Ratio Lower Upper
82 100
95 7.9 6.5 9.7
138 21.0 15.0 22.4



#26
2-Nitrophenol
Concen: 47.900 ng
RT: 9.799 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 139 Resp: 312592
Ion Ratio Lower Upper
139 100
109 29.1 28.8 43.2
65 41.2 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 54.579 ng
RT: 9.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:122 Resp: 525772
Ion Ratio Lower Upper
122 100
107 111.7 94.5 141.7
121 59.3 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 50.694 ng
RT: 10.093 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 93 Resp: 619036
Ion Ratio Lower Upper
93 100
95 32.9 25.9 38.9
123 13.2 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 48.079 ng
RT: 10.346 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:162 Resp: 508895
Ion Ratio Lower Upper
162 100
164 65.2 44.9 84.9
98 35.4 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 45.763 ng
RT: 10.540 min Scan# 1267
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:180 Resp: 515725
Ion Ratio Lower Upper
180 100
182 95.3 77.9 116.9
145 34.0 24.7 37.1









#31
Naphthalene
Concen: 45.111 ng
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:128 Resp: 1724152
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.6
127 13.3 10.7 16.1



#32
Benzoic acid
Concen: 47.659 ng
RT: 10.069 min Scan# 1187
Delta R.T. 0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:122 Resp: 348246
Ion Ratio Lower Upper
122 100
105 117.7 103.2 143.2
77 81.2 76.2 116.2









#33
4-Chloroaniline
Concen: 28.003 ng
RT: 10.857 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:	127	Resp:	413149
Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.1	39.1
65	26.0	22.5	33.7
92	18.3	15.0	22.6



#34
Hexachlorobutadiene
Concen: 44.235 ng
RT: 11.004 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	225	Resp:	274473
Ion	Ratio	Lower	Upper
225	100		
223	62.4	51.0	76.6
227	64.2	51.4	77.2









#35
Caprolactam
Concen: 49.913 ng
RT: 11.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:113 Resp: 166540
Ion Ratio Lower Upper
113 100
55 133.2 98.9 138.9
56 105.1 71.7 111.7



#36
4-Chloro-3-methylphenol
Concen: 48.301 ng
RT: 11.993 min Scan# 1514
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:107 Resp: 552071
Ion Ratio Lower Upper
107 100
144 27.7 21.1 31.7
142 83.4 66.0 99.0









#37
2-Methylnaphthalene
Concen: 45.307 ng
RT: 12.340 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:142 Resp: 1201466
Ion Ratio Lower Upper
142 100
141 89.6 71.8 107.6
115 33.7 31.6 47.4



#38
1-Methylnaphthalene
Concen: 43.560 ng
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:142 Resp: 1093215
Ion Ratio Lower Upper
142 100
141 93.8 73.3 109.9
116 3.6 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:164 Resp: 405611
Ion Ratio Lower Upper
164 100
162 101.6 80.2 120.2
160 44.8 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.143 ng
RT: 12.710 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:216 Resp: 474332
Ion Ratio Lower Upper
216 100
214 79.2 61.9 92.9
179 24.6 16.8 25.2
108 22.0 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 90.675 ng
RT: 12.681 min Scan# 1631
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:237	Resp:	258138
Ion Ratio	Lower	Upper
237	100	
235	62.2	44.5 84.5
272	13.4	0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 126.549 ng
RT: 16.010 min Scan# 2197
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:330	Resp:	611112
Ion Ratio	Lower	Upper
330	100	
332	96.8	77.8 116.8
141	37.3	25.4 38.2









#43
2,4,6-Trichlorophenol
Concen: 47.822 ng
RT: 12.957 min Scan# 1678
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:196 Resp: 348105
Ion Ratio Lower Upper
196 100
198 94.6 79.5 119.3
200 31.3 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 47.969 ng
RT: 13.045 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:196 Resp: 374626
Ion Ratio Lower Upper
196 100
198 95.5 74.7 112.1
97 52.0 37.4 56.0
132 32.6 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 87.403 ng
RT: 13.134 min Scan# 1708
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:172 Resp: 2398072
Ion Ratio Lower Upper
172 100
171 35.9 27.9 41.9
170 23.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 47.128 ng
RT: 13.345 min Scan# 1744
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:154 Resp: 1435099
Ion Ratio Lower Upper
154 100
153 41.2 20.4 60.4
76 11.4 0.0 30.8









#47
2-Chloronaphthalene
Concen: 46.091 ng
RT: 13.392 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:162 Resp: 1079748
Ion Ratio Lower Upper
162 100
127 41.1 31.4 47.0
164 32.5 26.6 39.8



#48
2-Nitroaniline
Concen: 50.433 ng
RT: 13.610 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion: 65 Resp: 293584
Ion Ratio Lower Upper
65 100
92 76.6 57.7 86.5
138 134.8 92.1 138.1









#49
Acenaphthylene
Concen: 46.888 ng
RT: 14.234 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:152 Resp: 1768676
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 13.5 10.6 16.0



#50
Dimethylphthalate
Concen: 46.237 ng
RT: 13.975 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:163 Resp: 1348607
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 10.1 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 45.766 ng
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:165 Resp: 300943
Ion Ratio Lower Upper
165 100
63 46.8 41.3 61.9
89 50.2 44.8 67.2



#52
Acenaphthene
Concen: 44.928 ng
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:154 Resp: 1030928
Ion Ratio Lower Upper
154 100
153 114.4 89.4 134.2
152 53.7 43.0 64.4









#53
3-Nitroaniline
Concen: 34.689 ng
RT: 14.439 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:138 Resp: 231105
Ion Ratio Lower Upper
138 100
108 10.9 8.4 12.6
92 101.8 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 88.507 ng
RT: 14.657 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:184 Resp: 309405
Ion Ratio Lower Upper
184 100
63 60.9 47.0 70.6
154 66.0 52.3 78.5









#55
Dibenzenofuran
Concen: 44.511 ng
RT: 14.910 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:168 Resp: 1589951
Ion Ratio Lower Upper
168 100
139 40.5 31.0 46.6
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 91.872 ng
RT: 14.769 min Scan# 1986
Delta R.T. -0.012 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:139 Resp: 477625
Ion Ratio Lower Upper
139 100
109 71.4 72.3 112.3#
65 81.9 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 47.557 ng
RT: 14.898 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:	165	Resp:	421180
Ion	Ratio	Lower	Upper
165	100		
63	43.8	32.3	48.5
89	68.3	55.9	83.9
182	2.7	2.2	3.4



#58
Fluorene
Concen: 45.343 ng
RT: 15.563 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	166	Resp:	1301200
Ion	Ratio	Lower	Upper
166	100		
165	99.4	78.0	117.0
167	13.4	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 46.496 ng
RT: 15.151 min Scan# 2051
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:232 Resp: 284136
Ion Ratio Lower Upper
232 100
131 57.0 32.2 48.4#
130 3.0 2.0 3.0
166 37.3 19.4 29.2#



#60
Diethylphthalate
Concen: 46.343 ng
RT: 15.334 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:149 Resp: 1386203
Ion Ratio Lower Upper
149 100
177 20.8 17.5 26.3
150 12.5 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 45.451 ng
RT: 15.551 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:204 Resp: 571798
Ion Ratio Lower Upper
204 100
206 32.3 26.6 40.0
141 72.8 48.6 73.0



#62
4-Nitroaniline
Concen: 43.991 ng
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36


Tgt Ion:138 Resp: 297607
Ion Ratio Lower Upper
138 100
92 46.6 31.2 71.2
108 75.1 78.6 118.6#








#63
Azobenzene
Concen: 45.119 ng
RT: 15.845 min Scan# 2169
Delta R.T. -0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Instrument :
BNA_P
ClientSampleId :
PB137592BS







Tgt Ion:	77	Resp:	1126693
Ion	Ratio	Lower	Upper
77	100		
182	27.0	6.0	46.0
105	23.4	0.0	38.8
51	28.5	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2410
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:	188	Resp:	747524
Ion	Ratio	Lower	Upper
188	100		
94	9.9	5.6	8.4#
80	10.8	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 41.000 ng
RT: 15.669 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:198 Resp: 196641
Ion Ratio Lower Upper
198 100
51 39.5 14.0 54.0
105 44.7 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 46.064 ng
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:169 Resp: 1121198
Ion Ratio Lower Upper
169 100
168 66.8 54.4 81.6
167 36.4 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 45.391 ng
RT: 16.451 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:248 Resp: 334493
Ion Ratio Lower Upper
248 100
250 96.7 79.8 119.8
141 90.4 57.8 86.6#



#68
Hexachlorobenzene
Concen: 45.386 ng
RT: 16.569 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:284 Resp: 396999
Ion Ratio Lower Upper
284 100
142 41.5 26.9 40.3#
249 30.3 24.5 36.7





#69

Atrazine

Concen: 47.311 ng

RT: 16.733 min Scan# 2320

Delta R.T. 0.000 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36





Instrument :

BNA_P

ClientSampleId :

PB137592BS

Tgt Ion:	200	Resp:	361533
Ion Ratio	Lower	Upper	
200	100		
173	28.9	4.1	44.1
215	46.0	27.7	67.7



#70

Pentachlorophenol

Concen: 106.917 ng




RT: 16.922 min Scan# 2352


Delta R.T. -0.006 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36

Tgt Ion:	266	Resp:	431630
Ion Ratio	Lower	Upper	
266	100		
268	62.6	51.6	77.4
264	62.1	50.3	75.5









#71
Phenanthrene
Concen: 46.052 ng
RT: 17.310 min Scan# 2418
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:178 Resp: 1944163
Ion Ratio Lower Upper
178 100
176 19.3 15.1 22.7
179 15.2 12.4 18.6



#72
Anthracene
Concen: 47.366 ng
RT: 17.398 min Scan# 2433
Delta R.T. -0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:178 Resp: 1958969
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.4 12.2 18.2









#73
 Carbazole
 Concen: 45.291 ng
 RT: 17.675 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BP006160.D
 Acq: 09 Jul 2021 14:36




Instrument :
 BNA_P
 ClientSampleId :
 PB137592BS


Tgt Ion:167 Resp: 1856159
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.1 25.7
 139 13.9 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 48.084 ng
 RT: 18.216 min Scan# 2572
 Delta R.T. -0.000 min
 Lab File: BP006160.D
 Acq: 09 Jul 2021 14:36

Tgt Ion:149 Resp: 2372025
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 4.7 4.5 6.7









#75
 Fluoranthene
 Concen: 45.568 ng
 RT: 19.274 min Scan# 2752
 Delta R.T. -0.000 min
 Lab File: BP006160.D
 Acq: 09 Jul 2021 14:36




Instrument :
 BNA_P
 ClientSampleId :
 PB137592BS


Tgt Ion:	202	Resp:	2108450
Ion	Ratio	Lower	Upper
202	100		
101	12.0	0.0	27.6
203	17.4	0.0	37.8



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.345 min Scan# 3104
 Delta R.T. -0.000 min
 Lab File: BP006160.D
 Acq: 09 Jul 2021 14:36

Tgt Ion:	240	Resp:	683139
Ion	Ratio	Lower	Upper
240	100		
120	14.4	6.5	9.7#
236	25.7	20.6	31.0









#77
Benzidine
Concen: 44.325 ng
RT: 19.457 min Scan# 2783
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:184 Resp: 897881
Ion Ratio Lower Upper
184 100
185 14.1 11.0 16.4
183 12.0 9.5 14.3



#78
Pyrene
Concen: 46.968 ng
RT: 19.627 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:202 Resp: 2168710
Ion Ratio Lower Upper
202 100
200 21.0 16.3 24.5
203 17.5 14.2 21.2





#79

Terphenyl-d14

Concen: 86.326 ng

RT: 19.816 min Scan# 2844

Delta R.T. -0.000 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36

Instrument :

BNA_P

ClientSampleId :

PB137592BS





Tgt Ion:244 Resp: 2991064

Ion Ratio Lower Upper

244 100

212 7.6 6.0 9.0

122 13.7 7.1 10.7#



#80

Butylbenzylphthalate

Concen: 48.641 ng

RT: 20.492 min Scan# 2959

Delta R.T. 0.006 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36




Tgt Ion:149 Resp: 1114033


Ion Ratio Lower Upper

149 100

91 61.4 54.6 81.8

206 18.9 16.7 25.1





#81

Benzo(a)anthracene

Concen: 45.228 ng

RT: 21.333 min Scan# 3102

Delta R.T. 0.006 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36





Tgt Ion:228 Resp: 1954172

Ion Ratio Lower Upper

228 100

226 27.0 21.6 32.4

229 19.1 15.8 23.6



#82

3,3'-Dichlorobenzidine

Concen: 49.164 ng

RT: 21.263 min Scan# 3090

Delta R.T. -0.000 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36




Tgt Ion:252 Resp: 623148


Ion Ratio Lower Upper

252 100

254 65.3 50.6 76.0

126 14.7 7.0 10.6#





#83

Chrysene

Concen: 45.214 ng

RT: 21.386 min Scan# 3111

Delta R.T. 0.006 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36





Tgt Ion:228 Resp: 1930596

Ion Ratio Lower Upper

228 100

226 29.7 23.8 35.6

229 19.9 15.4 23.2



#84

Bis(2-ethylhexyl)phthalate

Concen: 48.866 ng

RT: 21.245 min Scan# 3087

Delta R.T. 0.000 min

Lab File: BP006160.D

Acq: 09 Jul 2021 14:36




Tgt Ion:149 Resp: 1651915


Ion Ratio Lower Upper

149 100

167 26.8 21.9 32.9

279 3.8 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 48.424 ng
RT: 22.157 min Scan# 3242
Delta R.T. 0.000 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:149 Resp: 2883138
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 7.1 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.745 min Scan# 3512
Delta R.T. 0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:264 Resp: 732141
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 22.6 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 53.871 ng
RT: 26.280 min Scan# 3943
Delta R.T. 0.018 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:276 Resp: 2907819
Ion Ratio Lower Upper
276 100
138 31.2 15.3 22.9#
277 25.6 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 50.960 ng
RT: 23.015 min Scan# 3388
Delta R.T. 0.012 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:252 Resp: 2263205
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 12.5 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 47.458 ng
RT: 23.063 min Scan# 3396
Delta R.T. 0.006 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36




Instrument :
BNA_P
ClientSampleId :
PB137592BS


Tgt Ion:252 Resp: 2060100
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 12.5 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 51.954 ng
RT: 23.645 min Scan# 3495
Delta R.T. 0.012 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:252 Resp: 2165226
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 13.9 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 53.698 ng
RT: 26.286 min Scan# 3944
Delta R.T. 0.018 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Instrument :
BNA_P
ClientSampleId :
PB137592BS

Tgt Ion:278 Resp: 2495820
Ion Ratio Lower Upper
278 100
139 20.1 9.3 13.9#
279 24.1 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 54.821 ng
RT: 27.051 min Scan# 4074
Delta R.T. 0.018 min
Lab File: BP006160.D
Acq: 09 Jul 2021 14:36

Tgt Ion:276 Resp: 2507657
Ion Ratio Lower Upper
276 100
277 24.4 18.9 28.3
138 26.1 12.6 18.8#

