

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071025\  
 Data File : BP025086.D  
 Acq On : 10 Jul 2025 16:57  
 Operator : CG/JU  
 Sample : Q2524-01  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 OLD RETENTION BASIN 40TH

Quant Time: Jul 10 17:21:39 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP070325.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jul 03 16:05:44 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.437  | 152  | 602912   | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.178 | 136  | 2460428  | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.084 | 164  | 1607858  | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 16.895 | 188  | 3360998  | 20.000 | ng    | -0.01    |
| 76) Chrysene-d12            | 21.336 | 240  | 3719680  | 20.000 | ng    | 0.00     |
| 86) Perylene-d12            | 24.477 | 264  | 4074446  | 20.000 | ng    | -0.02    |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.096  | 112  | 2456865  | 72.203 | ng    | 0.00     |
| 7) Phenol-d6                | 6.631  | 99   | 3399212  | 76.550 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.566  | 82   | 2139789  | 57.838 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.613 | 330  | 1465486  | 77.154 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.689 | 172  | 6020432  | 58.007 | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.677 | 244  | 10659521 | 62.982 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071025\  
Data File : BP025086.D  
Acq On : 10 Jul 2025 16:57  
Operator : CG/JU  
Sample : Q2524-01  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
OLD RETENTION BASIN 40TH

Quant Time: Jul 10 17:21:39 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP070325.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jul 03 16:05:44 2025  
Response via : Initial Calibration

