

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010949.D
 Acq On : 11 Jul 2022 13:43
 Operator : CG/JU
 Sample : PB146140BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS140

Quant Time: Jul 11 14:19:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	426845	20.000	ng/u1	0.00
20) Naphthalene-d8	10.493	136	1660825	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	794828	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	1403340	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.233	240	1421992	20.000	ng/u1	# 0.00
88) Perylene-d12	23.592	264	1639765	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	79326	7.030	ng/uL	0.00
4) Pyridine-d5	3.605	84	907990	30.230	ng/u1	0.00
7) Phenol-d5	6.881	99	1154082	34.155	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	746491	33.781	ng/u1	0.00
11) 2-Chlorophenol-d4	7.234	132	969427	34.581	ng/u1	0.00
15) 4-Methylphenol-d8	8.416	113	874791	32.750	ng/u1	0.00
21) Nitrobenzene-d5	8.863	128	428658	40.544	ng/u1	0.00
24) 2-Nitrophenol-d4	9.581	143	396713	45.354	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	729565	33.756	ng/u1	0.00
31) 4-Chloroaniline-d4	10.640	131	1250110	35.434	ng/u1	0.00
46) Dimethylphthalate-d6	13.775	166	1757237	32.275	ng/u1	0.00
49) Acenaphthylene-d8	14.045	160	2347567	33.784	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	327988	38.568	ng/u1	0.00
60) Fluorene-d10	15.357	176	1541496	32.992	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	205995	44.306	ng/u1	0.00
73) Anthracene-d10	17.222	188	2182279	34.952	ng/u1	0.00
81) Pyrene-d10	19.469	212	2409110	31.097	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.445	264	2908590	35.895	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	155497	12.922	ng/uL#	86
5) Pyridine	3.628	79	870553	28.612	ng/u1#	73
6) Benzaldehyde	6.852	77	687783	38.578	ng/u1	95
8) Phenol	6.905	94	1129151	31.245	ng/u1#	85
10) Bis(2-Chloroethyl)ether	7.134	93	917227	31.601	ng/u1	91
12) 2-Chlorophenol	7.269	128	923884	31.927	ng/u1	92
13) 2-Methylphenol	8.152	108	802801	29.839	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.234	45	1237796	31.785	ng/u1#	98
16) Acetophenone	8.528	105	1254328	30.578	ng/u1#	89
17) N-Nitroso-di-n-propyla...	8.522	70	585119	28.209	ng/u1	90
18) 4-Methylphenol	8.481	108	854031	30.481	ng/u1	98
19) Hexachloroethane	8.775	117	366536	31.277	ng/u1#	83
22) Nitrobenzene	8.904	77	947543	35.913	ng/u1	93
23) Isophorone	9.434	82	1542401	27.876	ng/u1	98
25) 2-Nitrophenol	9.610	139	416146	39.286	ng/u1#	82
26) 2,4-Dimethylphenol	9.681	107	863608	29.841	ng/u1	95
27) Bis(2-Chloroethoxy)met...	9.922	93	1060947	30.396	ng/u1	97
29) 2,4-Dichlorophenol	10.146	162	693959	30.900	ng/u1	99
30) Naphthalene	10.546	128	2692823	31.109	ng/u1	100
32) 4-Chloroaniline	10.663	127	993333	28.418	ng/u1	100
33) Hexachlorobutadiene	10.828	225	414395	28.882	ng/u1	96
34) Caprolactam	11.451	113	180783	25.624	ng/u1	91
35) 4-Chloro-3-methylphenol	11.798	107	706892	29.042	ng/u1	94
36) 2-Methylnaphthalene	12.163	142	1619189	28.887	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010949.D
 Acq On : 11 Jul 2022 13:43
 Operator : CG/JU
 Sample : PB146140BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS140

Quant Time: Jul 11 14:19:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.387	142	1604818	28.819	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.534	216	726743	31.818	ng/ul	99
40) Hexachlorocyclopentadiene	12.510	237	429172	29.653	ng/ul	95
41) 2,4,6-Trichlorophenol	12.781	196	431220	32.747	ng/ul	99
42) 2,4,5-Trichlorophenol	12.851	196	473460	33.505	ng/ul	99
43) 1,1'-Biphenyl	13.187	154	1992600	31.845	ng/ul#	98
44) 2-Chloronaphthalene	13.222	162	1541894	32.712	ng/ul	98
45) 2-Nitroaniline	13.439	65	400785	40.076	ng/ul#	83
47) Dimethylphthalate	13.822	163	1654399	30.342	ng/ul	98
48) 2,6-Dinitrotoluene	13.939	165	305441	37.387	ng/ul	97
50) Acenaphthylene	14.075	152	2312885	30.391	ng/ul	98
51) 3-Nitroaniline	14.269	138	363382	37.101	ng/ul	90
52) Acenaphthene	14.422	153	1583383	32.245	ng/ul	99
53) 2,4-Dinitrophenol	14.475	184	119970	38.251	ng/ul	87
55) 4-Nitrophenol	14.586	109	240018	35.495	ng/ul#	79
56) Dibenzofuran	14.757	168	2103553	31.274	ng/ul	96
57) 2,4-Dinitrotoluene	14.728	165	433391	39.701	ng/ul	86
58) 2,3,4,6-Tetrachlorophenol	14.986	232	327700	32.383	ng/ul	89
59) Diethylphthalate	15.198	149	1565329	28.876	ng/ul	98
61) Fluorene	15.410	166	1645318	30.677	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.410	204	755117	30.076	ng/ul	95
63) 4-Nitroaniline	15.439	138	384804	42.493	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.492	198	199243	39.657	ng/ul	98
67) N-Nitrosodiphenylamine	15.628	169	1356346	31.911	ng/ul	98
68) 4-Bromophenyl-phenylether	16.310	248	429845	30.563	ng/ul	95
69) Hexachlorobenzene	16.416	284	486788	29.883	ng/ul	95
70) Atrazine	16.592	200	212678	15.092	ng/ul	98
71) Pentachlorophenol	16.769	266	234297	27.933	ng/ul	95
72) Phenanthrene	17.163	178	2443148	32.692	ng/ul	99
74) Anthracene	17.257	178	2466371	32.814	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.145	216	741935	32.039	ng/uL	97
76) Pentachlorobenzene	14.675	250	637481	32.097	ng/uL	100
77) Carbazole	17.533	167	2279522	33.692	ng/ul	99
78) Di-n-butylphthalate	18.104	149	2255961	27.982	ng/ul	99
80) Fluoranthene	19.145	202	2669586	28.019	ng/ul#	88
82) Pyrene	19.498	202	2819153	29.298	ng/ul#	83
83) Butylbenzylphthalate	20.392	149	1024322	30.244	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.163	252	932342	31.752	ng/ul#	98
85) Benzo(a)anthracene	21.221	228	2894929	32.552	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.163	149	1548162	28.866	ng/ul#	97
87) Chrysene	21.268	228	2794899	32.796	ng/ul	100
89) Di-n-octyl phthalate	22.074	149	2780914	28.572	ng/ul	100
90) Benzo(b)fluoranthene	22.874	252	3301465	32.376	ng/ul#	95
91) Benzo(k)fluoranthene	22.921	252	3250736	33.538	ng/ul#	95
93) Benzo(a)pyrene	23.492	252	3017227	30.417	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.039	276	4150535	34.675	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.062	278	3473436	33.830	ng/ul#	92
96) Benzo(g,h,i)perylene	26.792	276	3419677	33.586	ng/ul#	88





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 12.922 ng/uL
 RT: 3.234 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43




Instrument :
 BNA_P
 ClientSampleId :
 SLCS140


Tgt Ion: 88 Resp: 155497
 Ion Ratio Lower Upper
 88 100
 43 29.2 35.8 53.6#
 58 77.6 68.1 102.1



#5
 Pyridine
 Concen: 28.612 ng/uL
 RT: 3.628 min Scan# 92
 Delta R.T. -0.000 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43

Tgt Ion: 79 Resp: 870553
 Ion Ratio Lower Upper
 79 100
 52 75.1 83.8 125.8#
 51 35.5 41.0 61.4#









#6
Benzaldehyde
Concen: 38.578 ng/ul
RT: 6.852 min Scan# 640
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion: 77 Resp: 687783
Ion Ratio Lower Upper
77 100
105 99.8 76.0 114.0
106 95.5 73.2 109.8



#8
Phenol
Concen: 31.245 ng/ul
RT: 6.905 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43


Tgt Ion: 94 Resp: 1129151
Ion Ratio Lower Upper
94 100
65 27.6 27.4 41.2
66 35.7 37.4 56.2#








#10
Bis(2-Chloroethyl)ether
Concen: 31.601 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Instrument :
BNA_P
ClientSampleId :
SLCS140



Tgt Ion: 93 Resp: 917227




Ion	Ratio	Lower	Upper
93	100		
63	79.4	72.4	108.6
95	32.4	27.3	40.9




#12
2-Chlorophenol
Concen: 31.927 ng/ul
RT: 7.269 min Scan# 711
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 128 Resp: 923884

Ion	Ratio	Lower	Upper
128	100		
64	47.0	44.8	67.2
130	32.3	25.8	38.6









#13
2-Methylphenol
Concen: 29.839 ng/ul
RT: 8.152 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:108 Resp: 802801
Ion Ratio Lower Upper
108 100
107 88.8 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.785 ng/ul
RT: 8.234 min Scan# 875
Delta R.T. -0.012 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 45 Resp: 1237796
Ion Ratio Lower Upper
45 100
77 13.7 11.2 16.8
79 10.5 6.9 10.3#









#16
Acetophenone
Concen: 30.578 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:105 Resp: 1254328
Ion Ratio Lower Upper
105 100
77 76.8 67.5 101.3
51 27.0 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 28.209 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 70 Resp: 585119
Ion Ratio Lower Upper
70 100
42 57.8 53.8 80.8
101 10.4 7.8 11.6
130 22.2 15.4 23.2









#18
4-Methylphenol
Concen: 30.481 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:108 Resp: 854031
Ion Ratio Lower Upper
108 100
107 112.2 88.3 132.5



#19
Hexachloroethane
Concen: 31.277 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:117 Resp: 366536
Ion Ratio Lower Upper
117 100
201 89.8 85.7 128.5
199 57.1 57.4 86.0#









#22
Nitrobenzene
Concen: 35.913 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion: 77 Resp: 947543
Ion Ratio Lower Upper
77 100
123 44.9 31.5 47.3
65 13.3 11.4 17.2



#23
Isophorone
Concen: 27.876 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 82 Resp: 1542401
Ion Ratio Lower Upper
82 100
95 6.7 6.6 10.0
138 17.1 13.8 20.6









#25
2-Nitrophenol
Concen: 39.286 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:139 Resp: 416146
Ion Ratio Lower Upper
139 100
65 52.2 54.2 81.4#
109 29.1 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 29.841 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:107 Resp: 863608
Ion Ratio Lower Upper
107 100
121 51.6 39.1 58.7
122 87.5 66.1 99.1









#27
Bis(2-Chloroethoxy)methane
Concen: 30.396 ng/ul
RT: 9.922 min Scan# 1161
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion: 93 Resp: 1060947
Ion Ratio Lower Upper
93 100
95 31.9 26.6 39.8
123 11.6 7.8 11.8



#29
2,4-Dichlorophenol
Concen: 30.900 ng/ul
RT: 10.146 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 162 Resp: 693959
Ion Ratio Lower Upper
162 100
164 65.1 52.5 78.7
98 38.2 31.4 47.0









#30
Naphthalene
Concen: 31.109 ng/ul
RT: 10.546 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:128 Resp: 2692823
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 28.418 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:127 Resp: 993333
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.2









#33
Hexachlorobutadiene
Concen: 28.882 ng/ul
RT: 10.828 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:225 Resp: 414395
Ion Ratio Lower Upper
225 100
223 63.8 53.7 80.5
227 65.0 49.4 74.2



#34
Caprolactam
Concen: 25.624 ng/ul
RT: 11.451 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:113 Resp: 180783
Ion Ratio Lower Upper
113 100
55 174.9 153.8 230.8
56 131.1 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 29.042 ng/ul
RT: 11.798 min Scan# 1481
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:107 Resp: 706892
Ion Ratio Lower Upper
107 100
144 26.0 18.6 27.8
142 79.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 28.887 ng/ul
RT: 12.163 min Scan# 1543
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:142 Resp: 1619189
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1









#37
1-Methylnaphthalene
Concen: 28.819 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:142 Resp: 1604818
Ion Ratio Lower Upper
142 100
141 91.7 73.3 109.9
116 3.8 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.818 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:216 Resp: 726743
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.6
179 22.8 18.1 27.1
108 20.2 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 29.653 ng/ul
RT: 12.510 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:237 Resp: 429172
Ion Ratio Lower Upper
237 100
235 61.8 52.8 79.2
272 13.1 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 32.747 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:196 Resp: 431220
Ion Ratio Lower Upper
196 100
198 98.2 78.2 117.4
200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 33.505 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:196 Resp: 473460
Ion Ratio Lower Upper
196 100
198 94.9 77.0 115.4
200 30.2 24.8 37.2



#43
1,1'-Biphenyl
Concen: 31.845 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:154 Resp: 1992600
Ion Ratio Lower Upper
154 100
153 39.7 31.9 47.9
76 13.9 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 32.712 ng/ul
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:162 Resp: 1541894
Ion Ratio Lower Upper
162 100
164 32.3 25.5 38.3
127 41.6 31.9 47.9



#45
2-Nitroaniline
Concen: 40.076 ng/ul
RT: 13.439 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion: 65 Resp: 400785
Ion Ratio Lower Upper
65 100
92 68.3 49.0 73.6
138 103.6 66.4 99.6#









#47
Dimethylphthalate
Concen: 30.342 ng/ul
RT: 13.822 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:163 Resp: 1654399
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 37.387 ng/ul
RT: 13.939 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:165 Resp: 305441
Ion Ratio Lower Upper
165 100
89 62.7 52.9 79.3
121 24.2 19.4 29.2









#50
Acenaphthylene
Concen: 30.391 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:152 Resp: 2312885
Ion Ratio Lower Upper
152 100
151 19.9 16.5 24.7
153 13.4 10.1 15.1



#51
3-Nitroaniline
Concen: 37.101 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:138 Resp: 363382
Ion Ratio Lower Upper
138 100
108 12.1 11.0 16.4
92 117.9 104.6 156.8









#52
Acenaphthene
Concen: 32.245 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:153 Resp: 1583383
Ion Ratio Lower Upper
153 100
152 47.3 36.2 54.2
154 87.9 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 38.251 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:184 Resp: 119970
Ion Ratio Lower Upper
184 100
63 74.0 73.8 110.6
154 67.6 57.2 85.8









#55
4-Nitrophenol
Concen: 35.495 ng/ul
RT: 14.586 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:109 Resp: 240018
Ion Ratio Lower Upper
109 100
139 143.3 75.6 113.4#
65 135.6 107.6 161.4



#56
Dibenzofuran
Concen: 31.274 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:168 Resp: 2103553
Ion Ratio Lower Upper
168 100
139 38.6 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 39.701 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:165 Resp: 433391
Ion Ratio Lower Upper
165 100
63 47.7 47.0 70.6
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 32.383 ng/ul
RT: 14.986 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:232 Resp: 327700
Ion Ratio Lower Upper
232 100
131 52.7 35.8 53.8
130 2.4 1.8 2.8
166 34.7 23.7 35.5









#59
Diethylphthalate
Concen: 28.876 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:149 Resp: 1565329
Ion Ratio Lower Upper
149 100
177 21.0 17.6 26.4
150 12.3 10.3 15.5



#61
Fluorene
Concen: 30.677 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:166 Resp: 1645318
Ion Ratio Lower Upper
166 100
165 97.7 77.4 116.2
167 13.8 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 30.076 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:204 Resp: 755117
Ion Ratio Lower Upper
204 100
206 33.2 26.0 39.0
141 67.3 49.3 73.9



#63
4-Nitroaniline
Concen: 42.493 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:138 Resp: 384804
Ion Ratio Lower Upper
138 100
92 52.6 46.2 69.2
108 74.9 85.8 128.8#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 39.657 ng/ul
 RT: 15.492 min Scan# 2109
 Delta R.T. -0.000 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43




Instrument :
 BNA_P
 ClientSampleId :
 SLCS140


Tgt Ion:198 Resp: 199243
 Ion Ratio Lower Upper
 198 100
 182 4.9 4.3 6.5
 77 28.3 23.7 35.5



#67
 N-Nitrosodiphenylamine
 Concen: 31.911 ng/ul
 RT: 15.628 min Scan# 2132
 Delta R.T. -0.006 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43

Tgt Ion:169 Resp: 1356346
 Ion Ratio Lower Upper
 169 100
 168 66.9 52.6 78.8
 167 35.8 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 30.563 ng/ul
RT: 16.310 min Scan# 2248
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:248 Resp: 429845
Ion Ratio Lower Upper
248 100
250 97.4 80.8 121.2
141 85.3 63.9 95.9



#69
Hexachlorobenzene
Concen: 29.883 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:284 Resp: 486788
Ion Ratio Lower Upper
284 100
142 39.4 27.8 41.6
249 31.0 25.4 38.2









#70
Atrazine
Concen: 15.092 ng/ul
RT: 16.592 min Scan# 212678
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion	Ratio	Lower	Upper
200	100		
173	27.8	20.7	31.1
215	48.0	38.6	57.8



#71
Pentachlorophenol
Concen: 27.933 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion	Ratio	Lower	Upper
266	100		
264	60.3	51.4	77.2
268	62.9	53.5	80.3









#72
Phenanthrene
Concen: 32.692 ng/ul
RT: 17.163 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:178 Resp: 2443148
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 18.9 14.6 22.0



#74
Anthracene
Concen: 32.814 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:178 Resp: 2466371
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.039 ng/uL
RT: 13.145 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:216 Resp: 741935
Ion Ratio Lower Upper
216 100
214 78.3 65.0 97.4
218 47.6 39.1 58.7



#76
Pentachlorobenzene
Concen: 32.097 ng/uL
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:250 Resp: 637481
Ion Ratio Lower Upper
250 100
248 62.1 49.8 74.6
252 63.7 51.2 76.8









#77
 Carbazole
 Concen: 33.692 ng/ul
 RT: 17.533 min Scan# 2456
 Delta R.T. -0.000 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43




Instrument :
 BNA_P
 ClientSampleId :
 SLCS140


Tgt Ion:167 Resp: 2279522
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.9 25.3
 139 13.4 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 27.982 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. -0.000 min
 Lab File: BP010949.D
 Acq: 11 Jul 2022 13:43

Tgt Ion:149 Resp: 2255961
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.0 4.6 6.8









#80
Fluoranthene
Concen: 28.019 ng/ul
RT: 19.145 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:202 Resp: 2669586
Ion Ratio Lower Upper
202 100
101 13.8 7.1 10.7#
100 10.6 5.7 8.5#



#82
Pyrene
Concen: 29.298 ng/ul
RT: 19.498 min Scan# 2790
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:202 Resp: 2819153
Ion Ratio Lower Upper
202 100
101 16.7 7.9 11.9#
100 13.7 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 30.244 ng/ul
RT: 20.392 min Scan# 2942
Delta R.T. -0.006 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:149 Resp: 1024322
Ion Ratio Lower Upper
149 100
91 78.2 60.0 90.0
206 18.2 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 31.752 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:252 Resp: 932342
Ion Ratio Lower Upper
252 100
254 64.7 51.0 76.4
126 13.2 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 32.552 ng/ul

RT: 21.221 min Scan# 3083

Delta R.T. -0.000 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43

Instrument :

BNA_P

ClientSampleId :

SLCS140





Tgt Ion:228 Resp: 2894929

Ion Ratio Lower Upper

228 100

229 19.6 15.0 22.4

226 26.8 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 28.866 ng/ul

RT: 21.163 min Scan# 3073

Delta R.T. -0.000 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43




Tgt Ion:149 Resp: 1548162


Ion Ratio Lower Upper

149 100

167 26.9 22.6 33.8

279 3.9 4.2 6.4#





#87

Chrysene

Concen: 32.796 ng/ul

RT: 21.268 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43

Instrument :

BNA_P

ClientSampleId :

SLCS140





Tgt Ion:228 Resp: 2794899

Ion Ratio Lower Upper

228 100

226 29.7 23.6 35.4

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 28.572 ng/ul




RT: 22.074 min Scan# 3228


Delta R.T. -0.006 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43

Tgt Ion:149 Resp: 2780914





#90

Benzo(b)fluoranthene

Concen: 32.376 ng/u1

RT: 22.874 min Scan# 3364

Delta R.T. -0.000 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43

Instrument :

BNA_P

ClientSampleId :

SLCS140





Tgt Ion:252 Resp: 3301465

Ion Ratio Lower Upper

252 100

253 22.5 17.8 26.6

125 13.2 6.2 9.2#



#91

Benzo(k)fluoranthene

Concen: 33.538 ng/u1

RT: 22.921 min Scan# 3372

Delta R.T. -0.006 min

Lab File: BP010949.D

Acq: 11 Jul 2022 13:43




Tgt Ion:252 Resp: 3250736


Ion Ratio Lower Upper

252 100

253 21.9 17.0 25.4

125 13.3 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 30.417 ng/ul
RT: 23.492 min Scan# 3469
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43




Instrument :
BNA_P
ClientSampleId :
SLCS140


Tgt Ion:252 Resp: 3017227
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 14.5 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 34.675 ng/ul
RT: 26.039 min Scan# 3902
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:276 Resp: 4150535
Ion Ratio Lower Upper
276 100
138 29.0 14.6 22.0#
277 24.7 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 33.830 ng/ul
RT: 26.062 min Scan# 3905
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Instrument :
BNA_P
ClientSampleId :
SLCS140

Tgt Ion:278 Resp: 3473436
Ion Ratio Lower Upper
278 100
139 20.2 9.4 14.2#
279 23.8 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 33.586 ng/ul
RT: 26.792 min Scan# 4030
Delta R.T. -0.000 min
Lab File: BP010949.D
Acq: 11 Jul 2022 13:43

Tgt Ion:276 Resp: 3419677
Ion Ratio Lower Upper
276 100
138 26.9 12.7 19.1#
277 24.0 18.2 27.4

