

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010952.D
 Acq On : 11 Jul 2022 16:59
 Operator : CG/JU
 Sample : PB146134BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS134

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2022
 Supervised By :mohammad ahmed 07/12/2022

Quant Time: Jul 11 18:22:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	405047	20.000	ng/u1	0.00
20) Naphthalene-d8	10.493	136	1603070	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	845202	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	1714023	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.233	240	1760360	20.000	ng/u1	# 0.00
88) Perylene-d12	23.592	264	1889122	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	65837	6.149	ng/uL	0.00
4) Pyridine-d5	3.605	84	817472	28.681	ng/u1	0.00
7) Phenol-d5	6.875	99	999046	31.158	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.040	67	623818	29.749	ng/u1	0.00
11) 2-Chlorophenol-d4	7.234	132	824252	30.984	ng/u1	0.00
15) 4-Methylphenol-d8	8.416	113	762386	30.078	ng/u1	0.00
21) Nitrobenzene-d5	8.863	128	360152	35.291	ng/u1	0.00
24) 2-Nitrophenol-d4	9.581	143	338572	40.101	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	643040	30.824	ng/u1	0.00
31) 4-Chloroaniline-d4	10.640	131	704055	20.675	ng/u1	0.00
46) Dimethylphthalate-d6	13.775	166	1699873	29.360	ng/u1	0.00
49) Acenaphthylene-d8	14.046	160	2199784	29.770	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	347004	38.372	ng/u1	0.00
60) Fluorene-d10	15.357	176	1500701	30.205	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	224417	39.519	ng/u1	0.00
73) Anthracene-d10	17.222	188	2300316	30.164	ng/u1	0.00
81) Pyrene-d10	19.469	212	2723463	28.398	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.445	264	2918535	31.264	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	140175	12.276	ng/uL#	84
5) Pyridine	3.623	79	815333	28.239	ng/u1#	76
6) Benzaldehyde	6.846	77	632416	37.382	ng/u1	97
8) Phenol	6.905	94	1026853m	29.943	ng/u1	
10) Bis(2-Chloroethyl)ether	7.134	93	806542	29.283	ng/u1	91
12) 2-Chlorophenol	7.264	128	825432	30.059	ng/u1	93
13) 2-Methylphenol	8.152	108	730595	28.616	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.240	45	1059289	28.665	ng/u1#	97
16) Acetophenone	8.528	105	1128918	29.001	ng/u1#	88
17) N-Nitroso-di-n-propyla...	8.516	70	528327	26.842	ng/u1	90
18) 4-Methylphenol	8.481	108	782901	29.446	ng/u1	99
19) Hexachloroethane	8.769	117	324607	29.189	ng/u1#	80
22) Nitrobenzene	8.905	77	853167	33.501	ng/u1	91
23) Isophorone	9.434	82	1422212	26.630	ng/u1	97
25) 2-Nitrophenol	9.610	139	379869	37.154	ng/u1#	79
26) 2,4-Dimethylphenol	9.681	107	797912	28.564	ng/u1	94
27) Bis(2-Chloroethoxy)met...	9.916	93	953151	28.291	ng/u1	99
29) 2,4-Dichlorophenol	10.146	162	639248	29.490	ng/u1	99
30) Naphthalene	10.546	128	2439158	29.194	ng/u1	100
32) 4-Chloroaniline	10.663	127	962104	28.516	ng/u1	100
33) Hexachlorobutadiene	10.828	225	364731	26.336	ng/u1	96
34) Caprolactam	11.452	113	203708	29.914	ng/u1	88
35) 4-Chloro-3-methylphenol	11.799	107	688125	29.289	ng/u1	94
36) 2-Methylnaphthalene	12.163	142	1490264	27.545	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010952.D
 Acq On : 11 Jul 2022 16:59
 Operator : CG/JU
 Sample : PB146134BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS134


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2022
 Supervised By :mohammad ahmed 07/12/2022

Quant Time: Jul 11 18:22:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.381	142	1502739	27.958	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.534	216	676817	27.866	ng/ul	98
40) Hexachlorocyclopentadiene	12.510	237	396758	25.779	ng/ul	96
41) 2,4,6-Trichlorophenol	12.781	196	429141	30.646	ng/ul	98
42) 2,4,5-Trichlorophenol	12.851	196	471382	31.370	ng/ul	99
43) 1,1'-Biphenyl	13.187	154	1862890	27.998	ng/ul#	98
44) 2-Chloronaphthalene	13.222	162	1462886	29.186	ng/ul	97
45) 2-Nitroaniline	13.440	65	419862	39.481	ng/ul#	79
47) Dimethylphthalate	13.822	163	1696662	29.262	ng/ul	98
48) 2,6-Dinitrotoluene	13.940	165	320214	36.859	ng/ul	97
50) Acenaphthylene	14.075	152	2288629	28.280	ng/ul	98
51) 3-Nitroaniline	14.269	138	396241	38.045	ng/ul	89
52) Acenaphthene	14.422	153	1560471	29.884	ng/ul	99
53) 2,4-Dinitrophenol	14.475	184	141295	42.365	ng/ul#	83
55) 4-Nitrophenol	14.587	109	263662	36.668	ng/ul#	74
56) Dibenzofuran	14.757	168	2105776	29.441	ng/ul	96
57) 2,4-Dinitrotoluene	14.728	165	481513	41.480	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	14.987	232	367353	34.138	ng/ul	90
59) Diethylphthalate	15.198	149	1694397	29.394	ng/ul	98
61) Fluorene	15.410	166	1707039	29.931	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.410	204	765183	28.660	ng/ul	95
63) 4-Nitroaniline	15.440	138	436021	45.279	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.493	198	231262	37.687	ng/ul	98
67) N-Nitrosodiphenylamine	15.628	169	1444391	27.823	ng/ul	99
68) 4-Bromophenyl-phenylether	16.310	248	458799	26.709	ng/ul	95
69) Hexachlorobenzene	16.416	284	527524	26.514	ng/ul	96
70) Atrazine	16.592	200	466644	27.111	ng/ul	97
71) Pentachlorophenol	16.769	266	290598	28.365	ng/ul	98
72) Phenanthrene	17.163	178	2732659	29.938	ng/ul	98
74) Anthracene	17.257	178	2754875	30.009	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.146	216	699365	24.726	ng/uL	98
76) Pentachlorobenzene	14.669	250	634802	26.169	ng/uL	99
77) Carbazole	17.534	167	2644982	32.008	ng/ul	99
78) Di-n-butylphthalate	18.098	149	2818234	28.620	ng/ul	99
80) Fluoranthene	19.145	202	3181382	26.972	ng/ul#	87
82) Pyrene	19.498	202	3343594	28.069	ng/ul#	84
83) Butylbenzylphthalate	20.392	149	1366170	32.583	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.163	252	1089748	29.979	ng/ul#	99
85) Benzo(a)anthracene	21.222	228	3339248	30.331	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.163	149	2029415	30.565	ng/ul#	98
87) Chrysene	21.269	228	3175257	30.097	ng/ul	99
89) Di-n-octyl phthalate	22.075	149	3526163	31.446	ng/ul	100
90) Benzo(b)fluoranthene	22.874	252	3583035	30.499	ng/ul#	95
91) Benzo(k)fluoranthene	22.922	252	3527015	31.585	ng/ul#	95
93) Benzo(a)pyrene	23.492	252	3194701	27.955	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.039	276	4217594	30.585	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.063	278	3509137	29.667	ng/ul#	92
96) Benzo(g,h,i)perylene	26.786	276	3469376	29.576	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.276 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion: 88 Resp: 140179
Ion Ratio Lower Upper
88 100
43 28.6 35.8 53.6
58 74.1 68.1 102.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#5
Pyridine
Concen: 28.239 ng/uL
RT: 3.623 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion: 79 Resp: 815333
Ion Ratio Lower Upper
79 100
52 79.2 83.8 125.8#
51 36.2 41.0 61.4#





#6
Benzaldehyde
Concen: 37.382 ng/ul
RT: 6.846 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion: 77 Resp: 632410
Ion Ratio Lower Upper
77 100
105 98.6 76.0 114.0
106 93.8 73.2 109.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#8
Phenol
Concen: 29.943 ng/ul m
RT: 6.905 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion: 94 Resp: 1026853
Ion Ratio Lower Upper
94 100
65 27.6 27.4 41.2
66 35.6 37.4 56.2#





#10
Bis(2-Chloroethyl)ether
Concen: 29.283 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion: 93 Resp: 806542
Ion Ratio Lower Upper
93 100
63 79.2 72.4 108.6
95 32.7 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#12
2-Chlorophenol
Concen: 30.059 ng/ul
RT: 7.264 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59

Tgt Ion:128 Resp: 825432
Ion Ratio Lower Upper
128 100
64 48.4 44.8 67.2
130 32.7 25.8 38.6









#13
2-Methylphenol
Concen: 28.616 ng/ul
RT: 8.152 min Scan# 861
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59

Instrument :
BNA_P
ClientSampleId :
SLCS134

Tgt Ion:108 Resp: 73059
Ion Ratio Lower Upper
108 100
107 88.8 71.4 107.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.665 ng/ul
RT: 8.240 min Scan# 876
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion: 45 Resp: 1059289
Ion Ratio Lower Upper
45 100
77 14.0 11.2 16.8
79 11.0 6.9 10.3#





#16
Acetophenone
Concen: 29.001 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:105 Resp: 1128918
Ion Ratio Lower Upper
105 100
77 76.2 67.5 101.3
51 26.6 29.6 44.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#17
N-Nitroso-di-n-propylamine
Concen: 26.842 ng/ul
RT: 8.516 min Scan# 923
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion: 70 Resp: 528327
Ion Ratio Lower Upper
70 100
42 56.7 53.8 80.8
101 9.6 7.8 11.6
130 21.6 15.4 23.2





#18
4-Methylphenol
Concen: 29.446 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:108 Resp: 78290
Ion Ratio Lower Upper
108 100
107 109.2 88.3 132.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#19
Hexachloroethane
Concen: 29.189 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:117 Resp: 324607
Ion Ratio Lower Upper
117 100
201 86.8 85.7 128.5
199 54.7 57.4 86.0#





#22
Nitrobenzene
Concen: 33.501 ng/ul
RT: 8.905 min Scan# 989
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion: 77 Resp: 85316
Ion Ratio Lower Upper
77 100
123 46.6 31.5 47.3
65 13.4 11.4 17.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#23
Isophorone
Concen: 26.630 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion: 82 Resp: 1422212
Ion Ratio Lower Upper
82 100
95 6.8 6.6 10.0
138 18.1 13.8 20.6





#25
2-Nitrophenol
Concen: 37.154 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:139 Resp: 379869
Ion Ratio Lower Upper
139 100
65 49.1 54.2 81.4
109 28.0 30.6 45.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#26
2,4-Dimethylphenol
Concen: 28.564 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:107 Resp: 797912
Ion Ratio Lower Upper
107 100
121 52.8 39.1 58.7
122 88.4 66.1 99.1





#27
Bis(2-Chloroethoxy)methane
Concen: 28.291 ng/ul
RT: 9.916 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion: 93 Resp: 95315
Ion Ratio Lower Upper
93 100
95 33.0 26.6 39.8
123 11.4 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#29
2,4-Dichlorophenol
Concen: 29.490 ng/ul
RT: 10.146 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:162 Resp: 639248
Ion Ratio Lower Upper
162 100
164 64.7 52.5 78.7
98 38.0 31.4 47.0





#30
Naphthalene
Concen: 29.194 ng/ul
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:128 Resp: 2439158
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 12.8 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#32
4-Chloroaniline
Concen: 28.516 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:127 Resp: 962104
Ion Ratio Lower Upper
127 100
129 32.8 26.2 39.2





#33
Hexachlorobutadiene
Concen: 26.336 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:225 Resp: 364733
Ion Ratio Lower Upper
225 100
223 62.6 53.7 80.5
227 63.2 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#34
Caprolactam
Concen: 29.914 ng/ul
RT: 11.452 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:113 Resp: 203708
Ion Ratio Lower Upper
113 100
55 170.4 153.8 230.8
56 129.4 110.2 165.4





#35
4-Chloro-3-methylphenol
Concen: 29.289 ng/ul
RT: 11.799 min Scan# 1481
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:107 Resp: 688125
Ion Ratio Lower Upper
107 100
144 25.4 18.6 27.8
142 79.8 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#36
2-Methylnaphthalene
Concen: 27.545 ng/ul
RT: 12.163 min Scan# 1543
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59



Tgt Ion:142 Resp: 1490264
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1









#37
1-Methylnaphthalene
Concen: 27.958 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59

Instrument :
BNA_P
ClientSampleId :
SLCS134

Tgt Ion:142 Resp: 1502739
Ion Ratio Lower Upper
142 100
141 91.4 73.3 109.9
116 3.7 4.0 6.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 27.866 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:216 Resp: 676817
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.6
179 21.7 18.1 27.1
108 19.1 13.6 20.4





#40
Hexachlorocyclopentadiene
Concen: 25.779 ng/ul
RT: 12.510 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:237 Resp: 396758
Ion Ratio Lower Upper
237 100
235 63.2 52.8 79.2
272 12.5 11.5 17.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#41
2,4,6-Trichlorophenol
Concen: 30.646 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59



Tgt Ion:196 Resp: 429141
Ion Ratio Lower Upper
196 100
198 95.7 78.2 117.4
200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 31.370 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59

Instrument :
BNA_P
ClientSampleId :
SLCS134

Tgt Ion:196 Resp: 47138
Ion Ratio Lower Upper
196 100
198 94.6 77.0 115.4
200 31.3 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#43
1,1'-Biphenyl
Concen: 27.998 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:154 Resp: 1862890
Ion Ratio Lower Upper
154 100
153 40.1 31.9 47.9
76 14.2 9.2 13.8#





#44
2-Chloronaphthalene
Concen: 29.186 ng/ul
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:162 Resp: 146288
Ion Ratio Lower Upper
162 100
164 32.6 25.5 38.3
127 42.1 31.9 47.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#45
2-Nitroaniline
Concen: 39.481 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion: 65 Resp: 419862
Ion Ratio Lower Upper
65 100
92 70.1 49.0 73.6
138 108.7 66.4 99.6





#47
Dimethylphthalate
Concen: 29.262 ng/ul
RT: 13.822 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:163 Resp: 169666
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.2 7.6 11.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#48
2,6-Dinitrotoluene
Concen: 36.859 ng/ul
RT: 13.940 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:165 Resp: 320214
Ion Ratio Lower Upper
165 100
89 63.4 52.9 79.3
121 24.1 19.4 29.2





#50
Acenaphthylene
Concen: 28.280 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument : BNA_P
ClientSampleId : SLCS134



Tgt Ion:152 Resp: 2288629
Ion Ratio Lower Upper
152 100
151 20.0 16.5 24.7
153 13.3 10.1 15.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#51
3-Nitroaniline
Concen: 38.045 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:138 Resp: 396241
Ion Ratio Lower Upper
138 100
108 12.0 11.0 16.4
92 117.2 104.6 156.8





#52
Acenaphthene
Concen: 29.884 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:153 Resp: 156047
Ion Ratio Lower Upper
153 100
152 47.3 36.2 54.2
154 88.7 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#53
2,4-Dinitrophenol
Concen: 42.365 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:184 Resp: 141295
Ion Ratio Lower Upper
184 100
63 70.5 73.8 110.6#
154 63.0 57.2 85.8





#55
4-Nitrophenol
Concen: 36.668 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:109 Resp: 263662
Ion Ratio Lower Upper
109 100
139 147.2 75.6 113.4
65 142.0 107.6 161.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#56
Dibenzofuran
Concen: 29.441 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:168 Resp: 2105776
Ion Ratio Lower Upper
168 100
139 38.5 32.8 49.2





#57
2,4-Dinitrotoluene
Concen: 41.480 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:165 Resp: 48151
Ion Ratio Lower Upper
165 100
63 46.6 47.0 70.6
182 2.7 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 34.138 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:232 Resp: 367353
Ion Ratio Lower Upper
232 100
131 52.8 35.8 53.8
130 2.8 1.8 2.8
166 32.9 23.7 35.5





#59
Diethylphthalate
Concen: 29.394 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:149 Resp: 1694397
Ion Ratio Lower Upper
149 100
177 20.9 17.6 26.4
150 12.2 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#61
Fluorene
Concen: 29.931 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:166 Resp: 1707039
Ion Ratio Lower Upper
166 100
165 98.8 77.4 116.2
167 13.3 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 28.660 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:204 Resp: 76518
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 67.2 49.3 73.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#63
4-Nitroaniline
Concen: 45.279 ng/ul
RT: 15.440 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:138 Resp: 436021
Ion Ratio Lower Upper
138 100
92 52.5 46.2 69.2
108 75.7 85.8 128.8





#66
4,6-Dinitro-2-methylphenol
Concen: 37.687 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:198 Resp: 231262
Ion Ratio Lower Upper
198 100
182 5.3 4.3 6.5
77 28.4 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#67
N-Nitrosodiphenylamine
Concen: 27.823 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:169 Resp: 1444391
Ion Ratio Lower Upper
169 100
168 66.7 52.6 78.8
167 35.8 27.8 41.6





#68
4-Bromophenyl-phenylether
Concen: 26.709 ng/ul
RT: 16.310 min Scan# 2248
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:248 Resp: 458799
Ion Ratio Lower Upper
248 100
250 96.3 80.8 121.2
141 84.8 63.9 95.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#69
Hexachlorobenzene
Concen: 26.514 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:284 Resp: 527524
Ion Ratio Lower Upper
284 100
142 38.3 27.8 41.6
249 31.5 25.4 38.2





#70
Atrazine
Concen: 27.111 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:200 Resp: 466644
Ion Ratio Lower Upper
200 100
173 27.8 20.7 31.1
215 46.7 38.6 57.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#71
Pentachlorophenol
Concen: 28.365 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:266 Resp: 290598
Ion Ratio Lower Upper
266 100
264 62.2 51.4 77.2
268 65.9 53.5 80.3





#72
Phenanthrene
Concen: 29.938 ng/ul
RT: 17.163 min Scan# 2393
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:178 Resp: 2732659
Ion Ratio Lower Upper
178 100
179 15.2 12.7 19.1
176 19.3 14.6 22.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#74
Anthracene
Concen: 30.009 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:178 Resp: 2754875
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 18.2 14.5 21.7





#75
1,2,3,4-Tetrachlorobenzene
Concen: 24.726 ng/uL
RT: 13.146 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:216 Resp: 69936
Ion Ratio Lower Upper
216 100
214 79.1 65.0 97.4
218 47.8 39.1 58.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#76
Pentachlorobenzene
Concen: 26.169 ng/uL
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:250 Resp: 634802
Ion Ratio Lower Upper
250 100
248 63.5 49.8 74.6
252 63.4 51.2 76.8





#77
 Carbazole
 Concen: 32.008 ng/ul
 RT: 17.534 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BP010952.D
 Acq: 11 Jul 2022 16:59




Instrument :
 BNA_P
 ClientSampleId :
 SLCS134



Tgt Ion:167 Resp: 264498
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.9 25.3
 139 13.4 10.6 16.0




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
 Supervised By :mohammad ahmed 07/12/2022



#78
 Di-n-butylphthalate
 Concen: 28.620 ng/ul
 RT: 18.098 min Scan# 2552
 Delta R.T. -0.006 min
 Lab File: BP010952.D
 Acq: 11 Jul 2022 16:59


Tgt Ion:149 Resp: 2818234
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.3 4.6 6.8





#80
Fluoranthene
Concen: 26.972 ng/ul
RT: 19.145 min Scan# 2730
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:202 Resp: 318138
Ion Ratio Lower Upper
202 100
101 14.0 7.1 10.7
100 10.8 5.7 8.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#82
Pyrene
Concen: 28.069 ng/ul
RT: 19.498 min Scan# 2790
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:202 Resp: 3343594
Ion Ratio Lower Upper
202 100
101 16.4 7.9 11.9#
100 13.6 6.7 10.1#





#83
Butylbenzylphthalate
Concen: 32.583 ng/ul
RT: 20.392 min Scan# 2942
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:149 Resp: 1366170
Ion Ratio Lower Upper
149 100
91 76.3 60.0 90.0
206 18.1 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#84
3,3'-Dichlorobenzidine
Concen: 29.979 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Tgt Ion:252 Resp: 1089748
Ion Ratio Lower Upper
252 100
254 63.9 51.0 76.4
126 13.0 8.6 12.8





#85
Benzo(a)anthracene
Concen: 30.331 ng/ul
RT: 21.222 min Scan# 3083
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:228 Resp: 333924
Ion Ratio Lower Upper
228 100
229 19.6 15.0 22.4
226 26.9 20.6 31.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 30.565 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59



Tgt Ion:149 Resp: 2029415
Ion Ratio Lower Upper
149 100
167 27.1 22.6 33.8
279 3.9 4.2 6.4





#87

Chrysene

Concen: 30.097 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BP010952.D

Acq: 11 Jul 2022 16:59

Instrument :

BNA_P

ClientSampleId :

SLCS134

Tgt Ion:228 Resp: 317525

Ion Ratio Lower Upper

228 100





226 29.7 23.6 35.4

229 19.8 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#89

Di-n-octyl phthalate

Concen: 31.446 ng/ul




RT: 22.075 min Scan# 3228


Delta R.T. -0.006 min

Lab File: BP010952.D

Acq: 11 Jul 2022 16:59


Tgt Ion:149 Resp: 3526163





#90
Benzo(b)fluoranthene
Concen: 30.499 ng/ul
RT: 22.874 min Scan# 3364
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134



Tgt Ion:252 Resp: 3583035
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 13.4 6.2 9.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#91
Benzo(k)fluoranthene
Concen: 31.585 ng/ul
RT: 22.922 min Scan# 3372
Delta R.T. -0.006 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59


Tgt Ion:252 Resp: 3527015
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 13.0 6.2 9.4





#93
Benzo(a)pyrene
Concen: 27.955 ng/ul
RT: 23.492 min Scan# 3469
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59




Instrument :
BNA_P
ClientSampleId :
SLCS134




Tgt Ion:252 Resp: 319470
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 14.0 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 30.585 ng/ul
RT: 26.039 min Scan# 3902
Delta R.T. 0.000 min
Lab File: BP010952.D
Acq: 11 Jul 2022 16:59



Tgt Ion:276 Resp: 4217594
Ion Ratio Lower Upper
276 100
138 28.9 14.6 22.0
277 24.5 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 29.667 ng/ul

RT: 26.063 min Scan# 3905

Delta R.T. 0.000 min

Lab File: BP010952.D

Acq: 11 Jul 2022 16:59

Instrument :

BNA_P

ClientSampleId :

SLCS134

Tgt Ion:278 Resp: 350913

Ion Ratio Lower Upper

278 100





139 20.2 9.4 14.2

279 24.3 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/12/2022
Supervised By :mohammad ahmed 07/12/2022



#96

Benzo(g,h,i)perylene

Concen: 29.576 ng/ul

RT: 26.786 min Scan# 4029

Delta R.T. -0.006 min

Lab File: BP010952.D

Acq: 11 Jul 2022 16:59

Tgt Ion:276 Resp: 3469376

Ion Ratio Lower Upper

276 100

138 27.5 12.7 19.1

277 23.8 18.2 27.4

