

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010954.D
 Acq On : 11 Jul 2022 18:08
 Operator : CG/JU
 Sample : PB146131BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS131

Quant Time: Jul 12 01:05:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	470257	20.000	ng/u1	0.00
20) Naphthalene-d8	10.493	136	1865128	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.351	164	891479	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	1509545	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.233	240	1440401	20.000	ng/u1	# 0.00
88) Perylene-d12	23.586	264	1706101	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	85207	6.854	ng/uL	0.00
4) Pyridine-d5	3.605	84	1033388	31.228	ng/u1	0.00
7) Phenol-d5	6.875	99	1262669	33.919	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	818316	33.613	ng/u1	0.00
11) 2-Chlorophenol-d4	7.234	132	1046479	33.883	ng/u1	0.00
15) 4-Methylphenol-d8	8.416	113	960973	32.656	ng/u1	0.00
21) Nitrobenzene-d5	8.863	128	465315	39.190	ng/u1	0.00
24) 2-Nitrophenol-d4	9.581	143	438589	44.649	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	810944	33.411	ng/u1	0.00
31) 4-Chloroaniline-d4	10.640	131	843916	21.300	ng/u1	0.00
46) Dimethylphthalate-d6	13.775	166	1891853	30.980	ng/u1	0.00
49) Acenaphthylene-d8	14.045	160	2534511	32.519	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	346886	36.368	ng/u1	0.00
60) Fluorene-d10	15.351	176	1642493	31.342	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	220486	44.086	ng/u1	0.00
73) Anthracene-d10	17.216	188	2255883	33.589	ng/u1	0.00
81) Pyrene-d10	19.469	212	2413140	30.751	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.439	264	2889021	34.267	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	173325	13.074	ng/uL#	88
5) Pyridine	3.622	79	1017981	30.368	ng/u1#	77
6) Benzaldehyde	6.846	77	802374	40.851	ng/u1	97
8) Phenol	6.905	94	1303678	32.744	ng/u1#	86
10) Bis(2-Chloroethyl)ether	7.134	93	1051884	32.895	ng/u1	90
12) 2-Chlorophenol	7.263	128	1049934	32.933	ng/u1	94
13) 2-Methylphenol	8.152	108	922266	31.115	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.240	45	1423366	33.176	ng/u1#	97
16) Acetophenone	8.528	105	1437567	31.809	ng/u1#	89
17) N-Nitroso-di-n-propyla...	8.522	70	689425	30.169	ng/u1	90
18) 4-Methylphenol	8.481	108	983143	31.850	ng/u1	100
19) Hexachloroethane	8.775	117	419179	32.467	ng/u1#	81
22) Nitrobenzene	8.904	77	1090389	36.800	ng/u1	91
23) Isophorone	9.434	82	1815514	29.218	ng/u1	98
25) 2-Nitrophenol	9.610	139	486266	40.878	ng/u1#	82
26) 2,4-Dimethylphenol	9.681	107	1012202	31.144	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.922	93	1225740	31.270	ng/u1	98
29) 2,4-Dichlorophenol	10.146	162	796848	31.595	ng/u1	100
30) Naphthalene	10.546	128	3073395	31.616	ng/u1	100
32) 4-Chloroaniline	10.663	127	1151822	29.343	ng/u1	99
33) Hexachlorobutadiene	10.828	225	478790	29.715	ng/u1	95
34) Caprolactam	11.451	113	210235	26.535	ng/u1	89
35) 4-Chloro-3-methylphenol	11.798	107	822119	30.076	ng/u1	94
36) 2-Methylnaphthalene	12.163	142	1856088	29.486	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071122\
 Data File : BP010954.D
 Acq On : 11 Jul 2022 18:08
 Operator : CG/JU
 Sample : PB146131BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS131

Quant Time: Jul 12 01:05:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.381	142	1844916	29.502	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.534	216	837907	32.708	ng/ul	99
40) Hexachlorocyclopentadiene	12.510	237	517666	31.889	ng/ul	97
41) 2,4,6-Trichlorophenol	12.781	196	509806	34.517	ng/ul	98
42) 2,4,5-Trichlorophenol	12.851	196	552119	34.835	ng/ul	99
43) 1,1'-Biphenyl	13.187	154	2269577	32.339	ng/ul#	98
44) 2-Chloronaphthalene	13.222	162	1748098	33.066	ng/ul	98
45) 2-Nitroaniline	13.440	65	465208	41.474	ng/ul#	80
47) Dimethylphthalate	13.822	163	1869710	30.573	ng/ul	99
48) 2,6-Dinitrotoluene	13.940	165	351649	38.377	ng/ul	98
50) Acenaphthylene	14.075	152	2665614	31.229	ng/ul	98
51) 3-Nitroaniline	14.269	138	409512	37.278	ng/ul	90
52) Acenaphthene	14.416	153	1796780	32.624	ng/ul	98
53) 2,4-Dinitrophenol	14.475	184	143860	40.895	ng/ul#	82
55) 4-Nitrophenol	14.587	109	267636	35.288	ng/ul#	78
56) Dibenzofuran	14.757	168	2357692	31.252	ng/ul	98
57) 2,4-Dinitrotoluene	14.728	165	492354	40.212	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	14.986	232	393442	34.664	ng/ul#	90
59) Diethylphthalate	15.198	149	1768790	29.092	ng/ul	99
61) Fluorene	15.410	166	1853007	30.804	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.410	204	849359	30.162	ng/ul	97
63) 4-Nitroaniline	15.439	138	430091	42.345	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.492	198	226226	41.860	ng/ul	97
67) N-Nitrosodiphenylamine	15.628	169	1521839	33.285	ng/ul	98
68) 4-Bromophenyl-phenylether	16.310	248	483215	31.941	ng/ul	95
69) Hexachlorobenzene	16.416	284	542075	30.935	ng/ul	96
70) Atrazine	16.592	200	431995	28.498	ng/ul	97
71) Pentachlorophenol	16.763	266	278255	30.839	ng/ul	99
72) Phenanthrene	17.163	178	2678264	33.317	ng/ul	99
74) Anthracene	17.251	178	2692961	33.308	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.145	216	844461	33.900	ng/uL	96
76) Pentachlorobenzene	14.669	250	718748	33.643	ng/uL	98
77) Carbazole	17.533	167	2478718	34.059	ng/ul	99
78) Di-n-butylphthalate	18.098	149	2525688	29.123	ng/ul	99
80) Fluoranthene	19.145	202	2838445	29.410	ng/ul#	89
82) Pyrene	19.498	202	2972380	30.496	ng/ul#	84
83) Butylbenzylphthalate	20.392	149	1092855	31.855	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.157	252	963499	32.394	ng/ul#	97
85) Benzo(a)anthracene	21.216	228	3022409	33.551	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.157	149	1592974	29.321	ng/ul#	96
87) Chrysene	21.269	228	2860369	33.135	ng/ul	99
89) Di-n-octyl phthalate	22.074	149	2843241	28.076	ng/ul	100
90) Benzo(b)fluoranthene	22.868	252	3399723	32.043	ng/ul#	96
91) Benzo(k)fluoranthene	22.921	252	3439399	34.105	ng/ul#	95
93) Benzo(a)pyrene	23.486	252	3154986	30.569	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.033	276	4451578	35.744	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.057	278	3693036	34.571	ng/ul#	93
96) Benzo(g,h,i)perylene	26.780	276	3664646	34.592	ng/ul#	87





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.074 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion: 88 Resp: 173325
Ion Ratio Lower Upper
88 100
43 29.5 35.8 53.6#
58 79.6 68.1 102.1



#5
Pyridine
Concen: 30.368 ng/uL
RT: 3.622 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 79 Resp: 1017981
Ion Ratio Lower Upper
79 100
52 80.2 83.8 125.8#
51 36.6 41.0 61.4#









#6
Benzaldehyde
Concen: 40.851 ng/ul
RT: 6.846 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion: 77 Resp: 802374
Ion Ratio Lower Upper
77 100
105 98.2 76.0 114.0
106 93.4 73.2 109.8



#8
Phenol
Concen: 32.744 ng/ul
RT: 6.905 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 94 Resp: 1303678
Ion Ratio Lower Upper
94 100
65 27.7 27.4 41.2
66 35.8 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 32.895 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Instrument :
BNA_P
ClientSampleId :
SLCS131

Tgt Ion: 93 Resp: 1051884




Ion	Ratio	Lower	Upper
93	100		
63	78.1	72.4	108.6
95	32.2	27.3	40.9




#12
2-Chlorophenol
Concen: 32.933 ng/ul
RT: 7.263 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 128 Resp: 1049934

Ion	Ratio	Lower	Upper
128	100		
64	49.1	44.8	67.2
130	32.1	25.8	38.6









#13
2-Methylphenol
Concen: 31.115 ng/ul
RT: 8.152 min Scan# 861
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:108 Resp: 922266
Ion Ratio Lower Upper
108 100
107 90.3 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.176 ng/ul
RT: 8.240 min Scan# 876
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 45 Resp: 1423366
Ion Ratio Lower Upper
45 100
77 13.6 11.2 16.8
79 10.8 6.9 10.3#









#16
Acetophenone
Concen: 31.809 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:105 Resp: 1437567
Ion Ratio Lower Upper
105 100
77 76.9 67.5 101.3
51 27.3 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 30.169 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 70 Resp: 689425
Ion Ratio Lower Upper
70 100
42 57.0 53.8 80.8
101 9.6 7.8 11.6
130 21.6 15.4 23.2









#18
4-Methylphenol
Concen: 31.850 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:108 Resp: 983143
Ion Ratio Lower Upper
108 100
107 110.9 88.3 132.5



#19
Hexachloroethane
Concen: 32.467 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:117 Resp: 419179
Ion Ratio Lower Upper
117 100
201 89.1 85.7 128.5
199 54.8 57.4 86.0#









#22
Nitrobenzene
Concen: 36.800 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion: 77 Resp: 1090389
Ion Ratio Lower Upper
77 100
123 46.0 31.5 47.3
65 13.1 11.4 17.2



#23
Isophorone
Concen: 29.218 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 82 Resp: 1815514
Ion Ratio Lower Upper
82 100
95 6.7 6.6 10.0
138 17.5 13.8 20.6









#25
2-Nitrophenol
Concen: 40.878 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:139 Resp: 486266
Ion Ratio Lower Upper
139 100
65 51.9 54.2 81.4#
109 28.6 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 31.144 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:107 Resp: 1012202
Ion Ratio Lower Upper
107 100
121 51.9 39.1 58.7
122 86.6 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 31.270 ng/ul
 RT: 9.922 min Scan# 1161
 Delta R.T. 0.000 min
 Lab File: BP010954.D
 Acq: 11 Jul 2022 18:08




Instrument :
 BNA_P
 ClientSampleId :
 SLCS131


Tgt Ion: 93 Resp: 1225740
 Ion Ratio Lower Upper
 93 100
 95 32.4 26.6 39.8
 123 11.6 7.8 11.8



#29
 2,4-Dichlorophenol
 Concen: 31.595 ng/ul
 RT: 10.146 min Scan# 1200
 Delta R.T. 0.000 min
 Lab File: BP010954.D
 Acq: 11 Jul 2022 18:08

Tgt Ion: 162 Resp: 796848
 Ion Ratio Lower Upper
 162 100
 164 65.5 52.5 78.7
 98 38.8 31.4 47.0









#30
Naphthalene
Concen: 31.616 ng/ul
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:128 Resp: 3073395
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 29.343 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:127 Resp: 1151822
Ion Ratio Lower Upper
127 100
129 33.4 26.2 39.2









#33
Hexachlorobutadiene
Concen: 29.715 ng/ul
RT: 10.828 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:225 Resp: 478790
Ion Ratio Lower Upper
225 100
223 62.3 53.7 80.5
227 64.6 49.4 74.2



#34
Caprolactam
Concen: 26.535 ng/ul
RT: 11.451 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:113 Resp: 210235
Ion Ratio Lower Upper
113 100
55 173.1 153.8 230.8
56 128.8 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 30.076 ng/ul
RT: 11.798 min Scan# 1481
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:107 Resp: 822119
Ion Ratio Lower Upper
107 100
144 25.7 18.6 27.8
142 79.8 59.6 89.4



#36
2-Methylnaphthalene
Concen: 29.486 ng/ul
RT: 12.163 min Scan# 1543
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:142 Resp: 1856088
Ion Ratio Lower Upper
142 100
141 87.5 70.1 105.1





#37

1-Methylnaphthalene

Concen: 29.502 ng/ul

RT: 12.381 min Scan# 1581

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:142 Resp: 1844916

Ion Ratio Lower Upper

142 100

141 92.0 73.3 109.9

116 3.9 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 32.708 ng/ul

RT: 12.534 min Scan# 1606

Delta R.T. 0.000 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Tgt Ion:216 Resp: 837907




Ion Ratio Lower Upper


216 100

214 78.5 63.0 94.6

179 22.3 18.1 27.1

108 20.1 13.6 20.4





#40

Hexachlorocyclopentadiene

Concen: 31.889 ng/ul

RT: 12.510 min Scan# 1602

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:237 Resp: 517666

Ion Ratio Lower Upper

237 100

235 63.4 52.8 79.2

272 13.8 11.5 17.3



#41

2,4,6-Trichlorophenol

Concen: 34.517 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:196 Resp: 509806


Ion Ratio Lower Upper

196 100

198 96.3 78.2 117.4

200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 34.835 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:196 Resp: 552119
Ion Ratio Lower Upper
196 100
198 95.3 77.0 115.4
200 29.7 24.8 37.2



#43
1,1'-Biphenyl
Concen: 32.339 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:154 Resp: 2269577
Ion Ratio Lower Upper
154 100
153 39.8 31.9 47.9
76 14.2 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 33.066 ng/ul
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:162 Resp: 1748098
Ion Ratio Lower Upper
162 100
164 32.7 25.5 38.3
127 41.3 31.9 47.9



#45
2-Nitroaniline
Concen: 41.474 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion: 65 Resp: 465208
Ion Ratio Lower Upper
65 100
92 70.1 49.0 73.6
138 106.5 66.4 99.6#









#47
Dimethylphthalate
Concen: 30.573 ng/ul
RT: 13.822 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:163 Resp: 1869710
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.0 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 38.377 ng/ul
RT: 13.940 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:165 Resp: 351649
Ion Ratio Lower Upper
165 100
89 63.8 52.9 79.3
121 23.9 19.4 29.2









#50
Acenaphthylene
Concen: 31.229 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:152 Resp: 2665614
Ion Ratio Lower Upper
152 100
151 19.8 16.5 24.7
153 13.2 10.1 15.1



#51
3-Nitroaniline
Concen: 37.278 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:138 Resp: 409512
Ion Ratio Lower Upper
138 100
108 12.6 11.0 16.4
92 118.5 104.6 156.8









#52
Acenaphthene
Concen: 32.624 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:153 Resp: 1796780
Ion Ratio Lower Upper
153 100
152 47.5 36.2 54.2
154 87.7 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 40.895 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:184 Resp: 143860
Ion Ratio Lower Upper
184 100
63 70.2 73.8 110.6#
154 62.8 57.2 85.8









#55
4-Nitrophenol
Concen: 35.288 ng/ul
RT: 14.587 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:109 Resp: 267636
Ion Ratio Lower Upper
109 100
139 143.0 75.6 113.4#
65 136.3 107.6 161.4



#56
Dibenzofuran
Concen: 31.252 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:168 Resp: 2357692
Ion Ratio Lower Upper
168 100
139 39.8 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 40.212 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:165 Resp: 492354
Ion Ratio Lower Upper
165 100
63 46.4 47.0 70.6#
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 34.664 ng/ul
RT: 14.986 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:232 Resp: 393442
Ion Ratio Lower Upper
232 100
131 52.3 35.8 53.8
130 2.8 1.8 2.8#
166 33.7 23.7 35.5









#59
Diethylphthalate
Concen: 29.092 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:149 Resp: 1768790
Ion Ratio Lower Upper
149 100
177 21.3 17.6 26.4
150 12.4 10.3 15.5



#61
Fluorene
Concen: 30.804 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:166 Resp: 1853007
Ion Ratio Lower Upper
166 100
165 97.9 77.4 116.2
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 30.162 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:204 Resp: 849359
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 65.3 49.3 73.9



#63
4-Nitroaniline
Concen: 42.345 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:138 Resp: 430091
Ion Ratio Lower Upper
138 100
92 52.3 46.2 69.2
108 73.6 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 41.860 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:198 Resp: 226226
Ion Ratio Lower Upper
198 100
182 5.5 4.3 6.5
77 27.7 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 33.285 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:169 Resp: 1521839
Ion Ratio Lower Upper
169 100
168 67.0 52.6 78.8
167 35.9 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 31.941 ng/ul
RT: 16.310 min Scan# 2248
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:248 Resp: 483215
Ion Ratio Lower Upper
248 100
250 96.6 80.8 121.2
141 84.5 63.9 95.9



#69
Hexachlorobenzene
Concen: 30.935 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:284 Resp: 542075
Ion Ratio Lower Upper
284 100
142 37.9 27.8 41.6
249 30.8 25.4 38.2





#70

Atrazine

Concen: 28.498 ng/ul

RT: 16.592 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:200 Resp: 431995

Ion Ratio Lower Upper

200 100

173 28.5 20.7 31.1

215 47.1 38.6 57.8



#71

Pentachlorophenol

Concen: 30.839 ng/ul

RT: 16.763 min Scan# 2325

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:266 Resp: 278255


Ion Ratio Lower Upper

266 100

264 64.3 51.4 77.2

268 64.6 53.5 80.3





#72

Phenanthrene

Concen: 33.317 ng/u1

RT: 17.163 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:178 Resp: 2678264

Ion Ratio Lower Upper

178 100

179 15.6 12.7 19.1

176 19.1 14.6 22.0



#74

Anthracene

Concen: 33.308 ng/u1

RT: 17.251 min Scan# 2408

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:178 Resp: 2692961


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.6 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 33.900 ng/uL

RT: 13.145 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:216 Resp: 844461

Ion Ratio Lower Upper

216 100

214 76.7 65.0 97.4

218 47.4 39.1 58.7



#76

Pentachlorobenzene

Concen: 33.643 ng/uL

RT: 14.669 min Scan# 1969

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:250 Resp: 718748


Ion Ratio Lower Upper

250 100

248 63.8 49.8 74.6

252 62.8 51.2 76.8





#77

Carbazole

Concen: 34.059 ng/ul

RT: 17.533 min Scan# 2456

Delta R.T. 0.000 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:167 Resp: 2478718

Ion Ratio Lower Upper

167 100

166 21.5 16.9 25.3

139 13.2 10.6 16.0



#78

Di-n-butylphthalate

Concen: 29.123 ng/ul

RT: 18.098 min Scan# 2552

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:149 Resp: 2525688


Ion Ratio Lower Upper

149 100

150 9.1 7.1 10.7

104 5.3 4.6 6.8









#80
Fluoranthene
Concen: 29.410 ng/ul
RT: 19.145 min Scan# 2730
Delta R.T. 0.000 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:202 Resp: 2838445
Ion Ratio Lower Upper
202 100
101 13.4 7.1 10.7#
100 10.3 5.7 8.5#



#82
Pyrene
Concen: 30.496 ng/ul
RT: 19.498 min Scan# 2790
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:202 Resp: 2972380
Ion Ratio Lower Upper
202 100
101 16.5 7.9 11.9#
100 13.0 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 31.855 ng/ul

RT: 20.392 min Scan# 2942

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:149 Resp: 1092855

Ion Ratio Lower Upper

149 100

91 72.9 60.0 90.0

206 18.9 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 32.394 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:252 Resp: 963499


Ion Ratio Lower Upper

252 100

254 65.2 51.0 76.4

126 14.1 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 33.551 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:228 Resp: 3022409

Ion Ratio Lower Upper

228 100

229 20.1 15.0 22.4

226 27.0 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 29.321 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:149 Resp: 1592974


Ion Ratio Lower Upper

149 100

167 26.2 22.6 33.8

279 3.4 4.2 6.4#





#87

Chrysene

Concen: 33.135 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:228 Resp: 2860369

Ion Ratio Lower Upper

228 100

226 29.8 23.6 35.4

229 20.1 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 28.076 ng/ul




RT: 22.074 min Scan# 3228


Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Tgt Ion:149 Resp: 2843241









#90
Benzo(b)fluoranthene
Concen: 32.043 ng/ul
RT: 22.868 min Scan# 3363
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SLCS131


Tgt Ion:252 Resp: 3399723
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 13.2 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 34.105 ng/ul
RT: 22.921 min Scan# 3372
Delta R.T. -0.006 min
Lab File: BP010954.D
Acq: 11 Jul 2022 18:08

Tgt Ion:252 Resp: 3439399
Ion Ratio Lower Upper
252 100
253 21.8 17.0 25.4
125 12.6 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 30.569 ng/ul

RT: 23.486 min Scan# 3468

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SLCS131





Tgt Ion:252 Resp: 3154986

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 14.3 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.744 ng/ul

RT: 26.033 min Scan# 3901

Delta R.T. -0.006 min

Lab File: BP010954.D

Acq: 11 Jul 2022 18:08




Tgt Ion:276 Resp: 4451578


Ion Ratio Lower Upper

276 100

138 29.5 14.6 22.0#

277 25.0 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 34.571 ng/ul
 RT: 26.057 min Scan# 3905
 Delta R.T. -0.006 min
 Lab File: BP010954.D
 Acq: 11 Jul 2022 18:08

Instrument :
 BNA_P
 ClientSampleId :
 SLCS131

Tgt Ion:278 Resp: 3693036
 Ion Ratio Lower Upper
 278 100
 139 20.2 9.4 14.2#
 279 23.6 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 34.592 ng/ul
 RT: 26.780 min Scan# 4028
 Delta R.T. -0.012 min
 Lab File: BP010954.D
 Acq: 11 Jul 2022 18:08

Tgt Ion:276 Resp: 3664646
 Ion Ratio Lower Upper
 276 100
 138 27.4 12.7 19.1#
 277 24.3 18.2 27.4

