

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071221\
 Data File : BP006192.D
 Acq On : 12 Jul 2021 21:21
 Operator : CG/JU
 Sample : M3002-04 2X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TP-H

Quant Time: Jul 13 04:17:30 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 12 14:59:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.864	152	220701	20.000	ng	0.00
21) Naphthalene-d8	10.663	136	875645	20.000	ng	0.00
39) Acenaphthene-d10	14.499	164	455312	20.000	ng	0.00
64) Phenanthrene-d10	17.245	188	851898	20.000	ng	# 0.00
76) Chrysene-d12	21.328	240	800899	20.000	ng	# 0.00
86) Perylene-d12	23.710	264	922649	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.440	112	631017	49.196	ng	0.00
7) Phenol-d6	7.052	99	748702	44.129	ng	0.00
23) Nitrobenzene-d5	9.034	82	401835	28.987	ng	0.00
42) 2,4,6-Tribromophenol	15.993	330	236560	43.640	ng	0.00
45) 2-Fluorobiphenyl	13.122	172	839117	27.245	ng	0.00
79) Terphenyl-d14	19.798	244	963958	23.730	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071221\
Data File : BP006192.D
Acq On : 12 Jul 2021 21:21
Operator : CG/JU
Sample : M3002-04 2X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TP-H

Quant Time: Jul 13 04:17:30 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 12 14:59:51 2021
Response via : Initial Calibration

