

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071322\
 Data File : BP010961.D
 Acq On : 13 Jul 2022 11:15
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Quant Time: Jul 14 02:44:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	492864	20.000	ng/ul	0.00
20) Naphthalene-d8	10.498	136	2203607	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.363	164	1294747	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.128	188	2697442	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.245	240	2706590	20.000	ng/ul	# 0.00
88) Perylene-d12	23.604	264	2854292	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	94699	7.268	ng/uL	0.00
4) Pyridine-d5	3.605	84	636831	18.362	ng/ul	0.00
7) Phenol-d5	6.881	99	737377	18.900	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	478902	18.769	ng/ul	0.00
11) 2-Chlorophenol-d4	7.240	132	599135	18.509	ng/ul	0.00
15) 4-Methylphenol-d8	8.422	113	604002	19.584	ng/ul	0.00
21) Nitrobenzene-d5	8.869	128	282975	20.172	ng/ul	0.00
24) 2-Nitrophenol-d4	9.587	143	264831	22.819	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	511031	17.821	ng/ul	0.00
31) 4-Chloroaniline-d4	10.645	131	861336	18.401	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	1563760	17.632	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	1891485	16.710	ng/ul	0.00
54) 4-Nitrophenol-d4	14.575	143	297922	21.506	ng/ul	0.00
60) Fluorene-d10	15.357	176	1353816	17.787	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	203642	22.787	ng/ul	0.00
73) Anthracene-d10	17.227	188	2055376	17.126	ng/ul	0.00
81) Pyrene-d10	19.474	212	2414890	16.377	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.451	264	2410556	17.090	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	98171	7.065	ng/uL#	92
5) Pyridine	3.628	79	663758	18.893	ng/ul#	76
6) Benzaldehyde	6.852	77	445465	21.639	ng/ul	96
8) Phenol	6.904	94	802782	19.238	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.140	93	624733	18.641	ng/ul	89
12) 2-Chlorophenol	7.269	128	626704	18.756	ng/ul	92
13) 2-Methylphenol	8.152	108	581301	18.712	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.246	45	862082	19.172	ng/ul#	97
16) Acetophenone	8.534	105	933833	19.715	ng/ul#	89
17) N-Nitroso-di-n-propyla...	8.522	70	468083	19.544	ng/ul	90
18) 4-Methylphenol	8.487	108	646639	19.988	ng/ul	97
19) Hexachloroethane	8.775	117	233121	17.228	ng/ul#	79
22) Nitrobenzene	8.910	77	683248	19.517	ng/ul	92
23) Isophorone	9.440	82	1252970	17.067	ng/ul	98
25) 2-Nitrophenol	9.616	139	308725	21.966	ng/ul#	82
26) 2,4-Dimethylphenol	9.687	107	681686	17.753	ng/ul	96
27) Bis(2-Chloroethoxy)met...	9.922	93	816015	17.620	ng/ul	97
29) 2,4-Dichlorophenol	10.151	162	532904	17.884	ng/ul	99
30) Naphthalene	10.551	128	1988637	17.315	ng/ul	99
32) 4-Chloroaniline	10.669	127	858788	18.517	ng/ul	98
33) Hexachlorobutadiene	10.834	225	294144	15.451	ng/ul	95
34) Caprolactam	11.457	113	185206	19.785	ng/ul	87
35) 4-Chloro-3-methylphenol	11.804	107	617583	19.123	ng/ul	93
36) 2-Methylnaphthalene	12.169	142	1301561	17.501	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071322\
 Data File : BP010961.D
 Acq On : 13 Jul 2022 11:15
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Quant Time: Jul 14 02:44:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.392	142	1312071	17.758	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.540	216	587062	15.779	ng/ul	98
40) Hexachlorocyclopentadiene	12.516	237	314838	13.354	ng/ul	96
41) 2,4,6-Trichlorophenol	12.787	196	384858	17.941	ng/ul	98
42) 2,4,5-Trichlorophenol	12.857	196	419227	18.212	ng/ul	100
43) 1,1'-Biphenyl	13.192	154	1684836	16.530	ng/ul#	98
44) 2-Chloronaphthalene	13.228	162	1268900	16.526	ng/ul	97
45) 2-Nitroaniline	13.445	65	369930	22.708	ng/ul#	79
47) Dimethylphthalate	13.828	163	1576734	17.752	ng/ul	98
48) 2,6-Dinitrotoluene	13.945	165	287537	21.606	ng/ul	97
50) Acenaphthylene	14.081	152	2119381	17.096	ng/ul	98
51) 3-Nitroaniline	14.275	138	338164	21.195	ng/ul	89
52) Acenaphthene	14.428	153	1387748	17.349	ng/ul	98
53) 2,4-Dinitrophenol	14.481	184	126755	24.810	ng/ul#	83
55) 4-Nitrophenol	14.586	109	233673	21.214	ng/ul#	75
56) Dibenzofuran	14.763	168	1928350	17.600	ng/ul	97
57) 2,4-Dinitrotoluene	14.733	165	435472	24.489	ng/ul	86
58) 2,3,4,6-Tetrachlorophenol	14.992	232	328585	19.933	ng/ul#	89
59) Diethylphthalate	15.204	149	1587545	17.978	ng/ul	98
61) Fluorene	15.416	166	1553303	17.779	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.416	204	699560	17.105	ng/ul	96
63) 4-Nitroaniline	15.445	138	335160	22.720	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.498	198	214577	22.219	ng/ul	96
67) N-Nitrosodiphenylamine	15.633	169	1334090	16.329	ng/ul	98
68) 4-Bromophenyl-phenylether	16.316	248	409352	15.142	ng/ul	96
69) Hexachlorobenzene	16.422	284	482856	15.421	ng/ul	96
70) Atrazine	16.604	200	449067	16.578	ng/ul	98
71) Pentachlorophenol	16.769	266	270049	16.749	ng/ul	96
72) Phenanthrene	17.169	178	2498997	17.397	ng/ul	99
74) Anthracene	17.263	178	2501667	17.316	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.151	216	617759	13.878	ng/ul	97
76) Pentachlorobenzene	14.675	250	571658	14.974	ng/ul	99
77) Carbazole	17.539	167	2374434	18.258	ng/ul	99
78) Di-n-butylphthalate	18.104	149	2648497	17.090	ng/ul	99
80) Fluoranthene	19.151	202	2925171	16.130	ng/ul#	87
82) Pyrene	19.504	202	3024091	16.512	ng/ul#	83
83) Butylbenzylphthalate	20.398	149	1227509	19.041	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.168	252	944941	16.907	ng/ul#	98
85) Benzo(a)anthracene	21.227	228	2912265	17.204	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.168	149	1819266	17.821	ng/ul#	97
87) Chrysene	21.280	228	2853801	17.593	ng/ul	100
89) Di-n-octyl phthalate	22.080	149	3131563	18.484	ng/ul	100
90) Benzo(b)fluoranthene	22.880	252	3067346	17.281	ng/ul#	95
91) Benzo(k)fluoranthene	22.927	252	2979128	17.657	ng/ul#	94
93) Benzo(a)pyrene	23.498	252	2989295	17.312	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.051	276	3171389	15.221	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.074	278	2746038	15.365	ng/ul#	92
96) Benzo(g,h,i)perylene	26.803	276	2641029	14.901	ng/ul#	89





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.065 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 88 Resp: 98171
Ion Ratio Lower Upper
88 100
43 30.7 35.8 53.6#
58 84.3 68.1 102.1



#5
Pyridine
Concen: 18.893 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 79 Resp: 663758
Ion Ratio Lower Upper
79 100
52 78.5 83.8 125.8#
51 36.3 41.0 61.4#









#6
Benzaldehyde
Concen: 21.639 ng/ul
RT: 6.852 min Scan# 640
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 77 Resp: 445465
Ion Ratio Lower Upper
77 100
105 99.1 76.0 114.0
106 94.7 73.2 109.8



#8
Phenol
Concen: 19.238 ng/ul
RT: 6.904 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 94 Resp: 802782
Ion Ratio Lower Upper
94 100
65 27.7 27.4 41.2
66 35.6 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 18.641 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 93 Resp: 624733
Ion Ratio Lower Upper
93 100
63 78.3 72.4 108.6
95 31.8 27.3 40.9



#12
2-Chlorophenol
Concen: 18.756 ng/ul
RT: 7.269 min Scan# 711
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 128 Resp: 626704
Ion Ratio Lower Upper
128 100
64 47.0 44.8 67.2
130 31.8 25.8 38.6









#13
2-Methylphenol
Concen: 18.712 ng/ul
RT: 8.152 min Scan# 861
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:108 Resp: 581301
Ion Ratio Lower Upper
108 100
107 90.7 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.172 ng/ul
RT: 8.246 min Scan# 877
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 45 Resp: 862082
Ion Ratio Lower Upper
45 100
77 13.2 11.2 16.8
79 10.5 6.9 10.3#









#16
Acetophenone
Concen: 19.715 ng/ul
RT: 8.534 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:105 Resp: 933833
Ion Ratio Lower Upper
105 100
77 76.4 67.5 101.3
51 27.2 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 19.544 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 70 Resp: 468083
Ion Ratio Lower Upper
70 100
42 57.4 53.8 80.8
101 9.6 7.8 11.6
130 22.2 15.4 23.2









#18
4-Methylphenol
Concen: 19.988 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:108 Resp: 646639
Ion Ratio Lower Upper
108 100
107 107.1 88.3 132.5



#19
Hexachloroethane
Concen: 17.228 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:117 Resp: 233121
Ion Ratio Lower Upper
117 100
201 84.7 85.7 128.5#
199 55.0 57.4 86.0#









#22
Nitrobenzene
Concen: 19.517 ng/ul
RT: 8.910 min Scan# 990
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 77 Resp: 683248
Ion Ratio Lower Upper
77 100
123 46.0 31.5 47.3
65 13.5 11.4 17.2



#23
Isophorone
Concen: 17.067 ng/ul
RT: 9.440 min Scan# 1080
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 82 Resp: 1252970
Ion Ratio Lower Upper
82 100
95 6.7 6.6 10.0
138 17.5 13.8 20.6









#25
2-Nitrophenol
Concen: 21.966 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:139 Resp: 308725
Ion Ratio Lower Upper
139 100
65 52.5 54.2 81.4#
109 29.2 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 17.753 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:107 Resp: 681686
Ion Ratio Lower Upper
107 100
121 51.2 39.1 58.7
122 86.5 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.620 ng/ul
 RT: 9.922 min Scan# 1162
 Delta R.T. -0.000 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698


Tgt Ion:	93	Resp:	816015
Ion	Ratio	Lower	Upper
93	100		
95	32.1	26.6	39.8
123	11.6	7.8	11.8



#29
 2,4-Dichlorophenol
 Concen: 17.884 ng/ul
 RT: 10.151 min Scan# 1201
 Delta R.T. 0.006 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Tgt Ion:	162	Resp:	532904
Ion	Ratio	Lower	Upper
162	100		
164	65.3	52.5	78.7
98	37.9	31.4	47.0









#30
Naphthalene
Concen: 17.315 ng/ul
RT: 10.551 min Scan# 1269
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:128 Resp: 1988637
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 10.2 15.4



#32
4-Chloroaniline
Concen: 18.517 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:127 Resp: 858788
Ion Ratio Lower Upper
127 100
129 31.7 26.2 39.2









#33
Hexachlorobutadiene
Concen: 15.451 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:225 Resp: 294144
Ion Ratio Lower Upper
225 100
223 63.8 53.7 80.5
227 65.8 49.4 74.2



#34
Caprolactam
Concen: 19.785 ng/ul
RT: 11.457 min Scan# 1423
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:113 Resp: 185206
Ion Ratio Lower Upper
113 100
55 172.3 153.8 230.8
56 123.9 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 19.123 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:107 Resp: 617583
Ion Ratio Lower Upper
107 100
144 26.4 18.6 27.8
142 80.9 59.6 89.4



#36
2-Methylnaphthalene
Concen: 17.501 ng/ul
RT: 12.169 min Scan# 1544
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:142 Resp: 1301561
Ion Ratio Lower Upper
142 100
141 88.9 70.1 105.1









#37
1-Methylnaphthalene
Concen: 17.758 ng/ul
RT: 12.392 min Scan# 1582
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:142 Resp: 1312071
Ion Ratio Lower Upper
142 100
141 91.7 73.3 109.9
116 3.6 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 15.779 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:216 Resp: 587062
Ion Ratio Lower Upper
216 100
214 76.8 63.0 94.6
179 22.3 18.1 27.1
108 19.4 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 13.354 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:237 Resp: 314838
Ion Ratio Lower Upper
237 100
235 63.2 52.8 79.2
272 12.1 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 17.941 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:196 Resp: 384858
Ion Ratio Lower Upper
196 100
198 95.7 78.2 117.4
200 29.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 18.212 ng/ul
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:196 Resp: 419227
Ion Ratio Lower Upper
196 100
198 96.1 77.0 115.4
200 30.4 24.8 37.2



#43
1,1'-Biphenyl
Concen: 16.530 ng/ul
RT: 13.192 min Scan# 1718
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:154 Resp: 1684836
Ion Ratio Lower Upper
154 100
153 40.2 31.9 47.9
76 14.3 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 16.526 ng/ul
RT: 13.228 min Scan# 1724
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:162 Resp: 1268900
Ion Ratio Lower Upper
162 100
164 32.9 25.5 38.3
127 41.9 31.9 47.9



#45
2-Nitroaniline
Concen: 22.708 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion: 65 Resp: 369930
Ion Ratio Lower Upper
65 100
92 71.3 49.0 73.6
138 107.3 66.4 99.6#









#47
Dimethylphthalate
Concen: 17.752 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:163 Resp: 1576734
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.1 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 21.606 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:165 Resp: 287537
Ion Ratio Lower Upper
165 100
89 63.4 52.9 79.3
121 25.2 19.4 29.2









#50
Acenaphthylene
Concen: 17.096 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:152 Resp: 2119381
Ion Ratio Lower Upper
152 100
151 19.8 16.5 24.7
153 13.5 10.1 15.1



#51
3-Nitroaniline
Concen: 21.195 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:138 Resp: 338164
Ion Ratio Lower Upper
138 100
108 12.2 11.0 16.4
92 117.5 104.6 156.8





#52

Acenaphthene

Concen: 17.349 ng/ul

RT: 14.428 min Scan# 1928

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:153 Resp: 1387748

Ion Ratio Lower Upper

153 100

152 47.7 36.2 54.2

154 87.7 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 24.810 ng/ul

RT: 14.481 min Scan# 1937

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15




Tgt Ion:184 Resp: 126755


Ion Ratio Lower Upper

184 100

63 69.9 73.8 110.6#

154 63.9 57.2 85.8









#55
4-Nitrophenol
Concen: 21.214 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:109 Resp: 233673
Ion Ratio Lower Upper
109 100
139 143.8 75.6 113.4#
65 143.5 107.6 161.4



#56
Dibenzofuran
Concen: 17.600 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:168 Resp: 1928350
Ion Ratio Lower Upper
168 100
139 38.8 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 24.489 ng/ul
RT: 14.733 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:165 Resp: 435472
Ion Ratio Lower Upper
165 100
63 47.7 47.0 70.6
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.933 ng/ul
RT: 14.992 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:232 Resp: 328585
Ion Ratio Lower Upper
232 100
131 52.6 35.8 53.8
130 2.9 1.8 2.8#
166 34.7 23.7 35.5





#59

Diethylphthalate

Concen: 17.978 ng/ul

RT: 15.204 min Scan# 2060

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:149 Resp: 1587545

Ion Ratio Lower Upper

149 100

177 21.2 17.6 26.4

150 12.3 10.3 15.5



#61

Fluorene

Concen: 17.779 ng/ul

RT: 15.416 min Scan# 2096

Delta R.T. -0.000 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15




Tgt Ion:166 Resp: 1553303


Ion Ratio Lower Upper

166 100

165 99.7 77.4 116.2

167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 17.105 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:204 Resp: 699560
Ion Ratio Lower Upper
204 100
206 33.3 26.0 39.0
141 66.2 49.3 73.9



#63
4-Nitroaniline
Concen: 22.720 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:138 Resp: 335160
Ion Ratio Lower Upper
138 100
92 53.1 46.2 69.2
108 74.5 85.8 128.8#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 22.219 ng/ul
 RT: 15.498 min Scan# 2110
 Delta R.T. 0.006 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698


Tgt Ion:198 Resp: 214577
 Ion Ratio Lower Upper
 198 100
 182 5.2 4.3 6.5
 77 27.0 23.7 35.5



#67
 N-Nitrosodiphenylamine
 Concen: 16.329 ng/ul
 RT: 15.633 min Scan# 2133
 Delta R.T. -0.000 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Tgt Ion:169 Resp: 1334090
 Ion Ratio Lower Upper
 169 100
 168 67.0 52.6 78.8
 167 35.9 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 15.142 ng/ul
RT: 16.316 min Scan# 2249
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:248 Resp: 409352
Ion Ratio Lower Upper
248 100
250 100.3 80.8 121.2
141 86.4 63.9 95.9



#69
Hexachlorobenzene
Concen: 15.421 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:284 Resp: 482856
Ion Ratio Lower Upper
284 100
142 39.1 27.8 41.6
249 31.9 25.4 38.2









#70
Atrazine
Concen: 16.578 ng/ul
RT: 16.604 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:200 Resp: 449067
Ion Ratio Lower Upper
200 100
173 28.0 20.7 31.1
215 48.4 38.6 57.8



#71
Pentachlorophenol
Concen: 16.749 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:266 Resp: 270049
Ion Ratio Lower Upper
266 100
264 62.0 51.4 77.2
268 62.3 53.5 80.3









#72
Phenanthrene
Concen: 17.397 ng/ul
RT: 17.169 min Scan# 2394
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:178 Resp: 2498997
Ion Ratio Lower Upper
178 100
179 15.1 12.7 19.1
176 18.8 14.6 22.0



#74
Anthracene
Concen: 17.316 ng/ul
RT: 17.263 min Scan# 2410
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:178 Resp: 2501667
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 13.878 ng/uL
RT: 13.151 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:216 Resp: 617759
Ion Ratio Lower Upper
216 100
214 78.2 65.0 97.4
218 47.8 39.1 58.7



#76
Pentachlorobenzene
Concen: 14.974 ng/uL
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:250 Resp: 571658
Ion Ratio Lower Upper
250 100
248 63.4 49.8 74.6
252 63.2 51.2 76.8









#77
 Carbazole
 Concen: 18.258 ng/ul
 RT: 17.539 min Scan# 2457
 Delta R.T. 0.006 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Tgt Ion:167 Resp: 2374434




Ion	Ratio	Lower	Upper
167	100		
166	21.8	16.9	25.3
139	13.5	10.6	16.0




#78
 Di-n-butylphthalate
 Concen: 17.090 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. -0.000 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Tgt Ion:149 Resp: 2648497

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.1	10.7
104	5.2	4.6	6.8





#80

Fluoranthene

Concen: 16.130 ng/ul

RT: 19.151 min Scan# 2731

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:202 Resp: 2925171

Ion Ratio Lower Upper

202 100

101 14.2 7.1 10.7#

100 10.9 5.7 8.5#



#82

Pyrene

Concen: 16.512 ng/ul

RT: 19.504 min Scan# 2791

Delta R.T. -0.000 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15




Tgt Ion:202 Resp: 3024091


Ion Ratio Lower Upper

202 100

101 16.7 7.9 11.9#

100 13.6 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 19.041 ng/ul

RT: 20.398 min Scan# 2943

Delta R.T. -0.000 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:149 Resp: 1227509

Ion Ratio Lower Upper

149 100

91 76.1 60.0 90.0

206 18.5 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 16.907 ng/ul

RT: 21.168 min Scan# 3074

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15




Tgt Ion:252 Resp: 944941


Ion Ratio Lower Upper

252 100

254 65.0 51.0 76.4

126 13.2 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 17.204 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:228 Resp: 2912265
Ion Ratio Lower Upper
228 100
229 19.9 15.0 22.4
226 27.1 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 17.821 ng/ul
RT: 21.168 min Scan# 3074
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:149 Resp: 1819266
Ion Ratio Lower Upper
149 100
167 26.4 22.6 33.8
279 3.9 4.2 6.4









#87
Chrysene
Concen: 17.593 ng/ul
RT: 21.280 min Scan# 3093
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:228 Resp: 2853801
Ion Ratio Lower Upper
228 100
226 29.7 23.6 35.4
229 19.4 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.484 ng/ul
RT: 22.080 min Scan# 3229
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:149 Resp: 3131563









#90
Benzo(b)fluoranthene
Concen: 17.281 ng/ul
RT: 22.880 min Scan# 3365
Delta R.T. 0.006 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:252 Resp: 3067346
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 13.4 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 17.657 ng/ul
RT: 22.927 min Scan# 3373
Delta R.T. -0.000 min
Lab File: BP010961.D
Acq: 13 Jul 2022 11:15

Tgt Ion:252 Resp: 2979128
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 13.3 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 17.312 ng/ul

RT: 23.498 min Scan# 3470

Delta R.T. 0.006 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:252 Resp: 2989295

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 14.5 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 15.221 ng/ul

RT: 26.051 min Scan# 3904

Delta R.T. 0.012 min

Lab File: BP010961.D

Acq: 13 Jul 2022 11:15




Tgt Ion:276 Resp: 3171389


Ion Ratio Lower Upper

276 100

138 29.6 14.6 22.0#

277 24.9 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 15.365 ng/ul
 RT: 26.074 min Scan# 3908
 Delta R.T. 0.012 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Tgt Ion:278 Resp: 2746038
 Ion Ratio Lower Upper
 278 100
 139 20.4 9.4 14.2#
 279 23.5 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 14.901 ng/ul
 RT: 26.803 min Scan# 4032
 Delta R.T. 0.012 min
 Lab File: BP010961.D
 Acq: 13 Jul 2022 11:15

Tgt Ion:276 Resp: 2641029
 Ion Ratio Lower Upper
 276 100
 138 26.5 12.7 19.1#
 277 23.6 18.2 27.4

