

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071421\
 Data File : BP006227.D
 Acq On : 14 Jul 2021 20:09
 Operator : CG/JU
 Sample : M3026-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ABN-JG-1

Quant Time: Jul 15 03:11:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 12 14:59:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.851	152	285590	20.000	ng	-0.01
21) Naphthalene-d8	10.651	136	1153385	20.000	ng	-0.01
39) Acenaphthene-d10	14.486	164	613752	20.000	ng	-0.01
64) Phenanthrene-d10	17.239	188	1171444	20.000	ng	#-0.01
76) Chrysene-d12	21.315	240	1086688	20.000	ng	#-0.01
86) Perylene-d12	23.692	264	1118777	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	16471	0.992	ng	0.02
7) Phenol-d6	7.057	99	217077	9.888	ng	0.01
23) Nitrobenzene-d5	9.022	82	800313	43.829	ng	-0.01
42) 2,4,6-Tribromophenol	16.022	330	1649	0.226	ng	0.02
45) 2-Fluorobiphenyl	13.110	172	1681375	40.499	ng	-0.01
79) Terphenyl-d14	19.786	244	2198333	39.885	ng	-0.02
Target Compounds						
						Qvalue
75) Fluoranthene	19.245	202	154122	2.126	ng	95
78) Pyrene	19.598	202	147566	2.009	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071421\
Data File : BP006227.D
Acq On : 14 Jul 2021 20:09
Operator : CG/JU
Sample : M3026-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ABN-JG-1

Quant Time: Jul 15 03:11:21 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 12 14:59:51 2021
Response via : Initial Calibration

