

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071421\
 Data File : BP006215.D
 Acq On : 14 Jul 2021 13:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 13:35:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 12 14:59:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	42.024	-5.1	108	0.01
3	Pyridine	40.000	44.011	-10.0	108	0.01
4	n-Nitrosodimethylamine	40.000	43.859	-9.6	112	0.01
5 S	2-Fluorophenol	80.000	84.659	-5.8	106	0.00
6	Aniline	40.000	41.916	-4.8	106	0.00
7 S	Phenol-d6	80.000	84.122	-5.2	105	0.00
8	2-Chlorophenol	40.000	41.674	-4.2	104	0.01
9	Benzaldehyde	40.000	41.596	-4.0	104	0.00
10 C	Phenol	40.000	42.313	-5.8	106	0.00
11	bis(2-Chloroethyl)ether	40.000	42.252	-5.6	105	0.01
12	1,3-Dichlorobenzene	40.000	41.276	-3.2	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.880	-2.2	102	0.01
14	1,2-Dichlorobenzene	40.000	41.440	-3.6	103	0.00
15	Benzyl Alcohol	40.000	42.305	-5.8	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.793	-4.5	108	0.01
17	2-Methylphenol	40.000	42.416	-6.0	105	0.00
18	Hexachloroethane	40.000	41.618	-4.0	104	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	42.666	-6.7	108	0.00
20	3+4-Methylphenols	40.000	43.359	-8.4	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.01
22	Acetophenone	40.000	42.411	-6.0	103	0.01
23 S	Nitrobenzene-d5	80.000	86.152	-7.7	105	0.01
24	Nitrobenzene	40.000	43.007	-7.5	107	0.00
25	Isophorone	40.000	42.689	-6.7	104	0.00
26 C	2-Nitrophenol	40.000	44.354	-10.9	103	0.00
27	2,4-Dimethylphenol	40.000	43.030	-7.6	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.069	-7.7	105	0.00
29 C	2,4-Dichlorophenol	40.000	43.129	-7.8	102	0.00
30	1,2,4-Trichlorobenzene	40.000	41.766	-4.4	100	0.00
31	Naphthalene	40.000	41.633	-4.1	102	0.00
32	Benzoic acid	40.000	42.547	-6.4	111	0.00
33	4-Chloroaniline	40.000	43.175	-7.9	102	0.00
34 C	Hexachlorobutadiene	40.000	41.253	-3.1	97	0.00
35	Caprolactam	40.000	44.948	-12.4	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.768	-6.9	103	0.00
37	2-Methylnaphthalene	40.000	41.378	-3.4	101	0.01
38	1-Methylnaphthalene	40.000	41.258	-3.1	100	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	46.682	-16.7	97	0.01
41 P	Hexachlorocyclopentadiene	40.000	35.844	10.4	71	0.00
42 S	2,4,6-Tribromophenol	80.000	81.414	-1.8	81	0.01
43 C	2,4,6-Trichlorophenol	40.000	47.661	-19.2	99	0.00
44	2,4,5-Trichlorophenol	40.000	47.750	-19.4	99	0.00
45 S	2-Fluorobiphenyl	80.000	91.873	-14.8	99	0.00
46	1,1'-Biphenyl	40.000	44.502	-11.3	96	0.00
47	2-Chloronaphthalene	40.000	44.469	-11.2	96	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071421\
 Data File : BP006215.D
 Acq On : 14 Jul 2021 13:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 13:35:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 12 14:59:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.872	-17.2	100	0.00
49	Acenaphthylene	40.000	42.302	-5.8	91	0.00
50	Dimethylphthalate	40.000	41.589	-4.0	90	0.00
51	2,6-Dinitrotoluene	40.000	42.683	-6.7	90	0.00
52 C	Acenaphthene	40.000	41.847	-4.6	91	0.00
53	3-Nitroaniline	40.000	44.131	-10.3	91	0.00
54 P	2,4-Dinitrophenol	40.000	40.837	-2.1	94	0.00
55	Dibenzofuran	40.000	41.913	-4.8	90	0.00
56 P	4-Nitrophenol	40.000	38.120	4.7	86	0.00
57	2,4-Dinitrotoluene	40.000	44.170	-10.4	91	0.00
58	Fluorene	40.000	41.629	-4.1	90	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	42.055	-5.1	86	0.00
60	Diethylphthalate	40.000	42.130	-5.3	92	0.00
61	4-Chlorophenyl-phenylether	40.000	41.283	-3.2	87	0.00
62	4-Nitroaniline	40.000	41.528	-3.8	92	0.00
63	Azobenzene	40.000	43.532	-8.8	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.01
65	4,6-Dinitro-2-methylphenol	40.000	40.989	-2.5	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.905	-7.3	89	0.00
67	4-Bromophenyl-phenylether	40.000	41.523	-3.8	83	0.00
68	Hexachlorobenzene	40.000	40.855	-2.1	82	0.00
69	Atrazine	40.000	42.874	-7.2	88	0.00
70 C	Pentachlorophenol	40.000	42.074	-5.2	84	0.00
71	Phenanthrene	40.000	41.957	-4.9	88	0.00
72	Anthracene	40.000	42.574	-6.4	89	0.01
73	Carbazole	40.000	42.999	-7.5	90	0.00
74	Di-n-butylphthalate	40.000	43.582	-9.0	91	0.01
75 C	Fluoranthene	40.000	40.920	-2.3	85	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	81	0.01
77	Benzidine	40.000	43.645	-9.1	81	0.00
78	Pyrene	40.000	42.251	-5.6	85	0.01
79 S	Terphenyl-d14	80.000	84.946	-6.2	83	0.01
80	Butylbenzylphthalate	40.000	43.853	-9.6	89	0.01
81	Benzo(a)anthracene	40.000	42.076	-5.2	84	0.00
82	3,3'-Dichlorobenzidine	40.000	42.450	-6.1	82	0.01
83	Chrysene	40.000	41.757	-4.4	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.261	-10.7	91	0.00
85 c	Di-n-octyl phthalate	40.000	45.087	-12.7	94	0.01
86 I	Perylene-d12	20.000	20.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.427	-6.1	85	0.02
88	Benzo(b)fluoranthene	40.000	42.905	-7.3	84	0.01
89	Benzo(k)fluoranthene	40.000	41.773	-4.4	87	0.01
90 C	Benzo(a)pyrene	40.000	42.004	-5.0	85	0.01
91	Dibenzo(a,h)anthracene	40.000	42.606	-6.5	85	0.02
92	Benzo(g,h,i)perylene	40.000	41.888	-4.7	85	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071421\
Data File : BP006215.D
Acq On : 14 Jul 2021 13:05
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 13:35:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 12 14:59:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0