

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071521\
 Data File : BP006237.D
 Acq On : 15 Jul 2021 14:20
 Operator : CG/JU
 Sample : M3039-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SB-04-COMP

Quant Time: Jul 15 16:20:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 13:37:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	158739	20.000	ng	0.00
21) Naphthalene-d8	10.646	136	608673	20.000	ng	# 0.00
39) Acenaphthene-d10	14.481	164	328987	20.000	ng	0.00
64) Phenanthrene-d10	17.233	188	650242	20.000	ng	# 0.00
76) Chrysene-d12	21.310	240	678685	20.000	ng	# 0.00
86) Perylene-d12	23.686	264	783895	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	1180718	127.983	ng	0.00
7) Phenol-d6	7.034	99	1425989	116.856	ng	0.00
23) Nitrobenzene-d5	9.016	82	811441	84.207	ng	0.00
42) 2,4,6-Tribromophenol	15.981	330	506127	129.220	ng	0.00
45) 2-Fluorobiphenyl	13.104	172	1713231	76.986	ng	0.00
79) Terphenyl-d14	19.780	244	2546960	73.991	ng	0.00
Target Compounds						
60) Diethylphthalate	15.304	149	84571	3.486	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071521\
Data File : BP006237.D
Acq On : 15 Jul 2021 14:20
Operator : CG/JU
Sample : M3039-10
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SB-04-COMP

Quant Time: Jul 15 16:20:41 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 13:37:02 2021
Response via : Initial Calibration

