

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071521\  
 Data File : BP006244.D  
 Acq On : 15 Jul 2021 18:21  
 Operator : CG/JU  
 Sample : M3039-06  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SB-02-COMP

Quant Time: Jul 16 03:20:36 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP070721.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jul 15 13:37:02 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.840  | 152  | 163265   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.646 | 136  | 664505   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.481 | 164  | 399012   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.233 | 188  | 782075   | 20.000  | ng    | # 0.00   |
| 76) Chrysene-d12            | 21.310 | 240  | 772647   | 20.000  | ng    | # 0.00   |
| 86) Perylene-d12            | 23.686 | 264  | 842996   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.422  | 112  | 1090422  | 114.919 | ng    | 0.00     |
| 7) Phenol-d6                | 7.034  | 99   | 1366219  | 108.855 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.010  | 82   | 770856   | 73.274  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.975 | 330  | 529062   | 111.370 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.104 | 172  | 1727096  | 63.989  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.780 | 244  | 2444769  | 62.385  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071521\  
Data File : BP006244.D  
Acq On : 15 Jul 2021 18:21  
Operator : CG/JU  
Sample : M3039-06  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
SB-02-COMP

Quant Time: Jul 16 03:20:36 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP070721.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jul 15 13:37:02 2021  
Response via : Initial Calibration

