

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/16/2025	BNA_P	<u>BP025154.D</u>
Compound Name	Response	Retention Time
DDT	7110729	20.578
DDD	166967	20.107
DDE	9571	19.583
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
176538	7287267	2.42