


Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP071720\
 Data File : BP002784.D
 Acq On : 17 Jul 2020 09:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP071020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 13 10:54:03 2020



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	56683	PASS
68	69	0.00	2	1.7	928	PASS
69	198	0.00	100	36.7	54510	PASS
70	69	0.00	2	0.5	292	PASS
127	198	10	80	45.8	67974	PASS
197	198	0.00	2	0.4	643	PASS
198	198	100	100	100.0	148557	PASS
199	198	5	9	6.9	10243	PASS
275	198	10	60	25.3	37550	PASS
365	198	1	100	3.0	4408	PASS
441	198	0.01	100	10.4	15466	PASS
442	198	50	100	69.9	103845	PASS
443	442	15	24	19.0	19735	PASS