

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071922\
 Data File : BP011030.D
 Acq On : 19 Jul 2022 13:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704

Quant Time: Jul 19 23:39:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	384425	20.000	ng/ul	0.00
20) Naphthalene-d8	10.469	136	1632609	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.339	164	925813	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.110	188	1847306	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.233	240	1805562	20.000	ng/ul	# 0.00
88) Perylene-d12	23.586	264	1860011	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	76814	7.204	ng/uL	0.00
4) Pyridine-d5	3.593	84	487314	17.681	ng/ul	0.00
7) Phenol-d5	6.857	99	537779	17.042	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	369581	18.213	ng/ul	0.00
11) 2-Chlorophenol-d4	7.210	132	461114	18.084	ng/ul	0.00
15) 4-Methylphenol-d8	8.393	113	434893	17.215	ng/ul	0.00
21) Nitrobenzene-d5	8.840	128	210721	17.711	ng/ul	0.00
24) 2-Nitrophenol-d4	9.557	143	200040	16.992	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.098	165	373258	17.337	ng/ul	0.00
31) 4-Chloroaniline-d4	10.616	131	594281	16.743	ng/ul	0.00
46) Dimethylphthalate-d6	13.763	166	1119524	17.693	ng/ul	0.00
49) Acenaphthylene-d8	14.034	160	1334173	17.870	ng/ul	0.00
54) 4-Nitrophenol-d4	14.557	143	196719	15.361	ng/ul	0.00
60) Fluorene-d10	15.339	176	938381	17.447	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.463	200	136593	14.943	ng/ul	0.00
73) Anthracene-d10	17.210	188	1414939	18.017	ng/ul	0.00
81) Pyrene-d10	19.463	212	1622366	17.108	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.433	264	1581547	17.704	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.216	88	81516	7.286	ng/uL#	82
5) Pyridine	3.611	79	505476	17.796	ng/ul#	75
6) Benzaldehyde	6.828	77	336924	21.981	ng/ul	97
8) Phenol	6.881	94	584785	17.434	ng/ul	88
10) Bis(2-Chloroethyl)ether	7.116	93	472876	17.763	ng/ul	90
12) 2-Chlorophenol	7.246	128	475939	18.057	ng/ul	92
13) 2-Methylphenol	8.128	108	425452	17.152	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.216	45	674079	18.513	ng/ul#	97
16) Acetophenone	8.504	105	683458	17.872	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.493	70	343821	17.918	ng/ul	93
18) 4-Methylphenol	8.457	108	462393	17.459	ng/ul	98
19) Hexachloroethane	8.752	117	198507	19.031	ng/ul#	81
22) Nitrobenzene	8.881	77	507175	18.127	ng/ul	94
23) Isophorone	9.410	82	925034	17.665	ng/ul	98
25) 2-Nitrophenol	9.587	139	230907	17.473	ng/ul#	82
26) 2,4-Dimethylphenol	9.657	107	495852	17.744	ng/ul	96
27) Bis(2-Chloroethoxy)met...	9.899	93	602408	17.383	ng/ul	98
29) 2,4-Dichlorophenol	10.122	162	391154	17.553	ng/ul	99
30) Naphthalene	10.522	128	1475869	17.740	ng/ul	99
32) 4-Chloroaniline	10.640	127	592913	16.888	ng/ul	99
33) Hexachlorobutadiene	10.804	225	230625	17.779	ng/ul	95
34) Caprolactam	11.434	113	169845	21.748	ng/ul	93
35) 4-Chloro-3-methylphenol	11.781	107	427928	16.992	ng/ul	95
36) 2-Methylnaphthalene	12.145	142	946954	17.557	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071922\
 Data File : BP011030.D
 Acq On : 19 Jul 2022 13:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704

Quant Time: Jul 19 23:39:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.363	142	947205	17.521	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.516	216	427106	17.491	ng/ul	97
40) Hexachlorocyclopentadiene	12.492	237	253976	16.793	ng/ul	97
41) 2,4,6-Trichlorophenol	12.763	196	269407	16.754	ng/ul	97
42) 2,4,5-Trichlorophenol	12.834	196	299364	17.420	ng/ul	100
43) 1,1'-Biphenyl	13.169	154	1218735	17.643	ng/ul#	97
44) 2-Chloronaphthalene	13.204	162	904556	17.433	ng/ul	97
45) 2-Nitroaniline	13.422	65	262209	17.345	ng/ul	86
47) Dimethylphthalate	13.810	163	1115022	17.464	ng/ul	99
48) 2,6-Dinitrotoluene	13.928	165	205276	17.758	ng/ul	96
50) Acenaphthylene	14.057	152	1496243	17.861	ng/ul	98
51) 3-Nitroaniline	14.257	138	218720	15.749	ng/ul	95
52) Acenaphthene	14.404	153	978352	17.629	ng/ul	98
53) 2,4-Dinitrophenol	14.463	184	120101	17.759	ng/ul	86
55) 4-Nitrophenol	14.569	109	197787	20.173	ng/ul#	80
56) Dibenzofuran	14.745	168	1351930	17.655	ng/ul	98
57) 2,4-Dinitrotoluene	14.716	165	299235	18.148	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	14.975	232	227822	16.928	ng/ul#	87
59) Diethylphthalate	15.186	149	1158223	17.664	ng/ul	97
61) Fluorene	15.398	166	1093613	17.795	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.398	204	501142	17.713	ng/ul	96
63) 4-Nitroaniline	15.428	138	235719	18.468	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.481	198	142322	15.177	ng/ul	99
67) N-Nitrosodiphenylamine	15.616	169	929073	17.606	ng/ul	98
68) 4-Bromophenyl-phenylether	16.298	248	293872	17.639	ng/ul	93
69) Hexachlorobenzene	16.398	284	338935	17.528	ng/ul#	91
70) Atrazine	16.586	200	314380	17.222	ng/ul	97
71) Pentachlorophenol	16.757	266	211731	17.779	ng/ul	99
72) Phenanthrene	17.151	178	1713287	17.802	ng/ul	98
74) Anthracene	17.245	178	1738102	18.155	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.128	216	442091	16.717	ng/ul	97
76) Pentachlorobenzene	14.657	250	403983	17.426	ng/ul	98
77) Carbazole	17.522	167	1614310	18.016	ng/ul	99
78) Di-n-butylphthalate	18.092	149	1970513	17.593	ng/ul	99
80) Fluoranthene	19.139	202	1982356	17.874	ng/ul#	89
82) Pyrene	19.492	202	2046409	17.884	ng/ul#	84
83) Butylbenzylphthalate	20.386	149	900470	17.640	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.157	252	608915	19.046	ng/ul#	97
85) Benzo(a)anthracene	21.216	228	1937203	17.617	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.157	149	1373180	18.029	ng/ul#	96
87) Chrysene	21.268	228	1910624	17.696	ng/ul	100
89) Di-n-octyl phthalate	22.068	149	2306616	18.491	ng/ul	100
90) Benzo(b)fluoranthene	22.868	252	2044540	17.802	ng/ul#	96
91) Benzo(k)fluoranthene	22.915	252	1924739	17.842	ng/ul#	95
93) Benzo(a)pyrene	23.480	252	1970434	17.865	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.027	276	2241116	17.485	ng/ul#	88
95) Dibenzo(a,h)anthracene	26.045	278	1932510	17.424	ng/ul#	91
96) Benzo(g,h,i)perylene	26.768	276	1901610	17.419	ng/ul#	87





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.286 ng/uL
 RT: 3.216 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704


Tgt Ion: 88 Resp: 81516
 Ion Ratio Lower Upper
 88 100
 43 28.3 35.8 53.6#
 58 71.8 68.1 102.1



#5
 Pyridine
 Concen: 17.796 ng/uL
 RT: 3.611 min Scan# 89
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Tgt Ion: 79 Resp: 505476
 Ion Ratio Lower Upper
 79 100
 52 76.6 83.8 125.8#
 51 36.6 41.0 61.4#









#6
Benzaldehyde
Concen: 21.981 ng/ul
RT: 6.828 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion: 77 Resp: 336924
Ion Ratio Lower Upper
77 100
105 97.5 76.0 114.0
106 95.6 73.2 109.8



#8
Phenol
Concen: 17.434 ng/ul
RT: 6.881 min Scan# 645
Delta R.T. -0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion: 94 Resp: 584785
Ion Ratio Lower Upper
94 100
65 28.0 27.4 41.2
66 38.1 37.4 56.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 17.763 ng/ul
 RT: 7.116 min Scan# 685
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704

Tgt Ion: 93 Resp: 472876




Ion	Ratio	Lower	Upper
93	100		
63	78.3	72.4	108.6
95	32.4	27.3	40.9




#12
 2-Chlorophenol
 Concen: 18.057 ng/ul
 RT: 7.246 min Scan# 707
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Tgt Ion: 128 Resp: 475939

Ion	Ratio	Lower	Upper
128	100		
64	47.3	44.8	67.2
130	33.3	25.8	38.6









#13
2-Methylphenol
Concen: 17.152 ng/ul
RT: 8.128 min Scan# 857
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:108 Resp: 425452
Ion Ratio Lower Upper
108 100
107 90.8 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.513 ng/ul
RT: 8.216 min Scan# 872
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion: 45 Resp: 674079
Ion Ratio Lower Upper
45 100
77 14.2 11.2 16.8
79 10.7 6.9 10.3#









#16
Acetophenone
Concen: 17.872 ng/ul
RT: 8.504 min Scan# 919
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:105 Resp: 683458
Ion Ratio Lower Upper
105 100
77 75.9 67.5 101.3
51 27.1 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 17.918 ng/ul
RT: 8.493 min Scan# 919
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion: 70 Resp: 343821
Ion Ratio Lower Upper
70 100
42 60.6 53.8 80.8
101 9.8 7.8 11.6
130 21.2 15.4 23.2









#18
4-Methylphenol
Concen: 17.459 ng/ul
RT: 8.457 min Scan# 913
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:108 Resp: 462393
Ion Ratio Lower Upper
108 100
107 112.5 88.3 132.5



#19
Hexachloroethane
Concen: 19.031 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:117 Resp: 198507
Ion Ratio Lower Upper
117 100
201 88.4 85.7 128.5
199 54.6 57.4 86.0#









#22
Nitrobenzene
Concen: 18.127 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion: 77 Resp: 507175
Ion Ratio Lower Upper
77 100
123 44.0 31.5 47.3
65 13.1 11.4 17.2



#23
Isophorone
Concen: 17.665 ng/ul
RT: 9.410 min Scan# 1075
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion: 82 Resp: 925034
Ion Ratio Lower Upper
82 100
95 6.7 6.6 10.0
138 16.7 13.8 20.6









#25
2-Nitrophenol
Concen: 17.473 ng/ul
RT: 9.587 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:139 Resp: 230907
Ion Ratio Lower Upper
139 100
65 52.3 54.2 81.4#
109 28.5 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 17.744 ng/ul
RT: 9.657 min Scan# 1117
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:107 Resp: 495852
Ion Ratio Lower Upper
107 100
121 51.2 39.1 58.7
122 87.0 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 17.383 ng/ul

RT: 9.899 min Scan# 1158

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion: 93 Resp: 602408

Ion Ratio Lower Upper

93 100

95 32.8 26.6 39.8

123 11.6 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 17.553 ng/ul

RT: 10.122 min Scan# 1196

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:162 Resp: 391154


Ion Ratio Lower Upper

162 100

164 64.7 52.5 78.7

98 39.8 31.4 47.0









#30
Naphthalene
Concen: 17.740 ng/ul
RT: 10.522 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:128 Resp: 1475869
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 16.888 ng/ul
RT: 10.640 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:127 Resp: 592913
Ion Ratio Lower Upper
127 100
129 31.9 26.2 39.2





#33

Hexachlorobutadiene

Concen: 17.779 ng/ul

RT: 10.804 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:225 Resp: 230625

Ion Ratio Lower Upper

225 100

223 63.6 53.7 80.5

227 65.5 49.4 74.2



#34

Caprolactam

Concen: 21.748 ng/ul

RT: 11.434 min Scan# 1419

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:113 Resp: 169845


Ion Ratio Lower Upper

113 100

55 181.3 153.8 230.8

56 130.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 16.992 ng/ul
RT: 11.781 min Scan# 1478
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:107 Resp: 427928
Ion Ratio Lower Upper
107 100
144 25.2 18.6 27.8
142 78.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 17.557 ng/ul
RT: 12.145 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:142 Resp: 946954
Ion Ratio Lower Upper
142 100
141 89.6 70.1 105.1





#37

1-Methylnaphthalene

Concen: 17.521 ng/ul

RT: 12.363 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:142 Resp: 947205

Ion Ratio Lower Upper

142 100

141 92.9 73.3 109.9

116 3.6 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.491 ng/ul

RT: 12.516 min Scan# 1603

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Tgt Ion:216 Resp: 427106




Ion Ratio Lower Upper


216 100

214 77.4 63.0 94.6

179 21.5 18.1 27.1

108 20.1 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 16.793 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:237 Resp: 253976
Ion Ratio Lower Upper
237 100
235 63.1 52.8 79.2
272 13.9 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 16.754 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:196 Resp: 269407
Ion Ratio Lower Upper
196 100
198 94.3 78.2 117.4
200 29.9 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 17.420 ng/ul

RT: 12.834 min Scan# 1657

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:196 Resp: 299364

Ion Ratio Lower Upper

196 100

198 96.0 77.0 115.4

200 30.6 24.8 37.2



#43

1,1'-Biphenyl

Concen: 17.643 ng/ul

RT: 13.169 min Scan# 1714

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:154 Resp: 1218735


Ion Ratio Lower Upper

154 100

153 40.6 31.9 47.9

76 14.6 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 17.433 ng/ul
RT: 13.204 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:162 Resp: 904556
Ion Ratio Lower Upper
162 100
164 32.2 25.5 38.3
127 43.3 31.9 47.9



#45
2-Nitroaniline
Concen: 17.345 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion: 65 Resp: 262209
Ion Ratio Lower Upper
65 100
92 69.1 49.0 73.6
138 98.0 66.4 99.6









#47
Dimethylphthalate
Concen: 17.464 ng/ul
RT: 13.810 min Scan# 1823
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:163 Resp: 1115022
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 9.8 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 17.758 ng/ul
RT: 13.928 min Scan# 1843
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:165 Resp: 205276
Ion Ratio Lower Upper
165 100
89 62.0 52.9 79.3
121 23.5 19.4 29.2









#50
Acenaphthylene
Concen: 17.861 ng/ul
RT: 14.057 min Scan# 1865
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:152 Resp: 1496243
Ion Ratio Lower Upper
152 100
151 19.9 16.5 24.7
153 13.3 10.1 15.1



#51
3-Nitroaniline
Concen: 15.749 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:138 Resp: 218720
Ion Ratio Lower Upper
138 100
108 12.8 11.0 16.4
92 124.0 104.6 156.8









#52
Acenaphthene
Concen: 17.629 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:153 Resp: 978352
Ion Ratio Lower Upper
153 100
152 47.0 36.2 54.2
154 86.9 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 17.759 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:184 Resp: 120101
Ion Ratio Lower Upper
184 100
63 73.9 73.8 110.6
154 64.6 57.2 85.8









#55
4-Nitrophenol
Concen: 20.173 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:109 Resp: 197787
Ion Ratio Lower Upper
109 100
139 131.2 75.6 113.4#
65 143.9 107.6 161.4



#56
Dibenzofuran
Concen: 17.655 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:168 Resp: 1351930
Ion Ratio Lower Upper
168 100
139 40.0 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 18.148 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:165 Resp: 299235
Ion Ratio Lower Upper
165 100
63 49.9 47.0 70.6
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 16.928 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:232 Resp: 227822
Ion Ratio Lower Upper
232 100
131 55.6 35.8 53.8#
130 2.7 1.8 2.8
166 34.9 23.7 35.5









#59
Diethylphthalate
Concen: 17.664 ng/ul
RT: 15.186 min Scan# 2057
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:149 Resp: 1158223
Ion Ratio Lower Upper
149 100
177 20.5 17.6 26.4
150 12.0 10.3 15.5



#61
Fluorene
Concen: 17.795 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:166 Resp: 1093613
Ion Ratio Lower Upper
166 100
165 98.9 77.4 116.2
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 17.713 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:204 Resp: 501142
Ion Ratio Lower Upper
204 100
206 32.1 26.0 39.0
141 66.0 49.3 73.9



#63
4-Nitroaniline
Concen: 18.468 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:138 Resp: 235719
Ion Ratio Lower Upper
138 100
92 52.9 46.2 69.2
108 75.9 85.8 128.8#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 15.177 ng/ul
 RT: 15.481 min Scan# 2107
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704


Tgt Ion:198 Resp: 142322
 Ion Ratio Lower Upper
 198 100
 182 5.5 4.3 6.5
 77 29.9 23.7 35.5



#67
 N-Nitrosodiphenylamine
 Concen: 17.606 ng/ul
 RT: 15.616 min Scan# 2130
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Tgt Ion:169 Resp: 929073
 Ion Ratio Lower Upper
 169 100
 168 66.5 52.6 78.8
 167 36.4 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 17.639 ng/ul
RT: 16.298 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:248 Resp: 293872
Ion Ratio Lower Upper
248 100
250 97.9 80.8 121.2
141 89.5 63.9 95.9



#69
Hexachlorobenzene
Concen: 17.528 ng/ul
RT: 16.398 min Scan# 2263
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:284 Resp: 338935
Ion Ratio Lower Upper
284 100
142 43.6 27.8 41.6#
249 30.3 25.4 38.2





#70

Atrazine

Concen: 17.222 ng/ul

RT: 16.586 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:200 Resp: 314380

Ion Ratio Lower Upper

200 100

173 28.2 20.7 31.1

215 46.7 38.6 57.8



#71

Pentachlorophenol

Concen: 17.779 ng/ul

RT: 16.757 min Scan# 2324

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:266 Resp: 211731


Ion Ratio Lower Upper

266 100

264 64.2 51.4 77.2

268 65.6 53.5 80.3









#72
Phenanthrene
Concen: 17.802 ng/ul
RT: 17.151 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:178 Resp: 1713287
Ion Ratio Lower Upper
178 100
179 15.0 12.7 19.1
176 19.1 14.6 22.0



#74
Anthracene
Concen: 18.155 ng/ul
RT: 17.245 min Scan# 2407
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:178 Resp: 1738102
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.717 ng/uL
RT: 13.128 min Scan# 1707
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:216 Resp: 442091
Ion Ratio Lower Upper
216 100
214 77.9 65.0 97.4
218 47.8 39.1 58.7



#76
Pentachlorobenzene
Concen: 17.426 ng/uL
RT: 14.657 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:250 Resp: 403983
Ion Ratio Lower Upper
250 100
248 63.6 49.8 74.6
252 61.7 51.2 76.8









#77
 Carbazole
 Concen: 18.016 ng/ul
 RT: 17.522 min Scan# 2454
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704


Tgt Ion:167 Resp: 1614310
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.9 25.3
 139 13.6 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 17.593 ng/ul
 RT: 18.092 min Scan# 2551
 Delta R.T. -0.000 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Tgt Ion:149 Resp: 1970513
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.1 10.7
 104 5.2 4.6 6.8









#80
Fluoranthene
Concen: 17.874 ng/ul
RT: 19.139 min Scan# 2729
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:202 Resp: 1982356
Ion Ratio Lower Upper
202 100
101 13.4 7.1 10.7#
100 10.2 5.7 8.5#



#82
Pyrene
Concen: 17.884 ng/ul
RT: 19.492 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:202 Resp: 2046409
Ion Ratio Lower Upper
202 100
101 16.5 7.9 11.9#
100 13.7 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 17.640 ng/ul

RT: 20.386 min Scan# 2941

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:149 Resp: 900470

Ion Ratio Lower Upper

149 100

91 79.1 60.0 90.0

206 17.5 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 19.046 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:252 Resp: 608915


Ion Ratio Lower Upper

252 100

254 65.0 51.0 76.4

126 13.7 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 17.617 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:228 Resp: 1937203

Ion Ratio Lower Upper

228 100

229 19.8 15.0 22.4

226 27.4 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.029 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.000 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:149 Resp: 1373180


Ion Ratio Lower Upper

149 100

167 26.2 22.6 33.8

279 3.6 4.2 6.4









#87
Chrysene
Concen: 17.696 ng/ul
RT: 21.268 min Scan# 3091
Delta R.T. 0.006 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49




Instrument :
BNA_P
ClientSampleId :
SSTD020704


Tgt Ion:228 Resp: 1910624
Ion Ratio Lower Upper
228 100
226 29.5 23.6 35.4
229 19.4 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.491 ng/ul
RT: 22.068 min Scan# 3227
Delta R.T. -0.000 min
Lab File: BP011030.D
Acq: 19 Jul 2022 13:49

Tgt Ion:149 Resp: 2306616





#90

Benzo(b)fluoranthene

Concen: 17.802 ng/ul

RT: 22.868 min Scan# 3363

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:252 Resp: 2044540

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 13.1 6.2 9.2#



#91

Benzo(k)fluoranthene

Concen: 17.842 ng/ul

RT: 22.915 min Scan# 3371

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:252 Resp: 1924739


Ion Ratio Lower Upper

252 100

253 21.8 17.0 25.4

125 13.2 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 17.865 ng/ul

RT: 23.480 min Scan# 3467

Delta R.T. 0.006 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49

Instrument :

BNA_P

ClientSampleId :

SSTD020704





Tgt Ion:252 Resp: 1970434

Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.0

125 15.0 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.485 ng/ul

RT: 26.027 min Scan# 3900

Delta R.T. 0.012 min

Lab File: BP011030.D

Acq: 19 Jul 2022 13:49




Tgt Ion:276 Resp: 2241116


Ion Ratio Lower Upper

276 100

138 30.1 14.6 22.0#

277 24.8 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 17.424 ng/ul
 RT: 26.045 min Scan# 3903
 Delta R.T. 0.006 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020704

Tgt Ion:278 Resp: 1932510
 Ion Ratio Lower Upper
 278 100
 139 21.9 9.4 14.2#
 279 23.9 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 17.419 ng/ul
 RT: 26.768 min Scan# 4026
 Delta R.T. 0.006 min
 Lab File: BP011030.D
 Acq: 19 Jul 2022 13:49

Tgt Ion:276 Resp: 1901610
 Ion Ratio Lower Upper
 276 100
 138 28.0 12.7 19.1#
 277 24.2 18.2 27.4

