

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071922\
 Data File : BP011037.D
 Acq On : 19 Jul 2022 17:50
 Operator : CG/JU
 Sample : N3753-07
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 GBHP8

Quant Time: Jul 19 23:41:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	337360	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	1371553	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.339	164	774599	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	1540145	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.227	240	1493424	20.000	ng/u1	# 0.00
88) Perylene-d12	23.580	264	1519957	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	52142	5.572	ng/uL	0.00
4) Pyridine-d5	3.593	84	211853	8.759	ng/u1	0.00
7) Phenol-d5	6.857	99	198833	7.180	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	680027	38.188	ng/u1	0.00
11) 2-Chlorophenol-d4	7.210	132	637389	28.485	ng/u1	0.00
15) 4-Methylphenol-d8	8.392	113	375519	16.938	ng/u1	0.00
21) Nitrobenzene-d5	8.839	128	404135	40.433	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	364749	36.879	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.092	165	595669	32.933	ng/u1	0.00
31) 4-Chloroaniline-d4	10.616	131	889749	29.839	ng/u1	0.00
46) Dimethylphthalate-d6	13.757	166	2148719	40.587	ng/u1	0.00
49) Acenaphthylene-d8	14.027	160	2601455	41.647	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	41745	3.896	ng/u1	0.00
60) Fluorene-d10	15.339	176	1864572	41.435	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.463	200	247176	32.433	ng/u1	0.00
73) Anthracene-d10	17.204	188	3062418	46.772	ng/u1	0.00
81) Pyrene-d10	19.462	212	3495292	44.563	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.433	264	3455506	47.334	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071922\
Data File : BP011037.D
Acq On : 19 Jul 2022 17:50
Operator : CG/JU
Sample : N3753-07
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
GBHP8

Quant Time: Jul 19 23:41:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 19 03:25:54 2022
Response via : Initial Calibration

