

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072021\
 Data File : BP006256.D
 Acq On : 20 Jul 2021 18:12
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP072021

Quant Time: Jul 21 08:58:00 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	39.051	2.4	95	-0.02
3	Pyridine	40.000	40.784	-2.0	97	-0.02
4	n-Nitrosodimethylamine	40.000	40.356	-0.9	97	-0.01
5 S	2-Fluorophenol	80.000	79.114	1.1	96	-0.01
6	Aniline	40.000	40.467	-1.2	97	-0.01
7 S	Phenol-d6	80.000	80.489	-0.6	96	0.00
8	2-Chlorophenol	40.000	40.371	-0.9	98	-0.01
9	Benzaldehyde	40.000	39.357	1.6	93	-0.01
10 C	Phenol	40.000	40.087	-0.2	96	-0.01
11	bis(2-Chloroethyl)ether	40.000	40.800	-2.0	99	-0.01
12	1,3-Dichlorobenzene	40.000	39.559	1.1	97	0.00
13 C	1,4-Dichlorobenzene	40.000	39.600	1.0	96	0.00
14	1,2-Dichlorobenzene	40.000	39.579	1.1	97	-0.01
15	Benzyl Alcohol	40.000	40.833	-2.1	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.011	-2.5	100	-0.02
17	2-Methylphenol	40.000	40.725	-1.8	97	-0.01
18	Hexachloroethane	40.000	39.654	0.9	96	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	41.858	-4.6	99	0.00
20	3+4-Methylphenols	40.000	40.752	-1.9	97	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	39.991	0.0	99	-0.01
23 S	Nitrobenzene-d5	80.000	84.840	-6.1	101	0.00
24	Nitrobenzene	40.000	41.587	-4.0	100	0.00
25	Isophorone	40.000	41.235	-3.1	100	0.00
26 C	2-Nitrophenol	40.000	40.628	-1.6	105	0.00
27	2,4-Dimethylphenol	40.000	40.170	-0.4	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.071	-0.2	99	0.00
29 C	2,4-Dichlorophenol	40.000	40.646	-1.6	98	0.00
30	1,2,4-Trichlorobenzene	40.000	39.645	0.9	98	0.00
31	Naphthalene	40.000	39.754	0.6	99	0.00
32	Benzoic acid	40.000	43.398	-8.5	108	0.00
33	4-Chloroaniline	40.000	39.804	0.5	98	0.00
34 C	Hexachlorobutadiene	40.000	39.516	1.2	98	0.00
35	Caprolactam	40.000	41.008	-2.5	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.577	-3.9	101	0.00
37	2-Methylnaphthalene	40.000	40.097	-0.2	99	0.00
38	1-Methylnaphthalene	40.000	39.999	0.0	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.105	2.2	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.148	-0.4	97	0.00
42 S	2,4,6-Tribromophenol	80.000	83.388	-4.2	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.195	-3.0	101	0.00
44	2,4,5-Trichlorophenol	40.000	41.178	-2.9	101	0.00
45 S	2-Fluorobiphenyl	80.000	78.868	1.4	100	0.00
46	1,1'-Biphenyl	40.000	39.320	1.7	100	0.00
47	2-Chloronaphthalene	40.000	39.446	1.4	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072021\
 Data File : BP006256.D
 Acq On : 20 Jul 2021 18:12
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP072021

Quant Time: Jul 21 08:58:00 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.516	-1.3	107	0.00
49	Acenaphthylene	40.000	40.236	-0.6	101	0.00
50	Dimethylphthalate	40.000	39.839	0.4	102	0.00
51	2,6-Dinitrotoluene	40.000	40.674	-1.7	104	0.00
52 C	Acenaphthene	40.000	39.837	0.4	101	0.00
53	3-Nitroaniline	40.000	40.203	-0.5	104	0.00
54 P	2,4-Dinitrophenol	40.000	44.773	-11.9	109	0.00
55	Dibenzofuran	40.000	39.256	1.9	100	0.00
56 P	4-Nitrophenol	40.000	42.341	-5.9	105	0.00
57	2,4-Dinitrotoluene	40.000	40.917	-2.3	107	0.00
58	Fluorene	40.000	39.477	1.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.839	-4.6	103	0.00
60	Diethylphthalate	40.000	40.222	-0.6	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.630	0.9	101	0.00
62	4-Nitroaniline	40.000	39.743	0.6	103	0.00
63	Azobenzene	40.000	40.533	-1.3	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.305	-0.8	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.438	1.4	100	0.00
67	4-Bromophenyl-phenylether	40.000	39.106	2.2	98	0.00
68	Hexachlorobenzene	40.000	39.225	1.9	100	0.00
69	Atrazine	40.000	40.855	-2.1	101	0.00
70 C	Pentachlorophenol	40.000	42.520	-6.3	105	0.00
71	Phenanthrene	40.000	39.271	1.8	101	0.00
72	Anthracene	40.000	39.398	1.5	100	0.00
73	Carbazole	40.000	39.291	1.8	100	0.00
74	Di-n-butylphthalate	40.000	40.817	-2.0	103	0.00
75 C	Fluoranthene	40.000	39.290	1.8	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	39.884	0.3	95	0.00
78	Pyrene	40.000	40.415	-1.0	103	0.00
79 S	Terphenyl-d14	80.000	80.340	-0.4	103	0.00
80	Butylbenzylphthalate	40.000	42.005	-5.0	103	0.00
81	Benzo(a)anthracene	40.000	39.863	0.3	102	0.00
82	3,3'-Dichlorobenzidine	40.000	39.982	0.0	99	0.00
83	Chrysene	40.000	39.745	0.6	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.958	-4.9	103	0.00
85 c	Di-n-octyl phthalate	40.000	41.637	-4.1	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.686	0.8	103	0.00
88	Benzo(b)fluoranthene	40.000	39.984	0.0	102	0.00
89	Benzo(k)fluoranthene	40.000	39.948	0.1	103	0.00
90 C	Benzo(a)pyrene	40.000	39.822	0.4	102	0.00
91	Dibenzo(a,h)anthracene	40.000	39.718	0.7	103	0.00
92	Benzo(g,h,i)perylene	40.000	39.856	0.4	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072021\
Data File : BP006256.D
Acq On : 20 Jul 2021 18:12
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP072021

Quant Time: Jul 21 08:58:00 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 08:44:35 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0