

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP072120\  
 Data File : BP002808.D  
 Acq On : 21 Jul 2020 10:57  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040

Quant Time: Jul 21 13:45:49 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP071020.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 13 10:54:03 2020  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.55  | 152  | 210494   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 10.32 | 136  | 747197   | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10      | 14.20 | 164  | 412831   | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10      | 16.96 | 188  | 841470   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12          | 21.07 | 240  | 742443   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 23.26 | 264  | 849771   | 20.00 | ng    | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 5) 2-Fluorophenol           | 5.18  | 112  | 929574   | 74.51 | ng    | 0.00     |
| 7) Phenol-d6                | 6.76  | 99   | 1230335  | 69.10 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.70  | 82   | 1080485  | 77.34 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.71 | 330  | 393109   | 91.22 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.82 | 172  | 2083284  | 78.07 | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.55 | 244  | 2581443  | 79.72 | ng    | 0.00     |

| Target Compounds               | R.T.  | QIon | Response | Conc   | Units | Qvalue |
|--------------------------------|-------|------|----------|--------|-------|--------|
| 2) 1,4-Dioxane                 | 3.12  | 88   | 210626   | 38.723 | ng    | 99     |
| 3) Pyridine                    | 3.50  | 79   | 584566   | 36.062 | ng    | 99     |
| 4) n-Nitrosodimethylamine      | 3.42  | 42   | 238751   | 39.098 | ng    | 99     |
| 6) Aniline                     | 6.88  | 93   | 790189   | 35.056 | ng    | 99     |
| 8) 2-Chlorophenol              | 7.12  | 128  | 502275   | 36.185 | ng    | 99     |
| 9) Benzaldehyde                | 6.69  | 77   | 341896   | 35.342 | ng    | 98     |
| 10) Phenol                     | 6.78  | 94   | 638675   | 35.016 | ng    | 99     |
| 11) bis(2-Chloroethyl)ether    | 6.99  | 93   | 528640   | 34.846 | ng    | 99     |
| 12) 1,3-Dichlorobenzene        | 7.44  | 146  | 586614   | 36.855 | ng    | 99     |
| 13) 1,4-Dichlorobenzene        | 7.58  | 146  | 591710   | 37.043 | ng    | 99     |
| 14) 1,2-Dichlorobenzene        | 7.89  | 146  | 556920   | 36.154 | ng    | 99     |
| 15) Benzyl Alcohol             | 7.79  | 79   | 434789   | 36.491 | ng    | 99     |
| 16) 2,2'-oxybis(1-Chloropropan | 8.09  | 45   | 726069   | 33.278 | ng    | 100    |
| 17) 2-Methylphenol             | 8.01  | 107  | 452430   | 34.817 | ng    | 99     |
| 18) Hexachloroethane           | 8.61  | 117  | 220180   | 38.916 | ng    | 95     |
| 19) n-Nitroso-di-n-propylamine | 8.36  | 70   | 348528   | 32.143 | ng    | 99     |
| 20) 3+4-Methylphenols          | 8.34  | 107  | 587311   | 34.182 | ng    | 98     |
| 22) Acetophenone               | 8.37  | 105  | 679526   | 36.890 | ng    | # 98   |
| 24) Nitrobenzene               | 8.74  | 77   | 579498   | 38.301 | ng    | 98     |
| 25) Isophorone                 | 9.26  | 82   | 1033676  | 36.303 | ng    | 99     |
| 26) 2-Nitrophenol              | 9.45  | 139  | 275926   | 42.223 | ng    | 97     |
| 27) 2,4-Dimethylphenol         | 9.53  | 122  | 399605   | 37.665 | ng    | 99     |
| 28) bis(2-Chloroethoxy)methane | 9.75  | 93   | 637213   | 36.844 | ng    | 100    |
| 29) 2,4-Dichlorophenol         | 9.99  | 162  | 456986   | 39.069 | ng    | 99     |
| 30) 1,2,4-Trichlorobenzene     | 10.19 | 180  | 538706   | 40.209 | ng    | 99     |
| 31) Naphthalene                | 10.37 | 128  | 1453241  | 36.839 | ng    | 100    |
| 32) Benzoic acid               | 9.72  | 122  | 219863   | 29.907 | ng    | 99     |
| 33) 4-Chloroaniline            | 10.49 | 127  | 618943   | 36.615 | ng    | 98     |
| 34) Hexachlorobutadiene        | 10.66 | 225  | 341216   | 44.063 | ng    | 96     |
| 35) Caprolactam                | 11.27 | 113  | 135089   | 33.031 | ng    | 94     |
| 36) 4-Chloro-3-methylphenol    | 11.65 | 107  | 448074   | 35.538 | ng    | 99     |
| 37) 2-Methylnaphthalene        | 11.99 | 142  | 999122   | 35.891 | ng    | 99     |
| 38) 1-Methylnaphthalene        | 12.22 | 142  | 934767   | 35.503 | ng    | 100    |
| 40) 1,2,4,5-Tetrachlorobenzene | 12.37 | 216  | 557597   | 43.949 | ng    | 98     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP072120\  
 Data File : BP002808.D  
 Acq On : 21 Jul 2020 10:57  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040

Quant Time: Jul 21 13:45:49 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP071020.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 13 10:54:03 2020  
 Response via : Initial Calibration


| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 41) Hexachlorocyclopentadiene  | 12.36 | 237  | 249817   | 47.755 | ng    | 100      |
| 43) 2,4,6-Trichlorophenol      | 12.63 | 196  | 343080   | 42.246 | ng    | 96       |
| 44) 2,4,5-Trichlorophenol      | 12.71 | 196  | 381843   | 40.539 | ng    | 96       |
| 46) 1,1'-Biphenyl              | 13.03 | 154  | 1177952  | 38.274 | ng    | 99       |
| 47) 2-Chloronaphthalene        | 13.06 | 162  | 973095   | 38.927 | ng    | 99       |
| 48) 2-Nitroaniline             | 13.28 | 65   | 297757   | 38.534 | ng    | 98       |
| 49) Acenaphthylene             | 13.92 | 152  | 1487708  | 37.751 | ng    | 100      |
| 50) Dimethylphthalate          | 13.67 | 163  | 1152237  | 36.597 | ng    | 100      |
| 51) 2,6-Dinitrotoluene         | 13.79 | 165  | 262173   | 38.896 | ng    | 96       |
| 52) Acenaphthene               | 14.27 | 154  | 878859   | 37.316 | ng    | 97       |
| 53) 3-Nitroaniline             | 14.12 | 138  | 283334   | 37.571 | ng    | 100      |
| 54) 2,4-Dinitrophenol          | 14.35 | 184  | 108882   | 36.215 | ng    | 96       |
| 55) Dibenzofuran               | 14.60 | 168  | 1364468  | 36.123 | ng    | 97       |
| 56) 4-Nitrophenol              | 14.47 | 139  | 228418   | 40.421 | ng    | 98       |
| 57) 2,4-Dinitrotoluene         | 14.59 | 165  | 335563   | 37.802 | ng    | 96       |
| 58) Fluorene                   | 15.26 | 166  | 1000459  | 35.795 | ng    | 99       |
| 59) 2,3,4,6-Tetrachlorophenol  | 14.85 | 232  | 289103   | 38.938 | ng    | 94       |
| 60) Diethylphthalate           | 15.05 | 149  | 1102316  | 36.111 | ng    | 99       |
| 61) 4-Chlorophenyl-phenylether | 15.26 | 204  | 562888   | 37.903 | ng    | 98       |
| 62) 4-Nitroaniline             | 15.29 | 138  | 276864   | 34.761 | ng    | 96       |
| 63) Azobenzene                 | 15.56 | 77   | 1104619  | 35.194 | ng    | 98       |
| 65) 4,6-Dinitro-2-methylphenol | 15.37 | 198  | 186944   | 38.357 | ng    | 90       |
| 66) n-Nitrosodiphenylamine     | 15.48 | 169  | 947898   | 38.133 | ng    | 98       |
| 67) 4-Bromophenyl-phenylether  | 16.16 | 248  | 402258   | 43.885 | ng    | 93       |
| 68) Hexachlorobenzene          | 16.28 | 284  | 420078   | 43.498 | ng    | 94       |
| 69) Atrazine                   | 16.45 | 200  | 258417   | 38.080 | ng    | 97       |
| 70) Pentachlorophenol          | 16.63 | 266  | 173226   | 41.078 | ng    | 96       |
| 71) Phenanthrene               | 17.00 | 178  | 1701349  | 37.343 | ng    | 100      |
| 72) Anthracene                 | 17.10 | 178  | 1654088  | 36.997 | ng    | 99       |
| 73) Carbazole                  | 17.37 | 167  | 1609275  | 36.680 | ng    | 100      |
| 74) Di-n-butylphthalate        | 17.95 | 149  | 1803951  | 37.913 | ng    | 99       |
| 75) Fluoranthene               | 18.99 | 202  | 1967215  | 37.264 | ng    | 97       |
| 77) Benzidine                  | 19.17 | 184  | 690221   | 37.506 | ng    | 100      |
| 78) Pyrene                     | 19.35 | 202  | 1960464  | 38.245 | ng    | 99       |
| 80) Butylbenzylphthalate       | 20.23 | 149  | 780280   | 38.573 | ng    | 99       |
| 81) Benzo(a)anthracene         | 21.06 | 228  | 1858493  | 38.670 | ng    | 100      |
| 82) 3,3'-Dichlorobenzidine     | 20.99 | 252  | 659275   | 41.421 | ng    | 98       |
| 83) Chrysene                   | 21.11 | 228  | 1707707  | 37.328 | ng    | 98       |
| 84) Bis(2-ethylhexyl)phthalate | 21.00 | 149  | 1053040  | 37.429 | ng    | 99       |
| 85) Di-n-octyl phthalate       | 21.86 | 149  | 1952637  | 38.604 | ng    | 98       |
| 87) Indeno(1,2,3-cd)pyrene     | 25.48 | 276  | 2662522  | 43.351 | ng    | 95       |
| 88) Benzo(b)fluoranthene       | 22.61 | 252  | 2099745  | 39.793 | ng    | 98       |
| 89) Benzo(k)fluoranthene       | 22.65 | 252  | 1905462  | 36.662 | ng    | 98       |
| 90) Benzo(a)pyrene             | 23.17 | 252  | 1965659  | 39.311 | ng    | 98       |
| 91) Dibenzo(a,h)anthracene     | 25.48 | 278  | 2148843  | 43.383 | ng    | 98       |
| 92) Benzo(g,h,i)perylene       | 26.14 | 276  | 2233993  | 44.484 | ng    | 96       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP072120\  
 Data File : BP002808.D  
 Acq On : 21 Jul 2020 10:57  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleID :  
 SSTDCCC040

Quant Time: Jul 21 13:45:49 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP071020.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 13 10:54:03 2020  
 Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.55 min Scan# 792  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 150     | 160.0 | 123.7 | 185.5 |
| 115     | 59.1  | 47.2  | 70.8  |






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane  
Concen: 38.723 ng  
RT: 3.12 min Scan# 40  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 88      | 100   |       |       |
| 58      | 71.5  | 58.2  | 87.4  |
| 43      | 27.0  | 21.5  | 32.3  |






Abundance

Ion 88.00 (87.70 to 88.70): BP0

Ion 58.00 (57.70 to 58.70): BP0

Ion 43.00 (42.70 to 43.70): BP0





#3

Pvridine

Concen: 36.062 ng

RT: 3.50 min Scan# 104

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion: 79 Resp: 584566

Ion Ratio Lower Upper




79 100

52 66.9 54.2 81.2

51 33.8 27.3 40.9



Abundance Ion 79.00 (78.70 to 79.70): BP0  
Ion 52.00 (51.70 to 52.70): BP0  
Ion 51.00 (50.70 to 51.70): BP0



#4

n-Nitrosodimethylamine

Concen: 39.098 ng

RT: 3.42 min Scan# 91

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion: 42 Resp: 238751

Ion Ratio Lower Upper



42 100


74 143.9 116.3 174.5

44 6.6 5.9 8.9



Abundance Ion 42.00 (41.70 to 42.70): BP0  
Ion 74.00 (73.70 to 74.70): BP0  
Ion 44.00 (43.70 to 44.70): BP0





#5

2-Fluorophenol

Concen: 74.509 ng

RT: 5.18 min Scan# 389

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:112 Resp: 929574

Ion Ratio Lower Upper

112 100

64 57.3 45.6 68.4




63 30.0 23.5 35.3



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BP0

Ion 63.00 (62.70 to 63.70): BP0



#6

Aniline

Concen: 35.056 ng

RT: 6.88 min Scan# 679

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion: 93 Resp: 790189

Ion Ratio Lower Upper

93 100

66 38.3 30.1 45.1



65 19.8 15.4 23.2




Abundance Ion 93.00 (92.70 to 93.70): BP0

Ion 66.00 (65.70 to 66.70): BP0

Ion 65.00 (64.70 to 65.70): BP0





#7

Phenol-d6

Concen: 69.099 ng

RT: 6.76 min Scan# 658

Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :





BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion: 99 Resp: 1230335

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 99  | 100   |       |       |
| 42  | 16.9  | 13.4  | 20.2  |
| 71  | 32.2  | 24.6  | 36.8  |



#8

2-Chlorophenol

Concen: 36.185 ng

RT: 7.12 min Scan# 720




Delta R.T. -0.01 min


Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion: 128 Resp: 502275

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 128 | 100   |       |       |
| 130 | 31.8  | 12.2  | 52.2  |
| 64  | 49.7  | 28.8  | 68.8  |





#9

Benzaldehyde

Concen: 35.342 ng

RT: 6.69 min Scan# 647

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040





Tot Ion: 77 Resp: 341896

Ion Ratio Lower Upper

77 100

105 97.1 73.2 113.2

106 92.9 73.6 113.6



#10

Phenol

Concen: 35.016 ng

RT: 6.78 min Scan# 662

Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57




Tgt Ion: 94 Resp: 638675


Ion Ratio Lower Upper

94 100

65 27.0 7.0 47.0

66 34.7 13.7 53.7









#11  
bis(2-Chloroethyl)ether  
Concen: 34.846 ng  
RT: 6.99 min Scan# 697  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 63      | 74.2  | 55.1  | 95.1  |
| 95      | 32.3  | 12.7  | 52.7  |



#12  
1,3-Dichlorobenzene  
Concen: 36.855 ng  
RT: 7.44 min Scan# 774  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 65.2  | 52.6  | 78.8  |
| 75      | 31.9  | 25.2  | 37.8  |









#13  
 1,4-Dichlorobenzene  
 Concen: 37.043 ng  
 RT: 7.58 min Scan# 798  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 65.8  | 52.1  | 78.1  |
| 111     | 40.3  | 32.2  | 48.2  |



#14  
 1,2-Dichlorobenzene  
 Concen: 36.154 ng  
 RT: 7.89 min Scan# 851  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 64.7  | 51.0  | 76.6  |
| 111     | 42.5  | 33.9  | 50.9  |









#15  
Benzyl Alcohol  
Concen: 36.491 ng  
RT: 7.79 min Scan# 834  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 79      | 100   |       |       |
| 108     | 77.5  | 63.1  | 94.7  |
| 77      | 63.5  | 50.8  | 76.2  |



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 33.278 ng  
RT: 8.09 min Scan# 884  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 45      | 100   |       |       |
| 77      | 14.9  | 0.0   | 34.8  |
| 79      | 11.9  | 0.0   | 31.6  |









#17  
2-Methylphenol  
Concen: 34.817 ng  
RT: 8.01 min Scan# 871  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


|          |       |       |        |
|----------|-------|-------|--------|
| Tot Ion: | 107   | Resp: | 452430 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 112.6 | 89.9  | 134.9  |
| 77       | 46.7  | 36.6  | 54.8   |
| 79       | 44.4  | 34.9  | 52.3   |



#18  
Hexachloroethane  
Concen: 38.916 ng  
RT: 8.61 min Scan# 973  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 220180 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 96.3  | 76.6  | 115.0  |
| 201      | 110.2 | 80.3  | 120.5  |









#19  
n-Nitroso-di-n-propylamine  
Concen: 32.143 ng  
RT: 8.36 min Scan# 930  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


|          |       |       |        |
|----------|-------|-------|--------|
| Tot Ion: | 70    | Resp: | 348528 |
| Ion      | Ratio | Lower | Upper  |
| 70       | 100   |       |        |
| 42       | 54.4  | 44.2  | 66.4   |
| 101      | 9.4   | 7.7   | 11.5   |
| 130      | 20.1  | 15.3  | 22.9   |



#20  
3+4-Methylphenols  
Concen: 34.182 ng  
RT: 8.34 min Scan# 927  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 587311 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 89.8  | 69.5  | 109.5  |
| 77       | 37.7  | 14.3  | 54.3   |
| 79       | 28.1  | 7.7   | 47.7   |









#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.32 min Scan# 1264  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


|             |       |        |
|-------------|-------|--------|
| Tot Ion:136 | Resp: | 747197 |
| Ion Ratio   | Lower | Upper  |
| 136 100     |       |        |
| 137 10.6    | 8.6   | 13.0   |
| 54 8.1      | 6.9   | 10.3   |
| 68 6.9      | 5.7   | 8.5    |



#22  
Acetophenone  
Concen: 36.890 ng  
RT: 8.37 min Scan# 932  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

|     |         |       |        |
|-----|---------|-------|--------|
| Tgt | Ion:105 | Resp: | 679526 |
| Ion | Ratio   | Lower | Upper  |
| 105 | 100     |       |        |
| 71  | 4.0     | 4.4   | 6.6#   |
| 51  | 25.9    | 21.8  | 32.8   |
| 120 | 22.0    | 18.2  | 27.2   |









#23  
Nitrobenzene-d5  
Concen: 77.340 ng  
RT: 8.70 min Scan# 988  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion: 82 Resp: 1080485  
Ion Ratio Lower Upper  
82 100  
128 42.6 33.8 50.8  
54 47.6 38.8 58.2



#24  
Nitrobenzene  
Concen: 38.301 ng  
RT: 8.74 min Scan# 995  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion: 77 Resp: 579498  
Ion Ratio Lower Upper  
77 100  
123 45.3 35.0 52.6  
65 13.1 10.4 15.6









#25  
Isophorone  
Concen: 36.303 ng  
RT: 9.26 min Scan# 1084  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion: 82 Resp: 1033676  
Ion Ratio Lower Upper  
82 100  
95 6.5 5.2 7.8  
138 17.1 13.2 19.8



#26  
2-Nitrophenol  
Concen: 42.223 ng  
RT: 9.45 min Scan# 1115  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion: 139 Resp: 275926  
Ion Ratio Lower Upper  
139 100  
109 27.3 21.9 32.9  
65 48.2 41.3 61.9









#27  
2,4-Dimethylphenol  
Concen: 37.665 ng  
RT: 9.53 min Scan# 1129  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 122     | 100   |       |       |
| 107     | 114.9 | 92.2  | 138.4 |
| 121     | 58.7  | 45.8  | 68.6  |



#28  
bis(2-Chloroethoxy)methane  
Concen: 36.844 ng  
RT: 9.75 min Scan# 1166  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 95      | 32.3  | 25.8  | 38.6  |
| 123     | 11.2  | 8.9   | 13.3  |









#29  
2,4-Dichlorophenol  
Concen: 39.069 ng  
RT: 9.99 min Scan# 1207  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 162     | 100   |       |       |
| 164     | 65.2  | 44.5  | 84.5  |
| 98      | 36.4  | 15.2  | 55.2  |



#30  
1,2,4-Trichlorobenzene  
Concen: 40.209 ng  
RT: 10.19 min Scan# 1242  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 100   |       |       |
| 182     | 95.6  | 75.8  | 113.6 |
| 145     | 27.8  | 22.6  | 34.0  |









#31  
Naphthalene  
Concen: 36.839 ng  
RT: 10.37 min Scan# 1272  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:128 Resp: 1453241  
Ion Ratio Lower Upper  
128 100  
129 10.9 8.6 13.0  
127 13.0 10.6 15.8



#32  
Benzoic acid  
Concen: 29.907 ng  
RT: 9.72 min Scan# 1162  
Delta R.T. 0.02 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:122 Resp: 219863  
Ion Ratio Lower Upper  
122 100  
105 124.1 105.4 145.4  
77 96.5 78.5 118.5









#33  
4-Chloroaniline  
Concen: 36.615 ng  
RT: 10.49 min Scan# 1292  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


|          |       |       |        |
|----------|-------|-------|--------|
| Tot Ion: | 127   | Resp: | 618943 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 31.6  | 25.3  | 37.9   |
| 65       | 29.8  | 25.4  | 38.2   |
| 92       | 19.4  | 15.8  | 23.8   |



#34  
Hexachlorobutadiene  
Concen: 44.063 ng  
RT: 10.66 min Scan# 1322  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 341216 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 65.5  | 49.9  | 74.9   |
| 227      | 61.9  | 51.8  | 77.6   |









#35  
 Caprolactam  
 Concen: 33.031 ng  
 RT: 11.27 min Scan# 1425  
 Delta R.T. 0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040


Tot Ion:113 Resp: 135089  
 Ion Ratio Lower Upper  
 113 100  
 55 164.5 152.8 192.8  
 56 121.9 107.5 147.5



#36  
 4-Chloro-3-methylphenol  
 Concen: 35.538 ng  
 RT: 11.65 min Scan# 1489  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

Tgt Ion:107 Resp: 448074  
 Ion Ratio Lower Upper  
 107 100  
 144 26.0 20.6 30.8  
 142 78.6 63.4 95.2









#37  
 2-Methylnaphthalene  
 Concen: 35.891 ng  
 RT: 11.99 min Scan# 1548  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTCCCC040


Tot Ion:142 Resp: 999122  
 Ion Ratio Lower Upper  
 142 100  
 141 86.9 70.4 105.6  
 115 34.3 27.0 40.6



#38  
 1-Methylnaphthalene  
 Concen: 35.503 ng  
 RT: 12.22 min Scan# 1586  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

Tgt Ion:142 Resp: 934767  
 Ion Ratio Lower Upper  
 142 100  
 141 89.4 71.5 107.3  
 116 3.8 2.9 4.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.20 min Scan# 1924

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:164 Resp: 412831

Ion Ratio Lower Upper

164 100

162 100.2 80.6 121.0




160 45.9 36.4 54.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 43.949 ng

RT: 12.37 min Scan# 1613

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:216 Resp: 557597


Ion Ratio Lower Upper

216 100

214 77.7 63.4 95.2

179 19.2 16.2 24.2

108 16.5 14.8 22.2






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 47.755 ng

RT: 12.36 min Scan# 1610

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:237 Resp: 249817

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 237 | 100   |       |       |
| 235 | 62.4  | 42.3  | 82.3  |
| 272 | 13.4  | 0.0   | 32.5  |

237 100

235 62.4 42.3 82.3




272 13.4 0.0 32.5



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 91.223 ng

RT: 15.71 min Scan# 2180

Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion:330 Resp: 393109

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 330 | 100   |       |       |
| 332 | 95.8  | 76.5  | 114.7 |
| 141 | 28.7  | 28.8  | 43.2# |

330 100

332 95.8 76.5 114.7



141 28.7 28.8 43.2#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43  
 2,4,6-Trichlorophenol  
 Concen: 42.246 ng  
 RT: 12.63 min Scan# 1656  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040


| Tot Ion | 196    | 198   | 200  |
|---------|--------|-------|------|
| Resp:   | 343080 |       |      |
| Ratio   | 100    | 99.8  | 30.8 |
| Lower   |        | 75.8  | 24.5 |
| Upper   |        | 113.6 | 36.7 |



#44  
 2,4,5-Trichlorophenol  
 Concen: 40.539 ng  
 RT: 12.71 min Scan# 1670  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

| Tgt Ion | 196    | 198   | 97   | 132  |
|---------|--------|-------|------|------|
| Resp:   | 381843 |       |      |      |
| Ratio   | 100    | 97.9  | 44.3 | 24.5 |
| Lower   |        | 75.8  | 38.0 | 20.4 |
| Upper   |        | 113.6 | 57.0 | 30.6 |









#45  
2-Fluorobiphenyl  
Concen: 78.066 ng  
RT: 12.82 min Scan# 1688  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:172 Resp: 2083284  
Ion Ratio Lower Upper  
172 100  
171 35.2 28.1 42.1  
170 23.6 19.3 28.9



#46  
1,1'-Biphenyl  
Concen: 38.274 ng  
RT: 13.03 min Scan# 1724  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:154 Resp: 1177952  
Ion Ratio Lower Upper  
154 100  
153 39.6 19.8 59.8  
76 13.3 0.0 34.7









#47  
2-Chloronaphthalene  
Concen: 38.927 ng  
RT: 13.06 min Scan# 1730  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion: 162 Resp: 973095  
Ion Ratio Lower Upper  
162 100  
127 35.4 29.0 43.4  
164 33.0 26.8 40.2



#48  
2-Nitroaniline  
Concen: 38.534 ng  
RT: 13.28 min Scan# 1767  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion: 65 Resp: 297757  
Ion Ratio Lower Upper  
65 100  
92 72.3 57.0 85.6  
138 109.7 86.0 129.0





#49

Acenaphthylene

Concen: 37.751 ng

RT: 13.92 min Scan# 1875

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :


BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion:152 Resp: 1487708


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 152 | 100   |       |       |
| 151 | 20.4  | 16.2  | 24.4  |
| 153 | 13.7  | 10.9  | 16.3  |




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



Time-->



#50

Dimethylphthalate

Concen: 36.597 ng

RT: 13.67 min Scan# 1834


Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:163 Resp: 1152237


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 163 | 100   |       |       |
| 194 | 4.4   | 3.4   | 5.2   |
| 164 | 10.0  | 7.8   | 11.8  |




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E







Time-->



#51  
2,6-Dinitrotoluene  
Concen: 38.896 ng  
RT: 13.79 min Scan# 1853  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 165     | 100   |       |       |
| 63      | 50.4  | 43.2  | 64.8  |
| 89      | 53.7  | 44.6  | 67.0  |



#52  
Acenaphthene  
Concen: 37.316 ng  
RT: 14.27 min Scan# 1935  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 154     | 100   |       |       |
| 153     | 109.6 | 90.1  | 135.1 |
| 152     | 53.2  | 43.8  | 65.6  |









#53  
3-Nitroaniline  
Concen: 37.571 ng  
RT: 14.12 min Scan# 1909  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:138 Resp: 283334  
Ion Ratio Lower Upper  
138 100  
108 9.8 8.2 12.2  
92 117.5 94.3 141.5



#54  
2,4-Dinitrophenol  
Concen: 36.215 ng  
RT: 14.35 min Scan# 1948  
Delta R.T. 0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:184 Resp: 108882  
Ion Ratio Lower Upper  
184 100  
63 58.7 51.6 77.4  
154 55.5 45.0 67.6





#55

Dibenzofuran

Concen: 36.123 ng

RT: 14.60 min Scan# 1992

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:168 Resp: 1364468

Ion Ratio Lower Upper

168 100

139 35.7 27.1 40.7




169 14.2 10.8 16.2



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 40.421 ng

RT: 14.47 min Scan# 1969

Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion:139 Resp: 228418

Ion Ratio Lower Upper

139 100

109 70.9 48.1 88.1



65 98.0 77.9 117.9




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BP0









#57  
2,4-Dinitrotoluene  
Concen: 37.802 ng  
RT: 14.59 min Scan# 1990  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:165 Resp: 335563  
Ion Ratio Lower Upper  
165 100  
63 46.6 35.1 52.7  
89 69.7 58.4 87.6  
182 3.0 2.2 3.2



#58  
Fluorene  
Concen: 35.795 ng  
RT: 15.26 min Scan# 2103  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:166 Resp: 1000459  
Ion Ratio Lower Upper  
166 100  
165 96.4 77.7 116.5  
167 13.5 11.0 16.6









#59  
 2,3,4,6-Tetrachlorophenol  
 Concen: 38.938 ng  
 RT: 14.85 min Scan# 2034  
 Delta R.T. 0.00 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 232     | 100   |       |       |
| 131     | 38.0  | 34.3  | 51.5  |
| 130     | 2.0   | 1.8   | 2.6   |
| 166     | 27.6  | 24.0  | 36.0  |



#60  
 Diethylphthalate  
 Concen: 36.111 ng  
 RT: 15.05 min Scan# 2067  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 149     | 100   |       |       |
| 177     | 23.0  | 18.9  | 28.3  |
| 150     | 12.4  | 10.1  | 15.1  |









#61  
4-Chlorophenyl-phenylether  
Concen: 37.903 ng  
RT: 15.26 min Scan# 2103  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:204 Resp: 562888  
Ion Ratio Lower Upper  
204 100  
206 32.0 26.1 39.1  
141 50.6 41.8 62.8



#62  
4-Nitroaniline  
Concen: 34.761 ng  
RT: 15.29 min Scan# 2109  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:138 Resp: 276864  
Ion Ratio Lower Upper  
138 100  
92 53.1 32.0 72.0  
108 76.6 52.2 92.2









#63  
Azobenzene  
Concen: 35.194 ng  
RT: 15.56 min Scan# 2154  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion: | 77    | Resp: | 1104619 |
|----------|-------|-------|---------|
| Ion      | Ratio | Lower | Upper   |
| 77       | 100   |       |         |
| 182      | 27.1  | 5.1   | 45.1    |
| 105      | 21.2  | 0.8   | 40.8    |
| 51       | 26.9  | 7.6   | 47.6    |



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 16.96 min Scan# 2393  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion: | 188   | Resp: | 841470 |
|----------|-------|-------|--------|
| Ion      | Ratio | Lower | Upper  |
| 188      | 100   |       |        |
| 94       | 9.3   | 8.1   | 12.1   |
| 80       | 10.1  | 8.9   | 13.3   |





#65

4,6-Dinitro-2-methylphenol

Concen: 38.357 ng

RT: 15.37 min Scan# 2122

Delta R.T. 0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:198 Resp: 186944

Ion Ratio Lower Upper

198 100

51 38.8 27.1 67.1




105 40.0 24.9 64.9



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BPO

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 38.133 ng

RT: 15.48 min Scan# 2141

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion:169 Resp: 947898

Ion Ratio Lower Upper

169 100

168 65.3 53.4 80.0



167 35.2 29.2 43.8




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67  
4-Bromophenyl-phenylether  
Concen: 43.885 ng  
RT: 16.16 min Scan# 2256  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 248     | 100   |       |       |
| 250     | 94.6  | 76.7  | 115.1 |
| 141     | 58.4  | 55.9  | 83.9  |



#68  
Hexachlorobenzene  
Concen: 43.498 ng  
RT: 16.28 min Scan# 2277  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 28.5  | 26.9  | 40.3  |
| 249     | 29.2  | 24.6  | 37.0  |





#69

Atrazine

Concen: 38.080 ng

RT: 16.45 min Scan# 2305

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampled :

SSTDCCC040


Tot Ion:200 Resp: 258417

Ion Ratio Lower Upper

200 100

173 24.4 6.3 46.3




215 48.8 30.6 70.6



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 41.078 ng

RT: 16.63 min Scan# 2337

Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion:266 Resp: 173226

Ion Ratio Lower Upper

266 100

268 64.1 50.6 75.8



264 66.3 48.6 72.8




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71  
Phenanthrene  
Concen: 37.343 ng  
RT: 17.00 min Scan# 2400  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:178 Resp: 1701349  
Ion Ratio Lower Upper  
178 100  
176 20.0 15.9 23.9  
179 15.3 12.3 18.5



#72  
Anthracene  
Concen: 36.997 ng  
RT: 17.10 min Scan# 2416  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:178 Resp: 1654088  
Ion Ratio Lower Upper  
178 100  
176 18.9 15.6 23.4  
179 15.9 12.6 19.0





#73

Carbazole

Concen: 36.680 ng

RT: 17.37 min Scan# 2463

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :


BNA\_P

ClientSampled :

SSTDCCC040

Tot Ion:167 Resp: 1609275

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 167 | 100   |       |       |
| 166 | 21.5  | 17.1  | 25.7  |
| 139 | 12.2  | 9.6   | 14.4  |






Abundance

Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 37.913 ng

RT: 17.95 min Scan# 2560


Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:149 Resp: 1803951

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 150 | 9.8   | 7.5   | 11.3  |
| 104 | 5.0   | 4.0   | 6.0   |






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#75

Fluoranthene

Concen: 37.264 ng

RT: 18.99 min Scan# 2738

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :


BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion:202 Resp: 1967215


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 202 | 100   |       |       |
| 101 | 11.0  | 0.0   | 32.5  |
| 203 | 17.8  | 0.0   | 38.6  |




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



Time-->



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.07 min Scan# 3092


Delta R.T. 0.00 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:240 Resp: 742443


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 240 | 100   |       |       |
| 120 | 10.1  | 9.4   | 14.2  |
| 236 | 26.8  | 21.3  | 31.9  |




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



Time-->



#77

Benzidine

Concen: 37.506 ng

RT: 19.17 min Scan# 2769

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :


BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion:184 Resp: 690221

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 184 | 100   |       |       |
| 185 | 14.3  | 11.4  | 17.2  |
| 183 | 11.4  | 9.0   | 13.4  |






Abundance

Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.245 ng

RT: 19.35 min Scan# 2798


Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:202 Resp: 1960464

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 202 | 100   |       |       |
| 200 | 21.8  | 17.7  | 26.5  |
| 203 | 17.8  | 14.5  | 21.7  |






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 79.717 ng

RT: 19.55 min Scan# 2833

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040


Tot Ion:244 Resp: 2581443

Ion Ratio Lower Upper

244 100

212 7.7 6.0 9.0




122 9.5 8.9 13.3



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 38.573 ng

RT: 20.23 min Scan# 2949

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57


Tgt Ion:149 Resp: 780280

Ion Ratio Lower Upper

149 100

91 73.1 59.4 89.2



206 23.2 18.3 27.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BP0

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 38.670 ng

RT: 21.06 min Scan# 3089

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :


BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion:228 Resp: 1858493

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 228 | 100   |       |       |
| 226 | 28.5  | 22.6  | 34.0  |
| 229 | 19.9  | 16.0  | 24.0  |






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 41.421 ng

RT: 20.99 min Scan# 3078


Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:252 Resp: 659275

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 254 | 62.2  | 51.0  | 76.4  |
| 126 | 10.0  | 8.4   | 12.6  |






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 37.328 ng

RT: 21.11 min Scan# 3098

Delta R.T. -0.01 min

Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Instrument :





BNA\_P

ClientSampleId :

SSTDCCC040

Tot Ion:228 Resp: 1707707

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 228 | 100   |       |       |
| 226 | 31.8  | 24.6  | 37.0  |
| 229 | 20.7  | 15.7  | 23.5  |



#84

Bis(2-ethylhexyl)phthalate

Concen: 37.429 ng

RT: 21.00 min Scan# 3080




Delta R.T. -0.01 min


Lab File: BP002808.D

Acq: 21 Jul 2020 10:57

Tgt Ion:149 Resp: 1053040

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 167 | 27.8  | 22.4  | 33.6  |
| 279 | 5.1   | 3.5   | 5.3   |









#85  
Di-n-octyl phthalate  
Concen: 38.604 ng  
RT: 21.86 min Scan# 3226  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:149 Resp: 1952637  
Ion Ratio Lower Upper  
149 100  
167 1.6 1.3 1.9  
43 9.4 8.2 12.2



#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 23.26 min Scan# 3464  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:264 Resp: 849771  
Ion Ratio Lower Upper  
264 100  
260 24.2 19.3 28.9  
265 21.2 16.8 25.2









#87  
 Indeno(1,2,3-cd)pyrene  
 Concen: 43.351 ng  
 RT: 25.48 min Scan# 3841  
 Delta R.T. 0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDC040


Tot Ion:276 Resp: 2662522  
 Ion Ratio Lower Upper  
 276 100  
 138 21.5 20.8 31.2  
 277 25.6 20.0 30.0



#88  
 Benzo(b)fluoranthene  
 Concen: 39.793 ng  
 RT: 22.61 min Scan# 3353  
 Delta R.T. -0.01 min  
 Lab File: BP002808.D  
 Acq: 21 Jul 2020 10:57

Tgt Ion:252 Resp: 2099745  
 Ion Ratio Lower Upper  
 252 100  
 253 22.8 17.7 26.5  
 125 8.7 8.2 12.2









#89  
Benzo(k)fluoranthene  
Concen: 36.662 ng  
RT: 22.65 min Scan# 3360  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040


Tot Ion:252 Resp: 1905462  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.8 26.8  
125 8.9 8.3 12.5



#90  
Benzo(a)pyrene  
Concen: 39.311 ng  
RT: 23.17 min Scan# 3448  
Delta R.T. -0.01 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:252 Resp: 1965659  
Ion Ratio Lower Upper  
252 100  
253 21.3 17.6 26.4  
125 9.7 8.6 13.0









#91  
Dibenzo(a,h)anthracene  
Concen: 43.383 ng  
RT: 25.48 min Scan# 3841  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040

Tot Ion:278 Resp: 2148843  
Ion Ratio Lower Upper  
278 100  
139 14.3 13.1 19.7  
279 23.8 19.0 28.6



#92  
Benzo(g,h,i)perylene  
Concen: 44.484 ng  
RT: 26.14 min Scan# 3954  
Delta R.T. 0.00 min  
Lab File: BP002808.D  
Acq: 21 Jul 2020 10:57

Tgt Ion:276 Resp: 2233993  
Ion Ratio Lower Upper  
276 100  
277 24.3 19.1 28.7  
138 19.1 18.3 27.5

