

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
 Data File : BP006259.D
 Acq On : 21 Jul 2021 10:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 21 11:17:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	40.332	-0.8	92	-0.01
3	Pyridine	40.000	42.030	-5.1	94	-0.01
4	n-Nitrosodimethylamine	40.000	42.619	-6.5	96	0.00
5 S	2-Fluorophenol	80.000	81.498	-1.9	93	0.00
6	Aniline	40.000	41.422	-3.6	93	0.00
7 S	Phenol-d6	80.000	82.499	-3.1	93	0.00
8	2-Chlorophenol	40.000	41.048	-2.6	93	0.00
9	Benzaldehyde	40.000	41.427	-3.6	92	0.00
10 C	Phenol	40.000	41.356	-3.4	93	0.00
11	bis(2-Chloroethyl)ether	40.000	42.014	-5.0	95	0.00
12	1,3-Dichlorobenzene	40.000	40.360	-0.9	93	0.00
13 C	1,4-Dichlorobenzene	40.000	40.282	-0.7	92	0.00
14	1,2-Dichlorobenzene	40.000	40.466	-1.2	93	0.00
15	Benzyl Alcohol	40.000	41.832	-4.6	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.444	-8.6	100	0.00
17	2-Methylphenol	40.000	41.618	-4.0	93	0.00
18	Hexachloroethane	40.000	41.473	-3.7	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.087	-7.7	96	0.00
20	3+4-Methylphenols	40.000	42.113	-5.3	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	40.123	-0.3	94	0.00
23 S	Nitrobenzene-d5	80.000	86.255	-7.8	98	0.00
24	Nitrobenzene	40.000	42.319	-5.8	97	0.00
25	Isophorone	40.000	41.604	-4.0	96	0.00
26 C	2-Nitrophenol	40.000	39.174	2.1	97	0.00
27	2,4-Dimethylphenol	40.000	40.621	-1.6	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.665	-1.7	96	0.00
29 C	2,4-Dichlorophenol	40.000	40.776	-1.9	94	0.00
30	1,2,4-Trichlorobenzene	40.000	39.230	1.9	93	0.00
31	Naphthalene	40.000	39.977	0.1	95	0.00
32	Benzoic acid	40.000	41.895	-4.7	99	0.00
33	4-Chloroaniline	40.000	39.850	0.4	93	0.00
34 C	Hexachlorobutadiene	40.000	39.293	1.8	93	0.00
35	Caprolactam	40.000	41.038	-2.6	92	-0.01
36 C	4-Chloro-3-methylphenol	40.000	41.820	-4.6	97	0.00
37	2-Methylnaphthalene	40.000	40.108	-0.3	95	0.00
38	1-Methylnaphthalene	40.000	40.071	-0.2	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.030	2.4	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.179	-0.4	93	0.00
42 S	2,4,6-Tribromophenol	80.000	76.673	4.2	90	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.647	-1.6	95	0.00
44	2,4,5-Trichlorophenol	40.000	40.685	-1.7	95	0.00
45 S	2-Fluorobiphenyl	80.000	78.532	1.8	95	0.00
46	1,1'-Biphenyl	40.000	39.243	1.9	95	0.00
47	2-Chloronaphthalene	40.000	39.314	1.7	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
 Data File : BP006259.D
 Acq On : 21 Jul 2021 10:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 21 11:17:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.700	-4.3	106	0.00
49	Acenaphthylene	40.000	40.106	-0.3	96	0.00
50	Dimethylphthalate	40.000	39.731	0.7	97	0.00
51	2,6-Dinitrotoluene	40.000	40.124	-0.3	98	0.00
52 C	Acenaphthene	40.000	39.573	1.1	96	0.00
53	3-Nitroaniline	40.000	40.706	-1.8	100	0.00
54 P	2,4-Dinitrophenol	40.000	45.461	-13.7	106	0.00
55	Dibenzofuran	40.000	39.531	1.2	97	0.00
56 P	4-Nitrophenol	40.000	43.003	-7.5	102	0.00
57	2,4-Dinitrotoluene	40.000	40.912	-2.3	102	0.00
58	Fluorene	40.000	39.732	0.7	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.477	-3.7	98	0.00
60	Diethylphthalate	40.000	40.300	-0.7	98	0.00
61	4-Chlorophenyl-phenylether	40.000	39.419	1.5	96	0.00
62	4-Nitroaniline	40.000	40.433	-1.1	101	0.00
63	Azobenzene	40.000	41.867	-4.7	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.787	-2.0	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.466	1.3	97	0.00
67	4-Bromophenyl-phenylether	40.000	36.453	8.9	88	0.00
68	Hexachlorobenzene	40.000	38.773	3.1	96	0.00
69	Atrazine	40.000	39.834	0.4	96	0.00
70 C	Pentachlorophenol	40.000	41.537	-3.8	99	0.00
71	Phenanthrene	40.000	39.482	1.3	98	0.00
72	Anthracene	40.000	39.846	0.4	98	0.00
73	Carbazole	40.000	39.535	1.2	97	0.00
74	Di-n-butylphthalate	40.000	40.906	-2.3	99	0.00
75 C	Fluoranthene	40.000	39.447	1.4	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzydine	40.000	40.094	-0.2	94	0.00
78	Pyrene	40.000	39.840	0.4	100	0.00
79 S	Terphenyl-d14	80.000	79.294	0.9	100	0.00
80	Butylbenzylphthalate	40.000	41.655	-4.1	101	0.00
81	Benzo(a)anthracene	40.000	39.804	0.5	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.259	-0.6	98	0.00
83	Chrysene	40.000	39.925	0.2	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.192	-3.0	100	0.00
85 c	Di-n-octyl phthalate	40.000	41.495	-3.7	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.193	-0.5	102	0.00
88	Benzo(b)fluoranthene	40.000	40.467	-1.2	102	-0.01
89	Benzo(k)fluoranthene	40.000	39.831	0.4	101	0.00
90 C	Benzo(a)pyrene	40.000	40.275	-0.7	101	0.00
91	Dibenzo(a,h)anthracene	40.000	40.496	-1.2	103	0.00
92	Benzo(g,h,i)perylene	40.000	39.844	0.4	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
Data File : BP006259.D
Acq On : 21 Jul 2021 10:24
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 21 11:17:51 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 08:44:35 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0