

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
 Data File : BP006270.D
 Acq On : 21 Jul 2021 19:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 01:08:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	41.581	-4.0	97	-0.01
3	Pyridine	40.000	43.087	-7.7	98	-0.02
4	n-Nitrosodimethylamine	40.000	43.470	-8.7	99	0.00
5 S	2-Fluorophenol	80.000	82.286	-2.9	95	-0.01
6	Aniline	40.000	41.467	-3.7	95	-0.01
7 S	Phenol-d6	80.000	84.124	-5.2	96	0.00
8	2-Chlorophenol	40.000	41.519	-3.8	96	-0.01
9	Benzaldehyde	40.000	42.095	-5.2	95	0.00
10 C	Phenol	40.000	41.974	-4.9	96	-0.01
11	bis(2-Chloroethyl)ether	40.000	41.935	-4.8	97	-0.01
12	1,3-Dichlorobenzene	40.000	40.233	-0.6	94	0.00
13 C	1,4-Dichlorobenzene	40.000	40.102	-0.3	93	0.00
14	1,2-Dichlorobenzene	40.000	40.028	-0.1	94	-0.01
15	Benzyl Alcohol	40.000	42.271	-5.7	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.876	-12.2	105	0.00
17	2-Methylphenol	40.000	41.744	-4.4	95	0.00
18	Hexachloroethane	40.000	41.178	-2.9	95	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	43.554	-8.9	98	0.00
20	3+4-Methylphenols	40.000	41.737	-4.3	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.00
22	Acetophenone	40.000	40.813	-2.0	96	-0.01
23 S	Nitrobenzene-d5	80.000	88.933	-11.2	101	0.00
24	Nitrobenzene	40.000	43.860	-9.6	100	0.00
25	Isophorone	40.000	41.363	-3.4	96	-0.01
26 C	2-Nitrophenol	40.000	40.860	-2.1	101	0.00
27	2,4-Dimethylphenol	40.000	40.558	-1.4	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.067	-2.7	97	-0.01
29 C	2,4-Dichlorophenol	40.000	40.555	-1.4	93	0.00
30	1,2,4-Trichlorobenzene	40.000	38.921	2.7	92	0.00
31	Naphthalene	40.000	39.796	0.5	94	-0.01
32	Benzoic acid	40.000	43.544	-8.9	103	-0.01
33	4-Chloroaniline	40.000	39.864	0.3	93	0.00
34 C	Hexachlorobutadiene	40.000	38.833	2.9	91	0.00
35	Caprolactam	40.000	43.013	-7.5	96	-0.02
36 C	4-Chloro-3-methylphenol	40.000	41.809	-4.5	96	0.00
37	2-Methylnaphthalene	40.000	39.603	1.0	93	-0.01
38	1-Methylnaphthalene	40.000	39.696	0.8	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.870	2.8	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.297	-0.7	90	-0.01
42 S	2,4,6-Tribromophenol	80.000	68.691	14.1	78	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.788	-4.5	95	0.00
44	2,4,5-Trichlorophenol	40.000	42.014	-5.0	95	-0.01
45 S	2-Fluorobiphenyl	80.000	79.095	1.1	93	0.00
46	1,1'-Biphenyl	40.000	39.750	0.6	93	0.00
47	2-Chloronaphthalene	40.000	39.861	0.3	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
 Data File : BP006270.D
 Acq On : 21 Jul 2021 19:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 01:08:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 08:44:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.998	-10.0	109	0.00
49	Acenaphthylene	40.000	40.295	-0.7	93	0.00
50	Dimethylphthalate	40.000	40.042	-0.1	94	0.00
51	2,6-Dinitrotoluene	40.000	41.055	-2.6	97	0.00
52 C	Acenaphthene	40.000	39.920	0.2	94	0.00
53	3-Nitroaniline	40.000	41.479	-3.7	99	0.00
54 P	2,4-Dinitrophenol	40.000	49.127	-22.8	111	0.00
55	Dibenzofuran	40.000	39.624	0.9	94	0.00
56 P	4-Nitrophenol	40.000	44.446	-11.1	102	0.00
57	2,4-Dinitrotoluene	40.000	41.204	-3.0	99	0.00
58	Fluorene	40.000	39.519	1.2	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.071	-2.7	94	0.00
60	Diethylphthalate	40.000	40.566	-1.4	95	0.00
61	4-Chlorophenyl-phenylether	40.000	39.126	2.2	92	0.00
62	4-Nitroaniline	40.000	41.880	-4.7	101	0.00
63	Azobenzene	40.000	42.621	-6.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.242	-5.6	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.685	0.8	94	0.00
67	4-Bromophenyl-phenylether	40.000	32.904	17.7	77	0.00
68	Hexachlorobenzene	40.000	37.846	5.4	90	0.00
69	Atrazine	40.000	39.943	0.1	92	0.00
70 C	Pentachlorophenol	40.000	41.950	-4.9	96	0.00
71	Phenanthrene	40.000	39.654	0.9	95	0.00
72	Anthracene	40.000	39.638	0.9	94	0.00
73	Carbazole	40.000	39.943	0.1	95	0.00
74	Di-n-butylphthalate	40.000	40.698	-1.7	95	0.00
75 C	Fluoranthene	40.000	39.967	0.1	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	40.820	-2.1	94	0.00
78	Pyrene	40.000	39.355	1.6	97	0.00
79 S	Terphenyl-d14	80.000	78.446	1.9	97	0.00
80	Butylbenzylphthalate	40.000	41.520	-3.8	99	0.00
81	Benzo(a)anthracene	40.000	39.889	0.3	98	0.00
82	3,3'-Dichlorobenzidine	40.000	39.882	0.3	96	0.00
83	Chrysene	40.000	39.744	0.6	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.780	-4.5	100	-0.01
85 c	Di-n-octyl phthalate	40.000	42.050	-5.1	101	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	99	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.624	-1.6	103	-0.01
88	Benzo(b)fluoranthene	40.000	39.472	1.3	99	-0.01
89	Benzo(k)fluoranthene	40.000	41.194	-3.0	104	0.00
90 C	Benzo(a)pyrene	40.000	40.295	-0.7	101	0.00
91	Dibenzo(a,h)anthracene	40.000	40.571	-1.4	103	-0.01
92	Benzo(g,h,i)perylene	40.000	40.407	-1.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072121\
Data File : BP006270.D
Acq On : 21 Jul 2021 19:31
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 22 01:08:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 08:44:35 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0