

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072122\
 Data File : BP011070.D
 Acq On : 21 Jul 2022 19:38
 Operator : CG/JU
 Sample : N3709-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jul 21 23:25:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	348332	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.463	136	1353661	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.334	164	722804	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	1405551	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.222	240	1401743	20.000	ng/u1	# 0.00
88) Perylene-d12	23.574	264	1465376	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.176	96	40399	4.181	ng/uL	0.00
4) Pyridine-d5	3.587	84	489902	19.616	ng/u1	0.00
7) Phenol-d5	6.852	99	612084	21.406	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.017	67	409063	22.248	ng/u1	0.00
11) 2-Chlorophenol-d4	7.205	132	523186	22.645	ng/u1	0.00
15) 4-Methylphenol-d8	8.387	113	487266	21.287	ng/u1	-0.01
21) Nitrobenzene-d5	8.834	128	235662	23.889	ng/u1	0.00
24) 2-Nitrophenol-d4	9.552	143	232872	23.857	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.087	165	398319	22.313	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.610	131	610291	20.738	ng/u1	0.00
46) Dimethylphthalate-d6	13.751	166	1230391	24.906	ng/u1	-0.01
49) Acenaphthylene-d8	14.022	160	1540806	26.435	ng/u1	0.00
54) 4-Nitrophenol-d4	14.551	143	145145	14.517	ng/u1	0.00
60) Fluorene-d10	15.334	176	1049096	24.984	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	90836	13.060	ng/u1	0.00
73) Anthracene-d10	17.204	188	1557341	26.063	ng/u1	0.00
81) Pyrene-d10	19.457	212	1807659	24.554	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.422	264	1838185	26.118	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072122\
Data File : BP011070.D
Acq On : 21 Jul 2022 19:38
Operator : CG/JU
Sample : N3709-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Jul 21 23:25:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 19 03:25:54 2022
Response via : Initial Calibration

