

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072122\
 Data File : BP011074.D
 Acq On : 21 Jul 2022 22:31
 Operator : CG/JU
 Sample : PB146324BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK324

Quant Time: Jul 21 23:26:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	467728	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.463	136	1792757	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.334	164	910903	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.098	188	1667404	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.222	240	1582020	20.000	ng/u1	# 0.00
88) Perylene-d12	23.569	264	1697579	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.176	96	84464	6.510	ng/uL	0.00
4) Pyridine-d5	3.581	84	1020907	30.443	ng/u1	-0.01
7) Phenol-d5	6.846	99	1197641	31.193	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.017	67	789820	31.991	ng/u1	0.00
11) 2-Chlorophenol-d4	7.205	132	1011737	32.612	ng/u1	0.00
15) 4-Methylphenol-d8	8.387	113	916522	29.818	ng/u1	-0.01
21) Nitrobenzene-d5	8.834	128	455251	34.846	ng/u1	0.00
24) 2-Nitrophenol-d4	9.552	143	450160	34.822	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.087	165	739493	31.279	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.610	131	1136059	29.148	ng/u1	0.00
46) Dimethylphthalate-d6	13.751	166	2121643	34.079	ng/u1	-0.01
49) Acenaphthylene-d8	14.022	160	2645882	36.020	ng/u1	0.00
54) 4-Nitrophenol-d4	14.546	143	299730	23.788	ng/u1	-0.01
60) Fluorene-d10	15.334	176	1739067	32.863	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	216534	26.244	ng/u1	0.00
73) Anthracene-d10	17.198	188	2534240	35.751	ng/u1	0.00
81) Pyrene-d10	19.457	212	2795633	33.647	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.416	264	2973984	36.476	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072122\
Data File : BP011074.D
Acq On : 21 Jul 2022 22:31
Operator : CG/JU
Sample : PB146324BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK324

Quant Time: Jul 21 23:26:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 19 03:25:54 2022
Response via : Initial Calibration

