

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072221\
 Data File : BP006275.D
 Acq On : 22 Jul 2021 11:37
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Jul 22 12:51:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 12:49:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	255700	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	1001591	20.000	ng	# 0.00
39) Acenaphthene-d10	14.422	164	566309	20.000	ng	0.00
64) Phenanthrene-d10	17.175	188	1120674	20.000	ng	# 0.00
76) Chrysene-d12	21.274	240	1146439	20.000	ng	# 0.00
86) Perylene-d12	23.621	264	1214273	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.393	112	664932	43.325	ng	0.00
7) Phenol-d6	6.969	99	907217	43.235	ng	0.00
23) Nitrobenzene-d5	8.952	82	781345	46.975	ng	0.00
42) 2,4,6-Tribromophenol	15.916	330	200170	34.339	ng	0.00
45) 2-Fluorobiphenyl	13.051	172	1607680	43.211	ng	0.00
79) Terphenyl-d14	19.751	244	2351840	41.037	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.340	88	141796	22.095	ng	91
3) Pyridine	3.734	79	399328	22.353	ng	96
4) n-Nitrosodimethylamine	3.646	42	137632	22.720	ng	# 79
6) Aniline	7.134	93	488299	21.020	ng	97
8) 2-Chlorophenol	7.363	128	360049	21.246	ng	94
9) Benzaldehyde	6.946	77	278583	23.059	ng	94
10) Phenol	6.999	94	451556	21.691	ng	95
11) bis(2-Chloroethyl)ether	7.234	93	346974	21.520	ng	93
12) 1,3-Dichlorobenzene	7.687	146	387211	20.855	ng	98
13) 1,4-Dichlorobenzene	7.834	146	392844	20.871	ng	99
14) 1,2-Dichlorobenzene	8.146	146	375234	20.813	ng	97
15) Benzyl Alcohol	8.040	79	317970	21.049	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.328	45	382606	23.554	ng	93
17) 2-Methylphenol	8.240	107	287072	20.791	ng	98
18) Hexachloroethane	8.869	117	141210	21.103	ng	91
19) n-Nitroso-di-n-propyla...	8.610	70	276191	21.487	ng	92
20) 3+4-Methylphenols	8.569	107	385866	20.748	ng	98
22) Acetophenone	8.622	105	514818	21.297	ng	# 98
24) Nitrobenzene	8.993	77	411825	23.288	ng	93
25) Isophorone	9.522	82	701277	20.666	ng	95
26) 2-Nitrophenol	9.704	139	146845	22.108	ng	95
27) 2,4-Dimethylphenol	9.769	122	282551	21.045	ng	100
28) bis(2-Chloroethoxy)met...	10.010	93	409877	21.397	ng	98
29) 2,4-Dichlorophenol	10.228	162	283665	20.467	ng	95
30) 1,2,4-Trichlorobenzene	10.446	180	319102	20.363	ng	98
31) Naphthalene	10.634	128	1108494	21.085	ng	99
32) Benzoic acid	9.863	122	171067	20.533	ng	96
33) 4-Chloroaniline	10.746	127	434045	20.283	ng	97
34) Hexachlorobutadiene	10.922	225	186737	20.100	ng	100
35) Caprolactam	11.528	113	87772	21.045	ng	# 83
36) 4-Chloro-3-methylphenol	11.869	107	339092	21.021	ng	94
37) 2-Methylnaphthalene	12.245	142	748004	20.374	ng	97
38) 1-Methylnaphthalene	12.463	142	707672	20.381	ng	99
40) 1,2,4,5-Tetrachloroben...	12.610	216	319477	20.828	ng	# 97
41) Hexachlorocyclopentadiene	12.593	237	171968	20.477	ng	98
43) 2,4,6-Trichlorophenol	12.857	196	207138	21.530	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072221\
 Data File : BP006275.D
 Acq On : 22 Jul 2021 11:37
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Jul 22 12:51:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 12:49:34 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.922	196	229072	22.161	ng	# 92
46) 1,1'-Biphenyl	13.263	154	894021	21.418	ng	97
47) 2-Chloronaphthalene	13.298	162	686718	21.368	ng	97
48) 2-Nitroaniline	13.504	65	187132	24.778	ng	93
49) Acenaphthylene	14.145	152	1084874	21.205	ng	99
50) Dimethylphthalate	13.892	163	834053	20.990	ng	100
51) 2,6-Dinitrotoluene	14.004	165	176881	22.680	ng	87
52) Acenaphthene	14.487	154	669201	21.159	ng	98
53) 3-Nitroaniline	14.334	138	187627	22.708	ng	90
54) 2,4-Dinitrophenol	14.534	184	78457	23.715	ng	# 92
55) Dibenzofuran	14.822	168	1055833	21.144	ng	97
56) 4-Nitrophenol	14.634	139	157862	21.532	ng	88
57) 2,4-Dinitrotoluene	14.792	165	226641	22.525	ng	92
58) Fluorene	15.475	166	833218	20.830	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	187311	20.928	ng	# 78
60) Diethylphthalate	15.263	149	854626	20.838	ng	98
61) 4-Chlorophenyl-phenyle...	15.475	204	389853	20.605	ng	92
62) 4-Nitroaniline	15.492	138	195081	22.604	ng	85
63) Azobenzene	15.769	77	919450	22.265	ng	92
65) 4,6-Dinitro-2-methylph...	15.551	198	114538	21.968	ng	83
66) n-Nitrosodiphenylamine	15.692	169	730146	21.235	ng	99
67) 4-Bromophenyl-phenylether	16.375	248	197860	17.524	ng	91
68) Hexachlorobenzene	16.475	284	256407	19.805	ng	94
69) Atrazine	16.651	200	229555	20.440	ng	95
70) Pentachlorophenol	16.822	266	149297	20.591	ng	98
71) Phenanthrene	17.216	178	1294756	21.093	ng	99
72) Anthracene	17.310	178	1256982	20.994	ng	100
73) Carbazole	17.581	167	1269984	21.160	ng	99
74) Di-n-butylphthalate	18.157	149	1408537	20.561	ng	99
75) Fluoranthene	19.192	202	1464401	20.788	ng	94
77) Benzidine	19.380	184	538968	16.153	ng	99
78) Pyrene	19.545	202	1538389	20.322	ng	99
80) Butylbenzylphthalate	20.439	149	628825	19.608	ng	91
81) Benzo(a)anthracene	21.257	228	1544308	20.992	ng	100
82) 3,3'-Dichlorobenzidine	21.198	252	396343	19.456	ng	# 96
83) Chrysene	21.310	228	1488340	20.895	ng	99
84) Bis(2-ethylhexyl)phtha...	21.210	149	985686	20.392	ng	# 97
85) Di-n-octyl phthalate	22.121	149	1649974	20.347	ng	# 96
87) Indeno(1,2,3-cd)pyrene	26.057	276	1822892	21.190	ng	# 89
88) Benzo(b)fluoranthene	22.910	252	1618226	20.654	ng	# 96
89) Benzo(k)fluoranthene	22.957	252	1515443	20.773	ng	# 96
90) Benzo(a)pyrene	23.515	252	1436964	20.441	ng	# 94
91) Dibenzo(a,h)anthracene	26.080	278	1568959	21.085	ng	# 92
92) Benzo(g,h,i)perylene	26.804	276	1510738	20.858	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072221\
 Data File : BP006275.D
 Acq On : 22 Jul 2021 11:37
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Jul 22 12:51:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 12:49:34 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.799 min Scan# 801
 Delta R.T. -0.000 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:152 Resp: 255700
 Ion Ratio Lower Upper
 152 100
 150 151.4 125.6 188.4
 115 63.7 48.5 72.7



#2
 1,4-Dioxane
 Concen: 22.095 ng
 RT: 3.340 min Scan# 43
 Delta R.T. -0.006 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37

Tgt Ion: 88 Resp: 141796
 Ion Ratio Lower Upper
 88 100
 58 66.7 46.7 70.1
 43 24.1 20.3 30.5









#3
Pyridine
Concen: 22.353 ng
RT: 3.734 min Scan# 110
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 399328
Ion Ratio Lower Upper
79 100
52 53.8 41.7 62.5
51 27.1 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 22.720 ng
RT: 3.646 min Scan# 95
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 42 Resp: 137632
Ion Ratio Lower Upper
42 100
74 157.2 105.4 158.0
44 8.4 4.2 6.4#









#5
2-Fluorophenol
Concen: 43.325 ng
RT: 5.393 min Scan# 392
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 664932
Ion Ratio Lower Upper
112 100
64 56.9 41.4 62.2
63 29.0 23.8 35.8



#6
Aniline
Concen: 21.020 ng
RT: 7.134 min Scan# 688
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 93 Resp: 488299
Ion Ratio Lower Upper
93 100
66 39.1 29.7 44.5
65 19.9 15.3 22.9









#7
Phenol-d6
Concen: 43.235 ng
RT: 6.969 min Scan# 660
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 99 Resp: 907217
Ion Ratio Lower Upper
99 100
42 15.0 13.4 20.0
71 34.6 30.6 46.0



#8
2-Chlorophenol
Concen: 21.246 ng
RT: 7.363 min Scan# 727
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 128 Resp: 360049
Ion Ratio Lower Upper
128 100
130 31.7 12.4 52.4
64 46.3 20.3 60.3









#9
Benzaldehyde
Concen: 23.059 ng
RT: 6.946 min Scan# 656
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 278583
Ion Ratio Lower Upper
77 100
105 93.9 76.7 116.7
106 87.8 76.2 116.2



#10
Phenol
Concen: 21.691 ng
RT: 6.999 min Scan# 665
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37


Tgt Ion: 94 Resp: 451556
Ion Ratio Lower Upper
94 100
65 28.0 8.8 48.8
66 35.9 20.7 60.7








#11
bis(2-Chloroethyl)ether
Concen: 21.520 ng
RT: 7.234 min Scan# 705
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Instrument :
BNA_P
ClientSampleId :
SSTDIC020







Tgt Ion: 93 Resp: 346974
Ion Ratio Lower Upper
93 100
63 74.0 46.2 86.2
95 32.7 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 20.855 ng
RT: 7.687 min Scan# 782
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:146 Resp: 387211
Ion Ratio Lower Upper
146 100
148 63.0 51.5 77.3
75 33.6 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 20.871 ng
RT: 7.834 min Scan# 807
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:146 Resp: 392844
Ion Ratio Lower Upper
146 100
148 64.1 51.3 76.9
111 42.5 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 20.813 ng
RT: 8.146 min Scan# 860
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:146 Resp: 375234
Ion Ratio Lower Upper
146 100
148 63.3 50.3 75.5
111 45.7 33.7 50.5









#15
Benzyl Alcohol
Concen: 21.049 ng
RT: 8.040 min Scan# 842
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 317970
Ion Ratio Lower Upper
79 100
108 72.3 62.5 93.7
77 63.0 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 23.554 ng
RT: 8.328 min Scan# 891
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 45 Resp: 382606
Ion Ratio Lower Upper
45 100
77 18.6 2.0 42.0
79 14.2 0.0 36.7









#17
2-Methylphenol
Concen: 20.791 ng
RT: 8.240 min Scan# 876
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	287072
Ion	Ratio	Lower	Upper
107	100		
108	109.3	87.7	131.5
77	47.7	40.2	60.2
79	46.4	39.0	58.4



#18
Hexachloroethane
Concen: 21.103 ng
RT: 8.869 min Scan# 983
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:	117	Resp:	141210
Ion	Ratio	Lower	Upper
117	100		
119	94.3	76.2	114.2
201	87.6	83.4	125.0









#19

n-Nitroso-di-n-propylamine
 Concen: 21.487 ng
 RT: 8.610 min Scan# 939
 Delta R.T. -0.006 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion: 70 Resp: 276191

Ion	Ratio	Lower	Upper
70	100		
42	44.4	40.8	61.2
101	9.1	8.8	13.2
130	20.1	18.9	28.3







#20

3+4-Methylphenols
 Concen: 20.748 ng
 RT: 8.569 min Scan# 932
 Delta R.T. -0.006 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37

Tgt Ion: 107 Resp: 385866

Ion	Ratio	Lower	Upper
107	100		
108	89.7	71.0	111.0
77	34.9	16.6	56.6
79	30.0	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.581 min Scan# 1274
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:136 Resp: 1001591




Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	7.2	4.6	6.8#
68	7.0	3.6	5.4#




#22
Acetophenone
Concen: 21.297 ng
RT: 8.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:105 Resp: 514818

Ion	Ratio	Lower	Upper
105	100		
71	3.9	1.9	2.9#
51	21.6	18.4	27.6
120	21.8	17.9	26.9









#23
Nitrobenzene-d5
Concen: 46.975 ng
RT: 8.952 min Scan# 997
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 781345
Ion Ratio Lower Upper
82 100
128 39.9 35.7 53.5
54 41.0 35.0 52.4



#24
Nitrobenzene
Concen: 23.288 ng
RT: 8.993 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 77 Resp: 411825
Ion Ratio Lower Upper
77 100
123 42.3 38.2 57.2
65 12.6 11.4 17.2









#25
Isophorone
Concen: 20.666 ng
RT: 9.522 min Scan# 1094
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 701277
Ion Ratio Lower Upper
82 100
95 6.9 6.5 9.7
138 16.3 15.0 22.4



#26
2-Nitrophenol
Concen: 22.108 ng
RT: 9.704 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 139 Resp: 146845
Ion Ratio Lower Upper
139 100
109 30.8 28.8 43.2
65 50.0 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 21.045 ng
RT: 9.769 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 282551
Ion Ratio Lower Upper
122 100
107 118.5 94.5 141.7
121 59.7 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 21.397 ng
RT: 10.010 min Scan# 1177
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 93 Resp: 409877
Ion Ratio Lower Upper
93 100
95 31.7 25.9 38.9
123 11.3 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 20.467 ng
RT: 10.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 283665
Ion Ratio Lower Upper
162 100
164 64.5 44.9 84.9
98 41.8 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 20.363 ng
RT: 10.446 min Scan# 1251
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:180 Resp: 319102
Ion Ratio Lower Upper
180 100
182 96.0 77.9 116.9
145 32.8 24.7 37.1









#31
Naphthalene
Concen: 21.085 ng
RT: 10.634 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:128 Resp: 1108494
Ion Ratio Lower Upper
128 100
129 11.2 9.0 13.6
127 13.1 10.7 16.1



#32
Benzoic acid
Concen: 20.533 ng
RT: 9.863 min Scan# 1152
Delta R.T. -0.047 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:122 Resp: 171067
Ion Ratio Lower Upper
122 100
105 127.2 103.2 143.2
77 100.9 76.2 116.2









#33
4-Chloroaniline
Concen: 20.283 ng
RT: 10.746 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:127	Resp: 434045
Ion Ratio	Lower Upper
127 100	
129 31.4	26.1 39.1
65 30.8	22.5 33.7
92 19.6	15.0 22.6



#34
Hexachlorobutadiene
Concen: 20.100 ng
RT: 10.922 min Scan# 1332
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:225	Resp: 186737
Ion Ratio	Lower Upper
225 100	
223 63.6	51.0 76.6
227 64.8	51.4 77.2









#35
Caprolactam
Concen: 21.045 ng
RT: 11.528 min Scan# 1435
Delta R.T. -0.024 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:113 Resp: 87772
Ion Ratio Lower Upper
113 100
55 141.8 98.9 138.9#
56 103.1 71.7 111.7



#36
4-Chloro-3-methylphenol
Concen: 21.021 ng
RT: 11.869 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:107 Resp: 339092
Ion Ratio Lower Upper
107 100
144 24.4 21.1 31.7
142 77.1 66.0 99.0









#37
2-Methylnaphthalene
Concen: 20.374 ng
RT: 12.245 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 748004
Ion Ratio Lower Upper
142 100
141 91.1 71.8 107.6
115 36.4 31.6 47.4



#38
1-Methylnaphthalene
Concen: 20.381 ng
RT: 12.463 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:142 Resp: 707672
Ion Ratio Lower Upper
142 100
141 92.8 73.3 109.9
116 4.2 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.422 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:164 Resp: 566309
Ion Ratio Lower Upper
164 100
162 102.4 80.2 120.2
160 45.1 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.828 ng
RT: 12.610 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:216 Resp: 319477
Ion Ratio Lower Upper
216 100
214 78.2 61.9 92.9
179 23.0 16.8 25.2
108 22.5 14.9 22.3#









#41
Hexachlorocyclopentadiene
Concen: 20.477 ng
RT: 12.593 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	237	Resp:	171968
Ion	Ratio	Lower	Upper
237	100		
235	63.0	44.5	84.5
272	12.8	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 34.339 ng
RT: 15.916 min Scan# 2181
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:	330	Resp:	200170
Ion	Ratio	Lower	Upper
330	100		
332	95.4	77.8	116.8
141	41.6	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 21.530 ng
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 207138
Ion Ratio Lower Upper
196 100
198 95.2 79.5 119.3
200 30.9 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 22.161 ng
RT: 12.922 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:196 Resp: 229072
Ion Ratio Lower Upper
196 100
198 96.0 74.7 112.1
97 57.6 37.4 56.0#
132 30.7 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 43.211 ng
RT: 13.051 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:172 Resp: 1607680
Ion Ratio Lower Upper
172 100
171 35.5 27.9 41.9
170 23.2 18.6 27.8



#46
1,1'-Biphenyl
Concen: 21.418 ng
RT: 13.263 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:154 Resp: 894021
Ion Ratio Lower Upper
154 100
153 41.4 20.4 60.4
76 14.8 0.0 30.8









#47
2-Chloronaphthalene
Concen: 21.368 ng
RT: 13.298 min Scan# 1736
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 686718
Ion Ratio Lower Upper
162 100
127 42.3 31.4 47.0
164 32.9 26.6 39.8



#48
2-Nitroaniline
Concen: 24.778 ng
RT: 13.504 min Scan# 1771
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion: 65 Resp: 187132
Ion Ratio Lower Upper
65 100
92 72.5 57.7 86.5
138 103.3 92.1 138.1









#49
Acenaphthylene
Concen: 21.205 ng
RT: 14.145 min Scan# 1880
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:152 Resp: 1084874
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 13.1 10.6 16.0



#50
Dimethylphthalate
Concen: 20.990 ng
RT: 13.892 min Scan# 1837
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:163 Resp: 834053
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 22.680 ng
RT: 14.004 min Scan# 1856
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 176881
Ion Ratio Lower Upper
165 100
63 60.7 41.3 61.9
89 65.9 44.8 67.2



#52
Acenaphthene
Concen: 21.159 ng
RT: 14.487 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:154 Resp: 669201
Ion Ratio Lower Upper
154 100
153 114.9 89.4 134.2
152 54.1 43.0 64.4









#53
3-Nitroaniline
Concen: 22.708 ng
RT: 14.334 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:138 Resp: 187627
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 121.0 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 23.715 ng
RT: 14.534 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:184 Resp: 78457
Ion Ratio Lower Upper
184 100
63 70.8 47.0 70.6#
154 65.2 52.3 78.5









#55
Dibenzenofuran
Concen: 21.144 ng
RT: 14.822 min Scan# 1995
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 1055833
Ion Ratio Lower Upper
168 100
139 41.0 31.0 46.6
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 21.532 ng
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:139 Resp: 157862
Ion Ratio Lower Upper
139 100
109 76.5 72.3 112.3
65 99.3 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 22.525 ng
RT: 14.792 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	165	Resp:	226641
Ion Ratio	Lower	Upper	
165	100		
63	44.9	32.3	48.5
89	77.3	55.9	83.9
182	3.2	2.2	3.4



#58
Fluorene
Concen: 20.830 ng
RT: 15.475 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:	166	Resp:	833218
Ion Ratio	Lower	Upper	
166	100		
165	97.2	78.0	117.0
167	13.5	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.928 ng
RT: 15.045 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:232 Resp: 187311
Ion Ratio Lower Upper
232 100
131 55.2 32.2 48.4#
130 3.0 2.0 3.0#
166 34.4 19.4 29.2#



#60
Diethylphthalate
Concen: 20.838 ng
RT: 15.263 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:149 Resp: 854626
Ion Ratio Lower Upper
149 100
177 20.7 17.5 26.3
150 12.8 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 20.605 ng
RT: 15.475 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	204	Resp:	389853
Ion Ratio	Lower	Upper	
204	100		
206	32.7	26.6	40.0
141	69.8	48.6	73.0



#62
4-Nitroaniline
Concen: 22.604 ng
RT: 15.492 min Scan# 2109
Delta R.T. -0.012 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:	138	Resp:	195081
Ion Ratio	Lower	Upper	
138	100		
92	57.8	31.2	71.2
108	80.7	78.6	118.6









#63
Azobenzene
Concen: 22.265 ng
RT: 15.769 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	77	Resp:	919450
Ion	Ratio	Lower	Upper
77	100		
182	21.5	6.0	46.0
105	20.1	0.0	38.8
51	23.3	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.175 min Scan# 2395
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:	188	Resp:	1120674
Ion	Ratio	Lower	Upper
188	100		
94	11.4	5.6	8.4#
80	13.2	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 21.968 ng
RT: 15.551 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:198 Resp: 114538
Ion Ratio Lower Upper
198 100
51 42.1 14.0 54.0
105 51.5 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 21.235 ng
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:169 Resp: 730146
Ion Ratio Lower Upper
169 100
168 66.8 54.4 81.6
167 35.8 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 17.524 ng
RT: 16.375 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 197860
Ion Ratio Lower Upper
248 100
250 93.8 79.8 119.8
141 82.6 57.8 86.6



#68
Hexachlorobenzene
Concen: 19.805 ng
RT: 16.475 min Scan# 2276
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:284 Resp: 256407
Ion Ratio Lower Upper
284 100
142 39.6 26.9 40.3
249 31.5 24.5 36.7





#69

Atrazine

Concen: 20.440 ng

RT: 16.651 min Scan# 2306

Delta R.T. -0.006 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37

Tgt Ion:200 Resp: 229555

Ion Ratio Lower Upper

200 100

173 28.0 4.1 44.1





215 45.1 27.7 67.7

Instrument :

BNA_P

ClientSampleId :

SSTDIC020



#70

Pentachlorophenol

Concen: 20.591 ng

RT: 16.822 min Scan# 2335

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37




Tgt Ion:266 Resp: 149297


Ion Ratio Lower Upper

266 100

268 63.3 51.6 77.4

264 61.4 50.3 75.5









#71
Phenanthrene
Concen: 21.093 ng
RT: 17.216 min Scan# 2402
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 1294756
Ion Ratio Lower Upper
178 100
176 19.3 15.1 22.7
179 15.6 12.4 18.6



#72
Anthracene
Concen: 20.994 ng
RT: 17.310 min Scan# 2418
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:178 Resp: 1256982
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.2 12.2 18.2









#73
 Carbazole
 Concen: 21.160 ng
 RT: 17.581 min Scan# 2464
 Delta R.T. -0.006 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:167 Resp: 1269984
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 14.0 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 20.561 ng
 RT: 18.157 min Scan# 2562
 Delta R.T. -0.000 min
 Lab File: BP006275.D
 Acq: 22 Jul 2021 11:37

Tgt Ion:149 Resp: 1408537
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 5.3 4.5 6.7









#75
Fluoranthene
Concen: 20.788 ng
RT: 19.192 min Scan# 2738
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:202 Resp: 1464401
Ion Ratio Lower Upper
202 100
101 13.5 0.0 27.6
203 17.3 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.274 min Scan# 3092
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:240 Resp: 1146439
Ion Ratio Lower Upper
240 100
120 13.1 6.5 9.7#
236 25.8 20.6 31.0





#77

Benzidine

Concen: 16.153 ng

RT: 19.380 min Scan# 2770

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37

Tgt Ion:184 Resp: 538968

Ion Ratio Lower Upper

184 100

185 14.0 11.0 16.4





183 11.7 9.5 14.3

Instrument :

BNA_P

ClientSampleId :

SSTDIC020



#78

Pyrene

Concen: 20.322 ng

RT: 19.545 min Scan# 2798

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37




Tgt Ion:202 Resp: 1538389


Ion Ratio Lower Upper

202 100

200 21.1 16.3 24.5

203 17.4 14.2 21.2





#79

Terphenyl-d14

Concen: 41.037 ng

RT: 19.751 min Scan# 2833

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37

Instrument :

BNA_P

ClientSampleId :

SSTDIC020





Tgt Ion:244 Resp: 2351840

Ion Ratio Lower Upper

244 100

212 7.3 6.0 9.0

122 12.9 7.1 10.7#



#80

Butylbenzylphthalate

Concen: 19.608 ng

RT: 20.439 min Scan# 2950

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37




Tgt Ion:149 Resp: 628825


Ion Ratio Lower Upper

149 100

91 76.5 54.6 81.8

206 19.1 16.7 25.1









#81
Benzo(a)anthracene
Concen: 20.992 ng
RT: 21.257 min Scan# 3089
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 1544308
Ion Ratio Lower Upper
228 100
226 26.7 21.6 32.4
229 19.7 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 19.456 ng
RT: 21.198 min Scan# 3079
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:252 Resp: 396343
Ion Ratio Lower Upper
252 100
254 65.2 50.6 76.0
126 13.9 7.0 10.6#









#83
Chrysene
Concen: 20.895 ng
RT: 21.310 min Scan# 3098
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 1488340
Ion Ratio Lower Upper
228 100
226 30.1 23.8 35.6
229 19.9 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 20.392 ng
RT: 21.210 min Scan# 3081
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:149 Resp: 985686
Ion Ratio Lower Upper
149 100
167 25.9 21.9 32.9
279 4.2 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 20.347 ng
RT: 22.121 min Scan# 3236
Delta R.T. -0.000 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 1649974
Ion Ratio Lower Upper
149 100
167 1.4 1.4 2.0
43 8.5 5.6 8.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.621 min Scan# 3491
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:264 Resp: 1214273
Ion Ratio Lower Upper
264 100
260 24.0 18.6 28.0
265 22.3 17.6 26.4





#87

Indeno(1,2,3-cd)pyrene

Concen: 21.190 ng

RT: 26.057 min Scan# 3905

Delta R.T. -0.012 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37

Instrument :

BNA_P

ClientSampleId :

SSTDIC020





Tgt Ion:276 Resp: 1822892

Ion Ratio Lower Upper

276 100

138 30.8 15.3 22.9#

277 25.3 20.4 30.6



#88

Benzo(b)fluoranthene

Concen: 20.654 ng

RT: 22.910 min Scan# 3370

Delta R.T. -0.000 min

Lab File: BP006275.D

Acq: 22 Jul 2021 11:37




Tgt Ion:252 Resp: 1618226


Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 12.7 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 20.773 ng
RT: 22.957 min Scan# 3378
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:252 Resp: 1515443
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 12.3 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 20.441 ng
RT: 23.515 min Scan# 3473
Delta R.T. -0.006 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:252 Resp: 1436964
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 14.9 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 21.085 ng
RT: 26.080 min Scan# 3909
Delta R.T. -0.012 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 1568959
Ion Ratio Lower Upper
278 100
139 19.6 9.3 13.9#
279 24.1 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 20.858 ng
RT: 26.804 min Scan# 4032
Delta R.T. -0.012 min
Lab File: BP006275.D
Acq: 22 Jul 2021 11:37

Tgt Ion:276 Resp: 1510738
Ion Ratio Lower Upper
276 100
277 23.2 18.9 28.3
138 25.6 12.6 18.8#

