

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072425\
 Data File : BP025236.D
 Acq On : 24 Jul 2025 10:36
 Operator : CG/JU
 Sample : PB168969TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB168969TB

Quant Time: Jul 24 10:56:43 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 22 02:55:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.425	152	661066	20.000	ng	0.00
21) Naphthalene-d8	10.172	136	2473242	20.000	ng	0.00
39) Acenaphthene-d10	14.078	164	1495565	20.000	ng	0.00
64) Phenanthrene-d10	16.907	188	2901132	20.000	ng	0.01
76) Chrysene-d12	21.331	240	2893957	20.000	ng	0.00
86) Perylene-d12	24.466	264	3251982	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.102	112	4659586	119.096	ng	0.00
7) Phenol-d6	6.643	99	5503246	113.569	ng	0.00
23) Nitrobenzene-d5	8.561	82	3468795	77.943	ng	0.00
42) 2,4,6-Tribromophenol	15.607	330	2597658	117.819	ng	0.00
45) 2-Fluorobiphenyl	12.672	172	8324571	78.473	ng	-0.01
79) Terphenyl-d14	19.660	244	11590337	79.060	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072425\
Data File : BP025236.D
Acq On : 24 Jul 2025 10:36
Operator : CG/JU
Sample : PB168969TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB168969TB

Quant Time: Jul 24 10:56:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 22 02:55:21 2025
Response via : Initial Calibration

