

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072522\
 Data File : BP011098.D
 Acq On : 25 Jul 2022 18:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020713

Quant Time: Jul 25 23:22:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	493036	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.457	136	2300305	20.000	ng/u1	#-0.01
38) Acenaphthene-d10	14.328	164	1382586	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.098	188	2704697	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.216	240	1974235	20.000	ng/u1	#-0.01
88) Perylene-d12	23.551	264	1520499	20.000	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	106917	7.818	ng/uL	-0.02
4) Pyridine-d5	3.570	84	736682	20.840	ng/u1	-0.02
7) Phenol-d5	6.846	99	883876	21.839	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.011	67	612894	23.551	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.199	132	735541	22.492	ng/u1	-0.01
15) 4-Methylphenol-d8	8.387	113	774178	23.894	ng/u1	-0.01
21) Nitrobenzene-d5	8.828	128	386940	23.082	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.546	143	397967	23.992	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.087	165	672794	22.179	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.604	131	1143219	22.860	ng/u1	-0.01
46) Dimethylphthalate-d6	13.751	166	2126960	22.509	ng/u1	-0.01
49) Acenaphthylene-d8	14.016	160	2544523	22.822	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.551	143	384356	20.098	ng/u1	0.00
60) Fluorene-d10	15.328	176	1745910	21.737	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.457	200	298025	22.268	ng/u1	0.00
73) Anthracene-d10	17.198	188	2555766	22.227	ng/u1	0.00
81) Pyrene-d10	19.451	212	2679598	25.843	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.404	264	1612173	22.076	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	110489	7.700	ng/uL#	87
5) Pyridine	3.593	79	772489	21.205	ng/u1#	76
6) Benzaldehyde	6.816	77	162639	8.273	ng/u1	96
8) Phenol	6.869	94	948800	22.055	ng/u1	88
10) Bis(2-Chloroethyl)ether	7.099	93	781243	22.882	ng/u1	92
12) 2-Chlorophenol	7.228	128	759256	22.461	ng/u1	95
13) 2-Methylphenol	8.116	108	738565	23.215	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.199	45	1172125	25.100	ng/u1	98
16) Acetophenone	8.493	105	1203738	24.542	ng/u1#	92
17) N-Nitroso-di-n-propyla...	8.487	70	657424	26.714	ng/u1	93
18) 4-Methylphenol	8.446	108	811523	23.891	ng/u1	97
19) Hexachloroethane	8.734	117	305811	22.860	ng/u1#	77
22) Nitrobenzene	8.869	77	906245	22.988	ng/u1	93
23) Isophorone	9.405	82	1818243	24.644	ng/u1	98
25) 2-Nitrophenol	9.581	139	438383	23.545	ng/u1#	83
26) 2,4-Dimethylphenol	9.646	107	910158	23.116	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.887	93	1110915	22.752	ng/u1	99
29) 2,4-Dichlorophenol	10.110	162	686794	21.874	ng/u1	99
30) Naphthalene	10.510	128	2539105	21.661	ng/u1	100
32) 4-Chloroaniline	10.628	127	1144086	23.129	ng/u1	100
33) Hexachlorobutadiene	10.793	225	385487	21.092	ng/u1	95
34) Caprolactam	11.428	113	153930	13.989	ng/u1	98
35) 4-Chloro-3-methylphenol	11.769	107	853003	24.039	ng/u1	96
36) 2-Methylnaphthalene	12.134	142	1703767	22.420	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072522\
 Data File : BP011098.D
 Acq On : 25 Jul 2022 18:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020713


Quant Time: Jul 25 23:22:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.351	142	1727888	22.684	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.504	216	756026	20.733	ng/ul#	98
40) Hexachlorocyclopentadiene	12.481	237	472903	20.939	ng/ul	96
41) 2,4,6-Trichlorophenol	12.751	196	525571	21.886	ng/ul	100
42) 2,4,5-Trichlorophenol	12.822	196	566333	22.067	ng/ul	97
43) 1,1'-Biphenyl	13.157	154	2234285	21.658	ng/ul#	96
44) 2-Chloronaphthalene	13.198	162	1670988	21.565	ng/ul	97
47) Dimethylphthalate	13.798	163	2127633	22.314	ng/ul	98
48) 2,6-Dinitrotoluene	13.916	165	429813	24.899	ng/ul	100
50) Acenaphthylene	14.045	152	2807715	22.443	ng/ul	99
51) 3-Nitroaniline	14.245	138	304428	14.678	ng/ul	92
52) Acenaphthene	14.393	153	1815550	21.907	ng/ul	98
53) 2,4-Dinitrophenol	14.457	184	207334	20.529	ng/ul	88
55) 4-Nitrophenol	14.563	109	301777	20.610	ng/ul#	77
56) Dibenzofuran	14.734	168	2479253	21.680	ng/ul	97
57) 2,4-Dinitrotoluene	14.710	165	610849	24.808	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	14.963	232	448624	22.322	ng/ul#	87
59) Diethylphthalate	15.175	149	2237644	22.852	ng/ul	98
61) Fluorene	15.387	166	2005140	21.848	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.387	204	903202	21.377	ng/ul	92
63) 4-Nitroaniline	15.416	138	209532	10.993	ng/ul#	79
66) 4,6-Dinitro-2-methylph...	15.475	198	312875	22.788	ng/ul#	99
67) N-Nitrosodiphenylamine	15.604	169	1731245	22.408	ng/ul	98
68) 4-Bromophenyl-phenylether	16.287	248	542815	22.252	ng/ul	93
69) Hexachlorobenzene	16.392	284	611715	21.607	ng/ul	94
70) Atrazine	16.575	200	606778	22.703	ng/ul	97
71) Pentachlorophenol	16.745	266	363777	20.863	ng/ul	98
72) Phenanthrene	17.139	178	3055588	21.684	ng/ul	98
74) Anthracene	17.234	178	3080279	21.975	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.116	216	795940	20.557	ng/uL	98
76) Pentachlorobenzene	14.645	250	736936	21.712	ng/uL	99
77) Carbazole	17.510	167	2796541	21.317	ng/ul	99
78) Di-n-butylphthalate	18.075	149	3815552	23.267	ng/ul	99
80) Fluoranthene	19.122	202	3334838	27.499	ng/ul#	85
82) Pyrene	19.480	202	3309406	26.450	ng/ul#	83
83) Butylbenzylphthalate	20.369	149	1595709	28.589	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.139	252	726884	20.793	ng/ul#	99
85) Benzo(a)anthracene	21.198	228	2665432	22.169	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.139	149	2306223	27.692	ng/ul#	96
87) Chrysene	21.251	228	2547747	21.581	ng/ul	100
89) Di-n-octyl phthalate	22.045	149	3520422	34.523	ng/ul	100
90) Benzo(b)fluoranthene	22.839	252	2156870	22.974	ng/ul#	95
91) Benzo(k)fluoranthene	22.886	252	2063028	23.394	ng/ul#	94
93) Benzo(a)pyrene	23.451	252	1976973	21.927	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	25.974	276	2039286	19.462	ng/ul#	88
95) Dibenzo(a,h)anthracene	25.992	278	1750085	19.302	ng/ul#	91
96) Benzo(g,h,i)perylene	26.721	276	1739638	19.494	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 23:22:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 19 03:25:54 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.700 ng/uL
RT: 3.193 min Scan# 11
Delta R.T. -0.023 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion: 88 Resp: 110489
Ion Ratio Lower Upper
88 100
43 31.9 35.8 53.6#
58 76.3 68.1 102.1



#5
Pyridine
Concen: 21.205 ng/uL
RT: 3.593 min Scan# 86
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 79 Resp: 772489
Ion Ratio Lower Upper
79 100
52 78.9 83.8 125.8#
51 37.0 41.0 61.4#









#6
Benzaldehyde
Concen: 8.273 ng/ul
RT: 6.816 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion: 77 Resp: 162639
Ion Ratio Lower Upper
77 100
105 99.5 76.0 114.0
106 94.4 73.2 109.8



#8
Phenol
Concen: 22.055 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 94 Resp: 948800
Ion Ratio Lower Upper
94 100
65 29.2 27.4 41.2
66 37.5 37.4 56.2









#10
Bis(2-Chloroethyl)ether
Concen: 22.882 ng/ul
RT: 7.099 min Scan# 61
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion: 93 Resp: 781243
Ion Ratio Lower Upper
93 100
63 80.6 72.4 108.6
95 32.4 27.3 40.9



#12
2-Chlorophenol
Concen: 22.461 ng/ul
RT: 7.228 min Scan# 704
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 128 Resp: 759256
Ion Ratio Lower Upper
128 100
64 50.4 44.8 67.2
130 32.5 25.8 38.6









#13
2-Methylphenol
Concen: 23.215 ng/ul
RT: 8.116 min Scan# 81
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:108 Resp: 738565
Ion Ratio Lower Upper
108 100
107 90.8 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.100 ng/ul
RT: 8.199 min Scan# 869
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 45 Resp: 1172125
Ion Ratio Lower Upper
45 100
77 13.6 11.2 16.8
79 10.2 6.9 10.3









#16
Acetophenone
Concen: 24.542 ng/ul
RT: 8.493 min Scan# 919
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:105 Resp: 1203738
Ion Ratio Lower Upper
105 100
77 78.9 67.5 101.3
51 29.4 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 26.714 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 70 Resp: 657424
Ion Ratio Lower Upper
70 100
42 60.1 53.8 80.8
101 9.3 7.8 11.6
130 21.4 15.4 23.2









#18
4-Methylphenol
Concen: 23.891 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:108 Resp: 811523
Ion Ratio Lower Upper
108 100
107 113.8 88.3 132.5



#19
Hexachloroethane
Concen: 22.860 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:117 Resp: 305811
Ion Ratio Lower Upper
117 100
201 83.8 85.7 128.5#
199 52.3 57.4 86.0#









#22
Nitrobenzene
Concen: 22.988 ng/ul
RT: 8.869 min Scan# 90
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion: 77 Resp: 906245
Ion Ratio Lower Upper
77 100
123 44.6 31.5 47.3
65 13.5 11.4 17.2



#23
Isophorone
Concen: 24.644 ng/ul
RT: 9.405 min Scan# 1074
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion: 82 Resp: 1818243
Ion Ratio Lower Upper
82 100
95 7.0 6.6 10.0
138 17.5 13.8 20.6









#25
2-Nitrophenol
Concen: 23.545 ng/ul
RT: 9.581 min Scan# 1104
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:139 Resp: 438383
Ion Ratio Lower Upper
139 100
65 53.4 54.2 81.4#
109 28.6 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 23.116 ng/ul
RT: 9.646 min Scan# 1115
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:107 Resp: 910158
Ion Ratio Lower Upper
107 100
121 50.0 39.1 58.7
122 84.3 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 22.752 ng/ul

RT: 9.887 min Scan# 1110915

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :





BNA_P

ClientSampleId :

SSTD020713

Tgt Ion: 93 Resp: 1110915

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.6	39.8
123	11.2	7.8	11.8



#29

2,4-Dichlorophenol

Concen: 21.874 ng/ul

RT: 10.110 min Scan# 1194




Delta R.T. -0.012 min


Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Tgt Ion:162 Resp: 686794

Ion	Ratio	Lower	Upper
162	100		
164	64.9	52.5	78.7
98	40.5	31.4	47.0









#30
Naphthalene
Concen: 21.661 ng/ul
RT: 10.510 min Scan# 1263
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:128 Resp: 2539105
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 23.129 ng/ul
RT: 10.628 min Scan# 1282
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:127 Resp: 1144086
Ion Ratio Lower Upper
127 100
129 32.8 26.2 39.2









#33
Hexachlorobutadiene
Concen: 21.092 ng/ul
RT: 10.793 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:225 Resp: 385487
Ion Ratio Lower Upper
225 100
223 61.7 53.7 80.5
227 64.1 49.4 74.2



#34
Caprolactam
Concen: 13.989 ng/ul
RT: 11.428 min Scan# 1418
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:113 Resp: 153930
Ion Ratio Lower Upper
113 100
55 187.7 153.8 230.8
56 138.0 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 24.039 ng/ul
RT: 11.769 min Scan# 1477
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:107 Resp: 853003
Ion Ratio Lower Upper
107 100
144 25.5 18.6 27.8
142 77.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 22.420 ng/ul
RT: 12.134 min Scan# 1538
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:142 Resp: 1703767
Ion Ratio Lower Upper
142 100
141 88.8 70.1 105.1









#37
1-Methylnaphthalene
Concen: 22.684 ng/ul
RT: 12.351 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:142 Resp: 1727888
Ion Ratio Lower Upper
142 100
141 92.9 73.3 109.9
116 4.1 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.733 ng/ul
RT: 12.504 min Scan# 1601
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:216 Resp: 756026
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.6
179 23.3 18.1 27.1
108 20.9 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 20.939 ng/ul
RT: 12.481 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:237 Resp: 472903
Ion Ratio Lower Upper
237 100
235 63.0 52.8 79.2
272 12.2 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 21.886 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:196 Resp: 525571
Ion Ratio Lower Upper
196 100
198 97.7 78.2 117.4
200 30.9 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 22.067 ng/ul
RT: 12.822 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:196 Resp: 566333
Ion Ratio Lower Upper
196 100
198 93.1 77.0 115.4
200 29.8 24.8 37.2



#43
1,1'-Biphenyl
Concen: 21.658 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:154 Resp: 2234285
Ion Ratio Lower Upper
154 100
153 41.0 31.9 47.9
76 15.2 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 21.565 ng/ul

RT: 13.198 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:162 Resp: 1670988

Ion Ratio Lower Upper

162 100

164 33.1 25.5 38.3

127 42.4 31.9 47.9



#47

Dimethylphthalate

Concen: 22.314 ng/ul

RT: 13.798 min Scan# 1821

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:163 Resp: 2127633


Ion Ratio Lower Upper

163 100

194 3.7 3.4 5.2

164 10.3 7.6 11.4









#48
2,6-Dinitrotoluene
Concen: 24.899 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:165 Resp: 429813
Ion Ratio Lower Upper
165 100
89 65.9 52.9 79.3
121 23.9 19.4 29.2



#50
Acenaphthylene
Concen: 22.443 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:152 Resp: 2807715
Ion Ratio Lower Upper
152 100
151 20.3 16.5 24.7
153 13.3 10.1 15.1









#51
3-Nitroaniline
Concen: 14.678 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:138 Resp: 304428
Ion Ratio Lower Upper
138 100
108 12.1 11.0 16.4
92 120.9 104.6 156.8



#52
Acenaphthene
Concen: 21.907 ng/ul
RT: 14.393 min Scan# 1922
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:153 Resp: 1815550
Ion Ratio Lower Upper
153 100
152 46.9 36.2 54.2
154 87.3 70.8 106.2









#53
2,4-Dinitrophenol
Concen: 20.529 ng/ul
RT: 14.457 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:184 Resp: 207334
Ion Ratio Lower Upper
184 100
63 76.6 73.8 110.6
154 65.9 57.2 85.8



#55
4-Nitrophenol
Concen: 20.610 ng/ul
RT: 14.563 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:109 Resp: 301777
Ion Ratio Lower Upper
109 100
139 137.6 75.6 113.4#
65 143.9 107.6 161.4









#56
Dibenzofuran
Concen: 21.680 ng/ul
RT: 14.734 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:168 Resp: 2479253
Ion Ratio Lower Upper
168 100
139 39.4 32.8 49.2



#57
2,4-Dinitrotoluene
Concen: 24.808 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:165 Resp: 610849
Ion Ratio Lower Upper
165 100
63 49.8 47.0 70.6
182 2.7 2.2 3.2









#58
2,3,4,6-Tetrachlorophenol
Concen: 22.322 ng/ul
RT: 14.963 min Scan# 20
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:232 Resp: 448624
Ion Ratio Lower Upper
232 100
131 54.8 35.8 53.8#
130 2.8 1.8 2.8#
166 35.5 23.7 35.5



#59
Diethylphthalate
Concen: 22.852 ng/ul
RT: 15.175 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:149 Resp: 2237644
Ion Ratio Lower Upper
149 100
177 20.4 17.6 26.4
150 12.5 10.3 15.5









#61
Fluorene
Concen: 21.848 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:166 Resp: 2005140
Ion Ratio Lower Upper
166 100
165 99.0 77.4 116.2
167 13.2 10.8 16.2



#62
4-Chlorophenyl-phenylether
Concen: 21.377 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:204 Resp: 903202
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 71.1 49.3 73.9









#63
4-Nitroaniline
Concen: 10.993 ng/ul
RT: 15.416 min Scan# 2095
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:138 Resp: 209532
Ion Ratio Lower Upper
138 100
92 54.5 46.2 69.2
108 76.8 85.8 128.8#



#66
4,6-Dinitro-2-methylphenol
Concen: 22.788 ng/ul
RT: 15.475 min Scan# 2106
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:198 Resp: 312875
Ion Ratio Lower Upper
198 100
182 4.1 4.3 6.5#
77 29.3 23.7 35.5





#67

N-Nitrosodiphenylamine

Concen: 22.408 ng/ul

RT: 15.604 min Scan# 2129

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:169 Resp: 1731245

Ion Ratio Lower Upper

169 100

168 66.7 52.6 78.8

167 36.4 27.8 41.6



#68

4-Bromophenyl-phenylether

Concen: 22.252 ng/ul

RT: 16.287 min Scan# 2244

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:248 Resp: 542815


Ion Ratio Lower Upper

248 100

250 97.2 80.8 121.2

141 88.8 63.9 95.9









#69
Hexachlorobenzene
Concen: 21.607 ng/ul
RT: 16.392 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:284 Resp: 611715
Ion Ratio Lower Upper
284 100
142 40.9 27.8 41.6
249 32.0 25.4 38.2



#70
Atrazine
Concen: 22.703 ng/ul
RT: 16.575 min Scan# 2293
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:200 Resp: 606778
Ion Ratio Lower Upper
200 100
173 27.5 20.7 31.1
215 46.4 38.6 57.8





#71

Pentachlorophenol

Concen: 20.863 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:266 Resp: 363777

Ion Ratio Lower Upper

266 100

264 64.0 51.4 77.2

268 64.1 53.5 80.3



#72

Phenanthrene

Concen: 21.684 ng/ul

RT: 17.139 min Scan# 2389

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:178 Resp: 3055588


Ion Ratio Lower Upper

178 100

179 15.3 12.7 19.1

176 19.3 14.6 22.0









#74
 Anthracene
 Concen: 21.975 ng/uI
 RT: 17.234 min Scan# 2406
 Delta R.T. -0.006 min
 Lab File: BP011098.D
 Acq: 25 Jul 2022 18:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020713


Tgt Ion:178 Resp: 3080279
 Ion Ratio Lower Upper
 178 100
 179 15.3 12.2 18.2
 176 18.4 14.5 21.7



#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 20.557 ng/uL
 RT: 13.116 min Scan# 1705
 Delta R.T. -0.012 min
 Lab File: BP011098.D
 Acq: 25 Jul 2022 18:29

Tgt Ion:216 Resp: 795940
 Ion Ratio Lower Upper
 216 100
 214 78.6 65.0 97.4
 218 48.2 39.1 58.7









#76
 Pentachlorobenzene
 Concen: 21.712 ng/uL
 RT: 14.645 min Scan# 1965
 Delta R.T. -0.012 min
 Lab File: BP011098.D
 Acq: 25 Jul 2022 18:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020713


Tgt Ion:250 Resp: 736936
 Ion Ratio Lower Upper
 250 100
 248 63.1 49.8 74.6
 252 62.7 51.2 76.8



#77
 Carbazole
 Concen: 21.317 ng/uL
 RT: 17.510 min Scan# 2452
 Delta R.T. -0.012 min
 Lab File: BP011098.D
 Acq: 25 Jul 2022 18:29

Tgt Ion:167 Resp: 2796541
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.9 25.3
 139 13.7 10.6 16.0





#78

Di-n-butylphthalate

Concen: 23.267 ng/ul

RT: 18.075 min Scan# 2548

Delta R.T. -0.018 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:149 Resp: 3815552

Ion Ratio Lower Upper

149 100

150 9.3 7.1 10.7

104 5.3 4.6 6.8



#80

Fluoranthene

Concen: 27.499 ng/ul

RT: 19.122 min Scan# 2726

Delta R.T. -0.012 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:202 Resp: 3334838


Ion Ratio Lower Upper

202 100

101 14.6 7.1 10.7#

100 11.6 5.7 8.5#









#82
Pyrene
Concen: 26.450 ng/ul
RT: 19.480 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:202 Resp: 3309406
Ion Ratio Lower Upper
202 100
101 16.6 7.9 11.9#
100 13.8 6.7 10.1#



#83
Butylbenzylphthalate
Concen: 28.589 ng/ul
RT: 20.369 min Scan# 2938
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:149 Resp: 1595709
Ion Ratio Lower Upper
149 100
91 79.0 60.0 90.0
206 17.8 17.4 26.2









#84
3,3'-Dichlorobenzidine
Concen: 20.793 ng/ul
RT: 21.139 min Scan# 3071
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:252 Resp: 726884
Ion Ratio Lower Upper
252 100
254 63.7 51.0 76.4
126 14.4 8.6 12.8#



#85
Benzo(a)anthracene
Concen: 22.169 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:228 Resp: 2665432
Ion Ratio Lower Upper
228 100
229 19.4 15.0 22.4
226 26.8 20.6 31.0









#86
Bis(2-ethylhexyl)phthalate
Concen: 27.692 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. -0.018 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:149 Resp: 2306223
Ion Ratio Lower Upper
149 100
167 26.3 22.6 33.8
279 3.7 4.2 6.4#



#87
Chrysene
Concen: 21.581 ng/ul
RT: 21.251 min Scan# 3088
Delta R.T. -0.012 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:228 Resp: 2547747
Ion Ratio Lower Upper
228 100
226 29.5 23.6 35.4
229 19.4 15.6 23.4









#89
Di-n-octyl phthalate
Concen: 34.523 ng/ul
RT: 22.045 min Scan# 311
Delta R.T. -0.024 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29




Instrument :
BNA_P
ClientSampleId :
SSTD020713


Tgt Ion:149 Resp: 3520422



#90
Benzo(b)fluoranthene
Concen: 22.974 ng/ul
RT: 22.839 min Scan# 3358
Delta R.T. -0.024 min
Lab File: BP011098.D
Acq: 25 Jul 2022 18:29

Tgt Ion:252 Resp: 2156870
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 14.0 6.2 9.2#





#91

Benzo(k)fluoranthene

Concen: 23.394 ng/u1

RT: 22.886 min Scan# 31

Delta R.T. -0.024 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:252 Resp: 2063028

Ion Ratio Lower Upper

252 100

253 21.9 17.0 25.4

125 14.5 6.2 9.4#



#93

Benzo(a)pyrene

Concen: 21.927 ng/u1

RT: 23.451 min Scan# 3462

Delta R.T. -0.024 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:252 Resp: 1976973


Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 15.0 6.6 10.0#





#94

Indeno(1,2,3-cd)pyrene

Concen: 19.462 ng/ul

RT: 25.974 min Scan# 3897

Delta R.T. -0.041 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29

Instrument :

BNA_P

ClientSampleId :

SSTD020713





Tgt Ion:276 Resp: 2039286

Ion Ratio Lower Upper

276 100

138 30.4 14.6 22.0#

277 25.1 19.6 29.4



#95

Dibenzo(a,h)anthracene

Concen: 19.302 ng/ul

RT: 25.992 min Scan# 3894

Delta R.T. -0.047 min

Lab File: BP011098.D

Acq: 25 Jul 2022 18:29




Tgt Ion:278 Resp: 1750085


Ion Ratio Lower Upper

278 100

139 22.0 9.4 14.2#

279 23.5 18.9 28.3





#96
 Benzo(g,h,i)perylene
 Concen: 19.494 ng/ul
 RT: 26.721 min Scan# 4018
 Delta R.T. -0.041 min
 Lab File: BP011098.D
 Acq: 25 Jul 2022 18:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020713

Tgt Ion:276 Resp: 1739638
 Ion Ratio Lower Upper
 276 100
 138 29.0 12.7 19.1#
 277 24.1 18.2 27.4

