

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
 Data File : BP006331.D
 Acq On : 26 Jul 2021 10:43
 Operator : CG/JU
 Sample : PB137844BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK844

Quant Time: Jul 26 12:13:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 04:29:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	223397	20.000	ng/ul	0.00
20) Naphthalene-d8	10.575	136	953219	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.416	164	545178	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.169	188	1082293	20.000	ng/ul	0.00
79) Chrysene-d12	21.262	240	1012833	20.000	ng/ul	0.00
88) Perylene-d12	23.615	264	1014805	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	40900	6.834	ng/uL	0.00
4) Pyridine-d5	3.699	84	609196	34.288	ng/ul	0.00
7) Phenol-d5	6.957	99	699750	33.351	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	454312	34.801	ng/ul	0.00
11) 2-Chlorophenol-d4	7.316	132	542997	34.835	ng/ul	-0.01
15) 4-Methylphenol-d8	8.493	113	547992	33.212	ng/ul	0.00
21) Nitrobenzene-d5	8.940	128	257016	36.211	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.657	143	252313	37.231	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.192	165	442039	33.286	ng/ul	0.00
31) 4-Chloroaniline-d4	10.710	131	742520	33.962	ng/ul	0.00
46) Dimethylphthalate-d6	13.833	166	1478441	38.103	ng/ul	0.00
49) Acenaphthylene-d8	14.110	160	1745883	36.623	ng/ul	0.00
54) 4-Nitrophenol-d4	14.616	143	255977	33.225	ng/ul	0.00
60) Fluorene-d10	15.410	176	1206526	36.989	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.528	200	163084	27.030	ng/ul	0.00
73) Anthracene-d10	17.269	188	1926324	39.428	ng/ul	0.00
81) Pyrene-d10	19.510	212	2103445	40.377	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.462	264	2015673	39.086	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
Data File : BP006331.D
Acq On : 26 Jul 2021 10:43
Operator : CG/JU
Sample : PB137844BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK844

Quant Time: Jul 26 12:13:22 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 24 04:29:12 2021
Response via : Initial Calibration

