

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
 Data File : BP006340.D
 Acq On : 26 Jul 2021 16:15
 Operator : CG/JU
 Sample : M3062-03MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JLX88MSD

Quant Time: Jul 26 17:52:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 26 12:12:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	227460	20.000	ng/u1	0.00
20) Naphthalene-d8	10.575	136	977336	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.416	164	550760	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	1059271	20.000	ng/u1	0.00
79) Chrysene-d12	21.262	240	975559	20.000	ng/u1	0.00
88) Perylene-d12	23.609	264	997958	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	23403	3.841	ng/uL	0.01
4) Pyridine-d5	3.710	84	264856	14.641	ng/u1	0.00
7) Phenol-d5	6.963	99	424597	19.876	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	280809	21.126	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	329112	20.736	ng/u1	0.00
15) 4-Methylphenol-d8	8.498	113	337890	20.113	ng/u1	0.00
21) Nitrobenzene-d5	8.945	128	158386	21.765	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	148030	21.304	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.192	165	287279	21.098	ng/u1	0.00
31) 4-Chloroaniline-d4	10.716	131	420276	18.748	ng/u1	0.00
46) Dimethylphthalate-d6	13.833	166	937585	23.919	ng/u1	0.00
49) Acenaphthylene-d8	14.104	160	1089208	22.617	ng/u1	0.00
54) 4-Nitrophenol-d4	14.616	143	157755	20.269	ng/u1	0.00
60) Fluorene-d10	15.410	176	755989	22.941	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.527	200	117664	19.925	ng/u1	0.00
73) Anthracene-d10	17.269	188	1185815	24.799	ng/u1	0.00
81) Pyrene-d10	19.504	212	1166532	23.248	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.456	264	886800	17.486	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	55918	8.692	ng/uL	99
5) Pyridine	3.734	79	312817	16.581	ng/u1	98
6) Benzaldehyde	6.940	77	313348	26.884	ng/u1	97
8) Phenol	6.987	94	505501	23.149	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.222	93	403787	23.508	ng/u1	98
12) 2-Chlorophenol	7.351	128	382565	23.776	ng/u1	97
13) 2-Methylphenol	8.234	108	362824	22.769	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.316	45	464754	24.219	ng/u1	100
16) Acetophenone	8.610	105	617844	23.916	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.598	70	337146	24.946	ng/u1	98
18) 4-Methylphenol	8.557	108	400269	23.392	ng/u1	98
19) Hexachloroethane	8.863	117	121602	18.475	ng/u1	96
22) Nitrobenzene	8.987	77	483831	25.625	ng/u1	97
23) Isophorone	9.516	82	881945	26.129	ng/u1	99
25) 2-Nitrophenol	9.692	139	196068	25.588	ng/u1	98
26) 2,4-Dimethylphenol	9.757	107	380602	20.945	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.998	93	541468	24.980	ng/u1	99
29) 2,4-Dichlorophenol	10.222	162	325282	24.350	ng/u1	99
30) Naphthalene	10.622	128	1218207	24.024	ng/u1	100
32) 4-Chloroaniline	10.739	127	477978	21.645	ng/u1	100
33) Hexachlorobutadiene	10.910	225	192755	23.652	ng/u1	99
34) Caprolactam	11.516	113	120569	26.138	ng/u1	91
35) 4-Chloro-3-methylphenol	11.863	107	406051	25.462	ng/u1	99
36) 2-Methylnaphthalene	12.233	142	842206	24.239	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
 Data File : BP006340.D
 Acq On : 26 Jul 2021 16:15
 Operator : CG/JU
 Sample : M3062-03MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JLX88MSD

Quant Time: Jul 26 17:52:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 26 12:12:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.451	142	825593	23.606	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.604	216	354624	23.893	ng/ul	98
40) Hexachlorocyclopentadiene	12.581	237	154950	16.048	ng/ul	99
41) 2,4,6-Trichlorophenol	12.845	196	239642	26.056	ng/ul	97
42) 2,4,5-Trichlorophenol	12.910	196	259437	25.782	ng/ul	100
43) 1,1'-Biphenyl	13.251	154	1055768	25.572	ng/ul	100
44) 2-Chloronaphthalene	13.286	162	779618	25.010	ng/ul	99
45) 2-Nitroaniline	13.498	65	267204	29.018	ng/ul	97
47) Dimethylphthalate	13.880	163	1043104	27.110	ng/ul	100
48) 2,6-Dinitrotoluene	13.998	165	198831	28.267	ng/ul	93
50) Acenaphthylene	14.133	152	1214979	25.906	ng/ul	99
51) 3-Nitroaniline	14.322	138	216078	26.217	ng/ul	97
52) Acenaphthene	14.475	153	880587	25.591	ng/ul	99
53) 2,4-Dinitrophenol	14.527	184	84429	20.238	ng/ul	99
55) 4-Nitrophenol	14.627	109	176427	28.024	ng/ul	96
56) Dibenzofuran	14.816	168	1193877	25.872	ng/ul	98
57) 2,4-Dinitrotoluene	14.780	165	301055	29.949	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.039	232	215323	27.439	ng/ul	96
59) Diethylphthalate	15.251	149	1141743	28.649	ng/ul	100
61) Fluorene	15.463	166	998080	26.294	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.469	204	456242	25.781	ng/ul	98
63) 4-Nitroaniline	15.486	138	242671	29.978	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.539	198	152115	26.351	ng/ul	95
67) N-Nitrosodiphenylamine	15.680	169	871972	27.710	ng/ul	100
68) 4-Bromophenyl-phenylether	16.363	248	265307	31.275	ng/ul	96
69) Hexachlorobenzene	16.463	284	243193	21.376	ng/ul	95
70) Atrazine	16.639	200	327089	30.036	ng/ul	99
71) Pentachlorophenol	16.810	266	177411	24.805	ng/ul	96
72) Phenanthrene	17.210	178	1597999	28.094	ng/ul	99
74) Anthracene	17.298	178	1607556	27.873	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.210	216	346919	23.893	ng/uL	100
76) Pentachlorobenzene	14.727	250	337506	23.843	ng/uL	99
77) Carbazole	17.574	167	1387682	26.303	ng/ul	99
78) Di-n-butylphthalate	18.145	149	2036434	32.297	ng/ul	99
80) Fluoranthene	19.180	202	1869040	30.229	ng/ul	99
82) Pyrene	19.533	202	1571954	24.501	ng/ul	98
83) Butylbenzylphthalate	20.427	149	893618	33.946	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.192	252	556657	25.897	ng/ul	99
85) Benzo(a)anthracene	21.251	228	1811540	29.967	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	1372117	33.291	ng/ul	100
87) Chrysene	21.304	228	1733267	29.913	ng/ul	99
89) Di-n-octyl phthalate	22.109	149	2276429	32.733	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	1827323	28.828	ng/ul	100
91) Benzo(k)fluoranthene	22.945	252	1830903	30.434	ng/ul	100
93) Benzo(a)pyrene	23.509	252	954829	17.233	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.045	276	1585729	22.358	ng/ul	94
95) Dibenzo(a,h)anthracene	26.068	278	1745262	29.054	ng/ul	99
96) Benzo(g,h,i)perylene	26.774	276	72190	1.232	ng/ul	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.692 ng/uL
RT: 3.340 min Scan# 43
Delta R.T. 0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	88	Resp:	55918
Ion	Ratio	Lower	Upper
88	100		
43	27.5	22.4	33.6
58	66.6	54.2	81.4



#5
Pyridine
Concen: 16.581 ng/uL
RT: 3.734 min Scan# 110
Delta R.T. 0.011 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	79	Resp:	312817
Ion	Ratio	Lower	Upper
79	100		
52	61.0	50.6	75.8
51	29.2	22.8	34.2









#6
Benzaldehyde
Concen: 26.884 ng/ul
RT: 6.940 min Scan# 655
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion: 77 Resp: 313348
Ion Ratio Lower Upper
77 100
105 90.0 74.5 111.7
106 85.3 71.0 106.4



#8
Phenol
Concen: 23.149 ng/ul
RT: 6.987 min Scan# 663
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 94 Resp: 505501
Ion Ratio Lower Upper
94 100
65 28.8 22.6 33.8
66 38.2 29.5 44.3









#10
Bis(2-Chloroethyl)ether
Concen: 23.508 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampleId :
JLX88MSD

Tgt Ion: 93 Resp: 403787




Ion	Ratio	Lower	Upper
93	100		
63	77.9	60.3	90.5
95	32.5	25.8	38.8




#12
2-Chlorophenol
Concen: 23.776 ng/ul
RT: 7.351 min Scan# 725
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 128 Resp: 382565

Ion	Ratio	Lower	Upper
128	100		
64	49.9	37.9	56.9
130	32.6	25.4	38.2









#13
2-Methylphenol
Concen: 22.769 ng/ul
RT: 8.234 min Scan# 875
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:108 Resp: 362824
Ion Ratio Lower Upper
108 100
107 92.2 72.9 109.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.219 ng/ul
RT: 8.316 min Scan# 889
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 45 Resp: 464754
Ion Ratio Lower Upper
45 100
77 18.1 14.6 21.8
79 13.6 10.6 16.0









#16
Acetophenone
Concen: 23.916 ng/ul
RT: 8.610 min Scan# 939
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:105 Resp: 617844
Ion Ratio Lower Upper
105 100
77 83.5 65.0 97.4
51 23.7 17.9 26.9



#17
N-Nitroso-di-n-propylamine
Concen: 24.946 ng/ul
RT: 8.598 min Scan# 937
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 70 Resp: 337146
Ion Ratio Lower Upper
70 100
42 48.1 37.0 55.6
101 9.1 7.3 10.9
130 18.2 15.2 22.8









#18
4-Methylphenol
Concen: 23.392 ng/ul
RT: 8.557 min Scan# 930
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:108 Resp: 400269
Ion Ratio Lower Upper
108 100
107 116.1 91.1 136.7



#19
Hexachloroethane
Concen: 18.475 ng/ul
RT: 8.863 min Scan# 982
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:117 Resp: 121602
Ion Ratio Lower Upper
117 100
201 87.0 66.3 99.5
199 53.5 41.4 62.2









#22
Nitrobenzene
Concen: 25.625 ng/ul
RT: 8.987 min Scan# 1003
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion: 77 Resp: 483831
Ion Ratio Lower Upper
77 100
123 39.7 33.3 49.9
65 13.3 10.4 15.6



#23
Isophorone
Concen: 26.129 ng/ul
RT: 9.516 min Scan# 1093
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 82 Resp: 881945
Ion Ratio Lower Upper
82 100
95 7.0 5.9 8.9
138 16.0 13.0 19.6









#25
2-Nitrophenol
Concen: 25.588 ng/ul
RT: 9.692 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:139 Resp: 196068
Ion Ratio Lower Upper
139 100
65 54.5 44.8 67.2
109 31.0 25.6 38.4



#26
2,4-Dimethylphenol
Concen: 20.945 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:107 Resp: 380602
Ion Ratio Lower Upper
107 100
121 49.1 39.9 59.9
122 81.9 65.0 97.4









#27
Bis(2-Chloroethoxy)methane
Concen: 24.980 ng/ul
RT: 9.998 min Scan# 1175
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion: 93 Resp: 541468
Ion Ratio Lower Upper
93 100
95 32.4 26.1 39.1
123 10.2 8.7 13.1



#29
2,4-Dichlorophenol
Concen: 24.350 ng/ul
RT: 10.222 min Scan# 1213
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion: 162 Resp: 325282
Ion Ratio Lower Upper
162 100
164 65.3 51.9 77.9
98 42.3 32.5 48.7









#30
Naphthalene
Concen: 24.024 ng/ul
RT: 10.622 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:128 Resp: 1218207
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.2 10.7 16.1



#32
4-Chloroaniline
Concen: 21.645 ng/ul
RT: 10.739 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:127 Resp: 477978
Ion Ratio Lower Upper
127 100
129 31.9 25.7 38.5









#33
Hexachlorobutadiene
Concen: 23.652 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampleId :
JLX88MSD

Tgt Ion:225 Resp: 192755




Ion	Ratio	Lower	Upper
225	100		
223	62.1	50.3	75.5
227	63.6	51.8	77.6




#34
Caprolactam
Concen: 26.138 ng/ul
RT: 11.516 min Scan# 1433
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:113 Resp: 120569

Ion	Ratio	Lower	Upper
113	100		
55	158.3	112.2	168.2
56	113.8	89.2	133.8









#35
4-Chloro-3-methylphenol
Concen: 25.462 ng/ul
RT: 11.863 min Scan# 1492
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	107	Resp:	406051
Ion	Ratio	Lower	Upper
107	100		
144	23.8	18.3	27.5
142	73.9	58.5	87.7



#36
2-Methylnaphthalene
Concen: 24.239 ng/ul
RT: 12.233 min Scan# 1555
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	142	Resp:	842206
Ion	Ratio	Lower	Upper
142	100		
141	90.0	72.6	108.8









#37
1-Methylnaphthalene
Concen: 23.606 ng/ul
RT: 12.451 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampleId :
JLX88MSD

Tgt Ion:142 Resp: 825593




Ion	Ratio	Lower	Upper
142	100		
141	93.8	74.5	111.7
116	4.4	3.5	5.3




#39
1,2,4,5-Tetrachlorobenzene
Concen: 23.893 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:216 Resp: 354624

Ion	Ratio	Lower	Upper
216	100		
214	77.2	63.5	95.3
179	22.7	18.7	28.1
108	21.5	16.9	25.3









#40
Hexachlorocyclopentadiene
Concen: 16.048 ng/ul
RT: 12.581 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	237	Resp:	154950
Ion Ratio	Lower	Upper	
237	100		
235	63.0	49.4	74.2
272	11.8	9.2	13.8



#41
2,4,6-Trichlorophenol
Concen: 26.056 ng/ul
RT: 12.845 min Scan# 1659
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	196	Resp:	239642
Ion Ratio	Lower	Upper	
196	100		
198	91.3	75.6	113.4
200	30.9	24.2	36.2









#42
2,4,5-Trichlorophenol
Concen: 25.782 ng/ul
RT: 12.910 min Scan# 1670
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampleId :
JLX88MSD

Tgt Ion:196 Resp: 259437




Ion	Ratio	Lower	Upper
196	100		
198	95.2	76.1	114.1
200	30.3	24.2	36.4




#43
1,1'-Biphenyl
Concen: 25.572 ng/ul
RT: 13.251 min Scan# 1728
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:154 Resp: 1055768

Ion	Ratio	Lower	Upper
154	100		
153	41.2	33.0	49.4
76	15.9	12.3	18.5









#44
2-Chloronaphthalene
Concen: 25.010 ng/ul
RT: 13.286 min Scan# 1734
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	162	Resp:	779618
Ion	Ratio	Lower	Upper
162	100		
164	33.0	26.3	39.5
127	43.8	34.7	52.1



#45
2-Nitroaniline
Concen: 29.018 ng/ul
RT: 13.498 min Scan# 1770
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	65	Resp:	267204
Ion	Ratio	Lower	Upper
65	100		
92	69.7	55.8	83.8
138	99.8	76.3	114.5









#47
Dimethylphthalate
Concen: 27.110 ng/ul
RT: 13.880 min Scan# 1835
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:163 Resp: 1043104
Ion Ratio Lower Upper
163 100
194 3.6 3.0 4.6
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 28.267 ng/ul
RT: 13.998 min Scan# 1855
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:165 Resp: 198831
Ion Ratio Lower Upper
165 100
89 71.5 52.3 78.5
121 26.5 19.8 29.6









#50
Acenaphthylene
Concen: 25.906 ng/ul
RT: 14.133 min Scan# 1878
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:152 Resp: 1214979
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.2
153 12.9 10.6 15.8



#51
3-Nitroaniline
Concen: 26.217 ng/ul
RT: 14.322 min Scan# 1910
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:138 Resp: 216078
Ion Ratio Lower Upper
138 100
108 12.2 9.8 14.6
92 130.7 101.0 151.6









#52
Acenaphthene
Concen: 25.591 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampleId :
JLX88MSD

Tgt Ion:153 Resp: 880587




Ion	Ratio	Lower	Upper
153	100		
152	47.8	38.0	57.0
154	87.7	69.0	103.6




#53
2,4-Dinitrophenol
Concen: 20.238 ng/ul
RT: 14.527 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:184 Resp: 84429

Ion	Ratio	Lower	Upper
184	100		
63	77.8	63.0	94.4
154	68.8	54.6	82.0









#55
4-Nitrophenol
Concen: 28.024 ng/ul
RT: 14.627 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:109 Resp: 176427
Ion Ratio Lower Upper
109 100
139 125.3 106.2 159.2
65 138.1 108.4 162.6



#56
Dibenzofuran
Concen: 25.872 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:168 Resp: 1193877
Ion Ratio Lower Upper
168 100
139 41.8 32.6 49.0









#57
2,4-Dinitrotoluene
Concen: 29.949 ng/ul
RT: 14.780 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:165 Resp: 301055
Ion Ratio Lower Upper
165 100
63 51.6 37.4 56.0
182 2.7 2.6 3.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 27.439 ng/ul
RT: 15.039 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:232 Resp: 215323
Ion Ratio Lower Upper
232 100
131 58.7 43.8 65.8
130 2.7 2.6 3.8
166 36.1 28.5 42.7









#59
Diethylphthalate
Concen: 28.649 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:149 Resp: 1141743
Ion Ratio Lower Upper
149 100
177 20.2 16.1 24.1
150 12.4 9.8 14.6



#61
Fluorene
Concen: 26.294 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:166 Resp: 998080
Ion Ratio Lower Upper
166 100
165 100.5 79.2 118.8
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 25.781 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:204 Resp: 456242
Ion Ratio Lower Upper
204 100
206 33.2 25.5 38.3
141 70.2 57.0 85.4



#63
4-Nitroaniline
Concen: 29.978 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:138 Resp: 242671
Ion Ratio Lower Upper
138 100
92 60.0 46.5 69.7
108 88.9 67.8 101.6









#66
4,6-Dinitro-2-methylphenol
Concen: 26.351 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:198	Resp:	152115
Ion Ratio	Lower	Upper
198	100	
182	5.0	4.6 6.8
77	35.4	25.6 38.4



#67
N-Nitrosodiphenylamine
Concen: 27.710 ng/ul
RT: 15.680 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:169	Resp:	871972
Ion Ratio	Lower	Upper
169	100	
168	67.4	53.8 80.8
167	36.6	29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 31.275 ng/ul
RT: 16.363 min Scan# 2257
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampled :
JLX88MSD


Tgt Ion:	248	Resp:	265307
Ion Ratio	Lower	Upper	
248	100		
250	96.7	79.4	119.2
141	93.1	70.3	105.5



#69
Hexachlorobenzene
Concen: 21.376 ng/ul
RT: 16.463 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	284	Resp:	243193
Ion Ratio	Lower	Upper	
284	100		
142	49.0	35.5	53.3
249	34.4	26.7	40.1









#70
Atrazine
Concen: 30.036 ng/ul
RT: 16.639 min Scan# 2304
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:200 Resp: 327089
Ion Ratio Lower Upper
200 100
173 27.5 22.4 33.6
215 45.1 37.0 55.4



#71
Pentachlorophenol
Concen: 24.805 ng/ul
RT: 16.810 min Scan# 2333
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:266 Resp: 177411
Ion Ratio Lower Upper
266 100
264 64.5 49.6 74.4
268 66.9 50.7 76.1









#72
Phenanthrene
Concen: 28.094 ng/ul
RT: 17.210 min Scan# 2401
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:178 Resp: 1597999
Ion Ratio Lower Upper
178 100
179 15.3 12.0 18.0
176 19.1 15.0 22.6



#74
Anthracene
Concen: 27.873 ng/ul
RT: 17.298 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:178 Resp: 1607556
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.893 ng/uL
RT: 13.210 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:216 Resp: 346919
Ion Ratio Lower Upper
216 100
214 78.0 62.4 93.6
218 48.0 37.8 56.6



#76
Pentachlorobenzene
Concen: 23.843 ng/uL
RT: 14.727 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:250 Resp: 337506
Ion Ratio Lower Upper
250 100
248 63.2 51.3 76.9
252 64.1 51.0 76.6









#77
 Carbazole
 Concen: 26.303 ng/ul
 RT: 17.574 min Scan# 2463
 Delta R.T. -0.006 min
 Lab File: BP006340.D
 Acq: 26 Jul 2021 16:15




Instrument :
 BNA_P
 ClientSampleId :
 JLX88MSD


Tgt Ion:167 Resp: 1387682
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.2 25.8
 139 14.4 11.3 16.9



#78
 Di-n-butylphthalate
 Concen: 32.297 ng/ul
 RT: 18.145 min Scan# 2560
 Delta R.T. -0.006 min
 Lab File: BP006340.D
 Acq: 26 Jul 2021 16:15

Tgt Ion:149 Resp: 2036434
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.0
 104 5.6 4.4 6.6









#80
Fluoranthene
Concen: 30.229 ng/ul
RT: 19.180 min Scan# 2736
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:202 Resp: 1869040
Ion Ratio Lower Upper
202 100
101 14.4 11.0 16.6
100 11.3 8.8 13.2



#82
Pyrene
Concen: 24.501 ng/ul
RT: 19.533 min Scan# 2796
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:202 Resp: 1571954
Ion Ratio Lower Upper
202 100
101 17.2 13.0 19.4
100 14.4 11.0 16.4









#83
Butylbenzylphthalate
Concen: 33.946 ng/ul
RT: 20.427 min Scan# 2948
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	149	Resp:	893618
Ion	Ratio	Lower	Upper
149	100		
91	79.9	61.8	92.8
206	18.0	15.0	22.6



#84
3,3'-Dichlorobenzidine
Concen: 25.897 ng/ul
RT: 21.192 min Scan# 3078
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	252	Resp:	556657
Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.4	77.0
126	13.1	11.3	16.9









#85
Benzo(a)anthracene
Concen: 29.967 ng/ul
RT: 21.251 min Scan# 3088
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:228 Resp: 1811540
Ion Ratio Lower Upper
228 100
229 19.5 15.9 23.9
226 26.9 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 33.291 ng/ul
RT: 21.192 min Scan# 3078
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:149 Resp: 1372117
Ion Ratio Lower Upper
149 100
167 25.6 20.3 30.5
279 3.5 2.7 4.1









#87
Chrysene
Concen: 29.913 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:228 Resp: 1733267
Ion Ratio Lower Upper
228 100
226 30.2 24.1 36.1
229 19.3 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 32.733 ng/ul
RT: 22.109 min Scan# 3234
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:149 Resp: 2276429









#90
Benzo(b)fluoranthene
Concen: 28.828 ng/ul
RT: 22.898 min Scan# 3368
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:252 Resp: 1827323
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 13.3 10.6 16.0



#91
Benzo(k)fluoranthene
Concen: 30.434 ng/ul
RT: 22.945 min Scan# 3376
Delta R.T. -0.006 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:252 Resp: 1830903
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 13.2 10.3 15.5









#93
Benzo(a)pyrene
Concen: 17.233 ng/ul
RT: 23.509 min Scan# 3472
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15




Instrument :
BNA_P
ClientSampleId :
JLX88MSD


Tgt Ion:	252	Resp:	954829
Ion Ratio	Lower	Upper	
252	100		
253	21.9	17.6	26.4
125	14.7	11.8	17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 22.358 ng/ul
RT: 26.045 min Scan# 3903
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:	276	Resp:	1585729
Ion Ratio	Lower	Upper	
276	100		
138	31.9	21.8	32.6
277	26.0	19.6	29.4









#95
Dibenzo(a,h)anthracene
Concen: 29.054 ng/ul
RT: 26.068 min Scan# 3907
Delta R.T. -0.000 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Instrument :
BNA_P
ClientSampled :
JLX88MSD

Tgt Ion:278 Resp: 1745262
Ion Ratio Lower Upper
278 100
139 19.7 15.8 23.6
279 24.4 18.8 28.2



#96
Benzo(g,h,i)perylene
Concen: 1.232 ng/ul
RT: 26.774 min Scan# 4027
Delta R.T. -0.018 min
Lab File: BP006340.D
Acq: 26 Jul 2021 16:15

Tgt Ion:276 Resp: 72190
Ion Ratio Lower Upper
276 100
138 29.0 20.2 30.4
277 24.2 18.6 27.8

