

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
 Data File : BP006351.D
 Acq On : 26 Jul 2021 22:35
 Operator : CG/JU
 Sample : M3091-21
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0059

Quant Time: Jul 27 01:10:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 04:29:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	281920	20.000	ng/u1	0.00
20) Naphthalene-d8	10.569	136	1196263	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.410	164	679437	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.163	188	1324559	20.000	ng/u1	0.00
79) Chrysene-d12	21.257	240	1277087	20.000	ng/u1	-0.01
88) Perylene-d12	23.604	264	1316767	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	41653	5.515	ng/uL	0.01
4) Pyridine-d5	3.711	84	279922	12.485	ng/u1	0.00
7) Phenol-d5	6.958	99	202042	7.631	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	534017	32.414	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	512494	26.053	ng/u1	0.00
15) 4-Methylphenol-d8	8.493	113	356211	17.107	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	301097	33.803	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.657	143	286511	33.687	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.193	165	484008	29.041	ng/u1	0.00
31) 4-Chloroaniline-d4	10.710	131	769933	28.061	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	1662529	34.380	ng/u1	0.00
49) Acenaphthylene-d8	14.104	160	1996799	33.610	ng/u1	0.00
54) 4-Nitrophenol-d4	14.610	143	68094	7.092	ng/u1	0.00
60) Fluorene-d10	15.410	176	1387406	34.129	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.522	200	210764	28.543	ng/u1	-0.01
73) Anthracene-d10	17.263	188	2279063	38.116	ng/u1	0.00
81) Pyrene-d10	19.504	212	2550298	38.825	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.457	264	2496675	37.311	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072621\
Data File : BP006351.D
Acq On : 26 Jul 2021 22:35
Operator : CG/JU
Sample : M3091-21
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0059

Quant Time: Jul 27 01:10:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 24 04:29:12 2021
Response via : Initial Calibration

