

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072624\
 Data File : BP021172.D
 Acq On : 26 Jul 2024 15:51
 Operator : MA/JU
 Sample : P3280-15
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jul 26 22:46:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 25 00:15:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.651	152	167998	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	682764	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.292	164	446633	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	967439	20.000	ng/u1	0.01
79) Chrysene-d12	21.568	240	946242	20.000	ng/u1	0.01
88) Perylene-d12	24.892	264	1110362	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	15610	4.228	ng/uL	0.00
4) Pyridine-d5	3.604	84	77085	7.212	ng/u1	0.00
7) Phenol-d5	6.887	99	78612	5.740	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.998	67	182605	22.000	ng/u1	0.00
11) 2-Chlorophenol-d4	7.210	132	229361	20.325	ng/u1	0.00
15) 4-Methylphenol-d8	8.387	113	160368	14.098	ng/u1	0.00
21) Nitrobenzene-d5	8.804	128	120907	22.799	ng/u1	0.00
24) 2-Nitrophenol-d4	9.516	143	142545	23.485	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	244593	22.285	ng/u1	0.00
31) 4-Chloroaniline-d4	10.575	131	312396	21.485	ng/u1	0.00
46) Dimethylphthalate-d6	13.704	166	890991	27.442	ng/u1	0.00
49) Acenaphthylene-d8	13.986	160	902864	24.765	ng/u1	0.00
54) 4-Nitrophenol-d4	14.639	143	24386	5.279	ng/u1	0.04
60) Fluorene-d10	15.310	176	719185	27.000	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.474	200	137921	23.562	ng/u1	0.00
73) Anthracene-d10	17.216	188	1293713	30.108	ng/u1	0.01
81) Pyrene-d10	19.604	212	1598367	27.322	ng/u1	0.02
92) Benzo(a)pyrene-d12	24.656	264	1751741	31.988	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072624\
Data File : BP021172.D
Acq On : 26 Jul 2024 15:51
Operator : MA/JU
Sample : P3280-15
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Jul 26 22:46:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 25 00:15:24 2024
Response via : Initial Calibration

