

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072721\
 Data File : BP006369.D
 Acq On : 27 Jul 2021 18:23
 Operator : CG/JU
 Sample : PB137903BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK903

Quant Time: Jul 27 18:50:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 04:29:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	425324	20.000	ng/u1	0.00
20) Naphthalene-d8	10.569	136	1795237	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.410	164	1040645	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.163	188	2053890	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	1972200	20.000	ng/u1	0.00
88) Perylene-d12	23.610	264	1916291	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	69383	6.089	ng/uL	0.00
4) Pyridine-d5	3.699	84	992284	29.335	ng/u1	0.00
7) Phenol-d5	6.958	99	1169172	29.269	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	748219	30.104	ng/u1	0.00
11) 2-Chlorophenol-d4	7.316	132	918929	30.964	ng/u1	-0.01
15) 4-Methylphenol-d8	8.493	113	928767	29.566	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	445852	33.354	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.658	143	474867	37.205	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.187	165	753833	30.140	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.710	131	1228332	29.831	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	2463826	33.266	ng/u1	0.00
49) Acenaphthylene-d8	14.104	160	2950064	32.420	ng/u1	0.00
54) 4-Nitrophenol-d4	14.610	143	477510	32.470	ng/u1	0.00
60) Fluorene-d10	15.410	176	1964356	31.549	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.522	200	317994	27.772	ng/u1	-0.01
73) Anthracene-d10	17.263	188	3111566	33.560	ng/u1	0.00
81) Pyrene-d10	19.504	212	3444075	33.952	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.463	264	3311768	34.008	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072721\
Data File : BP006369.D
Acq On : 27 Jul 2021 18:23
Operator : CG/JU
Sample : PB137903BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK903

Quant Time: Jul 27 18:50:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 24 04:29:12 2021
Response via : Initial Calibration

