

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072721\
 Data File : BP006375.D
 Acq On : 27 Jul 2021 21:50
 Operator : CG/JU
 Sample : M3084-17
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0035

Quant Time: Jul 28 01:48:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 04:29:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	321258	20.000	ng/u1	0.00
20) Naphthalene-d8	10.569	136	1386968	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.410	164	825693	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.163	188	1623441	20.000	ng/u1	0.00
79) Chrysene-d12	21.257	240	1508658	20.000	ng/u1	-0.01
88) Perylene-d12	23.604	264	1424064	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	23879	2.775	ng/uL	0.00
4) Pyridine-d5	3.705	84	198325	7.762	ng/u1	0.00
7) Phenol-d5	6.958	99	440311	14.593	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	301261	16.047	ng/u1	0.00
11) 2-Chlorophenol-d4	7.316	132	358456	15.991	ng/u1	-0.01
15) 4-Methylphenol-d8	8.487	113	333920	14.073	ng/u1	-0.01
21) Nitrobenzene-d5	8.940	128	174596	16.906	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.658	143	172406	17.484	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.187	165	301463	15.601	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.710	131	458163	14.402	ng/u1	0.00
46) Dimethylphthalate-d6	13.828	166	1069575	18.200	ng/u1	-0.01
49) Acenaphthylene-d8	14.104	160	1244540	17.237	ng/u1	0.00
54) 4-Nitrophenol-d4	14.610	143	160435	13.749	ng/u1	0.00
60) Fluorene-d10	15.404	176	865886	17.527	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.522	200	105357	11.641	ng/u1	-0.01
73) Anthracene-d10	17.263	188	1341427	18.304	ng/u1	0.00
81) Pyrene-d10	19.504	212	1541635	19.867	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.451	264	1352938	18.696	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072721\
Data File : BP006375.D
Acq On : 27 Jul 2021 21:50
Operator : CG/JU
Sample : M3084-17
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0035

Quant Time: Jul 28 01:48:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP072421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 24 04:29:12 2021
Response via : Initial Calibration

