

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072722\
 Data File : BP011121.D
 Acq On : 27 Jul 2022 10:10
 Operator : CG/JU
 Sample : SSTDCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020716

Quant Time: Jul 28 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 26 14:31:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	378509	20.000	ng/u1	0.00
20) Naphthalene-d8	10.452	136	1747165	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.328	164	1015315	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.092	188	1775610	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.216	240	1095903	20.000	ng/u1	# 0.00
88) Perylene-d12	23.557	264	1125490	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	80012	7.621	ng/uL	0.00
4) Pyridine-d5	3.570	84	558245	20.571	ng/u1	0.00
7) Phenol-d5	6.840	99	650110	20.924	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	497945	24.923	ng/u1	0.00
11) 2-Chlorophenol-d4	7.193	132	533820	21.263	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	557491	22.413	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	288752	22.678	ng/u1	0.00
24) 2-Nitrophenol-d4	9.540	143	274751	21.808	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	484975	21.049	ng/u1	0.00
31) 4-Chloroaniline-d4	10.599	131	779239	20.515	ng/u1	0.00
46) Dimethylphthalate-d6	13.745	166	1474985	21.255	ng/u1	0.00
49) Acenaphthylene-d8	14.016	160	1824358	22.282	ng/u1	0.00
54) 4-Nitrophenol-d4	14.545	143	216203	15.394	ng/u1	0.00
60) Fluorene-d10	15.328	176	1221958	20.717	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	168682	19.198	ng/u1	0.00
73) Anthracene-d10	17.192	188	1619276	21.452	ng/u1	0.00
81) Pyrene-d10	19.451	212	1481912	25.747	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.410	264	1143227	21.149	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	84355	7.657	ng/uL#	93
5) Pyridine	3.587	79	580628	20.761	ng/u1	85
6) Benzaldehyde	6.811	77	216142	14.322	ng/u1	98
8) Phenol	6.863	94	705075	21.349	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.093	93	604416	23.059	ng/u1	96
12) 2-Chlorophenol	7.222	128	558269	21.512	ng/u1	99
13) 2-Methylphenol	8.110	108	530576	21.724	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.187	45	581366	16.216	ng/u1	96
16) Acetophenone	8.487	105	915910	24.324	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.481	70	526115	27.847	ng/u1	94
18) 4-Methylphenol	8.440	108	591020	22.664	ng/u1	97
19) Hexachloroethane	8.728	117	248534	24.200	ng/u1#	75
22) Nitrobenzene	8.863	77	731241	24.421	ng/u1	98
23) Isophorone	9.399	82	1402365	25.024	ng/u1	98
25) 2-Nitrophenol	9.575	139	305102	21.574	ng/u1	90
26) 2,4-Dimethylphenol	9.640	107	674953	22.569	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.881	93	855723	23.074	ng/u1	98
29) 2,4-Dichlorophenol	10.104	162	495221	20.766	ng/u1	98
30) Naphthalene	10.504	128	1915849	21.518	ng/u1	99
32) 4-Chloroaniline	10.622	127	775230	20.634	ng/u1	100
33) Hexachlorobutadiene	10.787	225	306860	22.105	ng/u1	95
34) Caprolactam	11.416	113	156495	18.725	ng/u1	89
35) 4-Chloro-3-methylphenol	11.763	107	594535	22.060	ng/u1	96
36) 2-Methylnaphthalene	12.128	142	1267672	21.963	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072722\
 Data File : BP011121.D
 Acq On : 27 Jul 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020716

Quant Time: Jul 28 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 26 14:31:36 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.351	142	1282864	22.174	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.498	216	571863	21.355	ng/ul#	97
40) Hexachlorocyclopentadiene	12.475	237	362222	21.839	ng/ul	97
41) 2,4,6-Trichlorophenol	12.746	196	364166	20.650	ng/ul	98
42) 2,4,5-Trichlorophenol	12.816	196	388246	20.600	ng/ul	98
43) 1,1'-Biphenyl	13.151	154	1653657	21.828	ng/ul#	97
44) 2-Chloronaphthalene	13.193	162	1220394	21.447	ng/ul	97
45) 2-Nitroaniline	13.410	65	403711	24.352	ng/ul	94
47) Dimethylphthalate	13.793	163	1480995	21.151	ng/ul	98
48) 2,6-Dinitrotoluene	13.916	165	287406	22.672	ng/ul	96
50) Acenaphthylene	14.045	152	2021204	22.000	ng/ul	99
51) 3-Nitroaniline	14.245	138	192372	12.631	ng/ul	95
52) Acenaphthene	14.393	153	1314532	21.599	ng/ul	98
53) 2,4-Dinitrophenol	14.451	184	114272	15.408	ng/ul	98
55) 4-Nitrophenol	14.557	109	188483	17.529	ng/ul#	79
56) Dibenzofuran	14.728	168	1791190	21.329	ng/ul	98
57) 2,4-Dinitrotoluene	14.704	165	389840	21.559	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.957	232	290977	19.715	ng/ul#	80
59) Diethylphthalate	15.169	149	1522546	21.173	ng/ul	97
61) Fluorene	15.381	166	1396807	20.725	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.387	204	649891	20.946	ng/ul	92
63) 4-Nitroaniline	15.416	138	110228	7.875	ng/ul	86
66) 4,6-Dinitro-2-methylph...	15.469	198	176313	19.561	ng/ul#	90
67) N-Nitrosodiphenylamine	15.604	169	1190775	23.477	ng/ul	97
68) 4-Bromophenyl-phenylether	16.287	248	371227	23.181	ng/ul	92
69) Hexachlorobenzene	16.387	284	413518	22.249	ng/ul#	87
70) Atrazine	16.569	200	363740	20.730	ng/ul	96
71) Pentachlorophenol	16.739	266	206916	18.077	ng/ul	98
72) Phenanthrene	17.139	178	1988329	21.494	ng/ul	98
74) Anthracene	17.228	178	1973509	21.446	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.110	216	591804	23.282	ng/ul	98
76) Pentachlorobenzene	14.645	250	536045	24.057	ng/ul	100
77) Carbazole	17.510	167	1655756	19.225	ng/ul	98
78) Di-n-butylphthalate	18.075	149	2215897	20.583	ng/ul	100
80) Fluoranthene	19.122	202	1857872	27.599	ng/ul#	87
82) Pyrene	19.475	202	1851193	26.654	ng/ul#	81
83) Butylbenzylphthalate	20.375	149	722806	23.329	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.139	252	383054	19.740	ng/ul#	98
85) Benzo(a)anthracene	21.198	228	1414581	21.195	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.139	149	1036049	22.411	ng/ul#	95
87) Chrysene	21.251	228	1406239	21.458	ng/ul	100
89) Di-n-octyl phthalate	22.051	149	1600668	21.206	ng/ul	100
90) Benzo(b)fluoranthene	22.845	252	1425728	20.516	ng/ul#	96
91) Benzo(k)fluoranthene	22.892	252	1373500	21.041	ng/ul#	94
93) Benzo(a)pyrene	23.457	252	1424562	21.345	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	25.986	276	1766332	22.774	ng/ul#	87
95) Dibenzo(a,h)anthracene	26.004	278	1515351	22.579	ng/ul#	92
96) Benzo(g,h,i)perylene	26.727	276	1523698	23.066	ng/ul#	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072722\
 Data File : BP011121.D
 Acq On : 27 Jul 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020716

Quant Time: Jul 28 00:00:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 26 14:31:36 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.657 ng/uL
RT: 3.193 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion: 88 Resp: 84355
Ion Ratio Lower Upper
88 100
43 35.6 35.8 53.6#
58 82.0 68.1 102.1



#5
Pyridine
Concen: 20.761 ng/uL
RT: 3.587 min Scan# 85
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion: 79 Resp: 580628
Ion Ratio Lower Upper
79 100
52 86.9 83.8 125.8
51 44.3 41.0 61.4









#6
Benzaldehyde
Concen: 14.322 ng/ul
RT: 6.811 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion: 77 Resp: 216142
Ion Ratio Lower Upper
77 100
105 92.7 76.0 114.0
106 90.2 73.2 109.8



#8
Phenol
Concen: 21.349 ng/ul
RT: 6.863 min Scan# 642
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion: 94 Resp: 705075
Ion Ratio Lower Upper
94 100
65 32.3 27.4 41.2
66 41.7 37.4 56.2









#10
Bis(2-Chloroethyl)ether
Concen: 23.059 ng/ul
RT: 7.093 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion: 93 Resp: 604416
Ion Ratio Lower Upper
93 100
63 85.8 72.4 108.6
95 32.7 27.3 40.9



#12
2-Chlorophenol
Concen: 21.512 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:128 Resp: 558269
Ion Ratio Lower Upper
128 100
64 54.9 44.8 67.2
130 32.5 25.8 38.6





#13

2-Methylphenol

Concen: 21.724 ng/ul

RT: 8.110 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :





SSTD020716

Tgt Ion:108 Resp: 530576

Ion Ratio Lower Upper

108 100

107 90.2 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.216 ng/ul

RT: 8.187 min Scan# 867

Delta R.T. -0.018 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10




Tgt Ion: 45 Resp: 581366


Ion Ratio Lower Upper

45 100

77 11.8 11.2 16.8

79 8.9 6.9 10.3









#16
Acetophenone
Concen: 24.324 ng/ul
RT: 8.487 min Scan# 919
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:105 Resp: 915910
Ion Ratio Lower Upper
105 100
77 84.2 67.5 101.3
51 35.3 29.6 44.4



#17
N-Nitroso-di-n-propylamine
Concen: 27.847 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion: 70 Resp: 526115
Ion Ratio Lower Upper
70 100
42 74.4 53.8 80.8
101 9.7 7.8 11.6
130 20.0 15.4 23.2









#18
4-Methylphenol
Concen: 22.664 ng/ul
RT: 8.440 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:108 Resp: 591020
Ion Ratio Lower Upper
108 100
107 113.6 88.3 132.5



#19
Hexachloroethane
Concen: 24.200 ng/ul
RT: 8.728 min Scan# 959
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:117 Resp: 248534
Ion Ratio Lower Upper
117 100
201 81.4 85.7 128.5#
199 51.1 57.4 86.0#









#22
Nitrobenzene
Concen: 24.421 ng/ul
RT: 8.863 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion: 77 Resp: 731241
Ion Ratio Lower Upper
77 100
123 41.2 31.5 47.3
65 14.1 11.4 17.2



#23
Isophorone
Concen: 25.024 ng/ul
RT: 9.399 min Scan# 1073
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion: 82 Resp: 1402365
Ion Ratio Lower Upper
82 100
95 7.2 6.6 10.0
138 16.3 13.8 20.6









#25
2-Nitrophenol
Concen: 21.574 ng/ul
RT: 9.575 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:139 Resp: 305102
Ion Ratio Lower Upper
139 100
65 59.7 54.2 81.4
109 31.9 30.6 45.8



#26
2,4-Dimethylphenol
Concen: 22.569 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:107 Resp: 674953
Ion Ratio Lower Upper
107 100
121 50.6 39.1 58.7
122 83.6 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 23.074 ng/ul

RT: 9.881 min Scan# 1156

Delta R.T. -0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion: 93 Resp: 855723

Ion Ratio Lower Upper

93 100

95 32.4 26.6 39.8

123 10.6 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 20.766 ng/ul

RT: 10.104 min Scan# 1193

Delta R.T. -0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10




Tgt Ion:162 Resp: 495221


Ion Ratio Lower Upper

162 100

164 65.4 52.5 78.7

98 42.0 31.4 47.0









#30
Naphthalene
Concen: 21.518 ng/ul
RT: 10.504 min Scan# 1261
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:128 Resp: 1915849
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.3 10.2 15.4



#32
4-Chloroaniline
Concen: 20.634 ng/ul
RT: 10.622 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:127 Resp: 775230
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2









#33
Hexachlorobutadiene
Concen: 22.105 ng/ul
RT: 10.787 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:225 Resp: 306860
Ion Ratio Lower Upper
225 100
223 60.4 53.7 80.5
227 63.3 49.4 74.2



#34
Caprolactam
Concen: 18.725 ng/ul
RT: 11.416 min Scan# 1416
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:113 Resp: 156495
Ion Ratio Lower Upper
113 100
55 205.8 153.8 230.8
56 154.9 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 22.060 ng/ul
RT: 11.763 min Scan# 1475
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:107 Resp: 594535
Ion Ratio Lower Upper
107 100
144 24.4 18.6 27.8
142 78.0 59.6 89.4



#36
2-Methylnaphthalene
Concen: 21.963 ng/ul
RT: 12.128 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:142 Resp: 1267672
Ion Ratio Lower Upper
142 100
141 90.5 70.1 105.1









#37
1-Methylnaphthalene
Concen: 22.174 ng/ul
RT: 12.351 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:142 Resp: 1282864
Ion Ratio Lower Upper
142 100
141 91.9 73.3 109.9
116 3.8 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.355 ng/ul
RT: 12.498 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:216 Resp: 571863
Ion Ratio Lower Upper
216 100
214 78.0 63.0 94.6
179 23.5 18.1 27.1
108 22.4 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 21.839 ng/ul
RT: 12.475 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:237 Resp: 362222
Ion Ratio Lower Upper
237 100
235 63.8 52.8 79.2
272 12.8 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 20.650 ng/ul
RT: 12.746 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:196 Resp: 364166
Ion Ratio Lower Upper
196 100
198 95.8 78.2 117.4
200 31.0 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.600 ng/ul
RT: 12.816 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:196 Resp: 388246
Ion Ratio Lower Upper
196 100
198 94.0 77.0 115.4
200 31.2 24.8 37.2



#43
1,1'-Biphenyl
Concen: 21.828 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:154 Resp: 1653657
Ion Ratio Lower Upper
154 100
153 40.9 31.9 47.9
76 15.4 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 21.447 ng/ul
RT: 13.193 min Scan# 1718
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:162 Resp: 1220394
Ion Ratio Lower Upper
162 100
164 32.7 25.5 38.3
127 43.0 31.9 47.9



#45
2-Nitroaniline
Concen: 24.352 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion: 65 Resp: 403711
Ion Ratio Lower Upper
65 100
92 64.2 49.0 73.6
138 90.2 66.4 99.6









#47
Dimethylphthalate
Concen: 21.151 ng/ul
RT: 13.793 min Scan# 1820
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:163 Resp: 1480995
Ion Ratio Lower Upper
163 100
194 3.6 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 22.672 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:165 Resp: 287406
Ion Ratio Lower Upper
165 100
89 70.2 52.9 79.3
121 24.2 19.4 29.2









#50
Acenaphthylene
Concen: 22.000 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:152 Resp: 2021204
Ion Ratio Lower Upper
152 100
151 20.1 16.5 24.7
153 13.1 10.1 15.1



#51
3-Nitroaniline
Concen: 12.631 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:138 Resp: 192372
Ion Ratio Lower Upper
138 100
108 11.4 11.0 16.4
92 124.9 104.6 156.8





#52

Acenaphthene

Concen: 21.599 ng/ul

RT: 14.393 min Scan# 1922

Delta R.T. 0.000 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion:153 Resp: 1314532

Ion Ratio Lower Upper

153 100

152 47.2 36.2 54.2

154 87.3 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 15.408 ng/ul

RT: 14.451 min Scan# 1932

Delta R.T. 0.000 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10




Tgt Ion:184 Resp: 114272


Ion Ratio Lower Upper

184 100

63 92.1 73.8 110.6

154 68.0 57.2 85.8









#55
4-Nitrophenol
Concen: 17.529 ng/ul
RT: 14.557 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:109 Resp: 188483
Ion Ratio Lower Upper
109 100
139 126.3 75.6 113.4#
65 148.8 107.6 161.4



#56
Dibenzofuran
Concen: 21.329 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:168 Resp: 1791190
Ion Ratio Lower Upper
168 100
139 42.0 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 21.559 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:165 Resp: 389840
Ion Ratio Lower Upper
165 100
63 58.9 47.0 70.6
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.715 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:232 Resp: 290977
Ion Ratio Lower Upper
232 100
131 60.9 35.8 53.8#
130 3.3 1.8 2.8#
166 36.9 23.7 35.5#









#59
Diethylphthalate
Concen: 21.173 ng/ul
RT: 15.169 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:149 Resp: 1522546
Ion Ratio Lower Upper
149 100
177 20.2 17.6 26.4
150 12.5 10.3 15.5



#61
Fluorene
Concen: 20.725 ng/ul
RT: 15.381 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:166 Resp: 1396807
Ion Ratio Lower Upper
166 100
165 97.9 77.4 116.2
167 13.9 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.946 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:204 Resp: 649891
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 71.1 49.3 73.9



#63
4-Nitroaniline
Concen: 7.875 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:138 Resp: 110228
Ion Ratio Lower Upper
138 100
92 60.0 46.2 69.2
108 86.5 85.8 128.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.561 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:198 Resp: 176313
Ion Ratio Lower Upper
198 100
182 5.7 4.3 6.5
77 35.6 23.7 35.5#



#67
N-Nitrosodiphenylamine
Concen: 23.477 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:169 Resp: 1190775
Ion Ratio Lower Upper
169 100
168 67.8 52.6 78.8
167 36.0 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 23.181 ng/ul
RT: 16.287 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:248 Resp: 371227
Ion Ratio Lower Upper
248 100
250 94.8 80.8 121.2
141 88.2 63.9 95.9



#69
Hexachlorobenzene
Concen: 22.249 ng/ul
RT: 16.387 min Scan# 2261
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:284 Resp: 413518
Ion Ratio Lower Upper
284 100
142 47.6 27.8 41.6#
249 32.8 25.4 38.2









#70
Atrazine
Concen: 20.730 ng/ul
RT: 16.569 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:200 Resp: 363740
Ion Ratio Lower Upper
200 100
173 28.7 20.7 31.1
215 46.4 38.6 57.8



#71
Pentachlorophenol
Concen: 18.077 ng/ul
RT: 16.739 min Scan# 2321
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:266 Resp: 206916
Ion Ratio Lower Upper
266 100
264 63.5 51.4 77.2
268 64.2 53.5 80.3





#72

Phenanthrene

Concen: 21.494 ng/ul

RT: 17.139 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion:178 Resp: 1988329

Ion Ratio Lower Upper

178 100

179 15.0 12.7 19.1

176 19.4 14.6 22.0



#74

Anthracene

Concen: 21.446 ng/ul

RT: 17.228 min Scan# 2404

Delta R.T. -0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10




Tgt Ion:178 Resp: 1973509


Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.2

176 19.3 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.282 ng/uL
RT: 13.110 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:216 Resp: 591804
Ion Ratio Lower Upper
216 100
214 79.5 65.0 97.4
218 47.8 39.1 58.7



#76
Pentachlorobenzene
Concen: 24.057 ng/uL
RT: 14.645 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:250 Resp: 536045
Ion Ratio Lower Upper
250 100
248 62.6 49.8 74.6
252 64.0 51.2 76.8









#77
 Carbazole
 Concen: 19.225 ng/ul
 RT: 17.510 min Scan# 2452
 Delta R.T. 0.000 min
 Lab File: BP011121.D
 Acq: 27 Jul 2022 10:10




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020716


Tgt Ion:167 Resp: 1655756
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.9 25.3
 139 14.0 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 20.583 ng/ul
 RT: 18.075 min Scan# 2548
 Delta R.T. 0.000 min
 Lab File: BP011121.D
 Acq: 27 Jul 2022 10:10

Tgt Ion:149 Resp: 2215897
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 5.8 4.6 6.8









#80
Fluoranthene
Concen: 27.599 ng/ul
RT: 19.122 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:202 Resp: 1857872
Ion Ratio Lower Upper
202 100
101 14.1 7.1 10.7#
100 11.0 5.7 8.5#



#82
Pyrene
Concen: 26.654 ng/ul
RT: 19.475 min Scan# 2786
Delta R.T. -0.006 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:202 Resp: 1851193
Ion Ratio Lower Upper
202 100
101 17.5 7.9 11.9#
100 14.6 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 23.329 ng/ul
RT: 20.375 min Scan# 2939
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:149 Resp: 722806
Ion Ratio Lower Upper
149 100
91 79.4 60.0 90.0
206 18.3 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 19.740 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:252 Resp: 383054
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.4
126 15.2 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 21.195 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:228 Resp: 1414581
Ion Ratio Lower Upper
228 100
229 19.8 15.0 22.4
226 26.5 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.411 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:149 Resp: 1036049
Ion Ratio Lower Upper
149 100
167 25.4 22.6 33.8
279 3.3 4.2 6.4#





#87

Chrysene

Concen: 21.458 ng/ul

RT: 21.251 min Scan# 3088

Delta R.T. 0.000 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion:228 Resp: 1406239

Ion Ratio Lower Upper

228 100

226 29.8 23.6 35.4

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 21.206 ng/ul




RT: 22.051 min Scan# 3224


Delta R.T. 0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Tgt Ion:149 Resp: 1600668









#90
Benzo(b)fluoranthene
Concen: 20.516 ng/ul
RT: 22.845 min Scan# 3359
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020716


Tgt Ion:252 Resp: 1425728
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 13.5 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 21.041 ng/ul
RT: 22.892 min Scan# 3367
Delta R.T. 0.000 min
Lab File: BP011121.D
Acq: 27 Jul 2022 10:10

Tgt Ion:252 Resp: 1373500
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 13.4 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 21.345 ng/ul

RT: 23.457 min Scan# 3462

Delta R.T. 0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion:252 Resp: 1424562

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 14.6 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.774 ng/ul

RT: 25.986 min Scan# 3893

Delta R.T. 0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10




Tgt Ion:276 Resp: 1766332


Ion Ratio Lower Upper

276 100

138 30.9 14.6 22.0#

277 25.0 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 22.579 ng/ul

RT: 26.004 min Scan# 3895

Delta R.T. 0.006 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020716





Tgt Ion:278 Resp: 1515351

Ion Ratio Lower Upper

278 100

139 20.9 9.4 14.2#

279 23.2 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 23.066 ng/ul

RT: 26.727 min Scan# 4019

Delta R.T. 0.000 min

Lab File: BP011121.D

Acq: 27 Jul 2022 10:10

Tgt Ion:276 Resp: 1523698

Ion Ratio Lower Upper

276 100

138 27.6 12.7 19.1#

277 23.3 18.2 27.4

