

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072722\
 Data File : BP011149.D
 Acq On : 28 Jul 2022 02:54
 Operator : CG/JU
 Sample : N3871-17MS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 GBJ15MS

Quant Time: Jul 28 03:32:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 00:46:32 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
 Supervised By :mohammad ahmed 07/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	292289	20.000	ng/u1	0.00
20) Naphthalene-d8	10.452	136	1247450	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.316	164	700372	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.087	188	1446616	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.204	240	1408218	20.000	ng/u1	# 0.00
88) Perylene-d12	23.539	264	1460860	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	32930	4.062	ng/uL	0.00
4) Pyridine-d5	3.575	84	185396	8.847	ng/u1	0.00
7) Phenol-d5	6.834	99	204028	8.504	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.999	67	640439	41.511	ng/u1	0.00
11) 2-Chlorophenol-d4	7.193	132	576943	29.759	ng/u1	0.00
15) 4-Methylphenol-d8	8.375	113	372156	19.375	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	376099	41.371	ng/u1	0.00
24) 2-Nitrophenol-d4	9.540	143	360287	40.052	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	597989	36.351	ng/u1	0.00
31) 4-Chloroaniline-d4	10.599	131	797497	29.406	ng/u1	0.00
46) Dimethylphthalate-d6	13.746	166	2191861	45.790	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	2582871	45.732	ng/u1	0.00
54) 4-Nitrophenol-d4	14.540	143	88984	9.185	ng/u1	0.00
60) Fluorene-d10	15.322	176	1823785	44.824	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.451	200	327885	45.805	ng/u1	0.00
73) Anthracene-d10	17.187	188	2839995	46.179	ng/u1	0.00
81) Pyrene-d10	19.439	212	3403086	46.013	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.392	264	3417929	48.714	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	47616	5.597	ng/uL#	86
5) Pyridine	3.593	79	215513	9.979	ng/u1#	81
6) Benzaldehyde	6.805	77	647490m	55.559	ng/u1	
8) Phenol	6.864	94	226181	8.869	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.093	93	797271	39.389	ng/u1	95
12) 2-Chlorophenol	7.222	128	605937	30.236	ng/u1	98
13) 2-Methylphenol	8.111	108	438395	23.244	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.193	45	1356382	48.995	ng/u1	97
16) Acetophenone	8.487	105	1196909	41.163	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.475	70	659659	45.214	ng/u1	98
18) 4-Methylphenol	8.440	108	414478	20.583	ng/u1	97
19) Hexachloroethane	8.722	117	342582	43.198	ng/u1#	72
22) Nitrobenzene	8.863	77	956989	44.764	ng/u1	97
23) Isophorone	9.393	82	1824791	45.606	ng/u1	97
25) 2-Nitrophenol	9.569	139	396131	39.232	ng/u1	90
26) 2,4-Dimethylphenol	9.640	107	591786	27.715	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.881	93	1092844	41.272	ng/u1	97
29) 2,4-Dichlorophenol	10.099	162	612957	35.999	ng/u1	97
30) Naphthalene	10.499	128	2488748	39.151	ng/u1	99
32) 4-Chloroaniline	10.622	127	815931	30.416	ng/u1	99
33) Hexachlorobutadiene	10.781	225	381641	38.505	ng/u1	96
34) Caprolactam	11.428	113	33317	5.583	ng/u1	85
35) 4-Chloro-3-methylphenol	11.752	107	694637	36.098	ng/u1	99
36) 2-Methylnaphthalene	12.122	142	1649337	40.022	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072722\
 Data File : BP011149.D
 Acq On : 28 Jul 2022 02:54
 Operator : CG/JU
 Sample : N3871-17MS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 GBJ15MS


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
 Supervised By :mohammad ahmed 07/28/2022

Quant Time: Jul 28 03:32:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 00:46:32 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.346	142	1671339	40.461	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.493	216	732297	39.644	ng/ul#	96
40) Hexachlorocyclopentadiene	12.469	237	367522	32.123	ng/ul	96
41) 2,4,6-Trichlorophenol	12.740	196	497896	40.929	ng/ul	99
42) 2,4,5-Trichlorophenol	12.810	196	553789	42.597	ng/ul	98
43) 1,1'-Biphenyl	13.151	154	2172811	41.579	ng/ul#	97
44) 2-Chloronaphthalene	13.187	162	1616688	41.188	ng/ul	95
45) 2-Nitroaniline	13.404	65	625700	54.713	ng/ul	94
47) Dimethylphthalate	13.793	163	2184018	45.217	ng/ul	98
48) 2,6-Dinitrotoluene	13.910	165	444388	50.818	ng/ul	96
50) Acenaphthylene	14.040	152	2760082	43.553	ng/ul	99
51) 3-Nitroaniline	14.240	138	452738	43.093	ng/ul	95
52) Acenaphthene	14.387	153	1833872	43.681	ng/ul	98
53) 2,4-Dinitrophenol	14.445	184	222910	43.571	ng/ul	98
55) 4-Nitrophenol	14.551	109	81343	10.967	ng/ul#	80
56) Dibenzofuran	14.722	168	2552497	44.062	ng/ul	98
57) 2,4-Dinitrotoluene	14.698	165	663819	53.219	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.951	232	455317	44.722	ng/ul#	81
59) Diethylphthalate	15.163	149	2411820	48.622	ng/ul	97
61) Fluorene	15.375	166	2084272	44.833	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.381	204	921146	43.039	ng/ul	91
63) 4-Nitroaniline	15.416	138	513958m	53.229	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.463	198	331763	45.177	ng/ul#	89
67) N-Nitrosodiphenylamine	15.598	169	1818116	43.997	ng/ul	98
68) 4-Bromophenyl-phenylether	16.281	248	577393	44.255	ng/ul	93
69) Hexachlorobenzene	16.381	284	672638	44.421	ng/ul#	90
70) Atrazine	16.569	200	624203	43.665	ng/ul	97
71) Pentachlorophenol	16.734	266	417415	44.759	ng/ul	98
72) Phenanthrene	17.128	178	3498924	46.425	ng/ul	98
74) Anthracene	17.222	178	3432617	45.786	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	749401	36.187	ng/uL	97
76) Pentachlorobenzene	14.640	250	757517	41.728	ng/uL	99
77) Carbazole	17.498	167	3358698	47.867	ng/ul	98
78) Di-n-butylphthalate	18.069	149	4398432	50.148	ng/ul	99
80) Fluoranthene	19.116	202	4134085	47.792	ng/ul#	86
82) Pyrene	19.469	202	4175963	46.791	ng/ul#	79
83) Butylbenzylphthalate	20.363	149	2056181	51.645	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.133	252	998721	40.053	ng/ul#	97
85) Benzo(a)anthracene	21.192	228	3962565	46.204	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.128	149	3096907	52.133	ng/ul#	94
87) Chrysene	21.245	228	3885641	46.142	ng/ul	99
89) Di-n-octyl phthalate	22.033	149	5299570	54.092	ng/ul	100
90) Benzo(b)fluoranthene	22.833	252	4398764	48.766	ng/ul#	95
91) Benzo(k)fluoranthene	22.880	252	3978670	46.959	ng/ul#	94
93) Benzo(a)pyrene	23.445	252	4095901	47.283	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	25.963	276	4867980	48.355	ng/ul#	86
95) Dibenzo(a,h)anthracene	25.986	278	4154487	47.691	ng/ul#	92
96) Benzo(g,h,i)perylene	26.704	276	4183663	48.794	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 5.597 ng/uL
RT: 3.193 min Scan# 18
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion: 88 Resp: 47616
Ion Ratio Lower Upper
88 100
43 33.7 35.8 53.6
58 73.3 68.1 102.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#5
Pyridine
Concen: 9.979 ng/uL
RT: 3.593 min Scan# 86
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion: 79 Resp: 215513
Ion Ratio Lower Upper
79 100
52 81.3 83.8 125.8
51 44.4 41.0 61.4









#6
Benzaldehyde
Concen: 55.559 ng/ul m
RT: 6.805 min Scan# 633
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Instrument :
BNA_P
ClientSampleId :
GBJ15MS

Tgt Ion: 77 Resp: 647490
Ion Ratio Lower Upper
77 100
105 93.0 76.0 114.0
106 86.8 73.2 109.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#8
Phenol
Concen: 8.869 ng/ul
RT: 6.864 min Scan# 642
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion: 94 Resp: 226181
Ion Ratio Lower Upper
94 100
65 31.9 27.4 41.2
66 40.4 37.4 56.2





#10
Bis(2-Chloroethyl)ether
Concen: 39.389 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion: 93 Resp: 79727
Ion Ratio Lower Upper
93 100
63 85.6 72.4 108.6
95 32.5 27.3 40.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#12
2-Chlorophenol
Concen: 30.236 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54



Tgt Ion:128 Resp: 605937
Ion Ratio Lower Upper
128 100
64 53.3 44.8 67.2
130 32.5 25.8 38.6





#13

2-Methylphenol

Concen: 23.244 ng/ul

RT: 8.111 min Scan# 854

Delta R.T. -0.000 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54

Instrument :

BNA_P

ClientSampleId :

GBJ15MS

Tgt Ion:108 Resp: 438395

Ion Ratio Lower Upper





108 100

107 88.1 71.4 107.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 48.995 ng/ul

RT: 8.193 min Scan# 868

Delta R.T. -0.006 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54




Tgt Ion: 45 Resp: 1356382


Ion Ratio Lower Upper

45 100

77 12.1 11.2 16.8


79 9.0 6.9 10.3





#16
Acetophenone
Concen: 41.163 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion: 105 Resp: 1196909
Ion Ratio Lower Upper
105 100
77 82.2 67.5 101.3
51 32.7 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#17
N-Nitroso-di-n-propylamine
Concen: 45.214 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion: 70 Resp: 659659
Ion Ratio Lower Upper
70 100
42 69.0 53.8 80.8
101 9.4 7.8 11.6
130 19.3 15.4 23.2





#18
4-Methylphenol
Concen: 20.583 ng/ul
RT: 8.440 min Scan# 910
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:108 Resp: 414478
Ion Ratio Lower Upper
108 100
107 113.5 88.3 132.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#19
Hexachloroethane
Concen: 43.198 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:117 Resp: 342582
Ion Ratio Lower Upper
117 100
201 77.0 85.7 128.5#
199 48.7 57.4 86.0#





#22
Nitrobenzene
Concen: 44.764 ng/ul
RT: 8.863 min Scan# 982
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion: 77 Resp: 956989
Ion Ratio Lower Upper
77 100
123 42.0 31.5 47.3
65 13.9 11.4 17.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#23
Isophorone
Concen: 45.606 ng/ul
RT: 9.393 min Scan# 1072
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion: 82 Resp: 1824791
Ion Ratio Lower Upper
82 100
95 6.8 6.6 10.0
138 16.4 13.8 20.6









#25
2-Nitrophenol
Concen: 39.232 ng/ul
RT: 9.569 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Instrument :
BNA_P
ClientSampleId :
GBJ15MS

Tgt Ion:139 Resp: 396133
Ion Ratio Lower Upper
139 100
65 59.8 54.2 81.4
109 32.2 30.6 45.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#26
2,4-Dimethylphenol
Concen: 27.715 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:107 Resp: 591786
Ion Ratio Lower Upper
107 100
121 51.0 39.1 58.7
122 85.4 66.1 99.1





#27
Bis(2-Chloroethoxy)methane
Concen: 41.272 ng/ul
RT: 9.881 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion: 93 Resp: 1092844
Ion Ratio Lower Upper
93 100
95 31.9 26.6 39.8
123 11.0 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#29
2,4-Dichlorophenol
Concen: 35.999 ng/ul
RT: 10.099 min Scan# 1192
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Tgt Ion:162 Resp: 612957
Ion Ratio Lower Upper
162 100
164 63.6 52.5 78.7
98 41.6 31.4 47.0





#30
Naphthalene
Concen: 39.151 ng/ul
RT: 10.499 min Scan# 1260
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:128 Resp: 2488748
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#32
4-Chloroaniline
Concen: 30.416 ng/ul
RT: 10.622 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:127 Resp: 815931
Ion Ratio Lower Upper
127 100
129 32.3 26.2 39.2





#33
Hexachlorobutadiene
Concen: 38.505 ng/uI
RT: 10.781 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:225 Resp: 38164
Ion Ratio Lower Upper
225 100
223 64.9 53.7 80.5
227 65.8 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#34
Caprolactam
Concen: 5.583 ng/uI
RT: 11.428 min Scan# 1418
Delta R.T. 0.012 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:113 Resp: 33317
Ion Ratio Lower Upper
113 100
55 213.0 153.8 230.8
56 157.2 110.2 165.4





#35
4-Chloro-3-methylphenol
Concen: 36.098 ng/ul
RT: 11.752 min Scan# 1474
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:107 Resp: 694637
Ion Ratio Lower Upper
107 100
144 24.8 18.6 27.8
142 74.8 59.6 89.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#36
2-Methylnaphthalene
Concen: 40.022 ng/ul
RT: 12.122 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:142 Resp: 1649337
Ion Ratio Lower Upper
142 100
141 89.9 70.1 105.1





#37
1-Methylnaphthalene
Concen: 40.461 ng/ul
RT: 12.346 min Scan# 1574
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:142 Resp: 1671339
Ion Ratio Lower Upper
142 100
141 93.3 73.3 109.9
116 4.1 4.0 6.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.644 ng/ul
RT: 12.493 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:216 Resp: 732297
Ion Ratio Lower Upper
216 100
214 76.8 63.0 94.6
179 23.1 18.1 27.1
108 22.7 13.6 20.4#





#40
Hexachlorocyclopentadiene
Concen: 32.123 ng/ul
RT: 12.469 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion:237 Resp: 367522
Ion Ratio Lower Upper
237 100
235 63.3 52.8 79.2
272 12.1 11.5 17.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#41
2,4,6-Trichlorophenol
Concen: 40.929 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Tgt Ion:196 Resp: 497896
Ion Ratio Lower Upper
196 100
198 96.2 78.2 117.4
200 30.7 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 42.597 ng/ul
RT: 12.810 min Scan# 1654
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:196 Resp: 553789
Ion Ratio Lower Upper
196 100
198 93.9 77.0 115.4
200 30.4 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#43
1,1'-Biphenyl
Concen: 41.579 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:154 Resp: 2172811
Ion Ratio Lower Upper
154 100
153 40.6 31.9 47.9
76 15.6 9.2 13.8#





#44
2-Chloronaphthalene
Concen: 41.188 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:162 Resp: 1616688
Ion Ratio Lower Upper
162 100
164 33.0 25.5 38.3
127 44.2 31.9 47.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#45
2-Nitroaniline
Concen: 54.713 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion: 65 Resp: 625700
Ion Ratio Lower Upper
65 100
92 63.8 49.0 73.6
138 90.2 66.4 99.6





#47
Dimethylphthalate
Concen: 45.217 ng/ul
RT: 13.793 min Scan# 1820
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:163 Resp: 2184018
Ion Ratio Lower Upper
163 100
194 3.6 3.4 5.2
164 10.1 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#48
2,6-Dinitrotoluene
Concen: 50.818 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:165 Resp: 444388
Ion Ratio Lower Upper
165 100
89 69.8 52.9 79.3
121 24.6 19.4 29.2





#50
Acenaphthylene
Concen: 43.553 ng/ul
RT: 14.040 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:152 Resp: 2760082
Ion Ratio Lower Upper
152 100
151 19.9 16.5 24.7
153 13.2 10.1 15.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#51
3-Nitroaniline
Concen: 43.093 ng/ul
RT: 14.240 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:138 Resp: 452738
Ion Ratio Lower Upper
138 100
108 11.1 11.0 16.4
92 125.0 104.6 156.8





#52
Acenaphthene
Concen: 43.681 ng/ul
RT: 14.387 min Scan# 1921
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion:153 Resp: 1833872
Ion Ratio Lower Upper
153 100
152 47.9 36.2 54.2
154 87.7 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#53
2,4-Dinitrophenol
Concen: 43.571 ng/ul
RT: 14.445 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Tgt Ion:184 Resp: 222910
Ion Ratio Lower Upper
184 100
63 91.6 73.8 110.6
154 68.9 57.2 85.8





#55
4-Nitrophenol
Concen: 10.967 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:109 Resp: 81343
Ion Ratio Lower Upper
109 100
139 121.3 75.6 113.4
65 150.9 107.6 161.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#56
Dibenzofuran
Concen: 44.062 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion:168 Resp: 2552497
Ion Ratio Lower Upper
168 100
139 42.0 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 53.219 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Instrument :
BNA_P
ClientSampleId :
GBJ15MS

Tgt Ion:165 Resp: 663819
Ion Ratio Lower Upper
165 100
63 58.0 47.0 70.6
182 2.6 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 44.722 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:232 Resp: 455317
Ion Ratio Lower Upper
232 100
131 59.8 35.8 53.8#
130 3.3 1.8 2.8#
166 36.6 23.7 35.5#





#59
Diethylphthalate
Concen: 48.622 ng/ul
RT: 15.163 min Scan# 2053
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion:149 Resp: 2411820
Ion Ratio Lower Upper
149 100
177 19.9 17.6 26.4
150 12.7 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#61
Fluorene
Concen: 44.833 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Tgt Ion:166 Resp: 2084272
Ion Ratio Lower Upper
166 100
165 98.4 77.4 116.2
167 13.8 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 43.039 ng/ul
RT: 15.381 min Scan# 2089
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:204 Resp: 921146
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 72.0 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#63
4-Nitroaniline
Concen: 53.229 ng/ul m
RT: 15.416 min Scan# 2096
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:138 Resp: 513958
Ion Ratio Lower Upper
138 100
92 53.7 46.2 69.2
108 82.4 85.8 128.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 45.177 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:198 Resp: 331763
Ion Ratio Lower Upper
198 100
182 5.3 4.3 6.5
77 36.9 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#67
N-Nitrosodiphenylamine
Concen: 43.997 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:169 Resp: 1818116
Ion Ratio Lower Upper
169 100
168 67.0 52.6 78.8
167 36.5 27.8 41.6





#68
4-Bromophenyl-phenylether
Concen: 44.255 ng/ul
RT: 16.281 min Scan# 2243
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:248 Resp: 577393
Ion Ratio Lower Upper
248 100
250 97.0 80.8 121.2
141 89.0 63.9 95.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#69
Hexachlorobenzene
Concen: 44.421 ng/ul
RT: 16.381 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion:284 Resp: 672638
Ion Ratio Lower Upper
284 100
142 45.8 27.8 41.6#
249 31.8 25.4 38.2





#70

Atrazine

Concen: 43.665 ng/ul

RT: 16.569 min Scan# 2291

Delta R.T. 0.006 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54

Instrument :

BNA_P

ClientSampleId :





GBJ15MS

Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.2	20.7	31.1
215	46.4	38.6	57.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#71

Pentachlorophenol

Concen: 44.759 ng/ul




RT: 16.734 min Scan# 2320


Delta R.T. -0.000 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54


Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.2	51.4	77.2
268	64.3	53.5	80.3





#72
Phenanthrene
Concen: 46.425 ng/ul
RT: 17.128 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion:178 Resp: 3498924
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 19.3 14.6 22.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#74
Anthracene
Concen: 45.786 ng/ul
RT: 17.222 min Scan# 2403
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54



Tgt Ion:178 Resp: 3432617
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.5 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.187 ng/uL
RT: 13.104 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Instrument :
BNA_P
ClientSampleId :
GBJ15MS

Tgt Ion:216 Resp: 749401
Ion Ratio Lower Upper
216 100
214 78.3 65.0 97.4
218 47.8 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#76
Pentachlorobenzene
Concen: 41.728 ng/uL
RT: 14.640 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:250 Resp: 757517
Ion Ratio Lower Upper
250 100
248 63.7 49.8 74.6
252 64.2 51.2 76.8





#77
 Carbazole
 Concen: 47.867 ng/ul
 RT: 17.498 min Scan# 2450
 Delta R.T. -0.000 min
 Lab File: BP011149.D
 Acq: 28 Jul 2022 02:54




Instrument :
 BNA_P
 ClientSampleId :
 GBJ15MS




Tgt Ion:167 Resp: 3358698
 Ion Ratio Lower Upper
 167 100
 166 21.9 16.9 25.3
 139 14.4 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
 Supervised By :mohammad ahmed 07/28/2022




#78
 Di-n-butylphthalate
 Concen: 50.148 ng/ul
 RT: 18.069 min Scan# 2547
 Delta R.T. -0.000 min
 Lab File: BP011149.D
 Acq: 28 Jul 2022 02:54




Tgt Ion:149 Resp: 4398432
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.8 4.6 6.8





#80
Fluoranthene
Concen: 47.792 ng/ul
RT: 19.116 min Scan# 2725
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS




Tgt Ion:202 Resp: 4134085
Ion Ratio Lower Upper
202 100
101 14.6 7.1 10.7
100 11.2 5.7 8.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022




#82
Pyrene
Concen: 46.791 ng/ul
RT: 19.469 min Scan# 2785
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Tgt Ion:202 Resp: 4175963
Ion Ratio Lower Upper
202 100
101 18.1 7.9 11.9#
100 15.0 6.7 10.1#





#83
Butylbenzylphthalate
Concen: 51.645 ng/ul
RT: 20.363 min Scan# 2937
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:149 Resp: 2056181
Ion Ratio Lower Upper
149 100
91 81.1 60.0 90.0
206 17.6 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#84
3,3'-Dichlorobenzidine
Concen: 40.053 ng/ul
RT: 21.133 min Scan# 3068
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion:252 Resp: 998721
Ion Ratio Lower Upper
252 100
254 65.2 51.0 76.4
126 14.8 8.6 12.8





#85

Benzo(a)anthracene

Concen: 46.204 ng/ul

RT: 21.192 min Scan# 3078

Delta R.T. -0.000 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54

Instrument :

BNA_P

ClientSampleId :

GBJ15MS

Tgt Ion:228 Resp: 396256

Ion Ratio Lower Upper

228 100





229 19.8 15.0 22.4

226 27.2 20.6 31.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#86

Bis(2-ethylhexyl)phthalate

Concen: 52.133 ng/ul

RT: 21.128 min Scan# 3067

Delta R.T. -0.000 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54




Tgt Ion:149 Resp: 3096907


Ion Ratio Lower Upper

149 100

167 25.2 22.6 33.8

279 3.0 4.2 6.4#





#87

Chrysene

Concen: 46.142 ng/ul

RT: 21.245 min Scan# 3086

Delta R.T. 0.006 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54

Instrument :

BNA_P

ClientSampleId :

GBJ15MS

Tgt Ion:228 Resp: 388564

Ion Ratio Lower Upper

228 100





226 30.3 23.6 35.4

229 20.1 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#89

Di-n-octyl phthalate

Concen: 54.092 ng/ul




RT: 22.033 min Scan# 3221


Delta R.T. -0.006 min

Lab File: BP011149.D

Acq: 28 Jul 2022 02:54


Tgt Ion:149 Resp: 5299570





#90
Benzo(b)fluoranthene
Concen: 48.766 ng/ul
RT: 22.833 min Scan# 3356
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:252 Resp: 4398764
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 13.8 6.2 9.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#91
Benzo(k)fluoranthene
Concen: 46.959 ng/ul
RT: 22.880 min Scan# 3365
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54


Tgt Ion:252 Resp: 3978670
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 13.6 6.2 9.4





#93
Benzo(a)pyrene
Concen: 47.283 ng/ul
RT: 23.445 min Scan# 3461
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54




Instrument :
BNA_P
ClientSampleId :
GBJ15MS



Tgt Ion:252 Resp: 409590
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 15.0 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 48.355 ng/ul
RT: 25.963 min Scan# 3889
Delta R.T. -0.000 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion:276 Resp: 4867980
Ion Ratio Lower Upper
276 100
138 31.8 14.6 22.0
277 25.4 19.6 29.4





#95
Dibenzo(a,h)anthracene
Concen: 47.691 ng/ul
RT: 25.986 min Scan# 3892
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Instrument :

BNA_P

ClientSampleId :





GBJ15MS

Tgt Ion:278 Resp: 415448
Ion Ratio Lower Upper
278 100
139 21.0 9.4 14.2
279 23.9 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 07/28/2022
Supervised By :mohammad ahmed 07/28/2022



#96
Benzo(g,h,i)perylene
Concen: 48.794 ng/ul
RT: 26.704 min Scan# 4015
Delta R.T. 0.006 min
Lab File: BP011149.D
Acq: 28 Jul 2022 02:54

Tgt Ion:276 Resp: 4183663
Ion Ratio Lower Upper
276 100
138 29.4 12.7 19.1
277 24.2 18.2 27.4

