

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011160.D
 Acq On : 28 Jul 2022 13:55
 Operator : CG/JU
 Sample : SST04027
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040627

Quant Time: Jul 28 14:30:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	328464	20.000	ng/ul	0.00
20) Naphthalene-d8	10.446	136	1602332	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.316	164	911277	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.086	188	1523414	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.210	240	945391	20.000	ng/ul	# 0.00
88) Perylene-d12	23.551	264	1014882	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	114207	12.535	ng/uL	0.00
4) Pyridine-d5	3.558	84	853105	36.226	ng/ul	0.00
7) Phenol-d5	6.834	99	1267284	47.002	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.999	67	902077	52.030	ng/ul	0.00
11) 2-Chlorophenol-d4	7.187	132	1016892	46.675	ng/ul	0.00
15) 4-Methylphenol-d8	8.375	113	1097941	50.866	ng/ul	0.00
21) Nitrobenzene-d5	8.816	128	547431	46.881	ng/ul	0.00
24) 2-Nitrophenol-d4	9.534	143	557395	48.241	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.069	165	952840	45.093	ng/ul	0.00
31) 4-Chloroaniline-d4	10.593	131	1495083	42.918	ng/ul	0.00
46) Dimethylphthalate-d6	13.740	166	2728538	43.809	ng/ul	0.00
49) Acenaphthylene-d8	14.010	160	3383330	46.041	ng/ul	0.00
54) 4-Nitrophenol-d4	14.545	143	446064	35.387	ng/ul	0.01
60) Fluorene-d10	15.322	176	2231843	42.158	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.451	200	336876	44.689	ng/ul	0.00
73) Anthracene-d10	17.186	188	2836458	43.797	ng/ul	0.00
81) Pyrene-d10	19.445	212	2490297	50.155	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.404	264	2210588	45.351	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	115350	12.066	ng/uL#	85
5) Pyridine	3.575	79	871343	35.903	ng/ul#	77
6) Benzaldehyde	6.799	77	753002	57.497	ng/ul	99
8) Phenol	6.858	94	1361225	47.496	ng/ul	92
10) Bis(2-Chloroethyl)ether	7.087	93	1113987	48.975	ng/ul	96
12) 2-Chlorophenol	7.216	128	1048701	46.567	ng/ul	98
13) 2-Methylphenol	8.105	108	1049904	49.537	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.193	45	1906349	61.277	ng/ul	96
16) Acetophenone	8.481	105	1706365	52.221	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.475	70	955665	58.289	ng/ul	99
18) 4-Methylphenol	8.440	108	1159325	51.231	ng/ul	99
19) Hexachloroethane	8.722	117	449892	50.481	ng/ul#	75
22) Nitrobenzene	8.857	77	1377927	50.178	ng/ul	97
23) Isophorone	9.393	82	2681950	52.184	ng/ul	98
25) 2-Nitrophenol	9.563	139	610496	47.071	ng/ul	90
26) 2,4-Dimethylphenol	9.634	107	1287595	46.946	ng/ul	98
27) Bis(2-Chloroethoxy)met...	9.875	93	1611185	47.372	ng/ul	98
29) 2,4-Dichlorophenol	10.099	162	967274	44.226	ng/ul	98
30) Naphthalene	10.493	128	3526813	43.193	ng/ul	99
32) 4-Chloroaniline	10.616	127	1469388	42.644	ng/ul	99
33) Hexachlorobutadiene	10.775	225	546001	42.887	ng/ul	96
34) Caprolactam	11.422	113	208263	27.171	ng/ul	94
35) 4-Chloro-3-methylphenol	11.757	107	1175276	47.549	ng/ul	98
36) 2-Methylnaphthalene	12.116	142	2368871	44.751	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011160.D
 Acq On : 28 Jul 2022 13:55
 Operator : CG/JU
 Sample : SST04027
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040627

Quant Time: Jul 28 14:30:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.340	142	2384759	44.945	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.493	216	1034682	43.050	ng/ul#	97
40) Hexachlorocyclopentadiene	12.469	237	692493	46.519	ng/ul	95
41) 2,4,6-Trichlorophenol	12.740	196	709193	44.806	ng/ul	99
42) 2,4,5-Trichlorophenol	12.810	196	765666	45.264	ng/ul	99
43) 1,1'-Biphenyl	13.145	154	3027608	44.527	ng/ul#	97
44) 2-Chloronaphthalene	13.187	162	2243620	43.931	ng/ul	97
45) 2-Nitroaniline	13.404	65	776807	52.206	ng/ul	92
47) Dimethylphthalate	13.792	163	2709879	43.120	ng/ul	99
48) 2,6-Dinitrotoluene	13.910	165	551516	48.473	ng/ul	99
50) Acenaphthylene	14.040	152	3681696	44.650	ng/ul	99
51) 3-Nitroaniline	14.240	138	564220	41.275	ng/ul	95
52) Acenaphthene	14.381	153	2400955	43.953	ng/ul	98
53) 2,4-Dinitrophenol	14.445	184	254515	38.235	ng/ul	97
55) 4-Nitrophenol	14.557	109	377740	39.141	ng/ul#	79
56) Dibenzofuran	14.722	168	3206369	42.539	ng/ul	98
57) 2,4-Dinitrotoluene	14.698	165	724112	44.617	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	14.951	232	552454	41.704	ng/ul#	82
59) Diethylphthalate	15.169	149	2768103	42.889	ng/ul	97
61) Fluorene	15.375	166	2498205	41.300	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.375	204	1132914	40.683	ng/ul#	87
63) 4-Nitroaniline	15.416	138	426995	33.988	ng/ul#	82
66) 4,6-Dinitro-2-methylph...	15.463	198	339902	43.952	ng/ul	92
67) N-Nitrosodiphenylamine	15.598	169	2097811	48.206	ng/ul	98
68) 4-Bromophenyl-phenylether	16.281	248	671318	48.860	ng/ul	93
69) Hexachlorobenzene	16.381	284	727915	45.648	ng/ul#	89
70) Atrazine	16.569	200	640018	42.515	ng/ul	96
71) Pentachlorophenol	16.733	266	404064	41.144	ng/ul	98
72) Phenanthrene	17.133	178	3458712	43.578	ng/ul	98
74) Anthracene	17.222	178	3425424	43.387	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	1092707	50.104	ng/ul	97
76) Pentachlorobenzene	14.634	250	956720	50.044	ng/ul	99
77) Carbazole	17.504	167	2865837	38.784	ng/ul	98
78) Di-n-butylphthalate	18.069	149	3811571	41.266	ng/ul	99
80) Fluoranthene	19.116	202	3144064	54.141	ng/ul#	84
82) Pyrene	19.475	202	3078244	51.377	ng/ul#	83
83) Butylbenzylphthalate	20.369	149	1266123	47.370	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.139	252	722950	43.187	ng/ul#	96
85) Benzo(a)anthracene	21.198	228	2532586	43.988	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.133	149	1859050	46.616	ng/ul#	95
87) Chrysene	21.245	228	2455914	43.442	ng/ul	98
89) Di-n-octyl phthalate	22.045	149	3080937	45.266	ng/ul	100
90) Benzo(b)fluoranthene	22.839	252	2743524	43.781	ng/ul#	95
91) Benzo(k)fluoranthene	22.886	252	2591062	44.020	ng/ul#	93
93) Benzo(a)pyrene	23.451	252	2708095	45.000	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	25.974	276	3380261	48.333	ng/ul#	88
95) Dibenzo(a,h)anthracene	25.998	278	2892961	47.803	ng/ul#	91
96) Benzo(g,h,i)perylene	26.721	276	2941368	49.380	ng/ul#	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011160.D
 Acq On : 28 Jul 2022 13:55
 Operator : CG/JU
 Sample : SST040627
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040627

Quant Time: Jul 28 14:30:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 12.066 ng/uL
RT: 3.181 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion: 88 Resp: 115350
Ion Ratio Lower Upper
88 100
43 31.1 35.8 53.6#
58 74.2 68.1 102.1



#5
Pyridine
Concen: 35.903 ng/uL
RT: 3.575 min Scan# 83
Delta R.T. -0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion: 79 Resp: 871343
Ion Ratio Lower Upper
79 100
52 78.2 83.8 125.8#
51 39.1 41.0 61.4#









#6
Benzaldehyde
Concen: 57.497 ng/ul
RT: 6.799 min Scan# 611
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion: 77 Resp: 753002
Ion Ratio Lower Upper
77 100
105 94.4 76.0 114.0
106 89.6 73.2 109.8



#8
Phenol
Concen: 47.496 ng/ul
RT: 6.858 min Scan# 641
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion: 94 Resp: 1361225
Ion Ratio Lower Upper
94 100
65 31.2 27.4 41.2
66 40.0 37.4 56.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 48.975 ng/ul
 RT: 7.087 min Scan# 680
 Delta R.T. 0.000 min
 Lab File: BP011160.D
 Acq: 28 Jul 2022 13:55

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040627

Tgt Ion: 93 Resp: 1113987




Ion	Ratio	Lower	Upper
93	100		
63	86.2	72.4	108.6
95	32.5	27.3	40.9




#12
 2-Chlorophenol
 Concen: 46.567 ng/ul
 RT: 7.216 min Scan# 702
 Delta R.T. 0.000 min
 Lab File: BP011160.D
 Acq: 28 Jul 2022 13:55

Tgt Ion: 128 Resp: 1048701

Ion	Ratio	Lower	Upper
128	100		
64	53.4	44.8	67.2
130	32.6	25.8	38.6









#13
2-Methylphenol
Concen: 49.537 ng/ul
RT: 8.105 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:108 Resp: 1049904
Ion Ratio Lower Upper
108 100
107 89.3 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 61.277 ng/ul
RT: 8.193 min Scan# 868
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion: 45 Resp: 1906349
Ion Ratio Lower Upper
45 100
77 12.3 11.2 16.8
79 9.5 6.9 10.3









#16
Acetophenone
Concen: 52.221 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:105 Resp: 1706365
Ion Ratio Lower Upper
105 100
77 82.0 67.5 101.3
51 33.1 29.6 44.4



#17
N-Nitroso-di-n-propylamine
Concen: 58.289 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion: 70 Resp: 955665
Ion Ratio Lower Upper
70 100
42 68.2 53.8 80.8
101 9.2 7.8 11.6
130 19.0 15.4 23.2









#18
4-Methylphenol
Concen: 51.231 ng/ul
RT: 8.440 min Scan# 910
Delta R.T. 0.012 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:108 Resp: 1159325
Ion Ratio Lower Upper
108 100
107 111.1 88.3 132.5



#19
Hexachloroethane
Concen: 50.481 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:117 Resp: 449892
Ion Ratio Lower Upper
117 100
201 80.6 85.7 128.5#
199 52.0 57.4 86.0#









#22
Nitrobenzene
Concen: 50.178 ng/ul
RT: 8.857 min Scan# 981
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion: 77 Resp: 1377927
Ion Ratio Lower Upper
77 100
123 41.5 31.5 47.3
65 14.0 11.4 17.2



#23
Isophorone
Concen: 52.184 ng/ul
RT: 9.393 min Scan# 1072
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion: 82 Resp: 2681950
Ion Ratio Lower Upper
82 100
95 7.0 6.6 10.0
138 16.5 13.8 20.6









#25
2-Nitrophenol
Concen: 47.071 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:139 Resp: 610496
Ion Ratio Lower Upper
139 100
65 61.0 54.2 81.4
109 30.8 30.6 45.8



#26
2,4-Dimethylphenol
Concen: 46.946 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:107 Resp: 1287595
Ion Ratio Lower Upper
107 100
121 50.0 39.1 58.7
122 84.6 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 47.372 ng/ul
 RT: 9.875 min Scan# 1154
 Delta R.T. 0.006 min
 Lab File: BP011160.D
 Acq: 28 Jul 2022 13:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040627


Tgt Ion: 93 Resp: 1611185
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.6 39.8
 123 10.8 7.8 11.8



#29
 2,4-Dichlorophenol
 Concen: 44.226 ng/ul
 RT: 10.099 min Scan# 1192
 Delta R.T. 0.006 min
 Lab File: BP011160.D
 Acq: 28 Jul 2022 13:55

Tgt Ion:162 Resp: 967274
 Ion Ratio Lower Upper
 162 100
 164 64.7 52.5 78.7
 98 40.9 31.4 47.0









#30
Naphthalene
Concen: 43.193 ng/ul
RT: 10.493 min Scan# 1259
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:128 Resp: 3526813
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.4 10.2 15.4



#32
4-Chloroaniline
Concen: 42.644 ng/ul
RT: 10.616 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:127 Resp: 1469388
Ion Ratio Lower Upper
127 100
129 32.0 26.2 39.2









#33
Hexachlorobutadiene
Concen: 42.887 ng/ul
RT: 10.775 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:225 Resp: 546001
Ion Ratio Lower Upper
225 100
223 62.2 53.7 80.5
227 62.6 49.4 74.2



#34
Caprolactam
Concen: 27.171 ng/ul
RT: 11.422 min Scan# 1417
Delta R.T. 0.012 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:113 Resp: 208263
Ion Ratio Lower Upper
113 100
55 201.9 153.8 230.8
56 144.2 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 47.549 ng/ul
RT: 11.757 min Scan# 1473
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:107 Resp: 1175276
Ion Ratio Lower Upper
107 100
144 24.7 18.6 27.8
142 75.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 44.751 ng/ul
RT: 12.116 min Scan# 1535
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:142 Resp: 2368871
Ion Ratio Lower Upper
142 100
141 89.1 70.1 105.1





#37

1-Methylnaphthalene

Concen: 44.945 ng/ul

RT: 12.340 min Scan# 1573

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:142 Resp: 2384759

Ion Ratio Lower Upper

142 100

141 91.9 73.3 109.9

116 4.1 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 43.050 ng/ul

RT: 12.493 min Scan# 1599

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Tgt Ion:216 Resp: 1034682




Ion Ratio Lower Upper


216 100

214 80.1 63.0 94.6

179 23.4 18.1 27.1

108 22.3 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 46.519 ng/ul
RT: 12.469 min Scan# 1594
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:237 Resp: 692493
Ion Ratio Lower Upper
237 100
235 62.1 52.8 79.2
272 12.4 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 44.806 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:196 Resp: 709193
Ion Ratio Lower Upper
196 100
198 97.2 78.2 117.4
200 30.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 45.264 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:196 Resp: 765666
Ion Ratio Lower Upper
196 100
198 95.0 77.0 115.4
200 31.1 24.8 37.2



#43
1,1'-Biphenyl
Concen: 44.527 ng/ul
RT: 13.145 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:154 Resp: 3027608
Ion Ratio Lower Upper
154 100
153 40.4 31.9 47.9
76 15.8 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 43.931 ng/ul

RT: 13.187 min Scan# 1716

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:162 Resp: 2243620

Ion Ratio Lower Upper

162 100

164 32.4 25.5 38.3

127 42.6 31.9 47.9



#45

2-Nitroaniline

Concen: 52.206 ng/ul

RT: 13.404 min Scan# 1754

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion: 65 Resp: 776807


Ion Ratio Lower Upper

65 100

92 64.6 49.0 73.6

138 92.3 66.4 99.6









#47
Dimethylphthalate
Concen: 43.120 ng/ul
RT: 13.792 min Scan# 1819
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:163 Resp: 2709879
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.0 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 48.473 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:165 Resp: 551516
Ion Ratio Lower Upper
165 100
89 67.1 52.9 79.3
121 24.3 19.4 29.2





#50

Acenaphthylene

Concen: 44.650 ng/ul

RT: 14.040 min Scan# 1861

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:152 Resp: 3681696

Ion Ratio Lower Upper

152 100

151 20.5 16.5 24.7

153 13.3 10.1 15.1



#51

3-Nitroaniline

Concen: 41.275 ng/ul

RT: 14.240 min Scan# 1896

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:138 Resp: 564220


Ion Ratio Lower Upper

138 100

108 11.1 11.0 16.4

92 124.9 104.6 156.8









#52
Acenaphthene
Concen: 43.953 ng/ul
RT: 14.381 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:153 Resp: 2400955
Ion Ratio Lower Upper
153 100
152 47.9 36.2 54.2
154 87.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 38.235 ng/ul
RT: 14.445 min Scan# 1931
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:184 Resp: 254515
Ion Ratio Lower Upper
184 100
63 90.0 73.8 110.6
154 68.0 57.2 85.8









#55
4-Nitrophenol
Concen: 39.141 ng/ul
RT: 14.557 min Scan# 1949
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:109 Resp: 377740
Ion Ratio Lower Upper
109 100
139 128.7 75.6 113.4#
65 147.7 107.6 161.4



#56
Dibenzofuran
Concen: 42.539 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:168 Resp: 3206369
Ion Ratio Lower Upper
168 100
139 42.2 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 44.617 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:165 Resp: 724112
Ion Ratio Lower Upper
165 100
63 56.7 47.0 70.6
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.704 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:232 Resp: 552454
Ion Ratio Lower Upper
232 100
131 59.1 35.8 53.8#
130 3.4 1.8 2.8#
166 37.1 23.7 35.5#









#59
Diethylphthalate
Concen: 42.889 ng/ul
RT: 15.169 min Scan# 2053
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:149 Resp: 2768103
Ion Ratio Lower Upper
149 100
177 20.0 17.6 26.4
150 12.1 10.3 15.5



#61
Fluorene
Concen: 41.300 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:166 Resp: 2498205
Ion Ratio Lower Upper
166 100
165 96.7 77.4 116.2
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 40.683 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:204 Resp: 1132914
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 76.0 49.3 73.9#



#63
4-Nitroaniline
Concen: 33.988 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.012 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:138 Resp: 426995
Ion Ratio Lower Upper
138 100
92 54.4 46.2 69.2
108 81.5 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 43.952 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:198 Resp: 339902
Ion Ratio Lower Upper
198 100
182 5.4 4.3 6.5
77 34.8 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 48.206 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:169 Resp: 2097811
Ion Ratio Lower Upper
169 100
168 67.3 52.6 78.8
167 36.5 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 48.860 ng/ul
RT: 16.281 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:248 Resp: 671318
Ion Ratio Lower Upper
248 100
250 96.3 80.8 121.2
141 89.4 63.9 95.9



#69
Hexachlorobenzene
Concen: 45.648 ng/ul
RT: 16.381 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:284 Resp: 727915
Ion Ratio Lower Upper
284 100
142 46.5 27.8 41.6#
249 32.4 25.4 38.2





#70

Atrazine

Concen: 42.515 ng/ul

RT: 16.569 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:200 Resp: 640018

Ion Ratio Lower Upper

200 100

173 28.3 20.7 31.1

215 45.7 38.6 57.8



#71

Pentachlorophenol

Concen: 41.144 ng/ul

RT: 16.733 min Scan# 2320

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:266 Resp: 404064


Ion Ratio Lower Upper

266 100

264 63.7 51.4 77.2

268 64.3 53.5 80.3





#72

Phenanthrene

Concen: 43.578 ng/ul

RT: 17.133 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:178 Resp: 3458712

Ion Ratio Lower Upper

178 100

179 15.3 12.7 19.1

176 19.1 14.6 22.0



#74

Anthracene

Concen: 43.387 ng/ul

RT: 17.222 min Scan# 2403

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:178 Resp: 3425424


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.2

176 18.6 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 50.104 ng/uL
RT: 13.104 min Scan# 1703
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:216 Resp: 1092707
Ion Ratio Lower Upper
216 100
214 78.2 65.0 97.4
218 48.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 50.044 ng/uL
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:250 Resp: 956720
Ion Ratio Lower Upper
250 100
248 62.8 49.8 74.6
252 63.5 51.2 76.8





#77

Carbazole

Concen: 38.784 ng/ul

RT: 17.504 min Scan# 2450

Delta R.T. 0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:167 Resp: 2865837

Ion Ratio Lower Upper

167 100

166 21.9 16.9 25.3

139 14.1 10.6 16.0



#78

Di-n-butylphthalate

Concen: 41.266 ng/ul

RT: 18.069 min Scan# 2547

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:149 Resp: 3811571


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 54.141 ng/ul
RT: 19.116 min Scan# 2725
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:202 Resp: 3144064
Ion Ratio Lower Upper
202 100
101 15.2 7.1 10.7#
100 11.8 5.7 8.5#



#82
Pyrene
Concen: 51.377 ng/ul
RT: 19.475 min Scan# 2786
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:202 Resp: 3078244
Ion Ratio Lower Upper
202 100
101 16.8 7.9 11.9#
100 13.7 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 47.370 ng/ul

RT: 20.369 min Scan# 2938

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:149 Resp: 1266123

Ion Ratio Lower Upper

149 100

91 81.2 60.0 90.0

206 17.6 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 43.187 ng/ul

RT: 21.139 min Scan# 3069

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:252 Resp: 722950


Ion Ratio Lower Upper

252 100

254 66.0 51.0 76.4

126 14.1 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 43.988 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:228 Resp: 2532586
Ion Ratio Lower Upper
228 100
229 19.8 15.0 22.4
226 27.3 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 46.616 ng/ul
RT: 21.133 min Scan# 3068
Delta R.T. -0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:149 Resp: 1859050
Ion Ratio Lower Upper
149 100
167 25.5 22.6 33.8
279 3.1 4.2 6.4#









#87
Chrysene
Concen: 43.442 ng/ul
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:228 Resp: 2455914
Ion Ratio Lower Upper
228 100
226 30.4 23.6 35.4
229 20.1 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 45.266 ng/ul
RT: 22.045 min Scan# 3223
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:149 Resp: 3080937









#90
Benzo(b)fluoranthene
Concen: 43.781 ng/ul
RT: 22.839 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD040627


Tgt Ion:252 Resp: 2743524
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 14.1 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 44.020 ng/ul
RT: 22.886 min Scan# 3366
Delta R.T. -0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:252 Resp: 2591062
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 14.2 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 45.000 ng/ul

RT: 23.451 min Scan# 3462

Delta R.T. 0.000 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD040627





Tgt Ion:252 Resp: 2708095

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 15.0 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 48.333 ng/ul

RT: 25.974 min Scan# 3891

Delta R.T. -0.006 min

Lab File: BP011160.D

Acq: 28 Jul 2022 13:55




Tgt Ion:276 Resp: 3380261


Ion Ratio Lower Upper

276 100

138 30.4 14.6 22.0#

277 24.7 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 47.803 ng/ul
RT: 25.998 min Scan# 3895
Delta R.T. -0.006 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Instrument :
BNA_P
ClientSampleId :
SSTD040627

Tgt Ion:278 Resp: 2892961
Ion Ratio Lower Upper
278 100
139 21.4 9.4 14.2#
279 24.1 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 49.380 ng/ul
RT: 26.721 min Scan# 4018
Delta R.T. 0.000 min
Lab File: BP011160.D
Acq: 28 Jul 2022 13:55

Tgt Ion:276 Resp: 2941368
Ion Ratio Lower Upper
276 100
138 28.7 12.7 19.1#
277 23.9 18.2 27.4

