

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011161.D
 Acq On : 28 Jul 2022 14:29
 Operator : CG/JU
 Sample : SSTD08028
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080628

Quant Time: Jul 28 15:43:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	348992	20.000	ng/ul	0.00
20) Naphthalene-d8	10.446	136	1538867	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.322	164	715938	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.092	188	1137306	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.216	240	963320	20.000	ng/ul	# 0.00
88) Perylene-d12	23.557	264	1102576	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	236434	24.424	ng/uL	0.00
4) Pyridine-d5	3.558	84	1731266	69.191	ng/ul	0.00
7) Phenol-d5	6.840	99	2581255	90.104	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	6.999	67	1795509	97.469	ng/ul	0.00
11) 2-Chlorophenol-d4	7.187	132	2083920	90.026	ng/ul	0.00
15) 4-Methylphenol-d8	8.381	113	2097796	91.471	ng/ul	0.01
21) Nitrobenzene-d5	8.822	128	1059540	94.480	ng/ul	0.01
24) 2-Nitrophenol-d4	9.540	143	1084012	97.687	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	1735147	85.502	ng/ul	0.00
31) 4-Chloroaniline-d4	10.599	131	2482698	74.209	ng/ul	0.00
46) Dimethylphthalate-d6	13.746	166	3927472	80.264	ng/ul	0.00
49) Acenaphthylene-d8	14.010	160	5164018	89.446	ng/ul	0.00
54) 4-Nitrophenol-d4	14.551	143	679940	68.659	ng/ul	0.02
60) Fluorene-d10	15.322	176	3180723	76.474	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	521447	92.657	ng/ul	0.01
73) Anthracene-d10	17.192	188	4099655	84.792	ng/ul	0.00
81) Pyrene-d10	19.451	212	4232919	83.665	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.410	264	4676367	88.307	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	246429	24.262	ng/uL#	84
5) Pyridine	3.576	79	1766599	68.510	ng/ul#	76
6) Benzaldehyde	6.805	77	1170450	84.115	ng/ul	100
8) Phenol	6.864	94	2739100	89.951	ng/ul	92
10) Bis(2-Chloroethyl)ether	7.093	93	2199573	91.014	ng/ul	96
12) 2-Chlorophenol	7.222	128	2139574	89.418	ng/ul	97
13) 2-Methylphenol	8.105	108	2051306	91.092	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.193	45	3679041	111.302	ng/ul	97
16) Acetophenone	8.487	105	3096067	89.178	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.493	70	1682886	96.607	ng/ul	100
18) 4-Methylphenol	8.452	108	2161402	89.895	ng/ul	99
19) Hexachloroethane	8.722	117	923966	97.577	ng/ul#	75
22) Nitrobenzene	8.863	77	2564989	97.258	ng/ul	96
23) Isophorone	9.399	82	4595823	93.111	ng/ul	98
25) 2-Nitrophenol	9.569	139	1163629	93.419	ng/ul	90
26) 2,4-Dimethylphenol	9.640	107	2314687	87.875	ng/ul	99
27) Bis(2-Chloroethoxy)met...	9.881	93	2921210	89.431	ng/ul	98
29) 2,4-Dichlorophenol	10.105	162	1739890	82.833	ng/ul	98
30) Naphthalene	10.499	128	6420429	81.874	ng/ul	99
32) 4-Chloroaniline	10.622	127	2446549	73.932	ng/ul	100
33) Hexachlorobutadiene	10.781	225	1049746	85.856	ng/ul	96
34) Caprolactam	11.446	113	286419	38.909	ng/ul	94
35) 4-Chloro-3-methylphenol	11.763	107	1932960	81.428	ng/ul	98
36) 2-Methylnaphthalene	12.122	142	4071310	80.084	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011161.D
 Acq On : 28 Jul 2022 14:29
 Operator : CG/JU
 Sample : SSTD08028
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD080628


Quant Time: Jul 28 15:43:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.346	142	4007757	78.649	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.493	216	1729686	91.602	ng/ul#	96
40) Hexachlorocyclopentadiene	12.469	237	1273082	108.855	ng/ul	96
41) 2,4,6-Trichlorophenol	12.746	196	1154611	92.851	ng/ul	98
42) 2,4,5-Trichlorophenol	12.816	196	1188981	89.467	ng/ul	100
43) 1,1'-Biphenyl	13.151	154	4856834	90.919	ng/ul#	97
44) 2-Chloronaphthalene	13.187	162	3588284	89.429	ng/ul	94
45) 2-Nitroaniline	13.410	65	1188445	101.662	ng/ul	91
47) Dimethylphthalate	13.793	163	3894161	78.870	ng/ul	98
48) 2,6-Dinitrotoluene	13.916	165	813835	91.043	ng/ul	99
50) Acenaphthylene	14.040	152	5542832	85.561	ng/ul	99
51) 3-Nitroaniline	14.246	138	792642	73.806	ng/ul#	93
52) Acenaphthene	14.387	153	3620969	84.374	ng/ul	99
53) 2,4-Dinitrophenol	14.451	184	407247	77.872	ng/ul	94
55) 4-Nitrophenol	14.563	109	567787	74.886	ng/ul#	79
56) Dibenzofuran	14.728	168	4690333	79.205	ng/ul	98
57) 2,4-Dinitrotoluene	14.704	165	1033025	81.018	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.957	232	803492	77.205	ng/ul#	84
59) Diethylphthalate	15.169	149	3847665	75.882	ng/ul	97
61) Fluorene	15.381	166	3537977	74.447	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.381	204	1616383	73.881	ng/ul#	88
63) 4-Nitroaniline	15.422	138	569775	57.727	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	15.469	198	522523	90.505	ng/ul	91
67) N-Nitrosodiphenylamine	15.598	169	2933680	90.301	ng/ul	97
68) 4-Bromophenyl-phenylether	16.281	248	933188	90.978	ng/ul#	89
69) Hexachlorobenzene	16.387	284	1015964	85.341	ng/ul#	89
70) Atrazine	16.575	200	924567	82.267	ng/ul	96
71) Pentachlorophenol	16.740	266	621651	84.789	ng/ul	97
72) Phenanthrene	17.134	178	4967249	83.832	ng/ul	98
74) Anthracene	17.228	178	4922310	83.513	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	1775217	109.034	ng/ul	99
76) Pentachlorobenzene	14.640	250	1403811	98.360	ng/ul	98
77) Carbazole	17.504	167	4373108	79.274	ng/ul	98
78) Di-n-butylphthalate	18.075	149	5768715	83.658	ng/ul	99
80) Fluoranthene	19.122	202	5198398	87.851	ng/ul#	85
82) Pyrene	19.481	202	5160730	84.532	ng/ul#	83
83) Butylbenzylphthalate	20.375	149	2607815	95.751	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.139	252	1513475	88.729	ng/ul#	97
85) Benzo(a)anthracene	21.198	228	4989732	85.052	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.139	149	3876189	95.386	ng/ul#	95
87) Chrysene	21.251	228	4823744	83.738	ng/ul	98
89) Di-n-octyl phthalate	22.051	149	6983502	94.442	ng/ul	100
90) Benzo(b)fluoranthene	22.845	252	5953146	87.444	ng/ul#	95
91) Benzo(k)fluoranthene	22.898	252	5230667	81.796	ng/ul#	93
93) Benzo(a)pyrene	23.463	252	5627647	86.075	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	25.998	276	6979606	91.860	ng/ul#	85
95) Dibenzo(a,h)anthracene	26.021	278	5890811	89.598	ng/ul#	92
96) Benzo(g,h,i)perylene	26.745	276	5995817	92.653	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 28 15:43:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 28 14:10:09 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 24.262 ng/uL
RT: 3.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion: 88 Resp: 246429
Ion Ratio Lower Upper
88 100
43 31.3 35.8 53.6#
58 73.0 68.1 102.1



#5
Pyridine
Concen: 68.510 ng/uL
RT: 3.576 min Scan# 83
Delta R.T. -0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion: 79 Resp: 1766599
Ion Ratio Lower Upper
79 100
52 76.9 83.8 125.8#
51 39.1 41.0 61.4#









#6
Benzaldehyde
Concen: 84.115 ng/ul
RT: 6.805 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion: 77 Resp: 1170450
Ion Ratio Lower Upper
77 100
105 94.8 76.0 114.0
106 90.9 73.2 109.8



#8
Phenol
Concen: 89.951 ng/ul
RT: 6.864 min Scan# 642
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion: 94 Resp: 2739100
Ion Ratio Lower Upper
94 100
65 31.1 27.4 41.2
66 40.8 37.4 56.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 91.014 ng/ul
 RT: 7.093 min Scan# 681
 Delta R.T. 0.006 min
 Lab File: BP011161.D
 Acq: 28 Jul 2022 14:29

Instrument :
 BNA_P
 ClientSampleId :
 SST080628

Tgt Ion: 93 Resp: 2199573




Ion	Ratio	Lower	Upper
93	100		
63	85.7	72.4	108.6
95	32.7	27.3	40.9




#12
 2-Chlorophenol
 Concen: 89.418 ng/ul
 RT: 7.222 min Scan# 703
 Delta R.T. 0.006 min
 Lab File: BP011161.D
 Acq: 28 Jul 2022 14:29

Tgt Ion: 128 Resp: 2139574

Ion	Ratio	Lower	Upper
128	100		
64	52.4	44.8	67.2
130	32.4	25.8	38.6









#13
2-Methylphenol
Concen: 91.092 ng/ul
RT: 8.105 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:108 Resp: 2051306
Ion Ratio Lower Upper
108 100
107 90.1 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 111.302 ng/ul
RT: 8.193 min Scan# 868
Delta R.T. 0.000 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion: 45 Resp: 3679041
Ion Ratio Lower Upper
45 100
77 12.6 11.2 16.8
79 9.8 6.9 10.3









#16
Acetophenone
Concen: 89.178 ng/ul
RT: 8.487 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:105 Resp: 3096067
Ion Ratio Lower Upper
105 100
77 82.2 67.5 101.3
51 33.6 29.6 44.4



#17
N-Nitroso-di-n-propylamine
Concen: 96.607 ng/ul
RT: 8.493 min Scan# 919
Delta R.T. 0.024 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion: 70 Resp: 1682886
Ion Ratio Lower Upper
70 100
42 67.5 53.8 80.8
101 9.7 7.8 11.6
130 19.9 15.4 23.2









#18
4-Methylphenol
Concen: 89.895 ng/ul
RT: 8.452 min Scan# 912
Delta R.T. 0.024 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:108 Resp: 2161402
Ion Ratio Lower Upper
108 100
107 109.8 88.3 132.5



#19
Hexachloroethane
Concen: 97.577 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:117 Resp: 923966
Ion Ratio Lower Upper
117 100
201 80.6 85.7 128.5#
199 50.5 57.4 86.0#









#22
Nitrobenzene
Concen: 97.258 ng/ul
RT: 8.863 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion: 77 Resp: 2564989
Ion Ratio Lower Upper
77 100
123 42.6 31.5 47.3
65 14.3 11.4 17.2



#23
Isophorone
Concen: 93.111 ng/ul
RT: 9.399 min Scan# 1073
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion: 82 Resp: 4595823
Ion Ratio Lower Upper
82 100
95 7.1 6.6 10.0
138 16.3 13.8 20.6









#25
2-Nitrophenol
Concen: 93.419 ng/ul
RT: 9.569 min Scan# 1101
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:139 Resp: 1163629
Ion Ratio Lower Upper
139 100
65 60.6 54.2 81.4
109 31.5 30.6 45.8



#26
2,4-Dimethylphenol
Concen: 87.875 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:107 Resp: 2314687
Ion Ratio Lower Upper
107 100
121 49.9 39.1 58.7
122 83.9 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 89.431 ng/ul

RT: 9.881 min Scan# 1155

Delta R.T. 0.012 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion: 93 Resp: 2921210

Ion Ratio Lower Upper

93 100

95 32.4 26.6 39.8

123 10.7 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 82.833 ng/ul

RT: 10.105 min Scan# 1193

Delta R.T. 0.012 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:162 Resp: 1739890


Ion Ratio Lower Upper

162 100

164 65.7 52.5 78.7

98 41.6 31.4 47.0









#30
Naphthalene
Concen: 81.874 ng/ul
RT: 10.499 min Scan# 1259
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:128 Resp: 6420429
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.5 10.2 15.4



#32
4-Chloroaniline
Concen: 73.932 ng/ul
RT: 10.622 min Scan# 1281
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:127 Resp: 2446549
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2





#33

Hexachlorobutadiene

Concen: 85.856 ng/ul

RT: 10.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:225 Resp: 1049746

Ion Ratio Lower Upper

225 100

223 63.3 53.7 80.5

227 64.7 49.4 74.2



#34

Caprolactam

Concen: 38.909 ng/ul

RT: 11.446 min Scan# 1421

Delta R.T. 0.035 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:113 Resp: 286419


Ion Ratio Lower Upper

113 100

55 199.7 153.8 230.8

56 147.7 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 81.428 ng/ul
RT: 11.763 min Scan# 1473
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:107 Resp: 1932960
Ion Ratio Lower Upper
107 100
144 24.7 18.6 27.8
142 76.3 59.6 89.4



#36
2-Methylnaphthalene
Concen: 80.084 ng/ul
RT: 12.122 min Scan# 1536
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:142 Resp: 4071310
Ion Ratio Lower Upper
142 100
141 89.1 70.1 105.1









#37
1-Methylnaphthalene
Concen: 78.649 ng/ul
RT: 12.346 min Scan# 1573
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:142 Resp: 4007757
Ion Ratio Lower Upper
142 100
141 93.2 73.3 109.9
116 4.0 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 91.602 ng/ul
RT: 12.493 min Scan# 1599
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:216 Resp: 1729686
Ion Ratio Lower Upper
216 100
214 80.0 63.0 94.6
179 24.0 18.1 27.1
108 24.1 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 108.855 ng/ul
RT: 12.469 min Scan# 1594
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:237 Resp: 1273082
Ion Ratio Lower Upper
237 100
235 62.8 52.8 79.2
272 12.0 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 92.851 ng/ul
RT: 12.746 min Scan# 1642
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:196 Resp: 1154611
Ion Ratio Lower Upper
196 100
198 94.9 78.2 117.4
200 31.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 89.467 ng/ul
RT: 12.816 min Scan# 11
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:196 Resp: 1188981
Ion Ratio Lower Upper
196 100
198 96.7 77.0 115.4
200 30.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 90.919 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:154 Resp: 4856834
Ion Ratio Lower Upper
154 100
153 40.4 31.9 47.9
76 15.6 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 89.429 ng/ul

RT: 13.187 min Scan# 1716

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:162 Resp: 3588284

Ion Ratio Lower Upper

162 100

164 33.2 25.5 38.3

127 45.1 31.9 47.9



#45

2-Nitroaniline

Concen: 101.662 ng/ul

RT: 13.410 min Scan# 1755

Delta R.T. 0.012 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion: 65 Resp: 1188445


Ion Ratio Lower Upper

65 100

92 64.7 49.0 73.6

138 94.1 66.4 99.6









#47
Dimethylphthalate
Concen: 78.870 ng/ul
RT: 13.793 min Scan# 1820
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:163 Resp: 3894161
Ion Ratio Lower Upper
163 100
194 3.6 3.4 5.2
164 10.3 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 91.043 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:165 Resp: 813835
Ion Ratio Lower Upper
165 100
89 67.3 52.9 79.3
121 23.7 19.4 29.2





#50

Acenaphthylene

Concen: 85.561 ng/ul

RT: 14.040 min Scan# 1861

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:152 Resp: 5542832

Ion Ratio Lower Upper

152 100

151 20.3 16.5 24.7

153 13.6 10.1 15.1



#51

3-Nitroaniline

Concen: 73.806 ng/ul

RT: 14.246 min Scan# 1897

Delta R.T. 0.012 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:138 Resp: 792642


Ion Ratio Lower Upper

138 100

108 10.7 11.0 16.4#

92 122.4 104.6 156.8









#52
Acenaphthene
Concen: 84.374 ng/ul
RT: 14.387 min Scan# 1920
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:153 Resp: 3620969
Ion Ratio Lower Upper
153 100
152 47.7 36.2 54.2
154 88.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 77.872 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:184 Resp: 407247
Ion Ratio Lower Upper
184 100
63 86.6 73.8 110.6
154 66.6 57.2 85.8









#55
4-Nitrophenol
Concen: 74.886 ng/ul
RT: 14.563 min Scan# 1950
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:109 Resp: 567787
Ion Ratio Lower Upper
109 100
139 127.0 75.6 113.4#
65 148.7 107.6 161.4



#56
Dibenzofuran
Concen: 79.205 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:168 Resp: 4690333
Ion Ratio Lower Upper
168 100
139 42.3 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 81.018 ng/ul
RT: 14.704 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:165 Resp: 1033025
Ion Ratio Lower Upper
165 100
63 57.7 47.0 70.6
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 77.205 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:232 Resp: 803492
Ion Ratio Lower Upper
232 100
131 58.1 35.8 53.8#
130 3.1 1.8 2.8#
166 35.5 23.7 35.5#









#59
Diethylphthalate
Concen: 75.882 ng/ul
RT: 15.169 min Scan# 2053
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:149 Resp: 3847665
Ion Ratio Lower Upper
149 100
177 20.0 17.6 26.4
150 13.0 10.3 15.5



#61
Fluorene
Concen: 74.447 ng/ul
RT: 15.381 min Scan# 2090
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:166 Resp: 3537977
Ion Ratio Lower Upper
166 100
165 98.6 77.4 116.2
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 73.881 ng/ul
RT: 15.381 min Scan# 2089
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:204 Resp: 1616383
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 75.0 49.3 73.9#



#63
4-Nitroaniline
Concen: 57.727 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. 0.018 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:138 Resp: 569775
Ion Ratio Lower Upper
138 100
92 54.6 46.2 69.2
108 81.7 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 90.505 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:198 Resp: 522523
Ion Ratio Lower Upper
198 100
182 4.8 4.3 6.5
77 35.2 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 90.301 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:169 Resp: 2933680
Ion Ratio Lower Upper
169 100
168 67.6 52.6 78.8
167 37.2 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 90.978 ng/ul
RT: 16.281 min Scan# 211
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:248 Resp: 933188
Ion Ratio Lower Upper
248 100
250 96.9 80.8 121.2
141 97.4 63.9 95.9#



#69
Hexachlorobenzene
Concen: 85.341 ng/ul
RT: 16.387 min Scan# 2261
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:284 Resp: 1015964
Ion Ratio Lower Upper
284 100
142 46.1 27.8 41.6#
249 33.1 25.4 38.2









#70
Atrazine
Concen: 82.267 ng/ul
RT: 16.575 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:200 Resp: 924567
Ion Ratio Lower Upper
200 100
173 27.6 20.7 31.1
215 45.5 38.6 57.8



#71
Pentachlorophenol
Concen: 84.789 ng/ul
RT: 16.740 min Scan# 2321
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:266 Resp: 621651
Ion Ratio Lower Upper
266 100
264 62.0 51.4 77.2
268 63.9 53.5 80.3









#72
Phenanthrene
Concen: 83.832 ng/ul
RT: 17.134 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:178 Resp: 4967249
Ion Ratio Lower Upper
178 100
179 15.4 12.7 19.1
176 19.5 14.6 22.0



#74
Anthracene
Concen: 83.513 ng/ul
RT: 17.228 min Scan# 2404
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:178 Resp: 4922310
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.0 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 109.034 ng/uL

RT: 13.104 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:216 Resp: 1775217

Ion Ratio Lower Upper

216 100

214 80.1 65.0 97.4

218 48.0 39.1 58.7



#76

Pentachlorobenzene

Concen: 98.360 ng/uL

RT: 14.640 min Scan# 1964

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:250 Resp: 1403811


Ion Ratio Lower Upper

250 100

248 63.4 49.8 74.6

252 62.7 51.2 76.8





#77

Carbazole

Concen: 79.274 ng/ul

RT: 17.504 min Scan# 2450

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:167 Resp: 4373108

Ion Ratio Lower Upper

167 100

166 21.8 16.9 25.3

139 14.6 10.6 16.0



#78

Di-n-butylphthalate

Concen: 83.658 ng/ul

RT: 18.075 min Scan# 2548

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:149 Resp: 5768715


Ion Ratio Lower Upper

149 100

150 9.4 7.1 10.7

104 5.9 4.6 6.8





#80

Fluoranthene

Concen: 87.851 ng/ul

RT: 19.122 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:202 Resp: 5198398

Ion Ratio Lower Upper

202 100

101 14.9 7.1 10.7#

100 11.5 5.7 8.5#



#82

Pyrene

Concen: 84.532 ng/ul

RT: 19.481 min Scan# 2787

Delta R.T. 0.006 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:202 Resp: 5160730


Ion Ratio Lower Upper

202 100

101 16.8 7.9 11.9#

100 13.7 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 95.751 ng/ul
RT: 20.375 min Scan# 2938
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:149 Resp: 2607815
Ion Ratio Lower Upper
149 100
91 77.7 60.0 90.0
206 18.4 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 88.729 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:252 Resp: 1513475
Ion Ratio Lower Upper
252 100
254 64.8 51.0 76.4
126 15.7 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 85.052 ng/ul

RT: 21.198 min Scan# 3079

Delta R.T. 0.000 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:228 Resp: 4989732

Ion Ratio Lower Upper

228 100

229 20.5 15.0 22.4

226 27.6 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 95.386 ng/ul

RT: 21.139 min Scan# 3069

Delta R.T. 0.000 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:149 Resp: 3876189


Ion Ratio Lower Upper

149 100

167 25.9 22.6 33.8

279 3.4 4.2 6.4#









#87
Chrysene
Concen: 83.738 ng/ul
RT: 21.251 min Scan# 3088
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:228 Resp: 4823744
Ion Ratio Lower Upper
228 100
226 30.7 23.6 35.4
229 20.4 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 94.442 ng/ul
RT: 22.051 min Scan# 3224
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:149 Resp: 6983502









#90
Benzo(b)fluoranthene
Concen: 87.444 ng/ul
RT: 22.845 min Scan# 3358
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD080628


Tgt Ion:252 Resp: 5953146
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 14.5 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 81.796 ng/ul
RT: 22.898 min Scan# 3368
Delta R.T. 0.006 min
Lab File: BP011161.D
Acq: 28 Jul 2022 14:29

Tgt Ion:252 Resp: 5230667
Ion Ratio Lower Upper
252 100
253 22.4 17.0 25.4
125 14.0 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 86.075 ng/ul

RT: 23.463 min Scan# 3464

Delta R.T. 0.012 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD080628





Tgt Ion:252 Resp: 5627647

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 15.5 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 91.860 ng/ul

RT: 25.998 min Scan# 3895

Delta R.T. 0.018 min

Lab File: BP011161.D

Acq: 28 Jul 2022 14:29




Tgt Ion:276 Resp: 6979606


Ion Ratio Lower Upper

276 100

138 33.0 14.6 22.0#

277 25.1 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 89.598 ng/ul
 RT: 26.021 min Scan# 3896
 Delta R.T. 0.018 min
 Lab File: BP011161.D
 Acq: 28 Jul 2022 14:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080628

Tgt Ion:278 Resp: 5890811
 Ion Ratio Lower Upper
 278 100
 139 21.0 9.4 14.2#
 279 23.7 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 92.653 ng/ul
 RT: 26.745 min Scan# 4022
 Delta R.T. 0.024 min
 Lab File: BP011161.D
 Acq: 28 Jul 2022 14:29

Tgt Ion:276 Resp: 5995817
 Ion Ratio Lower Upper
 276 100
 138 29.1 12.7 19.1#
 277 24.1 18.2 27.4

