

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011162.D
 Acq On : 28 Jul 2022 15:04
 Operator : CG/JU
 Sample : SSTD16029
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160629

Quant Time: Jul 28 15:43:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	273832	20.000	ng/u1	0.00
20) Naphthalene-d8	10.446	136	1223702	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.322	164	683437	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.092	188	1405108	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.221	240	1229569	20.000	ng/u1	# 0.01
88) Perylene-d12	23.568	264	1373360	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	391602	51.556	ng/uL	0.00
4) Pyridine-d5	3.558	84	2734026	139.258	ng/u1	0.00
7) Phenol-d5	6.846	99	3949860	175.722	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.005	67	2591209	179.272	ng/u1	0.01
11) 2-Chlorophenol-d4	7.193	132	3169124	174.484	ng/u1	0.01
15) 4-Methylphenol-d8	8.387	113	3152627	175.195	ng/u1	0.02
21) Nitrobenzene-d5	8.828	128	1602900	179.743	ng/u1	0.02
24) 2-Nitrophenol-d4	9.540	143	1678975	190.271	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	2667878	165.323	ng/u1	0.01
31) 4-Chloroaniline-d4	10.604	131	4292262	161.340	ng/u1	0.01
46) Dimethylphthalate-d6	13.751	166	7670964	164.224	ng/u1	0.01
49) Acenaphthylene-d8	14.016	160	9068149	164.539	ng/u1	0.01
54) 4-Nitrophenol-d4	14.569	143	1693940	179.185	ng/u1	0.04
60) Fluorene-d10	15.328	176	6205256	156.287	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.469	200	1361346	195.796	ng/u1	0.02
73) Anthracene-d10	17.204	188	9243484	154.743	ng/u1	0.02
81) Pyrene-d10	19.457	212	10285284	159.272	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.427	264	10955948	166.097	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	402566	50.512	ng/uL#	83
5) Pyridine	3.575	79	2774328	137.121	ng/u1#	76
6) Benzaldehyde	6.805	77	1168265	107.002	ng/u1	98
8) Phenol	6.869	94	4089671	171.166	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.099	93	3202215	168.870	ng/u1	95
12) 2-Chlorophenol	7.222	128	3202572	170.580	ng/u1	98
13) 2-Methylphenol	8.110	108	3063278	173.367	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.199	45	5184431	199.894	ng/u1	97
16) Acetophenone	8.493	105	4597872	168.786	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.504	70	2479246	181.386	ng/u1	97
18) 4-Methylphenol	8.463	108	3247366	172.132	ng/u1	99
19) Hexachloroethane	8.722	117	1382592	186.088	ng/u1#	73
22) Nitrobenzene	8.869	77	3830856	182.667	ng/u1	96
23) Isophorone	9.410	82	7177163	182.858	ng/u1	98
25) 2-Nitrophenol	9.575	139	1772527	178.953	ng/u1	90
26) 2,4-Dimethylphenol	9.646	107	3544050	169.199	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.887	93	4269774	164.382	ng/u1	98
29) 2,4-Dichlorophenol	10.110	162	2652294	158.792	ng/u1	98
30) Naphthalene	10.499	128	9558540	153.284	ng/u1	98
32) 4-Chloroaniline	10.628	127	4247956	161.429	ng/u1	99
33) Hexachlorobutadiene	10.781	225	1502499	154.535	ng/u1	96
34) Caprolactam	11.481	113	592062	101.143	ng/u1	96
35) 4-Chloro-3-methylphenol	11.775	107	3365547	178.292	ng/u1	99
36) 2-Methylnaphthalene	12.122	142	6293628	155.681	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072822\
 Data File : BP011162.D
 Acq On : 28 Jul 2022 15:04
 Operator : CG/JU
 Sample : SSTD16029
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160629

Quant Time: Jul 28 15:43:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 14:10:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.345	142	6268161	154.687	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.498	216	2739405	151.975	ng/ul#	98
40) Hexachlorocyclopentadiene	12.469	237	1928110	172.703	ng/ul	96
41) 2,4,6-Trichlorophenol	12.745	196	1993036	167.897	ng/ul	98
42) 2,4,5-Trichlorophenol	12.822	196	2131932	168.050	ng/ul	99
43) 1,1'-Biphenyl	13.151	154	7772081	152.411	ng/ul#	95
44) 2-Chloronaphthalene	13.192	162	5874306	153.365	ng/ul	93
45) 2-Nitroaniline	13.416	65	2467260	221.092	ng/ul	92
47) Dimethylphthalate	13.804	163	7505688	159.245	ng/ul	98
48) 2,6-Dinitrotoluene	13.928	165	1705341	199.848	ng/ul	99
50) Acenaphthylene	14.045	152	9635687	155.814	ng/ul	98
51) 3-Nitroaniline	14.257	138	1743246	170.039	ng/ul#	94
52) Acenaphthene	14.392	153	6437326	157.132	ng/ul	98
53) 2,4-Dinitrophenol	14.457	184	1094508	219.239	ng/ul	99
55) 4-Nitrophenol	14.587	109	1438240	198.711	ng/ul#	83
56) Dibenzofuran	14.734	168	8402954	148.648	ng/ul	96
57) 2,4-Dinitrotoluene	14.716	165	2360541	193.937	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	14.963	232	1686725	169.779	ng/ul#	80
59) Diethylphthalate	15.181	149	8201236	169.433	ng/ul	98
61) Fluorene	15.386	166	6837625	150.721	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.386	204	2999971	143.642	ng/ul#	87
63) 4-Nitroaniline	15.386	138	110658	11.745	ng/ul#	23
66) 4,6-Dinitro-2-methylph...	15.486	198	1328587	186.263	ng/ul#	92
67) N-Nitrosodiphenylamine	15.610	169	6119085	152.452	ng/ul	97
68) 4-Bromophenyl-phenylether	16.286	248	1978348	156.112	ng/ul#	89
69) Hexachlorobenzene	16.392	284	2227944	151.479	ng/ul#	89
70) Atrazine	16.586	200	2205527	158.843	ng/ul	95
71) Pentachlorophenol	16.745	266	1535833	169.553	ng/ul	97
72) Phenanthrene	17.139	178	11040923	150.822	ng/ul	97
74) Anthracene	17.239	178	10805726	148.391	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.110	216	2899912	144.166	ng/ul	97
76) Pentachlorobenzene	14.645	250	2570167	145.760	ng/ul	99
77) Carbazole	17.516	167	10606409	155.623	ng/ul	96
78) Di-n-butylphthalate	18.075	149	12486433	146.567	ng/ul#	93
80) Fluoranthene	19.127	202	12304300	162.911	ng/ul#	81
82) Pyrene	19.486	202	11873993	152.378	ng/ul#	74
83) Butylbenzylphthalate	20.374	149	6780741	195.057	ng/ul	90
84) 3,3'-Dichlorobenzidine	21.151	252	3347989	153.777	ng/ul#	98
85) Benzo(a)anthracene	21.210	228	11824710	157.912	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.145	149	9447855	182.151	ng/ul#	98
87) Chrysene	21.263	228	11043246	150.193	ng/ul	94
89) Di-n-octyl phthalate	22.051	149	14266711	154.896	ng/ul	100
90) Benzo(b)fluoranthene	22.863	252	13689430	161.434	ng/ul#	94
91) Benzo(k)fluoranthene	22.915	252	12147526	152.507	ng/ul#	93
93) Benzo(a)pyrene	23.486	252	12830229	157.547	ng/ul#	93
94) Indeno(1,2,3-cd)pyrene	26.027	276	15949902	168.531	ng/ul#	86
95) Dibenzo(a,h)anthracene	26.057	278	13183996	160.988	ng/ul#	91
96) Benzo(g,h,i)perylene	26.792	276	13743693	170.505	ng/ul#	87





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 50.512 ng/uL
RT: 3.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion: 88 Resp: 402566
Ion Ratio Lower Upper
88 100
43 30.4 35.8 53.6#
58 71.9 68.1 102.1



#5
Pyridine
Concen: 137.121 ng/uL
RT: 3.575 min Scan# 83
Delta R.T. -0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion: 79 Resp: 2774328
Ion Ratio Lower Upper
79 100
52 77.5 83.8 125.8#
51 38.7 41.0 61.4#









#6
Benzaldehyde
Concen: 107.002 ng/ul
RT: 6.805 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion: 77 Resp: 1168265
Ion Ratio Lower Upper
77 100
105 93.7 76.0 114.0
106 88.4 73.2 109.8



#8
Phenol
Concen: 171.166 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion: 94 Resp: 4089671
Ion Ratio Lower Upper
94 100
65 31.6 27.4 41.2
66 41.4 37.4 56.2









#10
Bis(2-Chloroethyl)ether
Concen: 168.870 ng/ul
RT: 7.099 min Scan# 61
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Instrument :
BNA_P
ClientSampleId :
SSTD160629

Tgt Ion: 93 Resp: 3202215




Ion	Ratio	Lower	Upper
93	100		
63	84.5	72.4	108.6
95	32.8	27.3	40.9




#12
2-Chlorophenol
Concen: 170.580 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:128 Resp: 3202572

Ion	Ratio	Lower	Upper
128	100		
64	53.8	44.8	67.2
130	32.3	25.8	38.6









#13
2-Methylphenol
Concen: 173.367 ng/ul
RT: 8.110 min Scan# 811
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:108 Resp: 3063278
Ion Ratio Lower Upper
108 100
107 89.2 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 199.894 ng/ul
RT: 8.199 min Scan# 869
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion: 45 Resp: 5184431
Ion Ratio Lower Upper
45 100
77 12.8 11.2 16.8
79 10.0 6.9 10.3









#16
Acetophenone
Concen: 168.786 ng/ul
RT: 8.493 min Scan# 917
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:105 Resp: 4597872
Ion Ratio Lower Upper
105 100
77 82.2 67.5 101.3
51 33.0 29.6 44.4



#17
N-Nitroso-di-n-propylamine
Concen: 181.386 ng/ul
RT: 8.504 min Scan# 921
Delta R.T. 0.035 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion: 70 Resp: 2479246
Ion Ratio Lower Upper
70 100
42 64.7 53.8 80.8
101 9.5 7.8 11.6
130 19.8 15.4 23.2









#18
4-Methylphenol
Concen: 172.132 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.035 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:108 Resp: 3247366
Ion Ratio Lower Upper
108 100
107 109.2 88.3 132.5



#19
Hexachloroethane
Concen: 186.088 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. -0.000 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:117 Resp: 1382592
Ion Ratio Lower Upper
117 100
201 79.5 85.7 128.5#
199 49.4 57.4 86.0#









#22
Nitrobenzene
Concen: 182.667 ng/ul
RT: 8.869 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion: 77 Resp: 3830856
Ion Ratio Lower Upper
77 100
123 42.8 31.5 47.3
65 14.3 11.4 17.2



#23
Isophorone
Concen: 182.858 ng/ul
RT: 9.410 min Scan# 1075
Delta R.T. 0.024 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion: 82 Resp: 7177163
Ion Ratio Lower Upper
82 100
95 7.2 6.6 10.0
138 16.3 13.8 20.6









#25
2-Nitrophenol
Concen: 178.953 ng/ul
RT: 9.575 min Scan# 1101
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:139 Resp: 1772527
Ion Ratio Lower Upper
139 100
65 59.5 54.2 81.4
109 31.9 30.6 45.8



#26
2,4-Dimethylphenol
Concen: 169.199 ng/ul
RT: 9.646 min Scan# 1115
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:107 Resp: 3544050
Ion Ratio Lower Upper
107 100
121 50.0 39.1 58.7
122 83.5 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 164.382 ng/ul

RT: 9.887 min Scan# 1153

Delta R.T. 0.018 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion: 93 Resp: 4269774

Ion Ratio Lower Upper

93 100

95 32.6 26.6 39.8

123 10.9 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 158.792 ng/ul

RT: 10.110 min Scan# 1194

Delta R.T. 0.018 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:162 Resp: 2652294


Ion Ratio Lower Upper

162 100

164 64.7 52.5 78.7

98 41.9 31.4 47.0









#30
Naphthalene
Concen: 153.284 ng/ul
RT: 10.499 min Scan# 1259
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:128 Resp: 9558540
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 13.6 10.2 15.4



#32
4-Chloroaniline
Concen: 161.429 ng/ul
RT: 10.628 min Scan# 1282
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:127 Resp: 4247956
Ion Ratio Lower Upper
127 100
129 32.0 26.2 39.2





#33

Hexachlorobutadiene

Concen: 154.535 ng/ul

RT: 10.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:225 Resp: 1502499

Ion Ratio Lower Upper

225 100

223 63.8 53.7 80.5

227 64.1 49.4 74.2



#34

Caprolactam

Concen: 101.143 ng/ul

RT: 11.481 min Scan# 1427

Delta R.T. 0.071 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:113 Resp: 592062


Ion Ratio Lower Upper

113 100

55 197.7 153.8 230.8

56 144.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 178.292 ng/ul
RT: 11.775 min Scan# 1473
Delta R.T. 0.024 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:107 Resp: 3365547
Ion Ratio Lower Upper
107 100
144 24.2 18.6 27.8
142 74.7 59.6 89.4



#36
2-Methylnaphthalene
Concen: 155.681 ng/ul
RT: 12.122 min Scan# 1536
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:142 Resp: 6293628
Ion Ratio Lower Upper
142 100
141 89.3 70.1 105.1





#37

1-Methylnaphthalene

Concen: 154.687 ng/ul

RT: 12.345 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:142 Resp: 6268161

Ion Ratio Lower Upper

142 100

141 91.6 73.3 109.9

116 4.3 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 151.975 ng/ul

RT: 12.498 min Scan# 1600

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Tgt Ion:216 Resp: 2739405




Ion Ratio Lower Upper


216 100

214 78.7 63.0 94.6

179 24.0 18.1 27.1

108 22.6 13.6 20.4#









#40
 Hexachlorocyclopentadiene
 Concen: 172.703 ng/ul
 RT: 12.469 min Scan# 1594
 Delta R.T. 0.006 min
 Lab File: BP011162.D
 Acq: 28 Jul 2022 15:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD160629


Tgt Ion:237 Resp: 1928110
 Ion Ratio Lower Upper
 237 100
 235 63.0 52.8 79.2
 272 11.8 11.5 17.3



#41
 2,4,6-Trichlorophenol
 Concen: 167.897 ng/ul
 RT: 12.745 min Scan# 1642
 Delta R.T. 0.006 min
 Lab File: BP011162.D
 Acq: 28 Jul 2022 15:04

Tgt Ion:196 Resp: 1993036
 Ion Ratio Lower Upper
 196 100
 198 95.5 78.2 117.4
 200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 168.050 ng/ul
RT: 12.822 min Scan# 1655
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:196 Resp: 2131932
Ion Ratio Lower Upper
196 100
198 95.3 77.0 115.4
200 31.0 24.8 37.2



#43
1,1'-Biphenyl
Concen: 152.411 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:154 Resp: 7772081
Ion Ratio Lower Upper
154 100
153 41.4 31.9 47.9
76 16.3 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 153.365 ng/ul

RT: 13.192 min Scan# 1718

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:162 Resp: 5874306

Ion Ratio Lower Upper

162 100

164 33.6 25.5 38.3

127 46.3 31.9 47.9



#45

2-Nitroaniline

Concen: 221.092 ng/ul

RT: 13.416 min Scan# 1756

Delta R.T. 0.018 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion: 65 Resp: 2467260


Ion Ratio Lower Upper

65 100

92 64.8 49.0 73.6

138 92.9 66.4 99.6









#47
Dimethylphthalate
Concen: 159.245 ng/ul
RT: 13.804 min Scan# 1822
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:163 Resp: 7505688
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.3 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 199.848 ng/ul
RT: 13.928 min Scan# 1843
Delta R.T. 0.024 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:165 Resp: 1705341
Ion Ratio Lower Upper
165 100
89 66.0 52.9 79.3
121 23.3 19.4 29.2





#50

Acenaphthylene

Concen: 155.814 ng/ul

RT: 14.045 min Scan# 1863

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:152 Resp: 9635687

Ion Ratio Lower Upper

152 100

151 21.1 16.5 24.7

153 14.1 10.1 15.1



#51

3-Nitroaniline

Concen: 170.039 ng/ul

RT: 14.257 min Scan# 1899

Delta R.T. 0.024 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:138 Resp: 1743246


Ion Ratio Lower Upper

138 100

108 10.7 11.0 16.4#

92 123.4 104.6 156.8









#52
Acenaphthene
Concen: 157.132 ng/ul
RT: 14.392 min Scan# 1922
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:153 Resp: 6437326
Ion Ratio Lower Upper
153 100
152 48.7 36.2 54.2
154 88.8 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 219.239 ng/ul
RT: 14.457 min Scan# 1933
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:184 Resp: 1094508
Ion Ratio Lower Upper
184 100
63 91.9 73.8 110.6
154 69.9 57.2 85.8









#55
4-Nitrophenol
Concen: 198.711 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.035 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:109 Resp: 1438240
Ion Ratio Lower Upper
109 100
139 123.1 75.6 113.4#
65 143.6 107.6 161.4



#56
Dibenzofuran
Concen: 148.648 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:168 Resp: 8402954
Ion Ratio Lower Upper
168 100
139 43.8 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 193.937 ng/ul
RT: 14.716 min Scan# 1974
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:165 Resp: 2360541
Ion Ratio Lower Upper
165 100
63 66.6 47.0 70.6
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 169.779 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:232 Resp: 1686725
Ion Ratio Lower Upper
232 100
131 60.9 35.8 53.8#
130 3.3 1.8 2.8#
166 37.3 23.7 35.5#









#59
Diethylphthalate
Concen: 169.433 ng/ul
RT: 15.181 min Scan# 2053
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:149 Resp: 8201236
Ion Ratio Lower Upper
149 100
177 20.3 17.6 26.4
150 12.8 10.3 15.5



#61
Fluorene
Concen: 150.721 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:166 Resp: 6837625
Ion Ratio Lower Upper
166 100
165 98.7 77.4 116.2
167 13.8 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 143.642 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:204 Resp: 2999971
Ion Ratio Lower Upper
204 100
206 33.2 26.0 39.0
141 76.9 49.3 73.9#



#63
4-Nitroaniline
Concen: 11.745 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. -0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:138 Resp: 110658
Ion Ratio Lower Upper
138 100
92 27.9 46.2 69.2#
108 5.8 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 186.263 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. 0.029 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:198 Resp: 1328587
Ion Ratio Lower Upper
198 100
182 3.9 4.3 6.5#
77 34.0 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 152.452 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:169 Resp: 6119085
Ion Ratio Lower Upper
169 100
168 67.5 52.6 78.8
167 37.0 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 156.112 ng/ul
RT: 16.286 min Scan# 211
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:248 Resp: 1978348
Ion Ratio Lower Upper
248 100
250 97.3 80.8 121.2
141 97.6 63.9 95.9#



#69
Hexachlorobenzene
Concen: 151.479 ng/ul
RT: 16.392 min Scan# 2262
Delta R.T. 0.012 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:284 Resp: 2227944
Ion Ratio Lower Upper
284 100
142 45.7 27.8 41.6#
249 33.5 25.4 38.2





#70

Atrazine

Concen: 158.843 ng/ul

RT: 16.586 min Scan# 21

Delta R.T. 0.024 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:200 Resp: 2205527

Ion Ratio Lower Upper

200 100

173 28.4 20.7 31.1

215 44.9 38.6 57.8



#71

Pentachlorophenol

Concen: 169.553 ng/ul

RT: 16.745 min Scan# 2322

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:266 Resp: 1535833


Ion Ratio Lower Upper

266 100

264 63.0 51.4 77.2

268 63.0 53.5 80.3





#72

Phenanthrene

Concen: 150.822 ng/ul

RT: 17.139 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:178 Resp:11040923

Ion Ratio Lower Upper

178 100

179 16.1 12.7 19.1

176 20.3 14.6 22.0



#74

Anthracene

Concen: 148.391 ng/ul

RT: 17.239 min Scan# 2406

Delta R.T. 0.018 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:178 Resp:10805726


Ion Ratio Lower Upper

178 100

179 16.2 12.2 18.2

176 19.8 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 144.166 ng/uL

RT: 13.110 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:216 Resp: 2899912

Ion Ratio Lower Upper

216 100

214 78.2 65.0 97.4

218 47.4 39.1 58.7



#76

Pentachlorobenzene

Concen: 145.760 ng/uL

RT: 14.645 min Scan# 1965

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:250 Resp: 2570167


Ion Ratio Lower Upper

250 100

248 63.5 49.8 74.6

252 64.6 51.2 76.8









#77
Carbazole
Concen: 155.623 ng/ul
RT: 17.516 min Scan# 2453
Delta R.T. 0.018 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:167 Resp:10606409
Ion Ratio Lower Upper
167 100
166 22.7 16.9 25.3
139 15.6 10.6 16.0



#78
Di-n-butylphthalate
Concen: 146.567 ng/ul
RT: 18.075 min Scan# 2548
Delta R.T. 0.006 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:149 Resp:12486433
Ion Ratio Lower Upper
149 100
150 11.6 7.1 10.7#
104 7.6 4.6 6.8#





#80

Fluoranthene

Concen: 162.911 ng/ul

RT: 19.127 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:202 Resp:12304300

Ion Ratio Lower Upper

202 100

101 16.7 7.1 10.7#

100 12.8 5.7 8.5#



#82

Pyrene

Concen: 152.378 ng/ul

RT: 19.486 min Scan# 2788

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:202 Resp:11873993


Ion Ratio Lower Upper

202 100

101 20.2 7.9 11.9#

100 16.5 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 195.057 ng/ul

RT: 20.374 min Scan# 2938

Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:149 Resp: 6780741

Ion Ratio Lower Upper

149 100

91 83.5 60.0 90.0

206 18.0 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 153.777 ng/ul

RT: 21.151 min Scan# 3071

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:252 Resp: 3347989


Ion Ratio Lower Upper

252 100

254 64.7 51.0 76.4

126 13.5 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 157.912 ng/ul

RT: 21.210 min Scan# 3081

Delta R.T. 0.012 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:228 Resp:11824710

Ion Ratio Lower Upper

228 100

229 21.1 15.0 22.4

226 29.5 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 182.151 ng/ul

RT: 21.145 min Scan# 3070

Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:149 Resp: 9447855


Ion Ratio Lower Upper

149 100

167 27.6 22.6 33.8

279 4.1 4.2 6.4#





#87

Chrysene

Concen: 150.193 ng/ul

RT: 21.263 min Scan# 3087

Delta R.T. 0.018 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:228 Resp:11043246

Ion Ratio Lower Upper

228 100

226 33.2 23.6 35.4

229 21.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 154.896 ng/ul




RT: 22.051 min Scan# 3224


Delta R.T. 0.006 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Tgt Ion:149 Resp:14266711









#90
Benzo(b)fluoranthene
Concen: 161.434 ng/ul
RT: 22.863 min Scan# 3362
Delta R.T. 0.024 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04




Instrument :
BNA_P
ClientSampleId :
SSTD160629


Tgt Ion:252 Resp:13689430
Ion Ratio Lower Upper
252 100
253 22.9 17.8 26.6
125 14.2 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 152.507 ng/ul
RT: 22.915 min Scan# 3371
Delta R.T. 0.024 min
Lab File: BP011162.D
Acq: 28 Jul 2022 15:04

Tgt Ion:252 Resp:12147526
Ion Ratio Lower Upper
252 100
253 22.9 17.0 25.4
125 13.8 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 157.547 ng/ul

RT: 23.486 min Scan# 34

Delta R.T. 0.035 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:252 Resp:12830229

Ion Ratio Lower Upper

252 100

253 22.6 17.4 26.0

125 15.6 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 168.531 ng/ul

RT: 26.027 min Scan# 3900

Delta R.T. 0.047 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04




Tgt Ion:276 Resp:15949902


Ion Ratio Lower Upper

276 100

138 32.4 14.6 22.0#

277 25.1 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 160.988 ng/ul

RT: 26.057 min Scan# 3905

Delta R.T. 0.053 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Instrument :

BNA_P

ClientSampleId :

SSTD160629





Tgt Ion:278 Resp:13183996

Ion Ratio Lower Upper

278 100

139 21.1 9.4 14.2#

279 24.3 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 170.505 ng/ul

RT: 26.792 min Scan# 4030

Delta R.T. 0.071 min

Lab File: BP011162.D

Acq: 28 Jul 2022 15:04

Tgt Ion:276 Resp:13743693

Ion Ratio Lower Upper

276 100

138 28.1 12.7 19.1#

277 23.7 18.2 27.4

