

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080122\
 Data File : BP011210.D
 Acq On : 01 Aug 2022 21:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 01 23:02:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 16:05:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.645	152	274954	20.000	ng/u1	0.00
20) Naphthalene-d8	10.439	136	1119320	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.316	164	576979	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.086	188	1146059	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.209	240	1128866	20.000	ng/u1	# 0.00
88) Perylene-d12	23.556	264	1192007	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	330	0.055	ng/uL	0.05
4) Pyridine-d5	3.557	84	239094	14.076	ng/u1	0.00
7) Phenol-d5	6.834	99	484142	20.173	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.992	67	351935	20.739	ng/u1	0.00
11) 2-Chlorophenol-d4	7.181	132	412027	21.786	ng/u1	0.00
15) 4-Methylphenol-d8	8.369	113	384110	19.566	ng/u1	0.00
21) Nitrobenzene-d5	8.810	128	191167	21.976	ng/u1	0.00
24) 2-Nitrophenol-d4	9.534	143	188009	22.563	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.069	165	319790	21.777	ng/u1	0.00
31) 4-Chloroaniline-d4	10.592	131	515546	21.657	ng/u1	0.00
46) Dimethylphthalate-d6	13.733	166	861596	21.918	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	1108514	22.440	ng/u1	0.00
54) 4-Nitrophenol-d4	14.545	143	161929	22.901	ng/u1	0.01
60) Fluorene-d10	15.316	176	737701	22.333	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.451	200	108181	18.725	ng/u1	0.00
73) Anthracene-d10	17.186	188	1121587	22.210	ng/u1	0.00
81) Pyrene-d10	19.445	212	1304395	21.504	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.403	264	1305139	22.044	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	35252	5.672	ng/uL#	79
5) Pyridine	3.581	79	245415	14.007	ng/u1#	71
6) Benzaldehyde	6.798	77	223156	19.277	ng/u1	97
8) Phenol	6.857	94	514351	19.977	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.087	93	430819	20.637	ng/u1	95
12) 2-Chlorophenol	7.216	128	424318	21.580	ng/u1	97
13) 2-Methylphenol	8.104	108	385872	20.220	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.187	45	728119	20.824	ng/u1	96
16) Acetophenone	8.475	105	617540	20.222	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.463	70	320554	19.857	ng/u1	100
18) 4-Methylphenol	8.434	108	406848	19.796	ng/u1	98
19) Hexachloroethane	8.716	117	189489	21.454	ng/u1#	74
22) Nitrobenzene	8.851	77	495275	22.145	ng/u1	98
23) Isophorone	9.381	82	885905	21.892	ng/u1	96
25) 2-Nitrophenol	9.563	139	209124	22.335	ng/u1	91
26) 2,4-Dimethylphenol	9.634	107	442435	21.996	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.869	93	542640	21.343	ng/u1	98
29) 2,4-Dichlorophenol	10.092	162	324075	21.339	ng/u1	98
30) Naphthalene	10.492	128	1279636	21.730	ng/u1	99
32) 4-Chloroaniline	10.616	127	519964	21.982	ng/u1	99
33) Hexachlorobutadiene	10.769	225	205686	22.175	ng/u1	96
34) Caprolactam	11.410	113	104657	20.721	ng/u1	88
35) 4-Chloro-3-methylphenol	11.757	107	369439	21.695	ng/u1	98
36) 2-Methylnaphthalene	12.116	142	784694	21.057	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080122\
 Data File : BP011210.D
 Acq On : 01 Aug 2022 21:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 01 23:02:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 16:05:22 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.339	142	791527	21.202	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.486	216	338292	21.074	ng/ul#	97
40) Hexachlorocyclopentadiene	12.463	237	187065	17.675	ng/ul	95
41) 2,4,6-Trichlorophenol	12.739	196	223267	22.149	ng/ul	97
42) 2,4,5-Trichlorophenol	12.810	196	238765	22.376	ng/ul	100
43) 1,1'-Biphenyl	13.139	154	991225	21.501	ng/ul#	97
44) 2-Chloronaphthalene	13.180	162	742188	21.535	ng/ul	95
45) 2-Nitroaniline	13.404	65	223642	22.327	ng/ul	95
47) Dimethylphthalate	13.786	163	862497	21.786	ng/ul	98
48) 2,6-Dinitrotoluene	13.904	165	155673	22.504	ng/ul	87
50) Acenaphthylene	14.033	152	1227859	22.250	ng/ul	99
51) 3-Nitroaniline	14.239	138	137286	18.376	ng/ul	98
52) Acenaphthene	14.380	153	791287	21.952	ng/ul	98
53) 2,4-Dinitrophenol	14.451	184	84189	21.579	ng/ul	97
55) 4-Nitrophenol	14.557	109	144905	23.850	ng/ul#	77
56) Dibenzofuran	14.716	168	1064635	21.728	ng/ul	98
57) 2,4-Dinitrotoluene	14.698	165	240497	24.918	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.951	232	182589	23.532	ng/ul#	81
59) Diethylphthalate	15.157	149	923232	23.333	ng/ul	96
61) Fluorene	15.374	166	861374	22.608	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.374	204	374285	21.585	ng/ul#	88
63) 4-Nitroaniline	15.410	138	102427	16.126	ng/ul	86
66) 4,6-Dinitro-2-methylph...	15.469	198	114865	19.577	ng/ul#	92
67) N-Nitrosodiphenylamine	15.592	169	722455	20.513	ng/ul	98
68) 4-Bromophenyl-phenylether	16.274	248	220173	20.020	ng/ul	90
69) Hexachlorobenzene	16.380	284	257219	20.445	ng/ul#	86
70) Atrazine	16.563	200	232490	20.991	ng/ul	96
71) Pentachlorophenol	16.733	266	154100	21.335	ng/ul	98
72) Phenanthrene	17.127	178	1350504	21.532	ng/ul	98
74) Anthracene	17.221	178	1375115	22.161	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	349894	17.709	ng/ul	97
76) Pentachlorobenzene	14.633	250	307732	18.898	ng/ul	100
77) Carbazole	17.504	167	1288912	23.446	ng/ul	98
78) Di-n-butylphthalate	18.068	149	1622103	24.674	ng/ul	100
80) Fluoranthene	19.115	202	1588674	21.202	ng/ul#	85
82) Pyrene	19.474	202	1623889	21.090	ng/ul#	83
83) Butylbenzylphthalate	20.368	149	780461	25.313	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.139	252	402342	20.434	ng/ul#	98
85) Benzo(a)anthracene	21.198	228	1552223	22.119	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.133	149	1205089	25.861	ng/ul#	94
87) Chrysene	21.245	228	1502522	21.538	ng/ul	99
89) Di-n-octyl phthalate	22.039	149	2147668	26.215	ng/ul	100
90) Benzo(b)fluoranthene	22.839	252	1603839	21.343	ng/ul#	94
91) Benzo(k)fluoranthene	22.892	252	1551150	21.750	ng/ul#	95
93) Benzo(a)pyrene	23.456	252	1587990	21.665	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	25.980	276	1870372	20.712	ng/ul#	87
95) Dibenzo(a,h)anthracene	26.003	278	1583731	20.448	ng/ul#	91
96) Benzo(g,h,i)perylene	26.727	276	1588524	20.339	ng/ul#	87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080122\
 Data File : BP011210.D
 Acq On : 01 Aug 2022 21:04
 Operator : CG/JU
 Sample : SSTCCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 01 23:02:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 28 16:05:22 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.672 ng/uL
RT: 3.181 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 35252
Ion Ratio Lower Upper
88 100
43 33.6 35.8 53.6#
58 64.2 68.1 102.1#



#5
Pyridine
Concen: 14.007 ng/uL
RT: 3.581 min Scan# 84
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion: 79 Resp: 245415
Ion Ratio Lower Upper
79 100
52 68.7 83.8 125.8#
51 39.4 41.0 61.4#









#6
Benzaldehyde
Concen: 19.277 ng/ul
RT: 6.798 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 223156
Ion Ratio Lower Upper
77 100
105 92.9 76.0 114.0
106 87.6 73.2 109.8



#8
Phenol
Concen: 19.977 ng/ul
RT: 6.857 min Scan# 641
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion: 94 Resp: 514351
Ion Ratio Lower Upper
94 100
65 31.4 27.4 41.2
66 40.8 37.4 56.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.637 ng/ul
RT: 7.087 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 430819
Ion Ratio Lower Upper
93 100
63 85.3 72.4 108.6
95 32.6 27.3 40.9



#12
2-Chlorophenol
Concen: 21.580 ng/ul
RT: 7.216 min Scan# 702
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion: 128 Resp: 424318
Ion Ratio Lower Upper
128 100
64 52.6 44.8 67.2
130 32.2 25.8 38.6





#13

2-Methylphenol

Concen: 20.220 ng/ul

RT: 8.104 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 385872

Ion Ratio Lower Upper

108 100

107 90.0 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.824 ng/ul

RT: 8.187 min Scan# 867

Delta R.T. -0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion: 45 Resp: 728119


Ion Ratio Lower Upper

45 100

77 12.5 11.2 16.8

79 9.7 6.9 10.3









#16
Acetophenone
Concen: 20.222 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 617540
Ion Ratio Lower Upper
105 100
77 81.5 67.5 101.3
51 34.0 29.6 44.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.857 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion: 70 Resp: 320554
Ion Ratio Lower Upper
70 100
42 67.2 53.8 80.8
101 9.6 7.8 11.6
130 19.1 15.4 23.2









#18
4-Methylphenol
Concen: 19.796 ng/ul
RT: 8.434 min Scan# 909
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 406848
Ion Ratio Lower Upper
108 100
107 112.4 88.3 132.5



#19
Hexachloroethane
Concen: 21.454 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:117 Resp: 189489
Ion Ratio Lower Upper
117 100
201 79.9 85.7 128.5#
199 50.4 57.4 86.0#









#22
Nitrobenzene
Concen: 22.145 ng/ul
RT: 8.851 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 495275
Ion Ratio Lower Upper
77 100
123 40.8 31.5 47.3
65 13.9 11.4 17.2



#23
Isophorone
Concen: 21.892 ng/ul
RT: 9.381 min Scan# 1070
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion: 82 Resp: 885905
Ion Ratio Lower Upper
82 100
95 7.0 6.6 10.0
138 15.5 13.8 20.6









#25
2-Nitrophenol
Concen: 22.335 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 209124
Ion Ratio Lower Upper
139 100
65 62.0 54.2 81.4
109 31.4 30.6 45.8



#26
2,4-Dimethylphenol
Concen: 21.996 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:107 Resp: 442435
Ion Ratio Lower Upper
107 100
121 49.0 39.1 58.7
122 83.0 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 21.343 ng/ul

RT: 9.869 min Scan# 1153

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 542640

Ion Ratio Lower Upper

93 100

95 32.0 26.6 39.8

123 10.7 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 21.339 ng/ul

RT: 10.092 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:162 Resp: 324075


Ion Ratio Lower Upper

162 100

164 64.9 52.5 78.7

98 42.4 31.4 47.0









#30
Naphthalene
Concen: 21.730 ng/ul
RT: 10.492 min Scan# 1259
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 1279636
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.5 10.2 15.4



#32
4-Chloroaniline
Concen: 21.982 ng/ul
RT: 10.616 min Scan# 1280
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04


Tgt Ion:127 Resp: 519964
Ion Ratio Lower Upper
127 100
129 32.4 26.2 39.2








#33
Hexachlorobutadiene
Concen: 22.175 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 205686
Ion Ratio Lower Upper
225 100
223 62.7 53.7 80.5
227 64.2 49.4 74.2



#34
Caprolactam
Concen: 20.721 ng/ul
RT: 11.410 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:113 Resp: 104657
Ion Ratio Lower Upper
113 100
55 209.0 153.8 230.8
56 152.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 21.695 ng/ul
RT: 11.757 min Scan# 1473
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 369439
Ion Ratio Lower Upper
107 100
144 24.5 18.6 27.8
142 76.1 59.6 89.4



#36
2-Methylnaphthalene
Concen: 21.057 ng/ul
RT: 12.116 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:142 Resp: 784694
Ion Ratio Lower Upper
142 100
141 90.3 70.1 105.1





#37

1-Methylnaphthalene

Concen: 21.202 ng/ul

RT: 12.339 min Scan# 1572

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:142 Resp: 791527

Ion Ratio Lower Upper

142 100

141 92.5 73.3 109.9

116 3.8 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.074 ng/ul

RT: 12.486 min Scan# 1598

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Tgt Ion:216 Resp: 338292




Ion Ratio Lower Upper


216 100

214 80.0 63.0 94.6

179 23.4 18.1 27.1

108 22.5 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 17.675 ng/ul
RT: 12.463 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 187065
Ion Ratio Lower Upper
237 100
235 62.2 52.8 79.2
272 12.6 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 22.149 ng/ul
RT: 12.739 min Scan# 1641
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:196 Resp: 223267
Ion Ratio Lower Upper
196 100
198 95.1 78.2 117.4
200 30.4 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 22.376 ng/ul

RT: 12.810 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:196 Resp: 238765

Ion Ratio Lower Upper

196 100

198 96.7 77.0 115.4

200 31.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 21.501 ng/ul

RT: 13.139 min Scan# 1709

Delta R.T. -0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:154 Resp: 991225


Ion Ratio Lower Upper

154 100

153 40.5 31.9 47.9

76 16.1 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 21.535 ng/ul

RT: 13.180 min Scan# 1716

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 742188

Ion Ratio Lower Upper

162 100

164 33.2 25.5 38.3

127 44.1 31.9 47.9



#45

2-Nitroaniline

Concen: 22.327 ng/ul

RT: 13.404 min Scan# 1754

Delta R.T. 0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion: 65 Resp: 223642


Ion Ratio Lower Upper

65 100

92 64.4 49.0 73.6

138 88.6 66.4 99.6









#47
Dimethylphthalate
Concen: 21.786 ng/ul
RT: 13.786 min Scan# 1819
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 862497
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 10.1 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 22.504 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:165 Resp: 155673
Ion Ratio Lower Upper
165 100
89 77.0 52.9 79.3
121 29.0 19.4 29.2









#50
Acenaphthylene
Concen: 22.250 ng/ul
RT: 14.033 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 1227859
Ion Ratio Lower Upper
152 100
151 20.2 16.5 24.7
153 13.1 10.1 15.1



#51
3-Nitroaniline
Concen: 18.376 ng/ul
RT: 14.239 min Scan# 1896
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:138 Resp: 137286
Ion Ratio Lower Upper
138 100
108 12.6 11.0 16.4
92 133.0 104.6 156.8









#52
Acenaphthene
Concen: 21.952 ng/ul
RT: 14.380 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 791287
Ion Ratio Lower Upper
153 100
152 47.8 36.2 54.2
154 87.4 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 21.579 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. 0.011 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:184 Resp: 84189
Ion Ratio Lower Upper
184 100
63 91.8 73.8 110.6
154 67.0 57.2 85.8









#55
4-Nitrophenol
Concen: 23.850 ng/ul
RT: 14.557 min Scan# 1949
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 144905
Ion Ratio Lower Upper
109 100
139 127.3 75.6 113.4#
65 152.6 107.6 161.4



#56
Dibenzofuran
Concen: 21.728 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:168 Resp: 1064635
Ion Ratio Lower Upper
168 100
139 42.4 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 24.918 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 240497
Ion Ratio Lower Upper
165 100
63 62.2 47.0 70.6
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.532 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:232 Resp: 182589
Ion Ratio Lower Upper
232 100
131 59.7 35.8 53.8#
130 3.1 1.8 2.8#
166 37.3 23.7 35.5#









#59
Diethylphthalate
Concen: 23.333 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 923232
Ion Ratio Lower Upper
149 100
177 19.2 17.6 26.4
150 12.3 10.3 15.5



#61
Fluorene
Concen: 22.608 ng/ul
RT: 15.374 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:166 Resp: 861374
Ion Ratio Lower Upper
166 100
165 98.6 77.4 116.2
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 21.585 ng/ul
RT: 15.374 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 374285
Ion Ratio Lower Upper
204 100
206 33.8 26.0 39.0
141 74.2 49.3 73.9#



#63
4-Nitroaniline
Concen: 16.126 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:138 Resp: 102427
Ion Ratio Lower Upper
138 100
92 60.4 46.2 69.2
108 87.2 85.8 128.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.577 ng/ul
RT: 15.469 min Scan# 2104
Delta R.T. 0.011 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 114865
Ion Ratio Lower Upper
198 100
182 7.0 4.3 6.5#
77 34.5 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 20.513 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:169 Resp: 722455
Ion Ratio Lower Upper
169 100
168 66.1 52.6 78.8
167 37.3 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 20.020 ng/ul
RT: 16.274 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 220173
Ion Ratio Lower Upper
248 100
250 95.5 80.8 121.2
141 94.2 63.9 95.9



#69
Hexachlorobenzene
Concen: 20.445 ng/ul
RT: 16.380 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:284 Resp: 257219
Ion Ratio Lower Upper
284 100
142 48.8 27.8 41.6#
249 33.2 25.4 38.2





#70

Atrazine

Concen: 20.991 ng/ul

RT: 16.563 min Scan# 2291

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 232490

Ion Ratio Lower Upper

200 100

173 27.8 20.7 31.1

215 45.8 38.6 57.8



#71

Pentachlorophenol

Concen: 21.335 ng/ul

RT: 16.733 min Scan# 2320

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:266 Resp: 154100


Ion Ratio Lower Upper

266 100

264 64.2 51.4 77.2

268 63.5 53.5 80.3





#72

Phenanthrene

Concen: 21.532 ng/ul

RT: 17.127 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 1350504

Ion Ratio Lower Upper

178 100

179 15.3 12.7 19.1

176 19.1 14.6 22.0



#74

Anthracene

Concen: 22.161 ng/ul

RT: 17.221 min Scan# 2403

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:178 Resp: 1375115


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 18.8 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.709 ng/uL
RT: 13.104 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 349894
Ion Ratio Lower Upper
216 100
214 77.9 65.0 97.4
218 49.9 39.1 58.7



#76
Pentachlorobenzene
Concen: 18.898 ng/uL
RT: 14.633 min Scan# 1963
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:250 Resp: 307732
Ion Ratio Lower Upper
250 100
248 62.9 49.8 74.6
252 64.1 51.2 76.8









#77
Carbazole
Concen: 23.446 ng/ul
RT: 17.504 min Scan# 2450
Delta R.T. 0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:167 Resp: 1288912
Ion Ratio Lower Upper
167 100
166 21.8 16.9 25.3
139 14.4 10.6 16.0



#78
Di-n-butylphthalate
Concen: 24.674 ng/ul
RT: 18.068 min Scan# 2547
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:149 Resp: 1622103
Ion Ratio Lower Upper
149 100
150 9.1 7.1 10.7
104 5.8 4.6 6.8









#80
Fluoranthene
Concen: 21.202 ng/ul
RT: 19.115 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 1588674
Ion Ratio Lower Upper
202 100
101 14.8 7.1 10.7#
100 11.7 5.7 8.5#



#82
Pyrene
Concen: 21.090 ng/ul
RT: 19.474 min Scan# 2786
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:202 Resp: 1623889
Ion Ratio Lower Upper
202 100
101 16.8 7.9 11.9#
100 13.7 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 25.313 ng/ul

RT: 20.368 min Scan# 2938

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 780461

Ion Ratio Lower Upper

149 100

91 77.3 60.0 90.0

206 18.0 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 20.434 ng/ul

RT: 21.139 min Scan# 3069

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:252 Resp: 402342


Ion Ratio Lower Upper

252 100

254 64.5 51.0 76.4

126 14.1 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 22.119 ng/ul
RT: 21.198 min Scan# 3078
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 1552223
Ion Ratio Lower Upper
228 100
229 19.4 15.0 22.4
226 26.8 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 25.861 ng/ul
RT: 21.133 min Scan# 3068
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:149 Resp: 1205089
Ion Ratio Lower Upper
149 100
167 25.0 22.6 33.8
279 3.4 4.2 6.4#









#87
Chrysene
Concen: 21.538 ng/ul
RT: 21.245 min Scan# 3087
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 1502522
Ion Ratio Lower Upper
228 100
226 29.9 23.6 35.4
229 19.8 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 26.215 ng/ul
RT: 22.039 min Scan# 3222
Delta R.T. -0.006 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:149 Resp: 2147668









#90
Benzo(b)fluoranthene
Concen: 21.343 ng/ul
RT: 22.839 min Scan# 3358
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 1603839
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.6
125 14.4 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 21.750 ng/ul
RT: 22.892 min Scan# 3367
Delta R.T. -0.000 min
Lab File: BP011210.D
Acq: 01 Aug 2022 21:04

Tgt Ion:252 Resp: 1551150
Ion Ratio Lower Upper
252 100
253 21.6 17.0 25.4
125 13.2 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 21.665 ng/ul

RT: 23.456 min Scan# 3462

Delta R.T. 0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 1587990

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 14.4 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.712 ng/ul

RT: 25.980 min Scan# 3892

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04




Tgt Ion:276 Resp: 1870372


Ion Ratio Lower Upper

276 100

138 31.0 14.6 22.0#

277 24.1 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 20.448 ng/ul

RT: 26.003 min Scan# 3896

Delta R.T. -0.000 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 1583731

Ion Ratio Lower Upper

278 100

139 21.5 9.4 14.2#

279 24.1 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.339 ng/ul

RT: 26.727 min Scan# 4019

Delta R.T. 0.006 min

Lab File: BP011210.D

Acq: 01 Aug 2022 21:04

Tgt Ion:276 Resp: 1588524

Ion Ratio Lower Upper

276 100

138 28.0 12.7 19.1#

277 23.6 18.2 27.4

