

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006451.D
 Acq On : 02 Aug 2021 14:51
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 02 16:07:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 15:59:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	213300	20.000	ng	0.00
21) Naphthalene-d8	10.546	136	880892	20.000	ng	# 0.00
39) Acenaphthene-d10	14.392	164	495445	20.000	ng	0.00
64) Phenanthrene-d10	17.145	188	968065	20.000	ng	# 0.00
76) Chrysene-d12	21.245	240	955463	20.000	ng	# 0.00
86) Perylene-d12	23.574	264	980677	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.364	112	575625	43.318	ng	0.00
7) Phenol-d6	6.940	99	816306	44.388	ng	0.00
23) Nitrobenzene-d5	8.916	82	781810	46.963	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	206138	45.663	ng	0.00
45) 2-Fluorobiphenyl	13.016	172	1382714	41.977	ng	0.00
79) Terphenyl-d14	19.716	244	1963136	41.572	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.311	88	129314	22.908	ng	91
3) Pyridine	3.705	79	367018	22.736	ng	94
4) n-Nitrosodimethylamine	3.617	42	141381	25.606	ng	# 86
6) Aniline	7.099	93	441342	22.077	ng	95
8) 2-Chlorophenol	7.328	128	311534	21.471	ng	90
9) Benzaldehyde	6.916	77	264346	24.917	ng	92
10) Phenol	6.963	94	410513	22.442	ng	98
11) bis(2-Chloroethyl)ether	7.199	93	314942	22.419	ng	89
12) 1,3-Dichlorobenzene	7.652	146	330113	21.334	ng	96
13) 1,4-Dichlorobenzene	7.799	146	330509	21.033	ng	97
14) 1,2-Dichlorobenzene	8.110	146	321535	21.380	ng	95
15) Benzyl Alcohol	8.005	79	304345	23.009	ng	93
16) 2,2'-oxybis(1-Chloropr...	8.293	45	375125	24.600	ng	91
17) 2-Methylphenol	8.205	107	261509	22.146	ng	98
18) Hexachloroethane	8.834	117	129890	22.731	ng	# 86
19) n-Nitroso-di-n-propyla...	8.575	70	273103	24.091	ng	# 95
20) 3+4-Methylphenols	8.534	107	352635	22.089	ng	98
22) Acetophenone	8.587	105	481064	22.087	ng	# 99
24) Nitrobenzene	8.957	77	406713	23.402	ng	89
25) Isophorone	9.487	82	690953	22.608	ng	94
26) 2-Nitrophenol	9.669	139	137126	20.618	ng	92
27) 2,4-Dimethylphenol	9.734	122	246487	20.599	ng	95
28) bis(2-Chloroethoxy)met...	9.975	93	378221	21.696	ng	# 97
29) 2,4-Dichlorophenol	10.193	162	248920	20.297	ng	93
30) 1,2,4-Trichlorobenzene	10.410	180	267180	19.758	ng	97
31) Naphthalene	10.599	128	972966	20.904	ng	99
32) Benzoic acid	9.834	122	164569	19.333	ng	91
33) 4-Chloroaniline	10.710	127	389411	20.921	ng	95
34) Hexachlorobutadiene	10.881	225	157709	19.880	ng	97
35) Caprolactam	11.493	113	84192	21.690	ng	# 73
36) 4-Chloro-3-methylphenol	11.834	107	310286	21.258	ng	91
37) 2-Methylnaphthalene	12.210	142	650158	20.343	ng	98
38) 1-Methylnaphthalene	12.428	142	615008	20.413	ng	97
40) 1,2,4,5-Tetrachloroben...	12.575	216	267910	20.058	ng	# 95
41) Hexachlorocyclopentadiene	12.557	237	144906	20.030	ng	99
43) 2,4,6-Trichlorophenol	12.822	196	177586	20.138	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006451.D
 Acq On : 02 Aug 2021 14:51
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 02 16:07:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 15:59:11 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.887	196	199434	20.626	ng	# 88
46) 1,1'-Biphenyl	13.228	154	776780	20.931	ng	95
47) 2-Chloronaphthalene	13.263	162	599307	21.026	ng	94
48) 2-Nitroaniline	13.475	65	197800	24.099	ng	87
49) Acenaphthylene	14.110	152	958632	21.233	ng	100
50) Dimethylphthalate	13.863	163	739853	21.276	ng	99
51) 2,6-Dinitrotoluene	13.975	165	163475	21.719	ng	# 79
52) Acenaphthene	14.457	154	588406	21.084	ng	97
53) 3-Nitroaniline	14.304	138	177095	21.911	ng	83
54) 2,4-Dinitrophenol	14.504	184	77917	20.539	ng	# 80
55) Dibenzofuran	14.792	168	922986	21.120	ng	94
56) 4-Nitrophenol	14.604	139	152814	22.225	ng	# 84
57) 2,4-Dinitrotoluene	14.763	165	214303	21.781	ng	# 83
58) Fluorene	15.445	166	732013	21.014	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.016	232	165214	20.577	ng	# 70
60) Diethylphthalate	15.228	149	778670	21.833	ng	97
61) 4-Chlorophenyl-phenyle...	15.445	204	328544	20.130	ng	90
62) 4-Nitroaniline	15.469	138	185448	22.135	ng	90
63) Azobenzene	15.734	77	916208	24.064	ng	91
65) 4,6-Dinitro-2-methylph...	15.522	198	110377	20.504	ng	75
66) n-Nitrosodiphenylamine	15.657	169	628051	20.934	ng	99
67) 4-Bromophenyl-phenylether	16.339	248	190554	22.287	ng	# 86
68) Hexachlorobenzene	16.445	284	215047	19.879	ng	# 88
69) Atrazine	16.622	200	193241	20.402	ng	96
70) Pentachlorophenol	16.786	266	132763	20.456	ng	98
71) Phenanthrene	17.186	178	1129336	21.145	ng	100
72) Anthracene	17.275	178	1098654	21.269	ng	99
73) Carbazole	17.551	167	1104923	21.380	ng	99
74) Di-n-butylphthalate	18.122	149	1257885	21.508	ng	99
75) Fluoranthene	19.157	202	1252302	20.951	ng	93
77) Benzidine	19.345	184	443624	17.788	ng	99
78) Pyrene	19.510	202	1310918	20.858	ng	99
80) Butylbenzylphthalate	20.404	149	558268	20.994	ng	# 85
81) Benzo(a)anthracene	21.227	228	1281870	20.895	ng	99
82) 3,3'-Dichlorobenzidine	21.169	252	337251	20.769	ng	# 94
83) Chrysene	21.280	228	1239321	20.886	ng	99
84) Bis(2-ethylhexyl)phtha...	21.169	149	878037	21.914	ng	# 95
85) Di-n-octyl phthalate	22.074	149	1418421	21.284	ng	# 93
87) Indeno(1,2,3-cd)pyrene	25.980	276	1428053	20.373	ng	# 88
88) Benzo(b)fluoranthene	22.863	252	1291716	20.490	ng	# 96
89) Benzo(k)fluoranthene	22.910	252	1268208	21.247	ng	# 96
90) Benzo(a)pyrene	23.468	252	1164366	20.618	ng	# 95
91) Dibenzo(a,h)anthracene	26.004	278	1243297	20.530	ng	# 92
92) Benzo(g,h,i)perylene	26.721	276	1191073	20.398	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006451.D
 Acq On : 02 Aug 2021 14:51
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 02 16:07:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 15:59:11 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.763 min Scan# 795
 Delta R.T. 0.000 min
 Lab File: BP006451.D
 Acq: 02 Aug 2021 14:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:152 Resp: 213300
 Ion Ratio Lower Upper
 152 100
 150 156.8 125.6 188.4
 115 65.6 48.5 72.7



#2
 1,4-Dioxane
 Concen: 22.908 ng
 RT: 3.311 min Scan# 38
 Delta R.T. 0.006 min
 Lab File: BP006451.D
 Acq: 02 Aug 2021 14:51

Tgt Ion: 88 Resp: 129314
 Ion Ratio Lower Upper
 88 100
 58 68.1 46.7 70.1
 43 26.1 20.3 30.5









#3
Pyridine
Concen: 22.736 ng
RT: 3.705 min Scan# 105
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 367018
Ion Ratio Lower Upper
79 100
52 58.6 41.7 62.5
51 29.8 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 25.606 ng
RT: 3.617 min Scan# 90
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 42 Resp: 141381
Ion Ratio Lower Upper
42 100
74 148.4 105.4 158.0
44 10.1 4.2 6.4









#5
2-Fluorophenol
Concen: 43.318 ng
RT: 5.364 min Scan# 387
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 575625
Ion Ratio Lower Upper
112 100
64 61.5 41.4 62.2
63 31.7 23.8 35.8



#6
Aniline
Concen: 22.077 ng
RT: 7.099 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 93 Resp: 441342
Ion Ratio Lower Upper
93 100
66 40.7 29.7 44.5
65 19.7 15.3 22.9









#7
Phenol-d6
Concen: 44.388 ng
RT: 6.940 min Scan# 655
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 99 Resp: 816306
Ion Ratio Lower Upper
99 100
42 17.1 13.4 20.0
71 36.3 30.6 46.0



#8
2-Chlorophenol
Concen: 21.471 ng
RT: 7.328 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 128 Resp: 311534
Ion Ratio Lower Upper
128 100
130 32.6 12.4 52.4
64 51.4 20.3 60.3









#9
Benzaldehyde
Concen: 24.917 ng
RT: 6.916 min Scan# 651
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 264346
Ion Ratio Lower Upper
77 100
105 89.8 76.7 116.7
106 87.2 76.2 116.2



#10
Phenol
Concen: 22.442 ng
RT: 6.963 min Scan# 659
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 94 Resp: 410513
Ion Ratio Lower Upper
94 100
65 29.8 8.8 48.8
66 39.2 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 22.419 ng
RT: 7.199 min Scan# 699
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 314942
Ion Ratio Lower Upper
93 100
63 78.0 46.2 86.2
95 31.0 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 21.334 ng
RT: 7.652 min Scan# 776
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 146 Resp: 330113
Ion Ratio Lower Upper
146 100
148 63.2 51.5 77.3
75 36.8 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 21.033 ng
RT: 7.799 min Scan# 801
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:146 Resp: 330509
Ion Ratio Lower Upper
146 100
148 65.0 51.3 76.9
111 44.8 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 21.380 ng
RT: 8.110 min Scan# 854
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:146 Resp: 321535
Ion Ratio Lower Upper
146 100
148 64.9 50.3 75.5
111 48.0 33.7 50.5









#15
Benzyl Alcohol
Concen: 23.009 ng
RT: 8.005 min Scan# 836
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion: 79 Resp: 304345
Ion Ratio Lower Upper
79 100
108 68.3 62.5 93.7
77 64.5 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 24.600 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 45 Resp: 375125
Ion Ratio Lower Upper
45 100
77 17.5 2.0 42.0
79 13.5 0.0 36.7









#17
2-Methylphenol
Concen: 22.146 ng
RT: 8.205 min Scan# 870
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	261509
Ion	Ratio	Lower	Upper
107	100		
108	106.8	87.7	131.5
77	50.9	40.2	60.2
79	48.9	39.0	58.4



#18
Hexachloroethane
Concen: 22.731 ng
RT: 8.834 min Scan# 977
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:	117	Resp:	129890
Ion	Ratio	Lower	Upper
117	100		
119	98.6	76.2	114.2
201	81.0	83.4	125.0









#19

n-Nitroso-di-n-propylamine
Concen: 24.091 ng
RT: 8.575 min Scan# 933
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	70	Resp:	273103
Ion	Ratio	Lower	Upper
70	100		
42	49.0	40.8	61.2
101	9.1	8.8	13.2
130	18.3	18.9	28.3




#20

3+4-Methylphenols
Concen: 22.089 ng
RT: 8.534 min Scan# 926
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:	107	Resp:	352635
Ion	Ratio	Lower	Upper
107	100		
108	88.5	71.0	111.0
77	36.3	16.6	56.6
79	31.8	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	136	Resp:	880892
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.6	12.8
54	8.1	4.6	6.8#
68	7.3	3.6	5.4#



#22
Acetophenone
Concen: 22.087 ng
RT: 8.587 min Scan# 935
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:	105	Resp:	481064
Ion Ratio	Lower	Upper	
105	100		
71	3.4	1.9	2.9#
51	23.8	18.4	27.6
120	22.0	17.9	26.9









#23
Nitrobenzene-d5
Concen: 46.963 ng
RT: 8.916 min Scan# 991
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 781810
Ion Ratio Lower Upper
82 100
128 38.1 35.7 53.5
54 42.4 35.0 52.4



#24
Nitrobenzene
Concen: 23.402 ng
RT: 8.957 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 77 Resp: 406713
Ion Ratio Lower Upper
77 100
123 39.1 38.2 57.2
65 12.8 11.4 17.2









#25
Isophorone
Concen: 22.608 ng
RT: 9.487 min Scan# 1088
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 690953
Ion Ratio Lower Upper
82 100
95 6.9 6.5 9.7
138 15.8 15.0 22.4



#26
2-Nitrophenol
Concen: 20.618 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 139 Resp: 137126
Ion Ratio Lower Upper
139 100
109 36.6 28.8 43.2
65 60.1 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 20.599 ng
RT: 9.734 min Scan# 1130
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 246487
Ion Ratio Lower Upper
122 100
107 124.7 94.5 141.7
121 60.6 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 21.696 ng
RT: 9.975 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 93 Resp: 378221
Ion Ratio Lower Upper
93 100
95 31.7 25.9 38.9
123 10.1 10.2 15.2#









#29
2,4-Dichlorophenol
Concen: 20.297 ng
RT: 10.193 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 248920
Ion Ratio Lower Upper
162 100
164 62.5 44.9 84.9
98 42.4 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 19.758 ng
RT: 10.410 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:180 Resp: 267180
Ion Ratio Lower Upper
180 100
182 95.7 77.9 116.9
145 35.2 24.7 37.1









#31
Naphthalene
Concen: 20.904 ng
RT: 10.599 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:128 Resp: 972966
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.6
127 13.3 10.7 16.1



#32
Benzoic acid
Concen: 19.333 ng
RT: 9.834 min Scan# 1147
Delta R.T. -0.023 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:122 Resp: 164569
Ion Ratio Lower Upper
122 100
105 131.8 103.2 143.2
77 107.0 76.2 116.2









#33
4-Chloroaniline
Concen: 20.921 ng
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	127	Resp:	389411
Ion	Ratio	Lower	Upper
127	100		
129	31.3	26.1	39.1
65	33.2	22.5	33.7
92	20.7	15.0	22.6



#34
Hexachlorobutadiene
Concen: 19.880 ng
RT: 10.881 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:	225	Resp:	157709
Ion	Ratio	Lower	Upper
225	100		
223	60.8	51.0	76.6
227	63.2	51.4	77.2









#35
Caprolactam
Concen: 21.690 ng
RT: 11.493 min Scan# 1429
Delta R.T. -0.012 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:113 Resp: 84192
Ion Ratio Lower Upper
113 100
55 153.7 98.9 138.9#
56 112.9 71.7 111.7#



#36
4-Chloro-3-methylphenol
Concen: 21.258 ng
RT: 11.834 min Scan# 1487
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:107 Resp: 310286
Ion Ratio Lower Upper
107 100
144 23.6 21.1 31.7
142 73.4 66.0 99.0









#37
2-Methylnaphthalene
Concen: 20.343 ng
RT: 12.210 min Scan# 1551
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 650158
Ion Ratio Lower Upper
142 100
141 88.4 71.8 107.6
115 37.5 31.6 47.4



#38
1-Methylnaphthalene
Concen: 20.413 ng
RT: 12.428 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:142 Resp: 615008
Ion Ratio Lower Upper
142 100
141 94.8 73.3 109.9
116 4.1 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.392 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	164	Resp:	495445
Ion Ratio	Lower	Upper	
164	100		
162	103.5	80.2	120.2
160	45.1	34.5	51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.058 ng
RT: 12.575 min Scan# 1613
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:	216	Resp:	267910
Ion Ratio	Lower	Upper	
216	100		
214	78.8	61.9	92.9
179	23.7	16.8	25.2
108	25.5	14.9	22.3









#41
Hexachlorocyclopentadiene
Concen: 20.030 ng
RT: 12.557 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:237 Resp: 144906
Ion Ratio Lower Upper
237 100
235 63.4 44.5 84.5
272 12.0 0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 45.663 ng
RT: 15.881 min Scan# 2175
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:330 Resp: 206138
Ion Ratio Lower Upper
330 100
332 97.1 77.8 116.8
141 48.9 25.4 38.2#









#43
2,4,6-Trichlorophenol
Concen: 20.138 ng
RT: 12.822 min Scan# 1655
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:196 Resp: 177586
Ion Ratio Lower Upper
196 100
198 94.2 79.5 119.3
200 32.1 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 20.626 ng
RT: 12.887 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:196 Resp: 199434
Ion Ratio Lower Upper
196 100
198 95.5 74.7 112.1
97 65.0 37.4 56.0#
132 35.2 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 41.977 ng
RT: 13.016 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:172 Resp: 1382714
Ion Ratio Lower Upper
172 100
171 35.6 27.9 41.9
170 23.7 18.6 27.8



#46
1,1'-Biphenyl
Concen: 20.931 ng
RT: 13.228 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:154 Resp: 776780
Ion Ratio Lower Upper
154 100
153 41.6 20.4 60.4
76 16.2 0.0 30.8









#47
2-Chloronaphthalene
Concen: 21.026 ng
RT: 13.263 min Scan# 1730
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 599307
Ion Ratio Lower Upper
162 100
127 46.1 31.4 47.0
164 32.9 26.6 39.8



#48
2-Nitroaniline
Concen: 24.099 ng
RT: 13.475 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion: 65 Resp: 197800
Ion Ratio Lower Upper
65 100
92 68.1 57.7 86.5
138 95.5 92.1 138.1





#49

Acenaphthylene

Concen: 21.233 ng

RT: 14.110 min Scan# 1874

Delta R.T. 0.000 min

Lab File: BP006451.D

Acq: 02 Aug 2021 14:51

Instrument :

BNA_P

ClientSampleId :

SSTDIC020





Tgt Ion:152 Resp: 958632

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.2

153 13.3 10.6 16.0



#50

Dimethylphthalate

Concen: 21.276 ng

RT: 13.863 min Scan# 1832

Delta R.T. 0.000 min

Lab File: BP006451.D

Acq: 02 Aug 2021 14:51




Tgt Ion:163 Resp: 739853


Ion Ratio Lower Upper

163 100

194 3.7 3.4 5.2

164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 21.719 ng
RT: 13.975 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 163475
Ion Ratio Lower Upper
165 100
63 68.3 41.3 61.9#
89 69.5 44.8 67.2#



#52
Acenaphthene
Concen: 21.084 ng
RT: 14.457 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:154 Resp: 588406
Ion Ratio Lower Upper
154 100
153 115.4 89.4 134.2
152 55.9 43.0 64.4









#53
3-Nitroaniline
Concen: 21.911 ng
RT: 14.304 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 177095
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 129.2 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 20.539 ng
RT: 14.504 min Scan# 1941
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:184 Resp: 77917
Ion Ratio Lower Upper
184 100
63 87.2 47.0 70.6#
154 68.3 52.3 78.5









#55
Dibenzofuran
Concen: 21.120 ng
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 922986
Ion Ratio Lower Upper
168 100
139 43.0 31.0 46.6
169 12.8 10.8 16.2



#56
4-Nitrophenol
Concen: 22.225 ng
RT: 14.604 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:139 Resp: 152814
Ion Ratio Lower Upper
139 100
109 84.7 72.3 112.3
65 115.8 72.5 112.5#









#57
2,4-Dinitrotoluene
Concen: 21.781 ng
RT: 14.763 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 214303
Ion Ratio Lower Upper
165 100
63 50.0 32.3 48.5#
89 85.7 55.9 83.9#
182 2.7 2.2 3.4



#58
Fluorene
Concen: 21.014 ng
RT: 15.445 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:166 Resp: 732013
Ion Ratio Lower Upper
166 100
165 98.2 78.0 117.0
167 13.1 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.577 ng
RT: 15.016 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:232 Resp: 165214
Ion Ratio Lower Upper
232 100
131 61.6 32.2 48.4#
130 3.4 2.0 3.0#
166 37.1 19.4 29.2#



#60
Diethylphthalate
Concen: 21.833 ng
RT: 15.228 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:149 Resp: 778670
Ion Ratio Lower Upper
149 100
177 19.9 17.5 26.3
150 12.6 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 20.130 ng
RT: 15.445 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 328544
Ion Ratio Lower Upper
204 100
206 32.2 26.6 40.0
141 71.3 48.6 73.0



#62
4-Nitroaniline
Concen: 22.135 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:138 Resp: 185448
Ion Ratio Lower Upper
138 100
92 57.6 31.2 71.2
108 88.3 78.6 118.6









#63
Azobenzene
Concen: 24.064 ng
RT: 15.734 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion	Ratio	Lower	Upper
77	100		
182	18.3	6.0	46.0
105	18.0	0.0	38.8
51	24.3	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.145 min Scan# 2390
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.6	5.6	8.4#
80	13.2	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 20.504 ng
RT: 15.522 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:198 Resp: 110377
Ion Ratio Lower Upper
198 100
51 48.0 14.0 54.0
105 55.5 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 20.934 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:169 Resp: 628051
Ion Ratio Lower Upper
169 100
168 68.6 54.4 81.6
167 37.2 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 22.287 ng
RT: 16.339 min Scan# 2253
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 190554
Ion Ratio Lower Upper
248 100
250 96.7 79.8 119.8
141 95.9 57.8 86.6#



#68
Hexachlorobenzene
Concen: 19.879 ng
RT: 16.445 min Scan# 2271
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:284 Resp: 215047
Ion Ratio Lower Upper
284 100
142 43.5 26.9 40.3#
249 33.9 24.5 36.7









#69
Atrazine
Concen: 20.402 ng
RT: 16.622 min Scan# 2301
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:200 Resp: 193241
Ion Ratio Lower Upper
200 100
173 27.3 4.1 44.1
215 46.2 27.7 67.7



#70
Pentachlorophenol
Concen: 20.456 ng
RT: 16.786 min Scan# 2329
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:266 Resp: 132763
Ion Ratio Lower Upper
266 100
268 63.2 51.6 77.4
264 63.9 50.3 75.5









#71
Phenanthrene
Concen: 21.145 ng
RT: 17.186 min Scan# 2397
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 1129336
Ion Ratio Lower Upper
178 100
176 18.7 15.1 22.7
179 15.2 12.4 18.6



#72
Anthracene
Concen: 21.269 ng
RT: 17.275 min Scan# 2412
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:178 Resp: 1098654
Ion Ratio Lower Upper
178 100
176 18.2 14.9 22.3
179 15.5 12.2 18.2









#73
 Carbazole
 Concen: 21.380 ng
 RT: 17.551 min Scan# 2459
 Delta R.T. 0.000 min
 Lab File: BP006451.D
 Acq: 02 Aug 2021 14:51

Instrument :
 BNA_P
ClientSampleId :
 SSTDICC020

Tgt Ion:167 Resp: 1104923




Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.1	25.7
139	14.6	10.7	16.1




#74
 Di-n-butylphthalate
 Concen: 21.508 ng
 RT: 18.122 min Scan# 2556
 Delta R.T. 0.006 min
 Lab File: BP006451.D
 Acq: 02 Aug 2021 14:51

Tgt Ion:149 Resp: 1257885

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	5.9	4.5	6.7









#75
Fluoranthene
Concen: 20.951 ng
RT: 19.157 min Scan# 2732
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:202 Resp: 1252302
Ion Ratio Lower Upper
202 100
101 15.1 0.0 27.6
203 17.3 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.245 min Scan# 3087
Delta R.T. 0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:240 Resp: 955463
Ion Ratio Lower Upper
240 100
120 12.7 6.5 9.7#
236 25.6 20.6 31.0









#77
Benzidine
Concen: 17.788 ng
RT: 19.345 min Scan# 2764
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:184 Resp: 443624
Ion Ratio Lower Upper
184 100
185 14.5 11.0 16.4
183 11.9 9.5 14.3



#78
Pyrene
Concen: 20.858 ng
RT: 19.510 min Scan# 2792
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:202 Resp: 1310918
Ion Ratio Lower Upper
202 100
200 20.9 16.3 24.5
203 17.4 14.2 21.2









#79
Terphenyl-d14
Concen: 41.572 ng
RT: 19.716 min Scan# 2827
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:244 Resp: 1963136
Ion Ratio Lower Upper
244 100
212 7.8 6.0 9.0
122 14.3 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 20.994 ng
RT: 20.404 min Scan# 2944
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:149 Resp: 558268
Ion Ratio Lower Upper
149 100
91 82.5 54.6 81.8#
206 17.7 16.7 25.1









#81
Benzo(a)anthracene
Concen: 20.895 ng
RT: 21.227 min Scan# 3084
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 1281870
Ion Ratio Lower Upper
228 100
226 27.4 21.6 32.4
229 19.3 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 20.769 ng
RT: 21.169 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:252 Resp: 337251
Ion Ratio Lower Upper
252 100
254 66.8 50.6 76.0
126 14.3 7.0 10.6#









#83
Chrysene
Concen: 20.886 ng
RT: 21.280 min Scan# 3093
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 1239321
Ion Ratio Lower Upper
228 100
226 30.3 23.8 35.6
229 19.3 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 21.914 ng
RT: 21.169 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:149 Resp: 878037
Ion Ratio Lower Upper
149 100
167 25.1 21.9 32.9
279 3.5 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 21.284 ng
RT: 22.074 min Scan# 3228
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 1418421
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 9.8 5.6 8.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.574 min Scan# 3483
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:264 Resp: 980677
Ion Ratio Lower Upper
264 100
260 24.2 18.6 28.0
265 21.9 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 20.373 ng
RT: 25.980 min Scan# 3892
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:276 Resp: 1428053
Ion Ratio Lower Upper
276 100
138 31.8 15.3 22.9#
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 20.490 ng
RT: 22.863 min Scan# 3362
Delta R.T. 0.000 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:252 Resp: 1291716
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 13.2 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 21.247 ng
RT: 22.910 min Scan# 3370
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:252 Resp: 1268208
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 13.2 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 20.618 ng
RT: 23.468 min Scan# 3465
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:252 Resp: 1164366
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 14.6 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 20.530 ng
RT: 26.004 min Scan# 3896
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 1243297
Ion Ratio Lower Upper
278 100
139 20.3 9.3 13.9#
279 23.8 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 20.398 ng
RT: 26.721 min Scan# 4018
Delta R.T. -0.006 min
Lab File: BP006451.D
Acq: 02 Aug 2021 14:51

Tgt Ion:276 Resp: 1191073
Ion Ratio Lower Upper
276 100
277 24.2 18.9 28.3
138 26.4 12.6 18.8#

