

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006456.D
 Acq On : 02 Aug 2021 18:22
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP080221

Quant Time: Aug 02 19:55:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 19:55:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.568	0.576	-1.4	131	0.00
3	Pyridine	1.641	1.684	-2.6	131	0.00
4	n-Nitrosodimethylamine	0.620	0.638	-2.9	131	0.00
5 S	2-Fluorophenol	1.282	1.327	-3.5	132	0.00
6	Aniline	1.979	2.023	-2.2	129	0.00
7 S	Phenol-d6	1.811	1.866	-3.0	130	0.00
8	2-Chlorophenol	1.387	1.412	-1.8	130	0.00
9	Benzaldehyde	1.135	1.103	2.8	134	0.00
10 C	Phenol	1.824	1.855	-1.7	129	0.00
11	bis(2-Chloroethyl)ether	1.396	1.411	-1.1	130	0.00
12	1,3-Dichlorobenzene	1.464	1.473	-0.6	131	-0.01
13 C	1,4-Dichlorobenzene	1.481	1.487	-0.4	130	0.00
14	1,2-Dichlorobenzene	1.423	1.420	0.2	131	0.00
15	Benzyl Alcohol	1.356	1.390	-2.5	129	0.00
16	2,2'-oxybis(1-Chloropropane	1.648	1.620	1.7	127	-0.01
17	2-Methylphenol	1.141	1.188	-4.1	129	0.00
18	Hexachloroethane	0.580	0.577	0.5	129	-0.01
19 P	n-Nitroso-di-n-propylamine	1.201	1.221	-1.7	126	0.00
20	3+4-Methylphenols	1.557	1.612	-3.5	130	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
22	Acetophenone	0.527	0.523	0.8	127	0.00
23 S	Nitrobenzene-d5	0.424	0.425	-0.2	126	0.00
24	Nitrobenzene	0.440	0.441	-0.2	126	0.00
25	Isophorone	0.757	0.772	-2.0	126	0.00
26 C	2-Nitrophenol	0.158	0.170	-7.6	132	0.00
27	2,4-Dimethylphenol	0.270	0.278	-3.0	126	0.00
28	bis(2-Chloroethoxy)methane	0.414	0.410	1.0	126	0.00
29 C	2,4-Dichlorophenol	0.273	0.283	-3.7	129	0.00
30	1,2,4-Trichlorobenzene	0.296	0.294	0.7	130	0.00
31	Naphthalene	1.065	1.061	0.4	128	0.00
32	Benzoic acid	0.205	0.219	-6.8	135	0.00
33	4-Chloroaniline	0.426	0.433	-1.6	124	0.00
34 C	Hexachlorobutadiene	0.173	0.173	0.0	130	0.00
35	Caprolactam	0.100	0.102	-2.0	120	0.00
36 C	4-Chloro-3-methylphenol	0.340	0.351	-3.2	124	0.00
37	2-Methylnaphthalene	0.715	0.720	-0.7	126	0.00
38	1-Methylnaphthalene	0.679	0.679	0.0	126	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	0.521	0.515	1.2	127	0.00
41 P	Hexachlorocyclopentadiene	0.276	0.285	-3.3	129	0.00
42 S	2,4,6-Tribromophenol	0.205	0.211	-2.9	121	0.00
43 C	2,4,6-Trichlorophenol	0.350	0.364	-4.0	127	0.00
44	2,4,5-Trichlorophenol	0.387	0.399	-3.1	125	0.00
45 S	2-Fluorobiphenyl	1.334	1.312	1.6	125	0.00
46	1,1'-Biphenyl	1.512	1.484	1.9	124	0.00
47	2-Chloronaphthalene	1.164	1.142	1.9	124	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006456.D
 Acq On : 02 Aug 2021 18:22
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP080221

Quant Time: Aug 02 19:55:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 19:55:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.390	0.408	-4.6	120	0.00
49	Acenaphthylene	1.850	1.852	-0.1	121	0.00
50	Dimethylphthalate	1.438	1.422	1.1	119	0.00
51	2,6-Dinitrotoluene	0.313	0.323	-3.2	120	0.00
52 C	Acenaphthene	1.141	1.110	2.7	121	0.00
53	3-Nitroaniline	0.347	0.362	-4.3	118	0.00
54 P	2,4-Dinitrophenol	0.170	0.176	-3.5	123	0.00
55	Dibenzofuran	1.787	1.737	2.8	120	0.00
56 P	4-Nitrophenol	0.315	0.320	-1.6	116	0.00
57	2,4-Dinitrotoluene	0.418	0.443	-6.0	119	0.00
58	Fluorene	1.419	1.401	1.3	120	0.00
59	2,3,4,6-Tetrachlorophenol	0.325	0.333	-2.5	122	0.00
60	Diethylphthalate	1.495	1.505	-0.7	120	0.00
61	4-Chlorophenyl-phenylether	0.645	0.632	2.0	121	0.00
62	4-Nitroaniline	0.367	0.384	-4.6	114	0.00
63	Azobenzene	1.726	1.732	-0.3	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.124	-3.3	120	0.00
66 c	n-Nitrosodiphenylamine	0.619	0.627	-1.3	119	0.00
67	4-Bromophenyl-phenylether	0.192	0.193	-0.5	119	0.00
68	Hexachlorobenzene	0.216	0.217	-0.5	121	0.00
69	Atrazine	0.191	0.199	-4.2	115	0.00
70 C	Pentachlorophenol	0.134	0.145	-8.2	120	0.00
71	Phenanthrene	1.122	1.096	2.3	114	0.00
72	Anthracene	1.075	1.086	-1.0	115	0.00
73	Carbazole	1.092	1.100	-0.7	113	0.00
74	Di-n-butylphthalate	1.233	1.306	-5.9	116	0.00
75 C	Fluoranthene	1.248	1.254	-0.5	112	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.492	0.560	-13.8	109	0.00
78	Pyrene	1.319	1.318	0.1	110	0.00
79 S	Terphenyl-d14	0.993	0.987	0.6	112	0.00
80	Butylbenzylphthalate	0.570	0.618	-8.4	113	0.00
81	Benzo(a)anthracene	1.288	1.273	1.2	108	0.00
82	3,3'-Dichlorobenzidine	0.341	0.357	-4.7	110	0.00
83	Chrysene	1.255	1.234	1.7	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.880	0.933	-6.0	113	0.00
85 c	Di-n-octyl phthalate	1.490	1.550	-4.0	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.435	1.472	-2.6	107	0.02
88	Benzo(b)fluoranthene	1.284	1.334	-3.9	111	0.01
89	Benzo(k)fluoranthene	1.245	1.234	0.9	104	0.00
90 C	Benzo(a)pyrene	1.149	1.188	-3.4	107	0.00
91	Dibenzo(a,h)anthracene	1.243	1.261	-1.4	106	0.01
92	Benzo(g,h,i)perylene	1.191	1.209	-1.5	106	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
Data File : BP006456.D
Acq On : 02 Aug 2021 18:22
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP080221

Quant Time: Aug 02 19:55:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 19:55:20 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0