

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006456.D
 Acq On : 02 Aug 2021 18:22
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP080221

Quant Time: Aug 02 19:55:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 19:55:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	40.532	-1.3	131	0.00
3	Pyridine	40.000	41.035	-2.6	131	0.00
4	n-Nitrosodimethylamine	40.000	41.144	-2.9	131	0.00
5 S	2-Fluorophenol	80.000	82.759	-3.4	132	0.00
6	Aniline	40.000	40.876	-2.2	129	0.00
7 S	Phenol-d6	80.000	82.439	-3.0	130	0.00
8	2-Chlorophenol	40.000	40.711	-1.8	130	0.00
9	Benzaldehyde	40.000	38.871	2.8	134	0.00
10 C	Phenol	40.000	40.677	-1.7	129	0.00
11	bis(2-Chloroethyl)ether	40.000	40.423	-1.1	130	0.00
12	1,3-Dichlorobenzene	40.000	40.233	-0.6	131	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.163	-0.4	130	0.00
14	1,2-Dichlorobenzene	40.000	39.912	0.2	131	0.00
15	Benzyl Alcohol	40.000	41.012	-2.5	129	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.331	1.7	127	-0.01
17	2-Methylphenol	40.000	41.615	-4.0	129	0.00
18	Hexachloroethane	40.000	39.778	0.6	129	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.695	-1.7	126	0.00
20	3+4-Methylphenols	40.000	41.419	-3.5	130	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	127	0.00
22	Acetophenone	40.000	39.684	0.8	127	0.00
23 S	Nitrobenzene-d5	80.000	80.064	-0.1	126	0.00
24	Nitrobenzene	40.000	40.076	-0.2	126	0.00
25	Isophorone	40.000	40.767	-1.9	126	0.00
26 C	2-Nitrophenol	40.000	43.167	-7.9	132	0.00
27	2,4-Dimethylphenol	40.000	41.172	-2.9	126	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.577	1.1	126	0.00
29 C	2,4-Dichlorophenol	40.000	41.455	-3.6	129	0.00
30	1,2,4-Trichlorobenzene	40.000	39.737	0.7	130	0.00
31	Naphthalene	40.000	39.841	0.4	128	0.00
32	Benzoic acid	40.000	42.776	-6.9	135	0.00
33	4-Chloroaniline	40.000	40.632	-1.6	124	0.00
34 C	Hexachlorobutadiene	40.000	39.884	0.3	130	0.00
35	Caprolactam	40.000	40.888	-2.2	120	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.223	-3.1	124	0.00
37	2-Methylnaphthalene	40.000	40.260	-0.6	126	0.00
38	1-Methylnaphthalene	40.000	40.012	-0.0	126	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.502	1.2	127	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.316	-3.3	129	0.00
42 S	2,4,6-Tribromophenol	80.000	82.514	-3.1	121	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.679	-4.2	127	0.00
44	2,4,5-Trichlorophenol	40.000	41.213	-3.0	125	0.00
45 S	2-Fluorobiphenyl	80.000	78.658	1.7	125	0.00
46	1,1'-Biphenyl	40.000	39.267	1.8	124	0.00
47	2-Chloronaphthalene	40.000	39.217	2.0	124	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006456.D
 Acq On : 02 Aug 2021 18:22
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP080221

Quant Time: Aug 02 19:55:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 19:55:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.835	-4.6	120	0.00
49	Acenaphthylene	40.000	40.040	-0.1	121	0.00
50	Dimethylphthalate	40.000	39.551	1.1	119	0.00
51	2,6-Dinitrotoluene	40.000	41.328	-3.3	120	0.00
52 C	Acenaphthene	40.000	38.916	2.7	121	0.00
53	3-Nitroaniline	40.000	41.710	-4.3	118	0.00
54 P	2,4-Dinitrophenol	40.000	41.409	-3.5	123	0.00
55	Dibenzofuran	40.000	38.884	2.8	120	0.00
56 P	4-Nitrophenol	40.000	40.589	-1.5	116	0.00
57	2,4-Dinitrotoluene	40.000	42.447	-6.1	119	0.00
58	Fluorene	40.000	39.496	1.3	120	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.032	-2.6	122	0.00
60	Diethylphthalate	40.000	40.261	-0.7	120	0.00
61	4-Chlorophenyl-phenylether	40.000	39.175	2.1	121	0.00
62	4-Nitroaniline	40.000	39.241	1.9	114	0.00
63	Azobenzene	40.000	40.138	-0.3	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.331	-3.3	120	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.508	-1.3	119	0.00
67	4-Bromophenyl-phenylether	40.000	40.294	-0.7	119	0.00
68	Hexachlorobenzene	40.000	40.236	-0.6	121	0.00
69	Atrazine	40.000	41.854	-4.6	115	0.00
70 C	Pentachlorophenol	40.000	43.160	-7.9	120	0.00
71	Phenanthrene	40.000	39.089	2.3	114	0.00
72	Anthracene	40.000	40.407	-1.0	115	0.00
73	Carbazole	40.000	40.312	-0.8	113	0.00
74	Di-n-butylphthalate	40.000	42.391	-6.0	116	0.00
75 C	Fluoranthene	40.000	40.188	-0.5	112	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	45.485	-13.7	109	0.00
78	Pyrene	40.000	39.990	0.0	110	0.00
79 S	Terphenyl-d14	80.000	79.460	0.7	112	0.00
80	Butylbenzylphthalate	40.000	40.008	-0.0	113	0.00
81	Benzo(a)anthracene	40.000	39.552	1.1	108	0.00
82	3,3'-Dichlorobenzidine	40.000	41.890	-4.7	110	0.00
83	Chrysene	40.000	39.360	1.6	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.396	-6.0	113	0.00
85 c	Di-n-octyl phthalate	40.000	41.607	-4.0	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.022	-2.6	107	0.02
88	Benzo(b)fluoranthene	40.000	41.549	-3.9	111	0.01
89	Benzo(k)fluoranthene	40.000	39.646	0.9	104	0.00
90 C	Benzo(a)pyrene	40.000	41.357	-3.4	107	0.00
91	Dibenzo(a,h)anthracene	40.000	40.583	-1.5	106	0.01
92	Benzo(g,h,i)perylene	40.000	40.591	-1.5	106	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
Data File : BP006456.D
Acq On : 02 Aug 2021 18:22
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP080221

Quant Time: Aug 02 19:55:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 19:55:20 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0