

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006462.D
 Acq On : 02 Aug 2021 21:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:30:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	39.324	1.7	103	0.00
3	Pyridine	40.000	40.812	-2.0	105	0.00
4	n-Nitrosodimethylamine	40.000	41.133	-2.8	106	0.00
5 S	2-Fluorophenol	80.000	80.537	-0.7	104	0.00
6	Aniline	40.000	41.989	-5.0	106	0.00
7 S	Phenol-d6	80.000	84.000	-5.0	107	0.00
8	2-Chlorophenol	40.000	41.252	-3.1	107	0.00
9	Benzaldehyde	40.000	38.793	3.0	107	0.00
10 C	Phenol	40.000	41.557	-3.9	106	0.00
11	bis(2-Chloroethyl)ether	40.000	41.173	-2.9	107	0.00
12	1,3-Dichlorobenzene	40.000	39.936	0.2	105	0.00
13 C	1,4-Dichlorobenzene	40.000	39.925	0.2	104	0.00
14	1,2-Dichlorobenzene	40.000	40.136	-0.3	106	0.00
15	Benzyl Alcohol	40.000	42.843	-7.1	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.360	-3.4	108	0.00
17	2-Methylphenol	40.000	42.823	-7.1	107	0.00
18	Hexachloroethane	40.000	40.241	-0.6	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.775	-9.4	110	0.00
20	3+4-Methylphenols	40.000	43.177	-7.9	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	39.922	0.2	108	0.00
23 S	Nitrobenzene-d5	80.000	80.953	-1.2	109	0.00
24	Nitrobenzene	40.000	40.645	-1.6	109	0.00
25	Isophorone	40.000	42.077	-5.2	110	0.00
26 C	2-Nitrophenol	40.000	42.159	-5.4	110	0.00
27	2,4-Dimethylphenol	40.000	41.679	-4.2	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.209	-0.5	109	0.00
29 C	2,4-Dichlorophenol	40.000	41.417	-3.5	109	0.00
30	1,2,4-Trichlorobenzene	40.000	39.351	1.6	109	0.00
31	Naphthalene	40.000	39.855	0.4	109	0.00
32	Benzoic acid	40.000	41.951	-4.9	113	0.00
33	4-Chloroaniline	40.000	41.605	-4.0	108	0.00
34 C	Hexachlorobutadiene	40.000	38.844	2.9	108	0.00
35	Caprolactam	40.000	44.240	-10.6	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.707	-9.3	111	0.00
37	2-Methylnaphthalene	40.000	41.369	-3.4	110	0.00
38	1-Methylnaphthalene	40.000	41.261	-3.2	110	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	111	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.000	5.0	110	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.080	2.3	110	0.00
42 S	2,4,6-Tribromophenol	80.000	82.750	-3.4	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.542	-1.4	111	0.00
44	2,4,5-Trichlorophenol	40.000	40.510	-1.3	111	0.00
45 S	2-Fluorobiphenyl	80.000	76.680	4.1	110	0.00
46	1,1'-Biphenyl	40.000	38.448	3.9	110	0.00
47	2-Chloronaphthalene	40.000	38.429	3.9	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006462.D
 Acq On : 02 Aug 2021 21:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:30:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.111	-7.8	112	0.00
49	Acenaphthylene	40.000	40.186	-0.5	110	0.00
50	Dimethylphthalate	40.000	40.185	-0.5	110	0.00
51	2,6-Dinitrotoluene	40.000	41.951	-4.9	110	0.00
52 C	Acenaphthene	40.000	39.234	1.9	110	0.00
53	3-Nitroaniline	40.000	43.550	-8.9	111	0.00
54 P	2,4-Dinitrophenol	40.000	40.113	-0.3	108	0.00
55	Dibenzofuran	40.000	39.522	1.2	110	0.00
56 P	4-Nitrophenol	40.000	42.648	-6.6	110	0.00
57	2,4-Dinitrotoluene	40.000	43.582	-9.0	111	0.00
58	Fluorene	40.000	40.091	-0.2	110	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.058	-2.6	110	0.00
60	Diethylphthalate	40.000	41.062	-2.7	110	0.00
61	4-Chlorophenyl-phenylether	40.000	39.532	1.2	110	0.00
62	4-Nitroaniline	40.000	41.771	-4.4	110	0.00
63	Azobenzene	40.000	42.032	-5.1	111	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.394	1.5	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.144	2.1	110	0.00
67	4-Bromophenyl-phenylether	40.000	38.088	4.8	108	0.00
68	Hexachlorobenzene	40.000	39.068	2.3	112	0.00
69	Atrazine	40.000	41.417	-3.5	109	0.00
70 C	Pentachlorophenol	40.000	41.905	-4.8	112	0.00
71	Phenanthrene	40.000	39.138	2.2	109	0.00
72	Anthracene	40.000	39.961	0.1	109	0.00
73	Carbazole	40.000	40.655	-1.6	109	0.00
74	Di-n-butylphthalate	40.000	41.495	-3.7	109	0.00
75 C	Fluoranthene	40.000	40.672	-1.7	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	42.996	-7.5	102	0.00
78	Pyrene	40.000	39.574	1.1	108	0.00
79 S	Terphenyl-d14	80.000	77.544	3.1	108	0.00
80	Butylbenzylphthalate	40.000	38.933	2.7	109	0.00
81	Benzo(a)anthracene	40.000	39.938	0.2	108	0.00
82	3,3'-Dichlorobenzidine	40.000	40.766	-1.9	106	0.00
83	Chrysene	40.000	39.599	1.0	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.245	-0.6	106	0.00
85 c	Di-n-octyl phthalate	40.000	39.716	0.7	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.385	-1.0	107	0.00
88	Benzo(b)fluoranthene	40.000	40.122	-0.3	109	0.00
89	Benzo(k)fluoranthene	40.000	39.828	0.4	106	0.00
90 C	Benzo(a)pyrene	40.000	40.977	-2.4	108	0.00
91	Dibenzo(a,h)anthracene	40.000	40.278	-0.7	107	0.00
92	Benzo(g,h,i)perylene	40.000	40.470	-1.2	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
Data File : BP006462.D
Acq On : 02 Aug 2021 21:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 03 05:30:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 05:26:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0