

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006464.D
 Acq On : 03 Aug 2021 00:03
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:27:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	0.00
2	1,4-Dioxane	40.000	38.642	3.4	129	0.00
3	Pyridine	40.000	39.660	0.9	130	0.00
4	n-Nitrosodimethylamine	40.000	42.615	-6.5	140	0.00
5 S	2-Fluorophenol	80.000	80.514	-0.6	133	0.00
6	Aniline	40.000	40.586	-1.5	132	0.00
7 S	Phenol-d6	80.000	81.350	-1.7	132	0.00
8	2-Chlorophenol	40.000	40.310	-0.8	133	0.00
9	Benzaldehyde	40.000	38.603	3.5	137	0.00
10 C	Phenol	40.000	40.526	-1.3	132	0.00
11	bis(2-Chloroethyl)ether	40.000	40.384	-1.0	134	0.00
12	1,3-Dichlorobenzene	40.000	39.637	0.9	133	0.00
13 C	1,4-Dichlorobenzene	40.000	39.837	0.4	133	0.00
14	1,2-Dichlorobenzene	40.000	39.636	0.9	134	0.00
15	Benzyl Alcohol	40.000	42.256	-5.6	137	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.622	-1.6	136	0.00
17	2-Methylphenol	40.000	41.688	-4.2	134	0.00
18	Hexachloroethane	40.000	41.483	-3.7	139	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.427	-6.1	136	0.00
20	3+4-Methylphenols	40.000	41.756	-4.4	135	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	133	0.00
22	Acetophenone	40.000	40.147	-0.4	134	0.00
23 S	Nitrobenzene-d5	80.000	82.373	-3.0	136	0.00
24	Nitrobenzene	40.000	41.130	-2.8	135	0.00
25	Isophorone	40.000	41.612	-4.0	135	0.00
26 C	2-Nitrophenol	40.000	42.167	-5.4	136	0.00
27	2,4-Dimethylphenol	40.000	41.592	-4.0	133	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.313	-0.8	134	0.00
29 C	2,4-Dichlorophenol	40.000	41.745	-4.4	136	0.00
30	1,2,4-Trichlorobenzene	40.000	40.185	-0.5	138	0.00
31	Naphthalene	40.000	40.028	-0.1	135	0.00
32	Benzoic acid	40.000	41.951	-4.9	139	0.02
33	4-Chloroaniline	40.000	41.076	-2.7	132	0.00
34 C	Hexachlorobutadiene	40.000	40.840	-2.1	140	0.00
35	Caprolactam	40.000	40.780	-2.0	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.551	-6.4	134	0.00
37	2-Methylnaphthalene	40.000	40.739	-1.8	134	0.00
38	1-Methylnaphthalene	40.000	40.825	-2.1	135	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.093	2.3	134	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.588	-4.0	139	0.00
42 S	2,4,6-Tribromophenol	80.000	83.081	-3.9	131	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.985	-2.5	133	0.00
44	2,4,5-Trichlorophenol	40.000	40.681	-1.7	132	0.00
45 S	2-Fluorobiphenyl	80.000	77.979	2.5	132	0.00
46	1,1'-Biphenyl	40.000	38.899	2.8	132	0.00
47	2-Chloronaphthalene	40.000	38.835	2.9	132	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006464.D
 Acq On : 03 Aug 2021 00:03
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:27:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.196	-8.0	133	0.00
49	Acenaphthylene	40.000	40.121	-0.3	130	0.00
50	Dimethylphthalate	40.000	39.803	0.5	129	0.00
51	2,6-Dinitrotoluene	40.000	41.117	-2.8	128	0.00
52 C	Acenaphthene	40.000	39.508	1.2	131	0.00
53	3-Nitroaniline	40.000	42.019	-5.0	127	0.00
54 P	2,4-Dinitrophenol	40.000	40.465	-1.2	129	0.00
55	Dibenzofuran	40.000	39.275	1.8	130	0.00
56 P	4-Nitrophenol	40.000	40.351	-0.9	124	0.00
57	2,4-Dinitrotoluene	40.000	42.676	-6.7	128	0.00
58	Fluorene	40.000	40.072	-0.2	130	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.066	-2.7	131	0.00
60	Diethylphthalate	40.000	41.089	-2.7	131	0.00
61	4-Chlorophenyl-phenylether	40.000	39.952	0.1	132	0.00
62	4-Nitroaniline	40.000	39.694	0.8	124	0.00
63	Azobenzene	40.000	42.680	-6.7	133	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.740	0.6	127	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.232	-0.6	130	0.00
67	4-Bromophenyl-phenylether	40.000	40.333	-0.8	131	0.00
68	Hexachlorobenzene	40.000	39.823	0.4	131	0.00
69	Atrazine	40.000	42.061	-5.2	127	0.00
70 C	Pentachlorophenol	40.000	42.524	-6.3	130	0.00
71	Phenanthrene	40.000	39.166	2.1	126	0.00
72	Anthracene	40.000	40.363	-0.9	126	0.00
73	Carbazole	40.000	40.247	-0.6	124	0.00
74	Di-n-butylphthalate	40.000	43.097	-7.7	130	0.00
75 C	Fluoranthene	40.000	39.832	0.4	122	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	43.501	-8.8	113	0.00
78	Pyrene	40.000	40.440	-1.1	121	0.00
79 S	Terphenyl-d14	80.000	79.954	0.1	123	0.00
80	Butylbenzylphthalate	40.000	40.404	-1.0	124	0.00
81	Benzo(a)anthracene	40.000	39.903	0.2	118	0.00
82	3,3'-Dichlorobenzidine	40.000	41.557	-3.9	118	0.00
83	Chrysene	40.000	39.665	0.8	118	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.829	-9.6	127	0.00
85 c	Di-n-octyl phthalate	40.000	43.083	-7.7	128	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.786	-2.0	117	0.00
88	Benzo(b)fluoranthene	40.000	40.217	-0.5	118	0.00
89	Benzo(k)fluoranthene	40.000	40.492	-1.2	117	0.00
90 C	Benzo(a)pyrene	40.000	41.012	-2.5	117	0.00
91	Dibenzo(a,h)anthracene	40.000	40.599	-1.5	117	0.00
92	Benzo(g,h,i)perylene	40.000	40.442	-1.1	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
Data File : BP006464.D
Acq On : 03 Aug 2021 00:03
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 03 05:27:59 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 05:26:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0