

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006470.D
 Acq On : 03 Aug 2021 03:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:28:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	136	0.00
2	1,4-Dioxane	40.000	41.687	-4.2	143	0.00
3	Pyridine	40.000	42.238	-5.6	142	0.00
4	n-Nitrosodimethylamine	40.000	42.834	-7.1	145	0.00
5 S	2-Fluorophenol	80.000	84.718	-5.9	143	0.00
6	Aniline	40.000	41.754	-4.4	139	0.00
7 S	Phenol-d6	80.000	84.425	-5.5	141	0.00
8	2-Chlorophenol	40.000	41.796	-4.5	142	0.00
9	Benzaldehyde	40.000	40.424	-1.1	147	0.00
10 C	Phenol	40.000	41.827	-4.6	140	0.00
11	bis(2-Chloroethyl)ether	40.000	41.110	-2.8	140	0.00
12	1,3-Dichlorobenzene	40.000	41.016	-2.5	142	0.00
13 C	1,4-Dichlorobenzene	40.000	41.161	-2.9	141	0.00
14	1,2-Dichlorobenzene	40.000	40.771	-1.9	141	0.00
15	Benzyl Alcohol	40.000	42.297	-5.7	140	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.952	-4.9	144	0.00
17	2-Methylphenol	40.000	42.449	-6.1	140	0.00
18	Hexachloroethane	40.000	41.539	-3.8	143	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.367	-5.9	139	0.00
20	3+4-Methylphenols	40.000	42.254	-5.6	140	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	133	0.00
22	Acetophenone	40.000	40.964	-2.4	137	0.00
23 S	Nitrobenzene-d5	80.000	84.558	-5.7	140	0.00
24	Nitrobenzene	40.000	42.180	-5.4	139	0.00
25	Isophorone	40.000	42.879	-7.2	139	0.00
26 C	2-Nitrophenol	40.000	44.203	-10.5	142	0.00
27	2,4-Dimethylphenol	40.000	42.782	-7.0	137	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.731	-4.3	139	0.00
29 C	2,4-Dichlorophenol	40.000	42.998	-7.5	140	0.00
30	1,2,4-Trichlorobenzene	40.000	41.087	-2.7	141	0.00
31	Naphthalene	40.000	41.302	-3.3	139	0.00
32	Benzoic acid	40.000	41.929	-4.8	139	0.02
33	4-Chloroaniline	40.000	42.331	-5.8	136	0.00
34 C	Hexachlorobutadiene	40.000	41.340	-3.4	142	0.00
35	Caprolactam	40.000	42.582	-6.5	131	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.956	-9.9	138	0.00
37	2-Methylnaphthalene	40.000	41.811	-4.5	137	0.00
38	1-Methylnaphthalene	40.000	41.649	-4.1	137	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	128	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.158	-2.9	137	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.818	-12.0	145	0.00
42 S	2,4,6-Tribromophenol	80.000	85.453	-6.8	131	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.237	-8.1	137	0.00
44	2,4,5-Trichlorophenol	40.000	42.952	-7.4	136	0.00
45 S	2-Fluorobiphenyl	80.000	81.833	-2.3	135	0.00
46	1,1'-Biphenyl	40.000	40.976	-2.4	135	0.00
47	2-Chloronaphthalene	40.000	41.168	-2.9	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
 Data File : BP006470.D
 Acq On : 03 Aug 2021 03:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 05:28:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 05:26:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.148	-12.9	135	0.00
49	Acenaphthylene	40.000	42.239	-5.6	133	0.00
50	Dimethylphthalate	40.000	41.697	-4.2	131	0.00
51	2,6-Dinitrotoluene	40.000	43.058	-7.6	131	0.00
52 C	Acenaphthene	40.000	41.242	-3.1	133	0.00
53	3-Nitroaniline	40.000	44.442	-11.1	131	0.00
54 P	2,4-Dinitrophenol	40.000	42.509	-6.3	132	0.00
55	Dibenzofuran	40.000	40.944	-2.4	132	0.00
56 P	4-Nitrophenol	40.000	42.303	-5.8	126	0.00
57	2,4-Dinitrotoluene	40.000	44.937	-12.3	131	0.00
58	Fluorene	40.000	41.601	-4.0	131	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.507	-6.3	132	0.00
60	Diethylphthalate	40.000	42.567	-6.4	132	0.00
61	4-Chlorophenyl-phenylether	40.000	41.415	-3.5	133	0.00
62	4-Nitroaniline	40.000	42.075	-5.2	128	0.00
63	Azobenzene	40.000	43.493	-8.7	132	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.252	-5.6	129	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.934	-4.8	129	0.00
67	4-Bromophenyl-phenylether	40.000	41.433	-3.6	129	0.00
68	Hexachlorobenzene	40.000	41.440	-3.6	131	0.00
69	Atrazine	40.000	44.191	-10.5	128	0.00
70 C	Pentachlorophenol	40.000	44.369	-10.9	130	0.00
71	Phenanthrene	40.000	41.019	-2.5	126	0.00
72	Anthracene	40.000	41.989	-5.0	126	0.00
73	Carbazole	40.000	42.154	-5.4	124	0.00
74	Di-n-butylphthalate	40.000	44.290	-10.7	128	0.00
75 C	Fluoranthene	40.000	41.463	-3.7	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	43.861	-9.7	108	0.00
78	Pyrene	40.000	42.350	-5.9	120	0.00
79 S	Terphenyl-d14	80.000	84.073	-5.1	122	0.00
80	Butylbenzylphthalate	40.000	42.629	-6.6	124	0.00
81	Benzo(a)anthracene	40.000	41.822	-4.6	117	0.00
82	3,3'-Dichlorobenzidine	40.000	43.775	-9.4	118	0.00
83	Chrysene	40.000	41.260	-3.1	116	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.262	-13.2	124	0.00
85 c	Di-n-octyl phthalate	40.000	44.161	-10.4	124	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.129	-2.8	113	0.00
88	Benzo(b)fluoranthene	40.000	42.166	-5.4	119	0.00
89	Benzo(k)fluoranthene	40.000	40.787	-2.0	113	0.00
90 C	Benzo(a)pyrene	40.000	42.226	-5.6	115	0.00
91	Dibenzo(a,h)anthracene	40.000	41.182	-3.0	114	-0.01
92	Benzo(g,h,i)perylene	40.000	41.347	-3.4	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080221\
Data File : BP006470.D
Acq On : 03 Aug 2021 03:27
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 03 05:28:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 05:26:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0