

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080222\
 Data File : BP011232.D
 Acq On : 02 Aug 2022 19:02
 Operator : CG/JU
 Sample : SSTD08035
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080635

Quant Time: Aug 02 23:18:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 02 23:11:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.605	152	210191	20.000	ng/u1	0.00
20) Naphthalene-d8	10.399	136	881051	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	441477	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.051	188	838929	20.000	ng/u1	0.00
79) Chrysene-d12	21.180	240	803076	20.000	ng/u1	0.00
88) Perylene-d12	23.498	264	882810	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.093	96	77480	16.881	ng/uL	0.00
4) Pyridine-d5	3.505	84	561672	43.257	ng/u1	0.00
7) Phenol-d5	6.793	99	1418300	77.307	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.952	67	1028905	79.312	ng/u1	0.00
11) 2-Chlorophenol-d4	7.140	132	1199059	82.934	ng/u1	0.00
15) 4-Methylphenol-d8	8.334	113	1151682	76.739	ng/u1	0.00
21) Nitrobenzene-d5	8.775	128	583068	85.154	ng/u1	0.00
24) 2-Nitrophenol-d4	9.493	143	568614	86.695	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.028	165	980794	84.854	ng/u1	0.00
31) 4-Chloroaniline-d4	10.551	131	1481380	79.057	ng/u1	0.00
46) Dimethylphthalate-d6	13.704	166	2575061	85.611	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	3235377	85.597	ng/u1	0.00
54) 4-Nitrophenol-d4	14.516	143	501912	92.769	ng/u1	0.01
60) Fluorene-d10	15.281	176	2128843	84.228	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.422	200	374098	88.460	ng/u1	0.00
73) Anthracene-d10	17.151	188	3091092	83.621	ng/u1	0.00
81) Pyrene-d10	19.410	212	3442087	79.765	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.351	264	3781231	86.234	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.128	88	77595	16.333	ng/uL#	94
5) Pyridine	3.522	79	570010	42.556	ng/u1	96
6) Benzaldehyde	6.758	77	668806	75.573	ng/u1	98
8) Phenol	6.822	94	1517249	77.085	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.046	93	1256279	78.720	ng/u1	98
12) 2-Chlorophenol	7.175	128	1234399	82.121	ng/u1	98
13) 2-Methylphenol	8.063	108	1148426	78.722	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.152	45	2110940	78.975	ng/u1	98
16) Acetophenone	8.440	105	1823249	78.100	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.440	70	914972	74.143	ng/u1	99
18) 4-Methylphenol	8.405	108	1216960	77.458	ng/u1	99
19) Hexachloroethane	8.675	117	585086	86.652	ng/u1	99
22) Nitrobenzene	8.816	77	1491172	84.704	ng/u1	96
23) Isophorone	9.352	82	2603939	81.751	ng/u1	99
25) 2-Nitrophenol	9.522	139	637986	86.566	ng/u1	96
26) 2,4-Dimethylphenol	9.593	107	1325648	83.731	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.834	93	1631391	81.517	ng/u1	98
29) 2,4-Dichlorophenol	10.057	162	1001443	83.773	ng/u1	98
30) Naphthalene	10.451	128	3792826	81.825	ng/u1	99
32) 4-Chloroaniline	10.575	127	1481762	79.583	ng/u1	99
33) Hexachlorobutadiene	10.728	225	620432	84.979	ng/u1	99
34) Caprolactam	11.393	113	168621	42.414	ng/u1	97
35) 4-Chloro-3-methylphenol	11.722	107	1146330	85.523	ng/u1	100
36) 2-Methylnaphthalene	12.075	142	2374139	80.938	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080222\
 Data File : BP011232.D
 Acq On : 02 Aug 2022 19:02
 Operator : CG/JU
 Sample : SSTD08035
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD080635


Quant Time: Aug 02 23:18:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 02 23:11:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.298	142	2365276	80.490	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.451	216	1023008	83.288	ng/ul	98
40) Hexachlorocyclopentadiene	12.422	237	669692	82.700	ng/ul	99
41) 2,4,6-Trichlorophenol	12.698	196	677507	87.842	ng/ul	100
42) 2,4,5-Trichlorophenol	12.775	196	729921	89.398	ng/ul	99
43) 1,1'-Biphenyl	13.104	154	2933387	83.160	ng/ul	99
44) 2-Chloronaphthalene	13.145	162	2170647	82.313	ng/ul	98
45) 2-Nitroaniline	13.369	65	709982	92.636	ng/ul	95
47) Dimethylphthalate	13.751	163	2557840	84.439	ng/ul	100
48) 2,6-Dinitrotoluene	13.875	165	512069	96.744	ng/ul	89
50) Acenaphthylene	13.998	152	3534528	83.706	ng/ul	99
51) 3-Nitroaniline	14.204	138	498715	87.242	ng/ul	94
52) Acenaphthene	14.345	153	2284800	82.839	ng/ul	98
53) 2,4-Dinitrophenol	14.416	184	268064	89.800	ng/ul	96
55) 4-Nitrophenol	14.528	109	436162	93.821	ng/ul	97
56) Dibenzofuran	14.687	168	3051092	81.383	ng/ul	100
57) 2,4-Dinitrotoluene	14.669	165	735078	99.536	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.916	232	538302	90.670	ng/ul	97
59) Diethylphthalate	15.134	149	2684026	88.652	ng/ul	99
61) Fluorene	15.339	166	2401256	82.370	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.339	204	1048300	79.010	ng/ul	96
63) 4-Nitroaniline	15.381	138	394976	81.271	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.434	198	377503	87.893	ng/ul#	98
67) N-Nitrosodiphenylamine	15.557	169	2033019	78.856	ng/ul	100
68) 4-Bromophenyl-phenylether	16.239	248	641492	79.685	ng/ul	98
69) Hexachlorobenzene	16.345	284	719049	78.079	ng/ul	98
70) Atrazine	16.534	200	666225	82.174	ng/ul	99
71) Pentachlorophenol	16.698	266	462683	87.508	ng/ul	99
72) Phenanthrene	17.092	178	3742599	81.515	ng/ul	99
74) Anthracene	17.186	178	3739908	82.336	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.063	216	1045284	72.274	ng/ul	99
76) Pentachlorobenzene	14.598	250	896440	75.205	ng/ul	99
77) Carbazole	17.469	167	3478999	86.452	ng/ul	99
78) Di-n-butylphthalate	18.033	149	4463313	92.747	ng/ul	100
80) Fluoranthene	19.080	202	4227298	79.304	ng/ul	99
82) Pyrene	19.439	202	4229408	77.212	ng/ul	99
83) Butylbenzylphthalate	20.333	149	1999351	91.151	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.104	252	1227002	87.597	ng/ul	99
85) Benzo(a)anthracene	21.163	228	4166350	83.456	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.104	149	3030697	91.423	ng/ul	100
87) Chrysene	21.216	228	4026495	81.131	ng/ul	98
89) Di-n-octyl phthalate	22.004	149	5488103	90.450	ng/ul	100
90) Benzo(b)fluoranthene	22.798	252	4616428	82.950	ng/ul	99
91) Benzo(k)fluoranthene	22.845	252	4491846	85.042	ng/ul	100
93) Benzo(a)pyrene	23.404	252	4588039	84.517	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.904	276	5566958	83.239	ng/ul	99
95) Dibenzo(a,h)anthracene	25.927	278	4668285	81.383	ng/ul	98
96) Benzo(g,h,i)perylene	26.645	276	4797677	82.944	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 02 23:18:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 02 23:11:33 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 16.333 ng/uL
 RT: 3.128 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080635


Tgt Ion: 88 Resp: 77595
 Ion Ratio Lower Upper
 88 100
 43 28.1 29.6 44.4#
 58 61.1 49.1 73.7



#5
 Pyridine
 Concen: 42.556 ng/uL
 RT: 3.522 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02

Tgt Ion: 79 Resp: 570010
 Ion Ratio Lower Upper
 79 100
 52 69.3 57.8 86.6
 51 36.3 31.5 47.3









#6
Benzaldehyde
Concen: 75.573 ng/ul
RT: 6.758 min Scan# 641
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion: 77 Resp: 668806
Ion Ratio Lower Upper
77 100
105 93.2 72.0 108.0
106 87.6 69.4 104.0



#8
Phenol
Concen: 77.085 ng/ul
RT: 6.822 min Scan# 652
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion: 94 Resp: 1517249
Ion Ratio Lower Upper
94 100
65 31.8 24.7 37.1
66 41.3 32.6 49.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 78.720 ng/ul
 RT: 7.046 min Scan# 690
 Delta R.T. 0.006 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080635

Tgt Ion: 93 Resp: 1256279




Ion	Ratio	Lower	Upper
93	100		
63	86.7	71.0	106.6
95	32.8	26.0	39.0




#12
 2-Chlorophenol
 Concen: 82.121 ng/ul
 RT: 7.175 min Scan# 712
 Delta R.T. 0.006 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02

Tgt Ion: 128 Resp: 1234399

Ion	Ratio	Lower	Upper
128	100		
64	52.4	43.8	65.6
130	32.5	25.4	38.2









#13
2-Methylphenol
Concen: 78.722 ng/ul
RT: 8.063 min Scan# 80
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:108 Resp: 1148426
Ion Ratio Lower Upper
108 100
107 90.5 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 78.975 ng/ul
RT: 8.152 min Scan# 878
Delta R.T. 0.018 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion: 45 Resp: 2110940
Ion Ratio Lower Upper
45 100
77 12.7 9.9 14.9
79 10.1 7.2 10.8









#16
Acetophenone
Concen: 78.100 ng/ul
RT: 8.440 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:105 Resp: 1823249
Ion Ratio Lower Upper
105 100
77 82.9 65.4 98.0
51 33.3 26.5 39.7



#17
N-Nitroso-di-n-propylamine
Concen: 74.143 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. 0.018 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion: 70 Resp: 914972
Ion Ratio Lower Upper
70 100
42 68.7 55.5 83.3
101 9.2 7.1 10.7
130 18.8 14.8 22.2









#18
4-Methylphenol
Concen: 77.458 ng/ul
RT: 8.405 min Scan# 919
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:108 Resp: 1216960
Ion Ratio Lower Upper
108 100
107 109.2 88.2 132.4



#19
Hexachloroethane
Concen: 86.652 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:117 Resp: 585086
Ion Ratio Lower Upper
117 100
201 78.9 62.2 93.2
199 50.1 39.8 59.6









#22
Nitrobenzene
Concen: 84.704 ng/ul
RT: 8.816 min Scan# 991
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion: 77 Resp: 1491172
Ion Ratio Lower Upper
77 100
123 41.7 31.4 47.0
65 14.2 10.6 15.8



#23
Isophorone
Concen: 81.751 ng/ul
RT: 9.352 min Scan# 1082
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion: 82 Resp: 2603939
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.8
138 16.0 12.6 18.8









#25
2-Nitrophenol
Concen: 86.566 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:	139	Resp:	637986
Ion Ratio	Lower	Upper	
139	100		
65	61.3	46.7	70.1
109	31.9	24.2	36.2



#26
2,4-Dimethylphenol
Concen: 83.731 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:	107	Resp:	1325648
Ion Ratio	Lower	Upper	
107	100		
121	48.9	40.3	60.5
122	83.2	67.7	101.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 81.517 ng/ul
 RT: 9.834 min Scan# 1164
 Delta R.T. 0.006 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080635

Tgt Ion: 93 Resp: 1631391




Ion	Ratio	Lower	Upper
93	100		
95	32.7	25.2	37.8
123	10.7	8.4	12.6




#29
 2,4-Dichlorophenol
 Concen: 83.773 ng/ul
 RT: 10.057 min Scan# 1202
 Delta R.T. 0.006 min
 Lab File: BP011232.D
 Acq: 02 Aug 2022 19:02

Tgt Ion:162 Resp: 1001443

Ion	Ratio	Lower	Upper
162	100		
164	64.6	52.6	79.0
98	41.5	34.2	51.4









#30
Naphthalene
Concen: 81.825 ng/ul
RT: 10.451 min Scan# 1268
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:128 Resp: 3792826
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 79.583 ng/ul
RT: 10.575 min Scan# 1290
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:127 Resp: 1481762
Ion Ratio Lower Upper
127 100
129 32.6 26.6 40.0









#33
Hexachlorobutadiene
Concen: 84.979 ng/ul
RT: 10.728 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:225 Resp: 620432
Ion Ratio Lower Upper
225 100
223 62.5 51.2 76.8
227 63.0 50.6 75.8



#34
Caprolactam
Concen: 42.414 ng/ul
RT: 11.393 min Scan# 1429
Delta R.T. 0.024 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:113 Resp: 168621
Ion Ratio Lower Upper
113 100
55 204.5 164.2 246.4
56 151.1 127.0 190.6









#35
4-Chloro-3-methylphenol
Concen: 85.523 ng/ul
RT: 11.722 min Scan# 1485
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:107 Resp: 1146330
Ion Ratio Lower Upper
107 100
144 24.5 19.0 28.4
142 74.9 59.8 89.8



#36
2-Methylnaphthalene
Concen: 80.938 ng/ul
RT: 12.075 min Scan# 1545
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:142 Resp: 2374139
Ion Ratio Lower Upper
142 100
141 90.3 71.8 107.8









#37
1-Methylnaphthalene
Concen: 80.490 ng/ul
RT: 12.298 min Scan# 1582
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:142 Resp: 2365276
Ion Ratio Lower Upper
142 100
141 93.0 74.1 111.1
116 4.2 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 83.288 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:216 Resp: 1023008
Ion Ratio Lower Upper
216 100
214 78.5 61.1 91.7
179 23.8 18.6 27.8
108 22.7 19.1 28.7









#40
Hexachlorocyclopentadiene
Concen: 82.700 ng/ul
RT: 12.422 min Scan# 1604
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:237 Resp: 669692
Ion Ratio Lower Upper
237 100
235 63.4 50.4 75.6
272 11.9 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 87.842 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:196 Resp: 677507
Ion Ratio Lower Upper
196 100
198 94.4 75.5 113.3
200 30.6 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 89.398 ng/ul
RT: 12.775 min Scan# 1664
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:196 Resp: 729921
Ion Ratio Lower Upper
196 100
198 96.3 76.8 115.2
200 30.5 25.0 37.6



#43
1,1'-Biphenyl
Concen: 83.160 ng/ul
RT: 13.104 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:154 Resp: 2933387
Ion Ratio Lower Upper
154 100
153 40.7 31.9 47.9
76 15.9 12.2 18.2





#44

2-Chloronaphthalene

Concen: 82.313 ng/ul

RT: 13.145 min Scan# 1726

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :

BNA_P

ClientSampleId :

SSTD080635





Tgt Ion:162 Resp: 2170647

Ion Ratio Lower Upper

162 100

164 32.7 25.6 38.4

127 44.7 34.6 51.8



#45

2-Nitroaniline

Concen: 92.636 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02




Tgt Ion: 65 Resp: 709982


Ion Ratio Lower Upper

65 100

92 65.4 51.4 77.2

138 92.0 67.4 101.0









#47
Dimethylphthalate
Concen: 84.439 ng/ul
RT: 13.751 min Scan# 1830
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:163 Resp: 2557840
Ion Ratio Lower Upper
163 100
194 3.6 2.9 4.3
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 96.744 ng/ul
RT: 13.875 min Scan# 1851
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:165 Resp: 512069
Ion Ratio Lower Upper
165 100
89 71.4 65.8 98.6
121 25.9 24.2 36.2









#50
Acenaphthylene
Concen: 83.706 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:152 Resp: 3534528
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.4 10.5 15.7



#51
3-Nitroaniline
Concen: 87.242 ng/ul
RT: 14.204 min Scan# 1907
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:138 Resp: 498715
Ion Ratio Lower Upper
138 100
108 10.8 9.9 14.9
92 125.9 106.8 160.2









#52
Acenaphthene
Concen: 82.839 ng/ul
RT: 14.345 min Scan# 1930
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:153 Resp: 2284800
Ion Ratio Lower Upper
153 100
152 48.8 39.5 59.3
154 88.1 68.6 103.0



#53
2,4-Dinitrophenol
Concen: 89.800 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:184 Resp: 268064
Ion Ratio Lower Upper
184 100
63 88.1 75.0 112.6
154 67.8 53.1 79.7









#55
4-Nitrophenol
Concen: 93.821 ng/ul
RT: 14.528 min Scan# 1962
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:109 Resp: 436162
Ion Ratio Lower Upper
109 100
139 122.1 100.6 150.8
65 148.5 122.6 184.0



#56
Dibenzofuran
Concen: 81.383 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:168 Resp: 3051092
Ion Ratio Lower Upper
168 100
139 42.5 34.2 51.2









#57
2,4-Dinitrotoluene
Concen: 99.536 ng/ul
RT: 14.669 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:165 Resp: 735078
Ion Ratio Lower Upper
165 100
63 66.0 50.6 75.8
182 2.4 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 90.670 ng/ul
RT: 14.916 min Scan# 2028
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:232 Resp: 538302
Ion Ratio Lower Upper
232 100
131 61.0 50.4 75.6
130 3.2 2.6 3.8
166 37.8 31.9 47.9









#59
Diethylphthalate
Concen: 88.652 ng/ul
RT: 15.134 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:149 Resp: 2684026
Ion Ratio Lower Upper
149 100
177 19.7 16.1 24.1
150 12.5 9.4 14.0



#61
Fluorene
Concen: 82.370 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:166 Resp: 2401256
Ion Ratio Lower Upper
166 100
165 98.0 77.7 116.5
167 13.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 79.010 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:204 Resp: 1048300
Ion Ratio Lower Upper
204 100
206 33.0 27.3 40.9
141 76.7 57.7 86.5



#63
4-Nitroaniline
Concen: 81.271 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:138 Resp: 394976
Ion Ratio Lower Upper
138 100
92 59.6 55.9 83.9
108 85.2 67.1 100.7









#66
4,6-Dinitro-2-methylphenol
Concen: 87.893 ng/ul
RT: 15.434 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:198 Resp: 377503
Ion Ratio Lower Upper
198 100
182 5.7 7.4 11.0#
77 36.0 28.6 42.8



#67
N-Nitrosodiphenylamine
Concen: 78.856 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:169 Resp: 2033019
Ion Ratio Lower Upper
169 100
168 66.5 53.6 80.4
167 37.4 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 79.685 ng/ul
RT: 16.239 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:248 Resp: 641492
Ion Ratio Lower Upper
248 100
250 97.2 77.3 115.9
141 98.0 76.4 114.6



#69
Hexachlorobenzene
Concen: 78.079 ng/ul
RT: 16.345 min Scan# 2271
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:284 Resp: 719049
Ion Ratio Lower Upper
284 100
142 48.7 39.5 59.3
249 33.9 25.1 37.7





#70

Atrazine

Concen: 82.174 ng/ul

RT: 16.534 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :

BNA_P

ClientSampleId :

SSTD080635





Tgt Ion:200 Resp: 666225

Ion Ratio Lower Upper

200 100

173 28.0 22.6 34.0

215 45.3 36.5 54.7



#71

Pentachlorophenol

Concen: 87.508 ng/ul

RT: 16.698 min Scan# 2331

Delta R.T. -0.000 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02




Tgt Ion:266 Resp: 462683


Ion Ratio Lower Upper

266 100

264 64.1 52.2 78.2

268 63.5 50.6 75.8









#72
Phenanthrene
Concen: 81.515 ng/ul
RT: 17.092 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:178 Resp: 3742599
Ion Ratio Lower Upper
178 100
179 15.8 12.1 18.1
176 19.6 15.4 23.2



#74
Anthracene
Concen: 82.336 ng/ul
RT: 17.186 min Scan# 2414
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:178 Resp: 3739908
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.1 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 72.274 ng/uL

RT: 13.063 min Scan# 1713

Delta R.T. -0.000 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :

BNA_P

ClientSampleId :

SSTD080635





Tgt Ion:216 Resp: 1045284

Ion Ratio Lower Upper

216 100

214 79.0 62.0 93.0

218 48.5 38.2 57.4



#76

Pentachlorobenzene

Concen: 75.205 ng/uL

RT: 14.598 min Scan# 1974

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02




Tgt Ion:250 Resp: 896440


Ion Ratio Lower Upper

250 100

248 62.0 49.2 73.8

252 62.6 50.8 76.2





#77

Carbazole

Concen: 86.452 ng/ul

RT: 17.469 min Scan# 2462

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :





BNA_P

ClientSampleId :

SSTD080635

Tgt Ion:167 Resp: 3478999

Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.1	25.7
139	14.9	11.5	17.3



#78

Di-n-butylphthalate

Concen: 92.747 ng/ul

RT: 18.033 min Scan# 2558




Delta R.T. -0.000 min


Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Tgt Ion:149 Resp: 4463313

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.0
104	6.3	5.1	7.7









#80
Fluoranthene
Concen: 79.304 ng/u1
RT: 19.080 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:202 Resp: 4227298
Ion Ratio Lower Upper
202 100
101 15.6 11.9 17.9
100 11.9 9.3 13.9



#82
Pyrene
Concen: 77.212 ng/u1
RT: 19.439 min Scan# 2797
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:202 Resp: 4229408
Ion Ratio Lower Upper
202 100
101 18.0 14.3 21.5
100 15.1 11.8 17.6









#83
Butylbenzylphthalate
Concen: 91.151 ng/ul
RT: 20.333 min Scan# 2949
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:149 Resp: 1999351
Ion Ratio Lower Upper
149 100
91 81.5 64.2 96.4
206 17.3 14.0 21.0



#84
3,3'-Dichlorobenzidine
Concen: 87.597 ng/ul
RT: 21.104 min Scan# 3080
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:252 Resp: 1227002
Ion Ratio Lower Upper
252 100
254 64.0 51.8 77.6
126 15.5 11.4 17.0









#85
Benzo(a)anthracene
Concen: 83.456 ng/ul
RT: 21.163 min Scan# 3089
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:228 Resp: 4166350
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.6
226 27.7 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 91.423 ng/ul
RT: 21.104 min Scan# 3080
Delta R.T. -0.000 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:149 Resp: 3030697
Ion Ratio Lower Upper
149 100
167 25.3 20.4 30.6
279 3.4 2.9 4.3





#87

Chrysene

Concen: 81.131 ng/ul

RT: 21.216 min Scan# 3099

Delta R.T. 0.006 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :

BNA_P

ClientSampleId :

SSTD080635





Tgt Ion:228 Resp: 4026495

Ion Ratio Lower Upper

228 100

226 30.8 23.8 35.6

229 20.0 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 90.450 ng/ul




RT: 22.004 min Scan# 3233


Delta R.T. -0.000 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Tgt Ion:149 Resp: 5488103









#90
Benzo(b)fluoranthene
Concen: 82.950 ng/ul
RT: 22.798 min Scan# 3367
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:252 Resp: 4616428
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 13.6 10.8 16.2



#91
Benzo(k)fluoranthene
Concen: 85.042 ng/ul
RT: 22.845 min Scan# 3376
Delta R.T. 0.006 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:252 Resp: 4491846
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 14.1 11.0 16.6









#93
Benzo(a)pyrene
Concen: 84.517 ng/ul
RT: 23.404 min Scan# 3469
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02




Instrument :
BNA_P
ClientSampleId :
SSTD080635


Tgt Ion:252 Resp: 4588039
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 15.1 12.5 18.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 83.239 ng/ul
RT: 25.904 min Scan# 3896
Delta R.T. 0.012 min
Lab File: BP011232.D
Acq: 02 Aug 2022 19:02

Tgt Ion:276 Resp: 5566958
Ion Ratio Lower Upper
276 100
138 32.3 25.5 38.3
277 25.1 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 81.383 ng/ul

RT: 25.927 min Scan# 3900

Delta R.T. 0.018 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Instrument :

BNA_P

ClientSampleId :

SSTD080635





Tgt Ion:278 Resp: 4668285

Ion Ratio Lower Upper

278 100

139 21.8 18.5 27.7

279 24.2 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 82.944 ng/ul

RT: 26.645 min Scan# 4022

Delta R.T. 0.018 min

Lab File: BP011232.D

Acq: 02 Aug 2022 19:02

Tgt Ion:276 Resp: 4797677

Ion Ratio Lower Upper

276 100

138 29.4 22.8 34.2

277 24.0 18.5 27.7

