

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080322\
 Data File : BP011250.D
 Acq On : 03 Aug 2022 18:07
 Operator : CG/JU
 Sample : N3900-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0C86

Quant Time: Aug 03 23:01:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 02 23:52:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.599	152	125216	20.000	ng/u1	0.00
20) Naphthalene-d8	10.393	136	504266	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.275	164	252577	20.000	ng/u1	# 0.00
64) Phenanthrene-d10	17.045	188	452839	20.000	ng/u1	0.00
79) Chrysene-d12	21.174	240	575798	20.000	ng/u1	0.00
88) Perylene-d12	23.492	264	670946	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.093	96	6282	4.056	ng/uL	0.00
4) Pyridine-d5	3.511	84	33480	8.015	ng/u1	0.00
7) Phenol-d5	6.787	99	40043	3.837	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.946	67	245735	32.635	ng/u1	0.00
11) 2-Chlorophenol-d4	7.134	132	160249	18.841	ng/u1	0.00
15) 4-Methylphenol-d8	8.322	113	81402	9.481	ng/u1	0.00
21) Nitrobenzene-d5	8.769	128	138811	38.271	ng/u1	0.00
24) 2-Nitrophenol-d4	9.487	143	86600	25.472	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	154395	23.814	ng/u1	0.00
31) 4-Chloroaniline-d4	10.546	131	241202	24.048	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	573537	33.154	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	793734	36.468	ng/u1	0.00
54) 4-Nitrophenol-d4	14.504	143	6743	2.292	ng/u1	0.00
60) Fluorene-d10	15.281	176	629485	42.765	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.410	200	54453	25.897	ng/u1	0.00
73) Anthracene-d10	17.145	188	1112035	55.876	ng/u1	0.00
81) Pyrene-d10	19.404	212	1240508	42.611	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.345	264	1448129	43.313	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080322\
Data File : BP011250.D
Acq On : 03 Aug 2022 18:07
Operator : CG/JU
Sample : N3900-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0C86

Quant Time: Aug 03 23:01:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 02 23:52:37 2022
Response via : Initial Calibration

