

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080322\
 Data File : BP011248.D
 Acq On : 03 Aug 2022 16:57
 Operator : CG/JU
 Sample : N3900-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0C96

Manual IntegrationsAPPROVED

Quant Time: Aug 03 23:00:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 02 23:52:37 2022
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 08/04/2022
 Supervised By :Sohil Jodhani 08/10/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.599	152	148981	20.000	ng/ul	0.00
20) Naphthalene-d8	10.393	136	631930	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.275	164	323975	20.000	ng/ul	# 0.00
64) Phenanthrene-d10	17.045	188	594410	20.000	ng/ul	0.00
79) Chrysene-d12	21.174	240	694148	20.000	ng/ul	0.00
88) Perylene-d12	23.492	264	782191	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.093	96	6652	3.610	ng/uL	0.00
4) Pyridine-d5	3.511	84	26240	5.280	ng/ul	0.00
7) Phenol-d5	6.781	99	49705	4.003	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.946	67	259817	29.001	ng/ul	0.00
11) 2-Chlorophenol-d4	7.134	132	179971	17.784	ng/ul	0.00
15) 4-Methylphenol-d8	8.322	113	94712	9.271	ng/ul	0.00
21) Nitrobenzene-d5	8.763	128	143172	31.499	ng/ul	0.00
24) 2-Nitrophenol-d4	9.487	143	91373	21.446	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	170902	21.035	ng/ul	0.00
31) 4-Chloroaniline-d4	10.546	131	235312	18.721	ng/ul	0.00
46) Dimethylphthalate-d6	13.698	166	577428	26.023	ng/ul	0.00
49) Acenaphthylene-d8	13.963	160	826991	29.622	ng/ul	0.00
54) 4-Nitrophenol-d4	14.498	143	6639m	1.759	ng/ul	0.00
60) Fluorene-d10	15.281	176	597768	31.661	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.410	200	48617	17.614	ng/ul	0.00
73) Anthracene-d10	17.145	188	1016099	38.895	ng/ul	0.00
81) Pyrene-d10	19.410	212	1112751	31.706	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.345	264	1263321	32.411	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080322\
Data File : BP011248.D
Acq On : 03 Aug 2022 16:57
Operator : CG/JU
Sample : N3900-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0C96

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 08/04/2022
Supervised By :Sohil Jodhani 08/10/2022

Quant Time: Aug 03 23:00:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 02 23:52:37 2022
Response via : Initial Calibration

